

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1216 SAS No.: Q1216 SDG NO.: Q1216
 Lab File ID: VY021046.D Date Analyzed: 02/04/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:47
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	154178	7.71	233780	8.62	206737	11.42
UPPER LIMIT	308356	8.213	467560	9.115	413474	11.92
LOWER LIMIT	77089	7.213	116890	8.115	103369	10.92
EPA SAMPLE NO.						
JPP-18.1-012825	91498	7.71	158532	8.62	140199	11.42
JPP-21.2-012825	157599	7.71	293437	8.62	274614	11.41
JPP-26.1-012825	158642	7.71	291874	8.62	271395	11.42
VY0204SBL01	173019	7.71	321305	8.62	292351	11.42
VY0204SBS01	159738	7.71	247448	8.62	212473	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.