

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1222 SAS No.: Q1222 SDG No.: Q1222

Instrument ID: MSVOA\_L Calibration Date/Time: 02/03/2025 10:48

Lab File ID: VL041958.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:55 16:05

GC Column: RTX-1 ID: 0.32 (mm)

| COMPOUND                | RRF   | RRF010 | MIN<br>RRF | %D    | MAX%D |
|-------------------------|-------|--------|------------|-------|-------|
| Vinyl Chloride          | 0.519 | 0.529  |            | 1.93  | 30    |
| 1,1-Dichloroethene      | 0.501 | 0.473  |            | -5.59 | 30    |
| cis-1,2-Dichloroethene  | 1.349 | 1.325  |            | -1.78 | 30    |
| 1,1,1-Trichloroethane   | 1.938 | 2.024  |            | 4.44  | 30    |
| Trichloroethene         | 0.364 | 0.408  |            | 12.09 | 30    |
| Tetrachloroethene       | 0.318 | 0.325  |            | 2.2   | 30    |
| 1-Bromo-4-Fluorobenzene | 0.788 | 0.780  |            | -1.01 | 30    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.