

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL011425

Review By	Mahesh Dadoda	Review On	1/15/2025 12:59:33 PM
Supervise By	Semsettin Yesilyurt	Supervise On	1/15/2025 1:00:29 PM
SubDirectory	VL011425	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL011425AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2569
Initial Calibration Stds	AP2571,AP2572,AP2573
CCC	AP2571
Internal Standard/PEM	AP2569
ICV/I.BLK	AP2574,MDL,AP2572,AP2573
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL041887.D	14 Jan 2025 07:50		SY/MD	Ok
2	VSTDICC0.5	VSTDICC0.5	VL041888.D	14 Jan 2025 11:40	not used	SY/MD	Not Ok
3	VSTDICCC010	VSTDICCC010	VL041889.D	14 Jan 2025 12:55		SY/MD	Ok,M
4	VSTDICC002	VSTDICC002	VL041890.D	14 Jan 2025 13:28		SY/MD	Ok,M
5	VSTDICC001	VSTDICC001	VL041891.D	14 Jan 2025 13:59		SY/MD	Ok,M
6	VSTDICC0.5	VSTDICC0.5	VL041892.D	14 Jan 2025 14:30		SY/MD	Ok,M
7	VSTDICC0.1	VSTDICC0.1	VL041893.D	14 Jan 2025 15:01		SY/MD	Ok,M
8	VSTDICC0.03	VSTDICC0.03	VL041894.D	14 Jan 2025 15:32		SY/MD	Ok,M
9	VSTDICC015	VSTDICC015	VL041895.D	14 Jan 2025 16:05		SY/MD	Ok,M
10	VSTDICV010	ICVVL011425	VL041896.D	14 Jan 2025 17:33		SY/MD	Ok,M
11	VL0114ABL01	VL0114ABL01	VL041897.D	14 Jan 2025 18:34		SY/MD	Ok
12	VL0114ABL02	VL0114ABL02	VL041898.D	14 Jan 2025 19:05		SY/MD	Ok
13	VL0114ABS01	VL0114ABS01	VL041899.D	14 Jan 2025 19:36		SY/MD	Ok,M
14	VIBLK	VIBLK	VL041900.D	14 Jan 2025 20:07		SY/MD	Ok
15	Q1073-04	IA1	VL041901.D	14 Jan 2025 20:40		SY/MD	Ok,M
16	Q1073-04DUP	IA1DUP	VL041902.D	14 Jan 2025 21:14		SY/MD	Ok,M
17	Q1073-05	IA2	VL041903.D	14 Jan 2025 21:47		SY/MD	Ok,M
18	Q1073-06	IA3	VL041904.D	14 Jan 2025 22:21		SY/MD	Ok,M

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL011425

Review By	Maresh Dadoda	Review On	1/15/2025 12:59:33 PM
Supervise By	Semsettin Yesilyurt	Supervise On	1/15/2025 1:00:29 PM
SubDirectory	VL011425	HP Acquire Method	TO15AIR.M HP Processing Method VL011425AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2569		
Initial Calibration Stds	AP2571,AP2572,AP2573		
CCC	AP2571		
Internal Standard/PEM	AP2569		
ICV/I.BLK	AP2574,MDL,AP2572,AP2573		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1073-01	SV1	VL041905.D	14 Jan 2025 22:51		SY/MD	Ok,M
20	Q1073-02	SV2	VL041906.D	14 Jan 2025 23:22		SY/MD	Ok,M
21	Q1073-03	SV3	VL041907.D	14 Jan 2025 23:52		SY/MD	Ok,M
22	VIBLK	VIBLK	VL041908.D	15 Jan 2025 00:22		SY/MD	Ok
23	MDL 0.4	MDL 0.4	VL041909.D	15 Jan 2025 00:52		SY/MD	Ok,M
24	MDL 0.4	MDL 0.4	VL041910.D	15 Jan 2025 01:23		SY/MD	Ok,M
25	MDL 0.1	MDL 0.1	VL041911.D	15 Jan 2025 01:53		SY/MD	Ok,M
26	MDL 0.1	MDL 0.1	VL041912.D	15 Jan 2025 02:23		SY/MD	Ok,M
27	MDL 0.03	MDL 0.03	VL041913.D	15 Jan 2025 02:53		SY/MD	Ok,M
28	MDL 0.03	MDL 0.03	VL041914.D	15 Jan 2025 03:23		SY/MD	Ok,M
29	VSTDCCC010	VSTDCCC010EC	VL041915.D	15 Jan 2025 08:37	out of tune	SY/MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL020325

Review By	Semsettin Yesilyurt	Review On	2/4/2025 9:24:56 AM
Supervise By	Maresh Dadoda	Supervise On	2/4/2025 3:44:53 PM
SubDirectory	VL020325	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL011425AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2569
Initial Calibration Stds	AP2571,AP2572,AP2573
CCC	AP2571
Internal Standard/PEM	AP2569
ICV/I.BLK	AP2574
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL041957.D	03 Feb 2025 09:47		SY/MD	Ok
2	VSTDCCC010	VSTDCCC010	VL041958.D	03 Feb 2025 10:48		SY/MD	Ok,M
3	VL0203ABL01	VL0203ABL01	VL041959.D	03 Feb 2025 11:34		SY/MD	Ok
4	VL0203ABL02	VL0203ABL02	VL041960.D	03 Feb 2025 12:05		SY/MD	Ok
5	VL0203ABS01	VL0203ABS01	VL041961.D	03 Feb 2025 12:57		SY/MD	Ok,M
6	Q1222-01	SV1	VL041962.D	03 Feb 2025 13:57		SY/MD	Ok,M
7	Q1222-01DL	SV1DL	VL041963.D	03 Feb 2025 14:28	Not Required	SY/MD	Not Ok
8	Q1256-01	SV1	VL041964.D	03 Feb 2025 14:59		SY/MD	Ok,M
9	Q1256-01DL	SV1DL	VL041965.D	03 Feb 2025 15:30	Not Required	SY/MD	Not Ok
10	QC012825 CAN 10439	QC012825 CAN 10439	VL041966.D	03 Feb 2025 16:02		SY/MD	Ok
11	Q1222-02	IA1	VL041967.D	03 Feb 2025 16:35		SY/MD	Ok,M
12	Q1222-02DUP	IA1DUP	VL041968.D	03 Feb 2025 17:08		SY/MD	Ok,M
13	Q1222-03	OA1	VL041969.D	03 Feb 2025 17:41		SY/MD	Ok,M
14	Q1256-02	IA1	VL041970.D	03 Feb 2025 18:15		SY/MD	Ok,M
15	Q1168-14 0.03	LOQ-AIR-02-QT1-2025	VL041971.D	03 Feb 2025 21:01	rerun	SY/MD	Not Ok
16	Q1168-13 0.03	LOD-MDL-AIR-01-QT1	VL041972.D	03 Feb 2025 21:32		SY/MD	Ok,M
17	Q1168-14 0.03	LOQ-AIR-02-QT1-2025	VL041973.D	03 Feb 2025 22:13		SY/MD	Ok,M
18	QC020325 CAN 10599	QC020325 CAN 10599	VL041974.D	04 Feb 2025 08:03		SY/MD	Ok

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL020325

Review By	Semsettin Yesilyurt	Review On	2/4/2025 9:24:56 AM
Supervise By	Maresh Dadoda	Supervise On	2/4/2025 3:44:53 PM
SubDirectory	VL020325	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL011425AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2569
Initial Calibration Stds	AP2571,AP2572,AP2573
CCC	AP2571
Internal Standard/PEM	AP2569
ICV/I.BLK	AP2574
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VSTDCCC010	VSTDCCC010EC	VL041975.D	04 Feb 2025 08:34		SY/MD	Ok,M
----	------------	--------------	------------	-------------------	--	-------	------

M : Manual Integration