

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51952 MA No.: SDG No.: A6301
Instrument ID: CV1 Start Date: 01/31/2025
Analytical Method: CVAA Run Batch: LB134514
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.050	0	0.2	0.21	3	2.5	2.4	-4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51952 MA No.: SDG No.: A6301
Instrument ID: CV1 Start Date: 01/31/2025
Analytical Method: CVAA Run Batch: LB134514
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.1	2	7.5	7.4	-1	10	10	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51952 MA No.: _____ SDG No.: A6301
Instrument ID: CV1 Start Date: 02/18/2025
Analytical Method: CVAA Run Batch: LB134741
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0	0	0.2	0.18	-10	2.5	2.5	-1

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51952 MA No.: SDG No.: A6301
Instrument ID: CV1 Start Date: 02/18/2025
Analytical Method: CVAA Run Batch: LB134741
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.0	1	7.5	7.7	3	10	9.8	-2