

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51952 MA No.: SDG No.: A6301
Instrument ID: CV1 Start Date : 01/31/2025
Analytical Method: CVAA Run Batch : LB134514

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Mercury	0.999788	0.000192	0.016048	Lin. Reg	NONE

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51952 MA No.: SDG No.: A6301
Instrument ID: CV1 Start Date : 02/18/2025
Analytical Method: CVAA Run Batch : LB134741

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Mercury	0.999378	0.000227	-0.011753	Lin. Reg	NONE

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51952 MA No.: SDG No.: A6301
Instrument ID: CV1 Start Date : 02/18/2025
Analytical Method: CVAA Run Batch : LB134741

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------