

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN012225.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jan 23 00:34:56 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN036010.D 0.2 =BN036011.D 0.4 =BN036012.D 0.8 =BN036013.D 1.6 =BN036014.D 3.2 =BN036015.D 5.0 =BN036016.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.452	0.460	0.469	0.477	0.447	0.408	0.417	0.447	5.81
3)	n-Nitrosodimet...	0.798	0.749	0.877	0.883	0.829	0.781	0.759	0.811	6.65
4) S	2-Fluorophenol	1.032	1.012	1.092	1.099	1.042	0.997	1.010	1.040	3.88
5) S	Phenol-d6	1.284	1.195	1.270	1.155	1.230	1.210	1.209	1.222	3.61
6)	bis(2-Chloroet...	1.024	0.979	1.056	0.929	0.993	0.952	0.952	0.984	4.53
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.377	0.356	0.399	0.333	0.397	0.388	0.394	0.378	6.55
9)	Naphthalene	1.149	1.141	1.250	1.137	1.184	1.141	1.131	1.162	3.68
10)	Hexachlorobuta...	0.383	0.369	0.404	0.371	0.388	0.359	0.353	0.375	4.74
11) SURR	2-Methylnaphth...	0.522	0.527	0.578	0.528	0.556	0.550	0.545	0.544	3.66
12)	2-Methylnaphth...	0.702	0.688	0.760	0.700	0.741	0.735	0.721	0.721	3.58
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.240	0.238	0.256	0.238	0.268	0.275	0.282	0.257	7.32
15) S	2-Fluorobiphenyl	1.806	1.736	1.934	1.787	1.819	1.693	1.724	1.786	4.47
16)	Acenaphthylene	1.835	1.826	2.011	1.840	1.940	1.889	1.936	1.897	3.65
17)	Acenaphthene	1.248	1.236	1.365	1.266	1.338	1.310	1.327	1.299	3.78
18)	Fluorene	1.583	1.482	1.633	1.550	1.739	1.703	1.700	1.627	5.76
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.071	0.081	0.095	0.089	0.101	0.107	0.108	0.093	14.73
21)	4-Bromophenyl-...	0.285	0.269	0.307	0.287	0.293	0.273	0.281	0.285	4.42
22)	Hexachlorobenzene	0.391	0.358	0.407	0.374	0.380	0.355	0.361	0.375	5.08
23)	Atrazine	0.185	0.194	0.218	0.204	0.216	0.209	0.215	0.206	6.05
24)	Pentachlorophenol	0.131	0.131	0.164	0.155	0.179	0.185	0.192	0.162	15.18
25)	Phenanthrene	1.154	1.158	1.302	1.172	1.226	1.182	1.219	1.202	4.33
26)	Anthracene	1.019	1.016	1.151	1.064	1.128	1.123	1.151	1.093	5.42
27) SURR	Fluoranthene-d10	1.005	1.006	1.111	0.994	0.959	1.078	1.101	1.036	5.75
28)	Fluoranthene	1.312	1.350	1.507	1.357	1.317	1.506	1.533	1.412	6.99
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.657	1.588	1.693	1.636	1.646	1.552	1.575	1.621	3.12
31) S	Terphenyl-d14	0.821	0.807	0.871	0.831	0.860	0.804	0.822	0.831	3.09
32)	Benzo(a)anthra...	1.445	1.403	1.503	1.411	1.513	1.448	1.433	1.451	2.93
33)	Chrysene	1.501	1.476	1.545	1.448	1.515	1.435	1.463	1.483	2.63
34)	Bis(2-ethylhex...	0.919	0.793	0.798	0.748	0.791	0.748	0.768	0.795	7.36
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN012225.M

36)	Indeno(1,2,3-c...	1.525	1.477	1.621	1.585	1.669	1.668	1.692	1.605	5.03
37)	Benzo(b)fluora...	1.443	1.380	1.497	1.429	1.475	1.444	1.510	1.454	3.03
38)	Benzo(k)fluora...	1.427	1.378	1.486	1.427	1.519	1.496	1.524	1.465	3.76
39) C	Benzo(a)pyrene	1.237	1.164	1.263	1.203	1.264	1.265	1.296	1.242	3.61
40)	Dibenzo(a,h)an...	1.187	1.169	1.290	1.279	1.337	1.338	1.356	1.279	5.86
41)	Benzo(g,h,i)pe...	1.338	1.308	1.426	1.387	1.438	1.428	1.436	1.394	3.75

(#) = Out of Range