

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1225 SAS No.: Q1225 SDG No.: Q1225

Instrument ID: BNA_N Calibration Date/Time: 02/06/2025 05:40

Lab File ID: BN036320.D Init. Calib. Date(s): 01/22/2025 01/22/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:02 14:36

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.575		5.7	20.0
Fluoranthene-d10	1.036	1.072		3.5	20.0
2-Fluorophenol	1.040	1.170		12.5	20.0
Phenol-d6	1.222	1.367		11.9	20.0
Nitrobenzene-d5	0.378	0.417		10.3	20.0
2-Fluorobiphenyl	1.786	1.611		-9.8	20.0
2,4,6-Tribromophenol	0.257	0.200		-22.2	20.0
Terphenyl-d14	0.831	0.874		5.2	20.0
1,4-Dioxane	0.447	0.518		15.9	20.0

All other compounds must meet a minimum RRF of 0.010.