

Instrument ID: BNA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # BN012225**

|              |          |                   |                                           |
|--------------|----------|-------------------|-------------------------------------------|
| Review By    | yogesh   | Review On         | 1/24/2025 8:01:29 AM                      |
| Supervise By | mohammad | Supervise On      | 1/24/2025 8:08:50 AM                      |
| SubDirectory | BN012225 | HP Acquire Method | BNA_N, 8270_HP Processing Method bn012225 |

| STD. NAME                | STD REF.#                                        |
|--------------------------|--------------------------------------------------|
| Tune/Reschk              | SP6717                                           |
| Initial Calibration Stds | SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657 |
| CCC                      | SP6661                                           |
| Internal Standard/PEM    | SP6682,1ul/100ul sample                          |
| ICV/I.BLK                | SP6684                                           |
| Surrogate Standard       |                                                  |
| MS/MSD Standard          |                                                  |
| LCS Standard             |                                                  |

| Sr# | SampleID    | ClientID    | Data File Name | Date-Time         | Comment                         | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|---------------------------------|----------|--------|
| 1   | DFTPP       | DFTPP       | BN036009.D     | 22 Jan 2025 09:44 |                                 | RC/JU    | Ok     |
| 2   | SSTDICC0.1  | SSTDICC0.1  | BN036010.D     | 22 Jan 2025 11:02 |                                 | RC/JU    | Ok     |
| 3   | SSTDICC0.2  | SSTDICC0.2  | BN036011.D     | 22 Jan 2025 11:38 |                                 | RC/JU    | Ok     |
| 4   | SSTDICCC0.4 | SSTDICCC0.4 | BN036012.D     | 22 Jan 2025 12:13 | The Calibration is Good For DOD | RC/JU    | Ok     |
| 5   | SSTDICC0.8  | SSTDICC0.8  | BN036013.D     | 22 Jan 2025 12:49 |                                 | RC/JU    | Ok     |
| 6   | SSTDICC1.6  | SSTDICC1.6  | BN036014.D     | 22 Jan 2025 13:25 |                                 | RC/JU    | Ok     |
| 7   | SSTDICC3.2  | SSTDICC3.2  | BN036015.D     | 22 Jan 2025 14:01 |                                 | RC/JU    | Ok     |
| 8   | SSTDICC5.0  | SSTDICC5.0  | BN036016.D     | 22 Jan 2025 14:36 |                                 | RC/JU    | Ok     |
| 9   | SSTDICV0.4  | ICVBN012225 | BN036017.D     | 22 Jan 2025 15:53 |                                 | RC/JU    | Ok     |

M : Manual Integration

Instrument ID: BNA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # BN020525**

|              |          |                   |                                           |
|--------------|----------|-------------------|-------------------------------------------|
| Review By    | yogesh   | Review On         | 2/6/2025 5:49:41 AM                       |
| Supervise By | mohammad | Supervise On      | 2/7/2025 8:32:26 AM                       |
| SubDirectory | BN020525 | HP Acquire Method | BNA_N, 8270_HP Processing Method bn012225 |

| STD. NAME                                                                                          | STD REF.#                                                  |
|----------------------------------------------------------------------------------------------------|------------------------------------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | SP6717<br>SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657 |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | SP6661<br>SP6682,1ul/100ul sample<br>SP6684                |

| Sr# | SampleID   | ClientID            | Data File Name | Date-Time         | Comment          | Operator | Status   |
|-----|------------|---------------------|----------------|-------------------|------------------|----------|----------|
| 1   | DFTPP      | DFTPP               | BN036302.D     | 05 Feb 2025 18:08 |                  | RC/JU    | Ok       |
| 2   | SSTDCCC0.4 | SSTDCCC0.4          | BN036303.D     | 05 Feb 2025 18:47 |                  | RC/JU    | Ok       |
| 3   | PB166419BL | PB166419BL          | BN036304.D     | 05 Feb 2025 19:23 |                  | RC/JU    | Ok       |
| 4   | PB166470BL | PB166470BL          | BN036305.D     | 05 Feb 2025 19:59 |                  | RC/JU    | Ok       |
| 5   | Q1173-06RE | VPB-192-HYD-2025012 | BN036306.D     | 05 Feb 2025 20:35 | Surrogate Fail   | RC/JU    | Confirms |
| 6   | Q1225-01   | RW5-SP100-20250129  | BN036307.D     | 05 Feb 2025 21:11 | Need 5X Dilution | RC/JU    | Dilution |
| 7   | Q1225-02   | RW5-SP201-20250129  | BN036308.D     | 05 Feb 2025 21:47 |                  | RC/JU    | Ok       |
| 8   | Q1225-03   | RW5-SP303-20250129  | BN036309.D     | 05 Feb 2025 22:22 |                  | RC/JU    | Ok       |
| 9   | Q1250-02   | BP-VPB-192-GW-420-4 | BN036310.D     | 05 Feb 2025 22:58 |                  | RC/JU    | Ok       |
| 10  | Q1250-03   | BP-VPB-192-GW-300-3 | BN036311.D     | 05 Feb 2025 23:34 |                  | RC/JU    | Ok       |
| 11  | Q1250-05   | BP-VPB-192-GW-340-3 | BN036312.D     | 06 Feb 2025 00:10 |                  | RC/JU    | Ok       |
| 12  | Q1250-09   | BP-VPB-192-DUP-2025 | BN036313.D     | 06 Feb 2025 00:46 |                  | RC/JU    | Ok       |
| 13  | Q1250-10   | BP-VPB-192-GW-380-3 | BN036314.D     | 06 Feb 2025 01:22 |                  | RC/JU    | Ok       |
| 14  | Q1249-01   | BP-VPB-192-GW-460-4 | BN036315.D     | 06 Feb 2025 01:58 |                  | RC/JU    | Ok,M     |
| 15  | Q1283-01   | VPB-192-HYD-2025013 | BN036316.D     | 06 Feb 2025 02:34 | Surrogate Fail   | RC/JU    | ReRun    |
| 16  | PB166470BS | PB166470BS          | BN036317.D     | 06 Feb 2025 03:10 |                  | RC/JU    | Ok       |
| 17  | SSTDCCC0.4 | SSTDCCC0.4EC        | BN036318.D     | 06 Feb 2025 03:45 |                  | RC/JU    | Ok       |
| 18  | DFTPP      | DFTPP               | BN036319.D     | 06 Feb 2025 05:01 |                  | RC/JU    | Ok       |

Instrument ID: BNA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # BN020525**

|                          |                                                  |                   |                                           |
|--------------------------|--------------------------------------------------|-------------------|-------------------------------------------|
| Review By                | yogesh                                           | Review On         | 2/6/2025 5:49:41 AM                       |
| Supervise By             | mohammad                                         | Supervise On      | 2/7/2025 8:32:26 AM                       |
| SubDirectory             | BN020525                                         | HP Acquire Method | BNA_N, 8270_HP Processing Method bn012225 |
| <b>STD. NAME</b>         | <b>STD REF.#</b>                                 |                   |                                           |
| Tune/Reschk              | SP6717                                           |                   |                                           |
| Initial Calibration Stds | SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657 |                   |                                           |
| CCC                      | SP6661                                           |                   |                                           |
| Internal Standard/PEM    | SP6682,1ul/100ul sample                          |                   |                                           |
| ICV/I.BLK                | SP6684                                           |                   |                                           |
| Surrogate Standard       |                                                  |                   |                                           |
| MS/MSD Standard          |                                                  |                   |                                           |
| LCS Standard             |                                                  |                   |                                           |

|    |             |                     |            |                   |                |       |          |
|----|-------------|---------------------|------------|-------------------|----------------|-------|----------|
| 19 | SSTDCCC0.4  | SSTDCCC0.4          | BN036320.D | 06 Feb 2025 05:40 |                | RC/JU | Ok       |
| 20 | PB166533BL  | PB166533BL          | BN036321.D | 06 Feb 2025 06:15 |                | RC/JU | Ok       |
| 21 | Q1283-03    | BP-VPB-192-EB-20250 | BN036322.D | 06 Feb 2025 06:51 |                | RC/JU | Ok       |
| 22 | Q1283-05    | BP-VPB-192-GW-500-5 | BN036323.D | 06 Feb 2025 07:27 |                | RC/JU | Ok,M     |
| 23 | PB166419BS  | PB166419BS          | BN036324.D | 06 Feb 2025 08:03 |                | RC/JU | Ok,M     |
| 24 | PB166470BSD | PB166470BSD         | BN036325.D | 06 Feb 2025 08:39 |                | RC/JU | Ok,M     |
| 25 | PB166533BS  | PB166533BS          | BN036326.D | 06 Feb 2025 09:15 |                | RC/JU | Ok       |
| 26 | PB166533BSD | PB166533BSD         | BN036327.D | 06 Feb 2025 09:51 |                | RC/JU | Ok       |
| 27 | PB166419BSD | PB166419BSD         | BN036328.D | 06 Feb 2025 10:27 |                | RC/JU | Ok       |
| 28 | PB166237BL  | PB166237BL          | BN036329.D | 06 Feb 2025 11:03 |                | RC/JU | Ok       |
| 29 | Q1225-01DL  | RW5-SP100-20250129  | BN036330.D | 06 Feb 2025 11:39 |                | RC/JU | Ok       |
| 30 | Q1283-01RE  | VPB-192-HYD-2025013 | BN036331.D | 06 Feb 2025 12:15 | Surrogate Fail | RC/JU | Confirms |
| 31 | SSTDCCC0.4  | SSTDCCC0.4EC        | BN036332.D | 06 Feb 2025 12:54 |                | RC/JU | Ok       |

M : Manual Integration