

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51900 MA No.: \_\_\_\_\_ SDG No.: ME2993  
Instrument ID: KONELAB Start Date: 02/06/2025  
Analytical Method: Spectrophotometry Run Batch: LB134600  
Concentration Units: ug/L

| Analyte | True | Found | %D | True | Found | %D | True | Found | %D |
|---------|------|-------|----|------|-------|----|------|-------|----|
| Cyanide | 0    | 0.0   | 0  | 5    | 4.6   | -8 | 10   | 9.1   | -9 |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51900 MA No.: \_\_\_\_\_ SDG No.: ME2993  
Instrument ID: KONELAB Start Date: 02/06/2025  
Analytical Method: Spectrophotometry Run Batch: LB134600  
Concentration Units: ug/L

| Analyte | True | Found | %D | True | Found | %D | True | Found | %D |
|---------|------|-------|----|------|-------|----|------|-------|----|
| Cyanide | 100  | 100   | 0  | 250  | 250   | 0  | 500  | 500   | 0  |