ME2993S

FORM 5A - IN MATRIX SPIKE SAMPLE RECOVERY

Lab Name:	Chemtech Consulting Group			Contract:	EPW14030				
Lab Code:	ACE	Case No.:	51900	MA No. :		SDG No.:	ME2993		
Matrix :	Water			Analytical	Method:	CVAA			
% Solids:									

Concentration Units (µg/L, mg/L or mg/kg dry weight): ug/L

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	۶R	Q
Mercury	75 - 125	1.3		0.042	J	1.0	126	*