

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No.: _____ SDG No.: ME2993
 Instrument ID: P8 Start Date: 02/13/2025
 Analytical Method: ICP-MS Run Batch: LB134712
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.0	0	20	22	10	1000	1000	0
Antimony	0	0.0	0	2	2.1	5	50	51	2
Arsenic	0	0.0	0	1	1.1	10	50	52	4
Barium	0	0.0	0	10	10	0	250	260	4
Beryllium	0	0.0	0	1	1.1	10	50	52	4
Cadmium	0	0.0	0	1	1.0	0	50	52	4
Calcium	0	0.0	0	500	550	10	5000	5400	8
Chromium	0	0.0	0	2	2.1	5	50	50	0
Cobalt	0	0.0	0	1	1.2	20	50	53	6
Copper	0	0.0	0	2	2.2	10	500	570	14
Iron	0	0.0	0	50	59	18	2500	2700	8
Lead	0	0.0	0	1	0.99	-1	250	260	4
Magnesium	0	0.0	0	500	550	10	5000	5600	12
Manganese	0	0.0	0	1	1.1	10	500	520	4
Nickel	0	0.0	0	1	1.0	0	50	54	8
Potassium	0	0.0	0	500	500	0	2500	2500	0
Selenium	0	0.0	0	5	5.4	8	50	54	8
Silver	0	0.0	0	1	1.0	0	50	52	4
Sodium	0	0.0	0	500	550	10	5000	5700	14
Thallium	0	0.0	0	1	0.99	-1	50	52	4
Vanadium	0	0.0	0	5	5.2	4	50	50	0
Zinc	0	0.0	0	5	5.6	12	500	530	6

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No.: _____ SDG No.: ME2993
 Instrument ID: P8 Start Date: 02/13/2025
 Analytical Method: ICP-MS Run Batch: LB134712
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	2500	2600	4	5000	5100	2	10000	9800	-2
Antimony	125	130	4	250	260	4	500	500	0
Arsenic	125	130	4	250	260	4	500	510	2
Barium	625	640	2	1250	1300	4	2500	2500	0
Beryllium	125	130	4	250	260	4	500	510	2
Cadmium	125	120	-4	250	250	0	500	510	2
Calcium	12500	13000	4	25000	26000	4	50000	52000	4
Chromium	125	120	-4	250	240	-4	500	510	2
Cobalt	125	130	4	250	270	8	500	510	2
Copper	1250	1400	12	2500	2600	4	5000	5100	2
Iron	6250	6500	4	12500	13000	4	25000	26000	4
Lead	625	630	1	1250	1300	4	2500	2500	0
Magnesium	12500	14000	12	25000	26000	4	50000	52000	4
Manganese	1250	1300	4	2500	2600	4	5000	5000	0
Nickel	125	140	12	250	250	0	500	480	-4
Potassium	6250	6100	-2	12500	13000	4	25000	25000	0
Selenium	125	130	4	250	260	4	500	510	2
Silver	125	130	4	250	260	4	500	510	2
Sodium	12500	14000	12	25000	27000	8	50000	53000	6
Thallium	125	130	4	250	250	0	500	500	0
Vanadium	125	120	-4	250	250	0	500	500	0
Zinc	1250	1300	4	2500	2500	0	5000	5200	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No.: _____ SDG No.: ME2993
 Instrument ID: P8 Start Date: 02/13/2025
 Analytical Method: ICP-MS Run Batch: LB134712
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	20000	20000	0	100000	100000	0			
Antimony	1000	1000	0						
Arsenic	1000	990	-1						
Barium	5000	5000	0						
Beryllium	1000	990	-1						
Cadmium	1000	1000	0						
Calcium	100000	100000	0	500000	500000	0			
Chromium	1000	1000	0						
Cobalt	1000	990	-1						
Copper	10000	9900	-1						
Iron	50000	51000	2	250000	250000	0			
Lead	5000	5000	0						
Magnesium	100000	100000	0	500000	500000	0			
Manganese	10000	9900	-1						
Nickel	1000	1000	0						
Potassium	50000	49000	-2	250000	250000	0			
Selenium	1000	990	-1						
Silver	1000	990	-1						
Sodium	100000	100000	0	500000	500000	0			
Thallium	1000	1000	0						
Vanadium	1000	1000	0						
Zinc	10000	9900	-1						