

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134712

Review By	Janvi Patel	Review On	2/14/2025 5:32:06 PM
Supervise By	Kareem Khairalla	Supervise On	2/18/2025 11:55:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84414,MP84423,MP84422,MP84420,MP84419,MP84418,MP84417,MP84416,MP84415		
ICV Standard	MP84439		
CCV Standard	MP84438		
ICSA Standard	MP84440,MP84441		
CRI Standard			
LCS Standard	MP84600		
Chk Standard	MP84444,MP84443		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE007	TUNE	02/13/25 11:12		Jaswal	OK
2	S00	S0	CAL1	02/13/25 11:45		Jaswal	OK
3	S02	S02	CAL3	02/13/25 11:48		Jaswal	OK
4	S03	S03	CAL4	02/13/25 11:51		Jaswal	OK
5	S04	S04	CAL5	02/13/25 11:54		Jaswal	OK
6	S05	S05	CAL6	02/13/25 11:57		Jaswal	OK
7	S06	S06	CAL7	02/13/25 12:00		Jaswal	OK
8	S07	S07	CAL8	02/13/25 12:03		Jaswal	OK
9	S08	S08	CAL9	02/13/25 12:06		Jaswal	OK
10	ICV007	ICV007	ICV	02/13/25 12:12	Instrument paused for calibration review	Jaswal	OK
11	ICB007	ICB007	ICB	02/13/25 12:16		Jaswal	OK
12	ICSA007	ICSA007	ICSA	02/13/25 12:19		Jaswal	OK
13	ICSAB007	ICSAB007	ICSAB	02/13/25 12:28	Instrument paused for review	Jaswal	OK
14	CCV027	CCV027	CCV	02/13/25 12:31		Jaswal	OK
15	CCB027	CCB027	CCB	02/13/25 12:34		Jaswal	OK
16	Q1255-08A	MDCZN8A	PS	02/13/25 12:38	PS for Se(0.1ml of MP84474 in 10ml sample)	Jaswal	OK
17	PB166671BL	PBW671	MB	02/13/25 12:41		Jaswal	OK
18	PB166671BS	LCS671	LCS	02/13/25 12:45		Jaswal	OK

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134712

Review By	Janvi Patel	Review On	2/14/2025 5:32:06 PM
Supervise By	Kareem Khairalla	Supervise On	2/18/2025 11:55:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84414,MP84423,MP84422,MP84420,MP84419,MP84418,MP84417,MP84416,MP84415		
ICV Standard	MP84439		
CCV Standard	MP84438		
ICSA Standard	MP84440,MP84441		
CRI Standard			
LCS Standard	MP84600		
Chk Standard	MP84444,MP84443		

19	Q1231-01	ME2993	SAM	02/13/25 12:48		Jaswal	OK
20	Q1231-02	ME2993D	DUP	02/13/25 12:51		Jaswal	OK
21	Q1231-01L	ME2993L	SD	02/13/25 12:54		Jaswal	OK
22	Q1231-03	ME2993S	MS	02/13/25 12:58		Jaswal	OK
23	Q1231-04	ME29C1	SAM	02/13/25 13:01		Jaswal	OK
24	Q1231-05	ME29C2	SAM	02/13/25 13:04		Jaswal	OK
25	Q1231-06	ME29C3	SAM	02/13/25 13:07		Jaswal	OK
26	Q1231-07	ME29C4	SAM	02/13/25 13:10		Jaswal	OK
27	PB166703BL	PBW703	MB	02/13/25 13:14		Jaswal	OK
28	PB166703BS	LCS703	LCS	02/13/25 13:17		Jaswal	OK
29	Q1339-01	ME29C5	SAM	02/13/25 13:20	Na high	Jaswal	Dilution
30	Q1339-02	ME29C7	SAM	02/13/25 13:24		Jaswal	OK
31	Q1339-03	ME29C6	SAM	02/13/25 13:27		Jaswal	OK
32	Q1339-04	ME29C9	SAM	02/13/25 13:30		Jaswal	OK
33	Q1339-05	ME29D1	SAM	02/13/25 13:33		Jaswal	OK
34	Q1339-06RE	ME29D2	SAM	02/13/25 13:37	NOT USE	Jaswal	Not Ok
35	Q1339-07	ME29D4	SAM	02/13/25 13:40		Jaswal	OK
36	Q1339-08RE	ME29D2D	DUP	02/13/25 13:43	NOT USE	Jaswal	Not Ok
37	Q1339-07LRE	ME29D4L	SD	02/13/25 13:47	NOT USE	Jaswal	Not Ok
38	Q1339-09RE	ME29D2S	MS	02/13/25 13:50	NOT USE	Jaswal	Not Ok

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134712

Review By	Janvi Patel	Review On	2/14/2025 5:32:06 PM
Supervise By	Kareem Khairalla	Supervise On	2/18/2025 11:55:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP84414,MP84423,MP84422,MP84420,MP84419,MP84418,MP84417,MP84416,MP84415
ICV Standard	MP84439
CCV Standard	MP84438
ICSA Standard	MP84440,MP84441
CRI Standard	
LCS Standard	MP84600
Chk Standard	MP84444,MP84443

39	Q1339-01DL	ME29C5	SAM	02/13/25 13:53	Report 2x for Na	Jaswal	Confirms
40	CCV028	CCV028	CCV	02/13/25 13:56		Jaswal	OK
41	CCB028	CCB028	CCB	02/13/25 13:59		Jaswal	OK
42	Q1231-01A	ME2993A	PS	02/13/25 14:59	PS for Se(0.1ml of MP84474 in 10ml sample)	Jaswal	OK
43	CCV029	CCV029	CCV	02/13/25 15:09		Jaswal	OK
44	CCB029	CCB029	CCB	02/13/25 15:12		Jaswal	OK
45	Q1339-06	ME29D2	SAM	02/13/25 16:16		Jaswal	OK
46	Q1339-08	ME29D2D	DUP	02/13/25 16:20		Jaswal	OK
47	Q1339-06L	ME29D2L	SD	02/13/25 16:23		Jaswal	OK
48	Q1339-09	ME29D2S	MS	02/13/25 16:26		Jaswal	OK
49	CCV030	CCV030	CCV	02/13/25 16:31		Jaswal	OK
50	CCB030	CCB030	CCB	02/13/25 16:34		Jaswal	OK