

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1232

MATRIX: Solid

METHOD: 8270E/3541

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2. GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5. GC/MS Calibration Requirements.		✓	

The Initial Calibration met the requirements .

The Continuous Calibration File ID BF141336.D met the requirements except for
4-Nitrophenol and Pentachlorophenol but associated QC within limits therefore no
corrective action taken.

The Continuous Calibration File ID BF141410.D met the requirements except for
4,6-Dinitro-2-methylphenol but no positive hit in associated sample therefore no
corrective action taken.

The Continuous Calibration File ID BF141531.D met the requirements except for
Di-n-octyl phthalate but no positive hit in associated sample therefore no corrective
action taken.

6. Blank Contamination - If yes, list compounds and concentrations in each blank:	✓
7. Surrogate Recoveries Meet Criteria	✓

If not met, list those compounds and their recoveries which fall outside the acceptable
ranges.

The Surrogate recoveries met the acceptable criteria except for JPP-42.1-012925 [2,4
and 6-Tribromophenol - 9%] as per method one surrogate is allowed to fail therefore no
corrective action taken.

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

	NA	NO	YES
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The MS {Q1239-10MS} with File ID: BF141345.D recoveries met the requirements for all compounds except for 2-Chlorophenol[75%], Acenaphthylene[78%], Isophorone[75%] and Naphthalene[71%] due to matrix interference.			
The MSD {Q1239-10MSD} with File ID: BF141346.D recoveries met the acceptable requirements except for Benzaldehyde[92%] due to matrix interference.			
The Blank Spike for {PB166418BS} with File ID: BF141441.D met requirements for all samples except for Acetophenone[100%], Hexachlorocyclopentadiene[191%] but no positive hits in associated samples therefore no corrective action taken.			
The RPD for {Q1239-10MSD} with File ID: BF141346.D met criteria except for Benzo(k)fluoranthene[22%] due to difference in results of MS and MSD.			
9. Internal Standard Area/Retention Time Shift Meet Criteria			✓
Comments:			
10. Extraction Holding Time Met			✓
If not met, list number of days exceeded for each sample:			
11. Analysis Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

Samples JPP-46.1-012925, JPP-42.1-012925 and JPP-51.1-012925 were initially diluted due to viscous and concentrated matrix.

Samples JPP-46.2-012925, JPP-42.2-012925 were diluted due to high concentrations.

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

NA NO YES

and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date