

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: RU2 Engineering, LLC **Level:** LOW **SDG No.:** Q1232
Contract: RUTW01 **Lab Code:** CHEM **Case No.:** Q1232 **SAS No.:** Q1232
Matrix: Water **Sample ID:** Q1235-08 **Client ID:** JPP-16.1-012925DUP
Percent Solids for Sample: NA **Duplicate ID** Q1235-08DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	333	J	330	J	1		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	9.60	J	11.0	J	13		P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: RU2 Engineering, LLC **Level:** LOW **SDG No.:** Q1232
Contract: RUTW01 **Lab Code:** CHEM **Case No.:** Q1232 **SAS No.:** Q1232
Matrix: Water **Sample ID:** Q1235-08MS **Client ID:** JPP-16.1-012925MSD
Percent Solids for Sample: NA **Duplicate ID** Q1235-08MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3990		3880		3		P
Barium	ug/L	20	1220		1260		3		P
Cadmium	ug/L	20	916		904		1		P
Chromium	ug/L	20	1920		1920		0		P
Lead	ug/L	20	4380		4340		1		P
Mercury	ug/L	20	44.2		41.2		7		CV
Selenium	ug/L	20	9960		9640		3		P
Silver	ug/L	20	363		360		1		P