

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB011525

Review By	yogesh	Review On	1/15/2025 12:15:24 PM
Supervise By	Ankita	Supervise On	1/16/2025 10:14:58 AM
SubDirectory	FB011525	HP Acquire Method	HP Processing Method FB011525
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24113,PP24114,PP24115,PP24116,PP24117		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24118		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB031307.D	15 Jan 2025 10:20		YP/AJ	Ok
2	10 GRO STD		FB031308.D	15 Jan 2025 10:47		YP/AJ	Ok
3	20 GRO STD		FB031309.D	15 Jan 2025 11:13		YP/AJ	Ok
4	50 GRO STD		FB031310.D	15 Jan 2025 11:40		YP/AJ	Ok
5	100 GRO STD		FB031311.D	15 Jan 2025 12:07		YP/AJ	Ok
6	FB011525GROICV		FB031312.D	15 Jan 2025 12:44		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB013125

Review By	yogesh	Review On	1/31/2025 1:02:12 PM
Supervise By	Ankita	Supervise On	2/3/2025 1:09:11 PM
SubDirectory	FB013125	HP Acquire Method	HP Processing Method FB011525

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24110,PP24113,PP24114,PP24115,PP24116,PP24117
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24167,PP24168,PP24169 PP24111,PP24118

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB031412.D	31 Jan 2025 9:03		YP/AJ	Ok
2	VBF0131S1		FB031413.D	31 Jan 2025 9:41		YP/AJ	Ok
3	VBF0131S2		FB031414.D	31 Jan 2025 10:08		YP/AJ	Ok
4	BSF0131S1		FB031415.D	31 Jan 2025 10:47		YP/AJ	Ok
5	Q1242-01		FB031416.D	31 Jan 2025 11:25	vial-A	YP/AJ	Ok
6	Q1241-01		FB031417.D	31 Jan 2025 11:52	vial-A	YP/AJ	Ok
7	Q1241-05		FB031418.D	31 Jan 2025 12:18	vial-A	YP/AJ	Ok
8	Q1241-09		FB031419.D	31 Jan 2025 12:45	vial-A ,not pureged	YP/AJ	Not Ok
9	BSF0131S2		FB031420.D	31 Jan 2025 13:12		YP/AJ	Ok
10	20 PPB GRO STD		FB031421.D	31 Jan 2025 13:38		YP/AJ	Ok
11	Q1241-13		FB031422.D	31 Jan 2025 14:32	vial-A	YP/AJ	Ok
12	Q1241-17		FB031423.D	31 Jan 2025 14:58	vial-A	YP/AJ	Ok
13	Q1232-01		FB031424.D	31 Jan 2025 15:25	vial-A ,surrogate fail	YP/AJ	ReRun
14	Q1232-05		FB031425.D	31 Jan 2025 15:52	vial-A ,surrogate fail	YP/AJ	ReRun
15	Q1232-09		FB031426.D	31 Jan 2025 16:18	vial-A ,not pureged	YP/AJ	Not Ok
16	Q1232-13		FB031427.D	31 Jan 2025 16:45	vial-A ,not pureged	YP/AJ	Not Ok
17	Q1232-17		FB031428.D	31 Jan 2025 17:12	vial-A ,not pureged	YP/AJ	Not Ok
18	Q1235-01		FB031429.D	31 Jan 2025 17:38	vial-A	YP/AJ	Ok

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB013125

Review By	yogesh	Review On	1/31/2025 1:02:12 PM
Supervise By	Ankita	Supervise On	2/3/2025 1:09:11 PM
SubDirectory	FB013125	HP Acquire Method	HP Processing Method
			FB011525
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24113,PP24114,PP24115,PP24116,PP24117		
CCC Internal Standard/PEM	PP24167,PP24168,PP24169		
ICV/I.BLK Surrogate Standard	PP24111,PP24118		
MS/MSD Standard			
LCS Standard			

19	Q1235-05		FB031430.D	31 Jan 2025 18:05	vial-A	YP/AJ	Ok
20	BSF0131S3		FB031431.D	31 Jan 2025 18:32		YP/AJ	Ok,M
21	20 PPB GRO STD		FB031432.D	31 Jan 2025 18:58		YP/AJ	Ok
22	Q1241-09		FB031433.D	31 Jan 2025 19:52	vial-B	YP/AJ	Ok
23	Q1241-09		FB031434.D	31 Jan 2025 20:19	vial-C , not required	YP/AJ	Not Ok
24	Q1232-01		FB031435.D	31 Jan 2025 20:45	vial-B	YP/AJ	Ok
25	Q1232-01		FB031436.D	31 Jan 2025 21:12	vial-C , not required	YP/AJ	Not Ok
26	BSF0131S4		FB031437.D	31 Jan 2025 21:39		YP/AJ	Ok,M
27	20 PPB GRO STD		FB031438.D	31 Jan 2025 22:32		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB020325

Review By	yogesh	Review On	2/3/2025 12:29:22 PM
Supervise By	Ankita	Supervise On	2/4/2025 9:01:22 AM
SubDirectory	FB020325	HP Acquire Method	HP Processing Method FB011525

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24110,PP24113,PP24114,PP24115,PP24116,PP24117
CCC Internal Standard/PEM	PP24171,PP24172,PP24173
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24118

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB031439.D	3 Feb 2025 11:08		YP/AJ	Ok
2	VBF0203S1		FB031440.D	3 Feb 2025 11:47		YP/AJ	Ok
3	VBF0203S2		FB031441.D	3 Feb 2025 12:14		YP/AJ	Ok
4	BSF0203S1		FB031442.D	3 Feb 2025 12:40		YP/AJ	Ok
5	Q1232-05		FB031443.D	3 Feb 2025 13:20	Vial -B	YP/AJ	Ok
6	Q1232-09		FB031444.D	3 Feb 2025 13:47	Vial -B	YP/AJ	Ok
7	Q1232-13		FB031445.D	3 Feb 2025 14:14	Vial -B	YP/AJ	Ok
8	Q1232-17		FB031446.D	3 Feb 2025 14:40	Vial -B	YP/AJ	Ok
9	BSF0203S2		FB031447.D	3 Feb 2025 15:07		YP/AJ	Ok
10	20 PPB GRO STD		FB031448.D	3 Feb 2025 15:34		YP/AJ	Ok,M

M : Manual Integration