

### CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 17:39 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.79       | 4.80      | 4.70      | 4.90 | 0.01       |
| Aroclor-1242-2 (2)   | 4.81       | 4.82      | 4.72      | 4.92 | 0.01       |
| Aroclor-1242-3 (3)   | 4.87       | 4.87      | 4.77      | 4.97 | 0.00       |
| Aroclor-1242-4 (4)   | 4.99       | 4.99      | 4.89      | 5.09 | 0.00       |
| Aroclor-1242-5 (5)   | 5.64       | 5.65      | 5.55      | 5.75 | 0.01       |
| Tetrachloro-m-xylene | 3.70       | 3.70      | 3.60      | 3.80 | 0.00       |
| Decachlorobiphenyl   | 8.76       | 8.76      | 8.66      | 8.86 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 17:39 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1242-1       | (1) | 4.78       | 4.78      | 4.68      | 4.88 | 0.00       |
| Aroclor-1242-2       | (2) | 4.80       | 4.80      | 4.70      | 4.90 | 0.00       |
| Aroclor-1242-3       | (3) | 4.98       | 4.98      | 4.88      | 5.08 | 0.01       |
| Aroclor-1242-4       | (4) | 5.06       | 5.06      | 4.96      | 5.16 | 0.00       |
| Aroclor-1242-5       | (5) | 5.58       | 5.58      | 5.48      | 5.68 | 0.00       |
| Tetrachloro-m-xylene |     | 3.70       | 3.70      | 3.60      | 3.80 | 0.00       |
| Decachlorobiphenyl   |     | 8.71       | 8.71      | 8.61      | 8.81 | 0.00       |