

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO109331.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.794	4.695	4.895	529.170	500.000	5.8
Aroclor-1016-2	4.814	4.714	4.914	527.480	500.000	5.5
Aroclor-1016-3	4.870	4.770	4.970	517.980	500.000	3.6
Aroclor-1016-4	4.991	4.891	5.091	521.800	500.000	4.4
Aroclor-1016-5	5.249	5.149	5.349	546.460	500.000	9.3
Aroclor-1260-1	6.291	6.190	6.390	448.550	500.000	-10.3
Aroclor-1260-2	6.479	6.379	6.579	456.840	500.000	-8.6
Aroclor-1260-3	6.848	6.747	6.947	472.180	500.000	-5.6
Aroclor-1260-4	7.108	7.008	7.208	435.580	500.000	-12.9
Aroclor-1260-5	7.350	7.249	7.449	423.730	500.000	-15.3
Decachlorobiphenyl	8.759	8.658	8.858	39.540	50.000	-20.9
Tetrachloro-m-xylene	3.699	3.600	3.800	53.070	50.000	6.1

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO109331.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.780	4.703	4.903	573.010	500.000	14.6
Aroclor-1016-2	4.800	4.722	4.922	500.780	500.000	0.2
Aroclor-1016-3	4.976	4.898	5.098	538.260	500.000	7.7
Aroclor-1016-4	5.017	4.940	5.140	496.750	500.000	-0.7
Aroclor-1016-5	5.231	5.153	5.353	550.950	500.000	10.2
Aroclor-1260-1	6.264	6.186	6.386	483.980	500.000	-3.2
Aroclor-1260-2	6.452	6.373	6.573	475.980	500.000	-4.8
Aroclor-1260-3	6.606	6.527	6.727	473.890	500.000	-5.2
Aroclor-1260-4	7.077	6.999	7.199	438.870	500.000	-12.2
Aroclor-1260-5	7.318	7.239	7.439	441.630	500.000	-11.7
Decachlorobiphenyl	8.711	8.633	8.833	41.590	50.000	-16.8
Tetrachloro-m-xylene	3.697	3.618	3.818	54.770	50.000	9.5