

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020425\
 Data File : P0109406.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2025 08:51
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 01:43:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0020325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 04 08:55:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	3.698	3.694	172.3E6	112.4E6	22.731	21.040
2) SA Decachlor...	8.757	8.708	149.1E6	75205180	23.010	24.133

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020425\
Data File : PO109406.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2025 08:51
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 01:43:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO020325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 04 08:55:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.698 min
Delta R.T.: 0.000 min
Response: 172278820
Conc: 22.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.694 min
Delta R.T.: 0.000 min
Response: 112372269
Conc: 21.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.757 min
Delta R.T.: 0.000 min
Response: 149134574
Conc: 23.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.708 min
Delta R.T.: 0.000 min
Response: 75205180
Conc: 24.13 ng/ml