

LAB CHRONICLE

OrderID:	Q1232	OrderDate:	1/30/2025 11:55:00 AM
Client:	RU2 Engineering, LLC	Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR
Contact:	Rutu Manani	Location:	E11,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1232-01	JPP-46.2-012925	SOIL	Diesel Range Organics Gasoline Range Organics	8015D 8015D	01/29/25	01/31/25	01/31/25 01/31/25	01/30/25
Q1232-03	JPP-46.2-012925	SOIL	PCB Pesticide-TCL	8082A 8081B	01/29/25	01/31/25 01/31/25	02/01/25 01/31/25	01/30/25
Q1232-03RE	JPP-46.2-012925RE	SOIL	PCB	8082A	01/29/25	01/31/25	02/04/25	01/30/25
Q1232-04	JPP-46.2-012925	TCLP	TCLP Herbicide TCLP Pesticide	8151A 8081B	01/29/25	01/31/25 01/31/25	02/01/25 02/03/25	01/30/25
Q1232-05	JPP-46.1-012925	SOIL	Diesel Range Organics Gasoline Range Organics	8015D 8015D	01/29/25	01/31/25	01/31/25 02/03/25	01/30/25
Q1232-07	JPP-46.1-012925	SOIL	PCB Pesticide-TCL	8082A 8081B	01/29/25	01/31/25 01/31/25	02/01/25 01/31/25	01/30/25
Q1232-07RE	JPP-46.1-012925RE	SOIL	PCB	8082A	01/29/25	01/31/25	02/04/25	01/30/25
Q1232-08	JPP-46.1-012925	TCLP	TCLP Herbicide TCLP Pesticide	8151A 8081B	01/29/25	01/31/25 01/31/25	02/01/25 02/03/25	01/30/25
Q1232-09	JPP-42.1-012925	SOIL	Diesel Range Organics Gasoline Range Organics	8015D 8015D	01/29/25	01/31/25	01/31/25 02/03/25	01/30/25

LAB CHRONICLE

Q1232-11	JPP-42.1-012925	SOIL			01/29/25		01/30/25
			PCB	8082A	01/31/25	02/01/25	
			Pesticide-TCL	8081B	01/31/25	01/31/25	
Q1232-11RE	JPP-42.1-012925RE	SOIL			01/29/25		01/30/25
			PCB	8082A	01/31/25	02/04/25	
Q1232-12	JPP-42.1-012925	TCLP			01/29/25		01/30/25
			TCLP Herbicide	8151A	01/31/25	02/01/25	
			TCLP Pesticide	8081B	01/31/25	02/03/25	
Q1232-13	JPP-42.2-012925	SOIL			01/29/25		01/30/25
			Diesel Range Organics	8015D	01/31/25	01/31/25	
			Gasoline Range Organics	8015D		02/03/25	
Q1232-15	JPP-42.2-012925	SOIL			01/29/25		01/30/25
			PCB	8082A	01/31/25	02/01/25	
			Pesticide-TCL	8081B	01/31/25	01/31/25	
Q1232-15RE	JPP-42.2-012925RE	SOIL			01/29/25		01/30/25
			PCB	8082A	01/31/25	02/04/25	
Q1232-16	JPP-42.2-012925	TCLP			01/29/25		01/30/25
			TCLP Herbicide	8151A	01/31/25	02/01/25	
			TCLP Pesticide	8081B	01/31/25	02/03/25	
Q1232-17	JPP-51.1-012925	SOIL			01/29/25		01/30/25
			Diesel Range Organics	8015D	01/31/25	01/31/25	
			Gasoline Range Organics	8015D		02/03/25	
Q1232-19	JPP-51.1-012925	SOIL			01/29/25		01/30/25
			PCB	8082A	01/31/25	02/01/25	
			Pesticide-TCL	8081B	01/31/25	01/31/25	
Q1232-19RE	JPP-51.1-012925RE	SOIL			01/29/25		01/30/25
			PCB	8082A	01/31/25	02/04/25	
Q1232-20	JPP-51.1-012925	TCLP			01/29/25		01/30/25
			TCLP Herbicide	8151A	01/31/25	02/01/25	
			TCLP Pesticide	8081B	01/31/25	02/03/25	

Hit Summary Sheet
SW-846

SDG No.: Q1232

Order ID: Q1232

Client: RU2 Engineering, LLC

Project ID: NYCDDC SANTWOBR Brooklyn Bri

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : JPP-46.1-012925								
Q1232-07	JPP-46.1-012925	SOIL	Aroclor-1254	32.2		2.90	18.1	ug/kg
			Total Concentration:	32.200				
Client ID : JPP-46.1-012925RE								
Q1232-07RE	JPP-46.1-012925RE	SOIL	Aroclor-1254	32.3		2.90	18.1	ug/kg
			Total Concentration:	32.300				
Client ID : JPP-42.1-012925								
Q1232-11	JPP-42.1-012925	SOIL	Aroclor-1254	16.2	J	3.10	19.2	ug/kg
			Total Concentration:	16.200				
Client ID : JPP-42.1-012925RE								
Q1232-11RE	JPP-42.1-012925RE	SOIL	Aroclor-1254	16.6	J	3.10	19.2	ug/kg
			Total Concentration:	16.600				



SAMPLE DATA

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.2-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-03	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	89.2
Sample Wt/Vol:	30.05 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109349.D	1	01/31/25 08:15	02/01/25 00:15	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	19.0	U	3.80	19.0	ug/kg
11104-28-2	Aroclor-1221	19.0	U	7.20	19.0	ug/kg
11141-16-5	Aroclor-1232	19.0	U	3.80	19.0	ug/kg
53469-21-9	Aroclor-1242	19.0	U	3.80	19.0	ug/kg
12672-29-6	Aroclor-1248	19.0	U	8.80	19.0	ug/kg
11097-69-1	Aroclor-1254	19.0	U	3.10	19.0	ug/kg
37324-23-5	Aroclor-1262	19.0	U	5.10	19.0	ug/kg
11100-14-4	Aroclor-1268	19.0	U	3.80	19.0	ug/kg
11096-82-5	Aroclor-1260	19.0	U	3.30	19.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.6		32 - 144	108%	SPK: 20
2051-24-3	Decachlorobiphenyl	14.7		32 - 175	73%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.2-012925RE	SDG No.:	Q1232
Lab Sample ID:	Q1232-03RE	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	89.2
Sample Wt/Vol:	30.05 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109407.D	1	01/31/25 08:15	02/04/25 09:09	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	19.0	U	3.80	19.0	ug/kg
11104-28-2	Aroclor-1221	19.0	U	7.20	19.0	ug/kg
11141-16-5	Aroclor-1232	19.0	U	3.80	19.0	ug/kg
53469-21-9	Aroclor-1242	19.0	U	3.80	19.0	ug/kg
12672-29-6	Aroclor-1248	19.0	U	8.80	19.0	ug/kg
11097-69-1	Aroclor-1254	19.0	U	3.10	19.0	ug/kg
37324-23-5	Aroclor-1262	19.0	U	5.10	19.0	ug/kg
11100-14-4	Aroclor-1268	19.0	U	3.80	19.0	ug/kg
11096-82-5	Aroclor-1260	19.0	U	3.30	19.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.6		32 - 144	108%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.4		32 - 175	87%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.1-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-07	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	94
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109352.D	1	01/31/25 08:15	02/01/25 01:09	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	18.1	U	3.60	18.1	ug/kg
11104-28-2	Aroclor-1221	18.1	U	6.80	18.1	ug/kg
11141-16-5	Aroclor-1232	18.1	U	3.60	18.1	ug/kg
53469-21-9	Aroclor-1242	18.1	U	3.60	18.1	ug/kg
12672-29-6	Aroclor-1248	18.1	U	8.40	18.1	ug/kg
11097-69-1	Aroclor-1254	32.2		2.90	18.1	ug/kg
37324-23-5	Aroclor-1262	18.1	U	4.90	18.1	ug/kg
11100-14-4	Aroclor-1268	18.1	U	3.60	18.1	ug/kg
11096-82-5	Aroclor-1260	18.1	U	3.10	18.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	24.1		32 - 144	121%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.7		32 - 175	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.1-012925RE	SDG No.:	Q1232
Lab Sample ID:	Q1232-07RE	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	94
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109410.D	1	01/31/25 08:15	02/04/25 10:04	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	18.1	U	3.60	18.1	ug/kg
11104-28-2	Aroclor-1221	18.1	U	6.80	18.1	ug/kg
11141-16-5	Aroclor-1232	18.1	U	3.60	18.1	ug/kg
53469-21-9	Aroclor-1242	18.1	U	3.60	18.1	ug/kg
12672-29-6	Aroclor-1248	18.1	U	8.40	18.1	ug/kg
11097-69-1	Aroclor-1254	32.3		2.90	18.1	ug/kg
37324-23-5	Aroclor-1262	18.1	U	4.90	18.1	ug/kg
11100-14-4	Aroclor-1268	18.1	U	3.60	18.1	ug/kg
11096-82-5	Aroclor-1260	18.1	U	3.10	18.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.6		32 - 144	118%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.8		32 - 175	104%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-42.1-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-11	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	88.1
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109353.D	1	01/31/25 08:15	02/01/25 01:27	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	19.2	U	3.80	19.2	ug/kg
11104-28-2	Aroclor-1221	19.2	U	7.30	19.2	ug/kg
11141-16-5	Aroclor-1232	19.2	U	3.80	19.2	ug/kg
53469-21-9	Aroclor-1242	19.2	U	3.80	19.2	ug/kg
12672-29-6	Aroclor-1248	19.2	U	8.90	19.2	ug/kg
11097-69-1	Aroclor-1254	16.2	J	3.10	19.2	ug/kg
37324-23-5	Aroclor-1262	19.2	U	5.20	19.2	ug/kg
11100-14-4	Aroclor-1268	19.2	U	3.90	19.2	ug/kg
11096-82-5	Aroclor-1260	19.2	U	3.30	19.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.8		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.9		32 - 175	89%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-42.1-012925RE	SDG No.:	Q1232
Lab Sample ID:	Q1232-11RE	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	88.1
Sample Wt/Vol:	30.08 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109411.D	1	01/31/25 08:15	02/04/25 10:22	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	19.2	U	3.80	19.2	ug/kg
11104-28-2	Aroclor-1221	19.2	U	7.30	19.2	ug/kg
11141-16-5	Aroclor-1232	19.2	U	3.80	19.2	ug/kg
53469-21-9	Aroclor-1242	19.2	U	3.80	19.2	ug/kg
12672-29-6	Aroclor-1248	19.2	U	8.90	19.2	ug/kg
11097-69-1	Aroclor-1254	16.6	J	3.10	19.2	ug/kg
37324-23-5	Aroclor-1262	19.2	U	5.20	19.2	ug/kg
11100-14-4	Aroclor-1268	19.2	U	3.90	19.2	ug/kg
11096-82-5	Aroclor-1260	19.2	U	3.30	19.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.8		32 - 144	104%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.6		32 - 175	98%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-42.2-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-15	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	93.2
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109354.D	1	01/31/25 08:15	02/01/25 01:46	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	18.2	U	3.60	18.2	ug/kg
11104-28-2	Aroclor-1221	18.2	U	6.90	18.2	ug/kg
11141-16-5	Aroclor-1232	18.2	U	3.60	18.2	ug/kg
53469-21-9	Aroclor-1242	18.2	U	3.60	18.2	ug/kg
12672-29-6	Aroclor-1248	18.2	U	8.40	18.2	ug/kg
11097-69-1	Aroclor-1254	18.2	U	2.90	18.2	ug/kg
37324-23-5	Aroclor-1262	18.2	U	4.90	18.2	ug/kg
11100-14-4	Aroclor-1268	18.2	U	3.70	18.2	ug/kg
11096-82-5	Aroclor-1260	18.2	U	3.10	18.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.6		32 - 144	118%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.0		32 - 175	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-42.2-012925RE	SDG No.:	Q1232
Lab Sample ID:	Q1232-15RE	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	93.2
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109412.D	1	01/31/25 08:15	02/04/25 10:40	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	18.2	U	3.60	18.2	ug/kg
11104-28-2	Aroclor-1221	18.2	U	6.90	18.2	ug/kg
11141-16-5	Aroclor-1232	18.2	U	3.60	18.2	ug/kg
53469-21-9	Aroclor-1242	18.2	U	3.60	18.2	ug/kg
12672-29-6	Aroclor-1248	18.2	U	8.40	18.2	ug/kg
11097-69-1	Aroclor-1254	18.2	U	2.90	18.2	ug/kg
37324-23-5	Aroclor-1262	18.2	U	4.90	18.2	ug/kg
11100-14-4	Aroclor-1268	18.2	U	3.70	18.2	ug/kg
11096-82-5	Aroclor-1260	18.2	U	3.10	18.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.3		32 - 144	111%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.3		32 - 175	96%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-51.1-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-19	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	95.1
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109355.D	1	01/31/25 08:15	02/01/25 02:04	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	17.9	U	3.60	17.9	ug/kg
11104-28-2	Aroclor-1221	17.9	U	6.70	17.9	ug/kg
11141-16-5	Aroclor-1232	17.9	U	3.60	17.9	ug/kg
53469-21-9	Aroclor-1242	17.9	U	3.60	17.9	ug/kg
12672-29-6	Aroclor-1248	17.9	U	8.30	17.9	ug/kg
11097-69-1	Aroclor-1254	17.9	U	2.90	17.9	ug/kg
37324-23-5	Aroclor-1262	17.9	U	4.80	17.9	ug/kg
11100-14-4	Aroclor-1268	17.9	U	3.60	17.9	ug/kg
11096-82-5	Aroclor-1260	17.9	U	3.10	17.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	25.5		32 - 144	127%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.7		32 - 175	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-51.1-012925RE	SDG No.:	Q1232
Lab Sample ID:	Q1232-19RE	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	95.1
Sample Wt/Vol:	30.04 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109413.D	1	01/31/25 08:15	02/04/25 10:59	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	17.9	U	3.60	17.9	ug/kg
11104-28-2	Aroclor-1221	17.9	U	6.70	17.9	ug/kg
11141-16-5	Aroclor-1232	17.9	U	3.60	17.9	ug/kg
53469-21-9	Aroclor-1242	17.9	U	3.60	17.9	ug/kg
12672-29-6	Aroclor-1248	17.9	U	8.30	17.9	ug/kg
11097-69-1	Aroclor-1254	17.9	U	2.90	17.9	ug/kg
37324-23-5	Aroclor-1262	17.9	U	4.80	17.9	ug/kg
11100-14-4	Aroclor-1268	17.9	U	3.60	17.9	ug/kg
11096-82-5	Aroclor-1260	17.9	U	3.10	17.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	24.4		32 - 144	122%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.8		32 - 175	104%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO108981.D	PIBLK-PO108981.D	Tetrachloro-m-xylene	1	20	21.9	109		60	140
		Decachlorobiphenyl	1	20	21.6	108		60	140
		Tetrachloro-m-xylene	2	20	20.6	103		60	140
		Decachlorobiphenyl	2	20	21.9	109		60	140
I.BLK-PO109321.D	PIBLK-PO109321.D	Tetrachloro-m-xylene	1	20	20.3	102		60	140
		Decachlorobiphenyl	1	20	17.8	89		60	140
		Tetrachloro-m-xylene	2	20	19.1	95		60	140
		Decachlorobiphenyl	2	20	17.8	89		60	140
PB166412BL	PB166412BL	Tetrachloro-m-xylene	1	20	24.1	120		32	144
		Decachlorobiphenyl	1	20	21.0	105		32	175
		Tetrachloro-m-xylene	2	20	22.7	114		32	144
		Decachlorobiphenyl	2	20	21.0	105		32	175
PB166412BS	PB166412BS	Tetrachloro-m-xylene	1	20	23.1	116		32	144
		Decachlorobiphenyl	1	20	21.8	109		32	175
		Tetrachloro-m-xylene	2	20	22.0	110		32	144
		Decachlorobiphenyl	2	20	21.4	107		32	175
I.BLK-PO109335.D	PIBLK-PO109335.D	Tetrachloro-m-xylene	1	20	20.1	101		60	140
		Decachlorobiphenyl	1	20	16.1	81		60	140
		Tetrachloro-m-xylene	2	20	19.7	99		60	140
		Decachlorobiphenyl	2	20	17.1	86		60	140
I.BLK-PO109348.D	PIBLK-PO109348.D	Tetrachloro-m-xylene	1	20	20.2	101		60	140
		Decachlorobiphenyl	1	20	17.1	85		60	140
		Tetrachloro-m-xylene	2	20	19.8	99		60	140
		Decachlorobiphenyl	2	20	17.5	88		60	140
Q1232-03	JPP-46.2-012925	Tetrachloro-m-xylene	1	20	21.6	108		32	144
		Decachlorobiphenyl	1	20	13.9	70		32	175
		Tetrachloro-m-xylene	2	20	19.7	99		32	144
		Decachlorobiphenyl	2	20	14.7	73		32	175
Q1232-03MS	JPP-46.2-012925MS	Tetrachloro-m-xylene	1	20	22.5	112		32	144
		Decachlorobiphenyl	1	20	18.5	93		32	175
		Tetrachloro-m-xylene	2	20	22.2	111		32	144
		Decachlorobiphenyl	2	20	19.9	100		32	175
Q1232-03MSD	JPP-46.2-012925MSD	Tetrachloro-m-xylene	1	20	22.3	112		32	144
		Decachlorobiphenyl	1	20	18.4	92		32	175
		Tetrachloro-m-xylene	2	20	22.2	111		32	144
		Decachlorobiphenyl	2	20	20.0	100		32	175
Q1232-07	JPP-46.1-012925	Tetrachloro-m-xylene	1	20	24.1	121		32	144
		Decachlorobiphenyl	1	20	17.1	86		32	175
		Tetrachloro-m-xylene	2	20	23.0	115		32	144
		Decachlorobiphenyl	2	20	18.7	93		32	175
Q1232-11	JPP-42.1-012925	Tetrachloro-m-xylene	1	20	21.8	109		32	144

Surrogate Summary

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1232-11	JPP-42.1-012925	Decachlorobiphenyl	1	20	16.6	83		32	175
		Tetrachloro-m-xylene	2	20	20.3	101		32	144
Q1232-15	JPP-42.2-012925	Decachlorobiphenyl	2	20	17.9	89		32	175
		Tetrachloro-m-xylene	1	20	22.7	113		32	144
		Decachlorobiphenyl	1	20	16.4	82		32	175
		Tetrachloro-m-xylene	2	20	23.6	118		32	144
		Decachlorobiphenyl	2	20	18.0	90		32	175
		Tetrachloro-m-xylene	1	20	25.5	127		32	144
Q1232-19	JPP-51.1-012925	Decachlorobiphenyl	1	20	17.5	88		32	175
		Tetrachloro-m-xylene	2	20	23.4	117		32	144
		Decachlorobiphenyl	2	20	18.7	93		32	175
		Tetrachloro-m-xylene	1	20	19.8	99		60	140
I.BLK-PO109365.D	PIBLK-PO109365.D	Decachlorobiphenyl	1	20	13.4	67		60	140
		Tetrachloro-m-xylene	2	20	19.9	100		60	140
		Decachlorobiphenyl	2	20	15.1	75		60	140
		Tetrachloro-m-xylene	1	20	20.8	104		60	140
I.BLK-PO109369.D	PIBLK-PO109369.D	Decachlorobiphenyl	1	20	21.4	107		60	140
		Tetrachloro-m-xylene	2	20	20.1	101		60	140
		Decachlorobiphenyl	2	20	21.6	108		60	140
		Tetrachloro-m-xylene	1	20	22.7	114		60	140
I.BLK-PO109406.D	PIBLK-PO109406.D	Decachlorobiphenyl	1	20	23.0	115		60	140
		Tetrachloro-m-xylene	2	20	21.0	105		60	140
		Decachlorobiphenyl	2	20	24.1	121		60	140
		Tetrachloro-m-xylene	1	20	21.6	108		32	144
Q1232-03RE	JPP-46.2-012925RE	Decachlorobiphenyl	1	20	16.3	81		32	175
		Tetrachloro-m-xylene	2	20	18.4	92		32	144
		Decachlorobiphenyl	2	20	17.4	87		32	175
		Tetrachloro-m-xylene	1	20	21.0	105		32	144
Q1232-03MSRE	JPP-46.2-012925MSRE	Decachlorobiphenyl	1	20	21.1	105		32	175
		Tetrachloro-m-xylene	2	20	20.4	102		32	144
		Decachlorobiphenyl	2	20	22.9	114		32	175
		Tetrachloro-m-xylene	1	20	20.8	104		32	144
Q1232-03MSDRE	JPP-46.2-012925MSDRE	Decachlorobiphenyl	1	20	20.5	102		32	175
		Tetrachloro-m-xylene	2	20	20.4	102		32	144
		Decachlorobiphenyl	2	20	22.8	114		32	175
		Tetrachloro-m-xylene	1	20	23.6	118		32	144
Q1232-07RE	JPP-46.1-012925RE	Decachlorobiphenyl	1	20	19.4	97		32	175
		Tetrachloro-m-xylene	2	20	21.8	109		32	144
		Decachlorobiphenyl	2	20	20.8	104		32	175
		Tetrachloro-m-xylene	1	20	20.8	104		32	144
Q1232-11RE	JPP-42.1-012925RE	Decachlorobiphenyl	1	20	18.4	92		32	175

Surrogate Summary

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1232-11RE	JPP-42.1-012925RE	Tetrachloro-m-xylene	2	20	19.1	95		32	144
		Decachlorobiphenyl	2	20	19.6	98		32	175
Q1232-15RE	JPP-42.2-012925RE	Tetrachloro-m-xylene	1	20	22.3	111		32	144
		Decachlorobiphenyl	1	20	18.1	91		32	175
		Tetrachloro-m-xylene	2	20	21.5	108		32	144
		Decachlorobiphenyl	2	20	19.3	96		32	175
Q1232-19RE	JPP-51.1-012925RE	Tetrachloro-m-xylene	1	20	24.4	122		32	144
		Decachlorobiphenyl	1	20	19.4	97		32	175
		Tetrachloro-m-xylene	2	20	22.0	110		32	144
		Decachlorobiphenyl	2	20	20.8	104		32	175
I.BLK-PO109423.D	PIBLK-PO109423.D	Tetrachloro-m-xylene	1	20	22.1	111		60	140
		Decachlorobiphenyl	1	20	15.5	78		60	140
		Tetrachloro-m-xylene	2	20	21.1	105		60	140
		Decachlorobiphenyl	2	20	17.8	89		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8082A

DataFile : PO109350.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q1232-03MS	JPP-46.2-012925MS											
	AR1016	186.5	0	185	ug/kg	99				55	146	
	AR1260	186.5	0	175	ug/kg	94				45	144	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8082A

DataFile : PO109351.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q1232-03MSD	JPP-46.2-012925MSD											
	AR1016	186.5	0	183	ug/kg	98		1		55	146	20
	AR1260	186.5	0	173	ug/kg	93		1		45	144	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8082A

DataFile : PO109408.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: JPP-46.2-012925MSRE												
Q1232-03MSRE	AR1016	186.5	0	170	ug/kg	91				55	146	
	AR1260	186.5	0	170	ug/kg	91				45	144	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8082A

DataFile : PO109409.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: JPP-46.2-012925MSDRE												
Q1232-03MSDRE	AR1016	186.5	0	167	ug/kg	90		0		55	146	20
	AR1260	186.5	0	167	ug/kg	90		0		45	144	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8082A Datafile : PO109324.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB166412BS	AR1016	166.5	162	ug/kg	97				71	120		
	AR1260	166.5	149	ug/kg	89				65	130		

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB166412BL

Lab Name: CHEMTECH

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232

SAS No.: Q1232 SDG NO.: Q1232

Lab Sample ID: PB166412BL

Lab File ID: PO109323.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 01/31/2025

Date Analyzed (1): 01/31/2025

Date Analyzed (2): 01/31/2025

Time Analyzed (1): 11:35

Time Analyzed (2): 11:35

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB166412BS	PB166412BS	PO109324.D	01/31/2025	01/31/2025
JPP-46.2-012925	Q1232-03	PO109349.D	02/01/2025	02/01/2025
JPP-46.2-012925MS	Q1232-03MS	PO109350.D	02/01/2025	02/01/2025
JPP-46.2-012925MSD	Q1232-03MSD	PO109351.D	02/01/2025	02/01/2025
JPP-46.1-012925	Q1232-07	PO109352.D	02/01/2025	02/01/2025
JPP-42.1-012925	Q1232-11	PO109353.D	02/01/2025	02/01/2025
JPP-42.2-012925	Q1232-15	PO109354.D	02/01/2025	02/01/2025
JPP-51.1-012925	Q1232-19	PO109355.D	02/01/2025	02/01/2025
JPP-46.2-012925MSRE	Q1232-03MSRE	PO109408.D	02/04/2025	02/04/2025
JPP-46.2-012925MSDRE	Q1232-03MSDRE	PO109409.D	02/04/2025	02/04/2025

COMMENTS: _____



CALIBRATION SUMMARY

A
B
C
D
E
F
G

Calibration Times: 17:36 01:50

LAB FILE ID:	RT 1000 =	<u>PO108982.D</u>	RT 750 =	<u>PO108983.D</u>
	RT 500 =	PO108984.D	RT 250 =	PO108985.D
			RT 050 =	PO108986.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.76	8.76	8.76	8.76	8.76	8.76	8.66	8.86
Tetrachloro-m-xylene	3.70	3.70	3.70	3.70	3.70	3.70	3.60	3.80

A
B
C
D
E
F
G

A
B
C
D
E
F
G

Calibration Times: 17:36 01:50

ID: 0.32 (mm)

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.71	8.71	8.71	8.71	8.71	8.71	8.61	8.81
Tetrachloro-m-xylene	3.70	3.70	3.70	3.70	3.70	3.70	3.60	3.80

A
B
C
D
E
F
G

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Instrument ID: ECD_O

Calibration Date(s): 01/21/2025 01/22/2025

Calibration Times: 17:36 01:50

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO108982.D</u>		CF 750 = <u>PO108983.D</u>			
CF 500 = <u>PO108984.D</u>		CF 250 = <u>PO108985.D</u>		CF 050 = <u>PO108986.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	233065045	244410019	248250902	263810212	271896440	252286524
Aroclor-1016-2	(2)	325363679	338972765	344308556	357996712	357782160	344884774
Aroclor-1016-3	(3)	223049536	234801713	240773660	257313424	264644780	244116623
Aroclor-1016-4	(4)	175974311	184579665	189100190	198987084	205504880	190829226
Aroclor-1016-5	(5)	190134930	199311205	207066494	222947156	224348000	208761557
Aroclor-1260-1	(1)	348800771	364882765	375462458	399932624	416699380	381155600
Aroclor-1260-2	(2)	430525248	448276513	461774946	488220872	519105560	469580628
Aroclor-1260-3	(3)	359859412	378331397	386511394	408169100	424938760	391562013
Aroclor-1260-4	(4)	332481764	347292652	354683668	372951676	382377400	357957432
Aroclor-1260-5	(5)	803151986	830038244	833771178	853559688	852457760	834595771
Decachlorobiphenyl		6434348190	6661231787	6830111140	7188416880	7524624800	6927746559
Tetrachloro-m-xylene		7601276640	7501278947	7481675880	7644109480	7553803600	7556428909
Aroclor-1242-1	(1)	195196952	201389908	214304986	226003920	225308640	212440881
Aroclor-1242-2	(2)	268098195	278641713	289244872	304071824	293574460	286726213
Aroclor-1242-3	(3)	186201466	194711956	206440658	220830688	215362160	204709386
Aroclor-1242-4	(4)	146339247	152438169	160415562	170237596	164916200	158869355
Aroclor-1242-5	(5)	154300112	158580243	167963632	178401740	175909460	167031037
Decachlorobiphenyl		6093757170	6273163067	6555969800	6873050240	6988220200	6556832095
Tetrachloro-m-xylene		7302831450	7496516067	7696349800	7451780120	6586958200	7306887127
Aroclor-1248-1	(1)	145246997	150453889	157781164	170519664	173855120	159571367
Aroclor-1248-2	(2)	197265768	209686637	218963792	239048120	252482700	223489403
Aroclor-1248-3	(3)	248188978	257855635	270690240	293270108	299294900	273859972
Aroclor-1248-4	(4)	351583188	362624208	379215116	402583304	426725120	384546187
Aroclor-1248-5	(5)	245666962	251879323	262640736	279614828	288543680	265669106
Decachlorobiphenyl		6072639940	6272001520	6560139460	6944778840	7129210800	6595754112
Tetrachloro-m-xylene		7184593920	7370395840	7484351680	7653720840	7312573400	7401127136
Aroclor-1254-1	(1)	374847163	389452248	410697836	433454068	459205320	413531327
Aroclor-1254-2	(2)	325342531	339278107	358066768	381289080	405152620	361825821
Aroclor-1254-3	(3)	522141746	540493293	562618572	587400960	607950820	564121078
Aroclor-1254-4	(4)	328382284	332039301	348673262	361037032	352970440	344620464
Aroclor-1254-5	(5)	475331089	489980032	515035558	537655848	547482260	513096957
Decachlorobiphenyl		6154630290	6360969147	6655043100	6960677720	7204237800	6667111611
Tetrachloro-m-xylene		7266635340	7453080907	7677404800	7686251640	7337350200	7484144577
Aroclor-1268-1	(1)	1033690630	1036942464	1067670602	1080273708	1104982020	1064711885

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	952087675	951155156	979561020	979388044	983362060	969110791	2
Aroclor-1268-3	(3)	786406202	784932265	807859298	814292396	834332900	805564612	3
Aroclor-1268-4	(4)	327369336	328395845	341390382	351699060	363994160	342569757	5
Aroclor-1268-5	(5)	2419252053	2385425329	2430042174	2398845144	2323360200	2391384980	2
Decachlorobiphenyl		11036128070	11026162653	11389192500	11664983320	12135593800	11450412069	4
Tetrachloro-m-xylene		7546638300	7604080347	7812040940	7844017680	7509481400	7663251733	2

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1232 **SAS No.:** Q1232 **SDG NO.:** Q1232

Instrument ID: ECD_O **Calibration Date(s):** 01/21/2025 01/22/2025
Calibration Times: 17:36 01:50

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO108982.D</u>		CF 750 = <u>PO108983.D</u>			
CF 500 = <u>PO108984.D</u>		CF 250 = <u>PO108985.D</u>		CF 050 = <u>PO108986.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	145518803	142573049	166649690	172900172	181948120	11
Aroclor-1016-2	(2)	223129267	236622256	232122136	250011316	249296120	5
Aroclor-1016-3	(3)	121270732	126119179	128693642	138175620	137629920	6
Aroclor-1016-4	(4)	98703058	105288716	109289440	120650588	118660940	8
Aroclor-1016-5	(5)	130939362	136319948	141694888	155164716	153595760	7
Aroclor-1260-1	(1)	229857010	239931019	247217984	265407312	277830920	8
Aroclor-1260-2	(2)	274616892	283537992	291890314	314812824	339164440	9
Aroclor-1260-3	(3)	255547157	264550967	272530098	288370904	310931080	8
Aroclor-1260-4	(4)	210034482	216033343	222462830	233977972	245747280	6
Aroclor-1260-5	(5)	481178354	492665984	499979908	514637484	514938920	3
Decachlorobiphenyl		3156544800	3289680693	3391332080	3623383520	3732817400	7
Tetrachloro-m-xylene		5315746230	5437498093	5428207440	5462079080	5158066000	2
Aroclor-1242-1	(1)	130294727	135534199	142873858	151389892	154466220	7
Aroclor-1242-2	(2)	183791222	188978371	198913930	207089908	207312700	5
Aroclor-1242-3	(3)	100083291	104057155	110676450	116018292	114379700	6
Aroclor-1242-4	(4)	101694040	106737525	114342192	123447428	121997980	8
Aroclor-1242-5	(5)	124175329	128045065	135895568	145256248	150561360	8
Decachlorobiphenyl		2975923130	3104041173	3249737920	3444762080	3552503800	7
Tetrachloro-m-xylene		5146305370	5269389160	5422425040	5452266880	4942424000	4
Aroclor-1248-1	(1)	97508588	101198119	106540372	114549420	116802080	8
Aroclor-1248-2	(2)	135954819	143343636	151605960	164947764	174030220	10
Aroclor-1248-3	(3)	145968318	152945292	161442194	176334456	190180940	11
Aroclor-1248-4	(4)	171715835	178859555	187881420	202639424	216432800	9
Aroclor-1248-5	(5)	167823607	172638881	180053168	192035492	203508660	8
Decachlorobiphenyl		3009595400	3126029333	3260555920	3479534800	3551272200	7
Tetrachloro-m-xylene		5072748910	5224226253	5283938200	5402921120	5021628000	3
Aroclor-1254-1	(1)	253547823	264686405	278940988	295164372	313075680	8
Aroclor-1254-2	(2)	223099297	232757513	247383274	262896088	285176280	10
Aroclor-1254-3	(3)	362340991	375750797	393412880	409817176	422789620	6
Aroclor-1254-4	(4)	207964031	212728924	222445680	231289600	227332320	4
Aroclor-1254-5	(5)	302636627	313319767	329146572	343115096	345356740	6
Decachlorobiphenyl		3009418030	3125275440	3274408080	3481058080	3592269800	7
Tetrachloro-m-xylene		5157782560	5318511973	5462257400	5454366480	5168021000	3
Aroclor-1268-1	(1)	605633372	613614913	625929052	644237020	658094440	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	555793036	560357999	571244564	587032548	581198500	571125329	2
Aroclor-1268-3	(3)	443116844	448542784	458393734	471576124	481232120	460572321	3
Aroclor-1268-4	(4)	171795401	175640587	181667170	188528320	184928840	180512064	4
Aroclor-1268-5	(5)	1182764439	1191499568	1205466938	1214359780	1189190120	1196656169	1
Decachlorobiphenyl		5296716160	5425177360	5590471540	5802552800	5997123200	5622408212	5
Tetrachloro-m-xylene		5377448880	5418716973	5531134800	5564772080	5234786600	5425371867	2

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Instrument ID: ECD_O Date(s) Analyzed: 01/21/2025 01/22/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.91	3.81	4.01	100101000
		2	4.00	3.90	4.10	74407200
		3	4.08	3.98	4.18	208014000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.08	3.98	4.18	165122000
		2	4.57	4.47	4.67	90981600
		3	4.81	4.71	4.91	160152000
		4	4.99	4.89	5.09	88001400
		5	5.03	4.93	5.13	64818200
Aroclor-1262	500	1	6.85	6.75	6.95	534016000
		2	7.35	7.25	7.45	919466000
		3	7.64	7.54	7.74	364572000
		4	7.70	7.60	7.80	687626000
		5	8.20	8.10	8.30	302118000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Instrument ID: ECD_O Date(s) Analyzed: 01/21/2025 01/22/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.91	3.81	4.01	65489800
		2	4.00	3.90	4.10	49387600
		3	4.07	3.97	4.17	145160000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.07	3.97	4.17	115508000
		2	4.80	4.70	4.90	108913000
		3	4.98	4.88	5.08	59967000
		4	5.06	4.96	5.16	57262000
		5	5.23	5.13	5.33	60672000
Aroclor-1262	500	1	6.82	6.72	6.92	340358000
		2	7.32	7.22	7.42	554604000
		3	7.60	7.50	7.70	214440000
		4	7.67	7.57	7.77	391888000
		5	8.16	8.06	8.26	159461000

A
B
C
D
E
F
G

Calibration Times: 18:17 02:30

LAB FILE ID:	RT 1000 =	<u>PO109370.D</u>	RT 750 =	<u>PO109371.D</u>
	RT 500 =	PO109372.D	RT 250 =	PO109373.D
			RT 050 =	PO109374.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.76	8.76	8.76	8.76	8.76	8.76	8.66	8.86
Tetrachloro-m-xylene	3.70	3.70	3.70	3.70	3.70	3.70	3.60	3.80

A
B
C
D
E
F
G

A
B
C
D
E
F
G

Calibration Times: 18:17 02:30

LAB FILE ID:	RT 1000 =	<u>PO109370.D</u>	RT 750 =	<u>PO109371.D</u>
	RT 500 =	PO109372.D	RT 250 =	PO109373.D
			RT 050 =	PO109374.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.71	8.71	8.71	8.71	8.71	8.71	8.61	8.81
Tetrachloro-m-xylene	3.70	3.70	3.70	3.70	3.69	3.70	3.60	3.80

A
B
C
D
E
F
G

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1232 **SAS No.:** Q1232 **SDG NO.:** Q1232

Instrument ID: ECD_O **Calibration Date(s):** 02/03/2025 02/04/2025
Calibration Times: 18:17 02:30

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO109370.D</u>		CF 750 = <u>PO109371.D</u>			
CF 500 = <u>PO109372.D</u>		CF 250 = <u>PO109373.D</u>		CF 050 = <u>PO109374.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	235317786	240387277	252616446	270730944	277374780	255285447
Aroclor-1016-2	(2)	317456650	333812187	341175896	359252668	360710180	342481516
Aroclor-1016-3	(3)	222362207	229514455	240138458	257895316	262445680	242471223
Aroclor-1016-4	(4)	177068080	182121103	189950302	202822896	208610740	192114624
Aroclor-1016-5	(5)	188235094	193849961	202793534	220495896	221464260	205367749
Aroclor-1260-1	(1)	342293090	352624020	366276234	392547596	412690840	373286356
Aroclor-1260-2	(2)	423470897	430072589	448025360	476511620	512643040	458144701
Aroclor-1260-3	(3)	353148761	363197761	373408456	401006340	424222880	382996840
Aroclor-1260-4	(4)	326685301	334799956	345585554	367152080	376870660	350218710
Aroclor-1260-5	(5)	776260672	789582677	809655652	839954564	862435840	815577881
Decachlorobiphenyl		6054948770	6170848653	6385471300	6759256200	7035652800	6481235545
Tetrachloro-m-xylene		7537789720	7637854440	7762340380	7614954320	7341391600	7578866092
Aroclor-1242-1	(1)	197334599	204940145	212220936	230084932	227421240	214400370
Aroclor-1242-2	(2)	265786510	278371144	287816958	305404188	295561060	286587972
Aroclor-1242-3	(3)	185058250	194571396	202706248	219955100	215411620	203540523
Aroclor-1242-4	(4)	146700154	154078684	158957240	170855512	171944280	160507174
Aroclor-1242-5	(5)	153223867	159773192	165198400	181134604	184818260	168829665
Decachlorobiphenyl		5647313730	5850539653	6047566620	6386728720	6474172400	6081264225
Tetrachloro-m-xylene		7256950730	7471164160	7603441480	7790630080	7409049200	7506247130
Aroclor-1248-1	(1)	146962778	151915439	162826108	170942184	170822880	160693878
Aroclor-1248-2	(2)	199227110	207000031	222476092	237749956	255259300	224342498
Aroclor-1248-3	(3)	247483851	255500071	274857826	289200096	301995900	273807549
Aroclor-1248-4	(4)	352065611	361039059	382944132	400046444	414632680	382145585
Aroclor-1248-5	(5)	247656524	252385875	269070708	280954660	286070540	267227661
Decachlorobiphenyl		5609834240	5759565067	6027167560	6251093240	6308340800	5991200181
Tetrachloro-m-xylene		7204476060	7307013053	7583233240	7674308840	7202293800	7394264999
Aroclor-1254-1	(1)	389077966	394823324	411471450	435522448	465269420	419232922
Aroclor-1254-2	(2)	339809413	345654540	362716334	386275232	432365500	373364204
Aroclor-1254-3	(3)	560442568	567643307	589401626	614138652	657432700	597811771
Aroclor-1254-4	(4)	341037643	333768951	338607716	349872980	327782320	338213922
Aroclor-1254-5	(5)	489637976	493484969	513693524	532708460	530251720	511955330
Decachlorobiphenyl		5858870260	5885298800	6106847120	6350391840	6409478800	6122177364
Tetrachloro-m-xylene		7558480160	7569338440	7756963700	7879721960	7331379200	7619176692
Aroclor-1268-1	(1)	1060325183	1043050687	1065797526	1096750544	1040315180	1061247824

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	978377814	960952537	981985162	1008226020	943849700	974678247	2
Aroclor-1268-3	(3)	802971546	788973609	806486708	831416156	789167280	803803060	2
Aroclor-1268-4	(4)	317503517	316627091	334247174	358558444	384945380	342376321	9
Aroclor-1268-5	(5)	2278699068	2228567457	2261614568	2279048372	2049911940	2219568281	4
Decachlorobiphenyl		10306107350	10157126853	10447096540	10874980000	10427163200	10442494789	3
Tetrachloro-m-xylene		7845541500	7700418680	7908692460	8012095560	7236914800	7740732600	4

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Instrument ID: ECD_O

Calibration Date(s): 02/03/2025 02/04/2025

Calibration Times: 18:17 02:30

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO109370.D</u>		CF 750 = <u>PO109371.D</u>			
CF 500 = <u>PO109372.D</u>		CF 250 = <u>PO109373.D</u>		CF 050 = <u>PO109374.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	153264943	158113879	164537820	175109820	188488020	167902896
Aroclor-1016-2	(2)	215030936	222535907	230791366	244561788	247372860	232058571
Aroclor-1016-3	(3)	117766190	120601281	126178226	135206824	140293280	128009160
Aroclor-1016-4	(4)	96826511	100706524	107134058	116795180	127492600	109790975
Aroclor-1016-5	(5)	128361625	131809323	138948806	150436768	151836120	140278528
Aroclor-1260-1	(1)	225807591	232124708	242393268	261236200	274037320	247119817
Aroclor-1260-2	(2)	269418607	275070071	287381706	308186452	320338600	292079087
Aroclor-1260-3	(3)	255727834	263061365	276236828	304521580	322477280	284404977
Aroclor-1260-4	(4)	202687973	207427676	215633844	229721196	245640420	220222222
Aroclor-1260-5	(5)	468691821	473960392	485910284	503467072	521629300	490731774
Decachlorobiphenyl		2869026620	2934686067	3081400200	3274572920	3421855800	3116308321
Tetrachloro-m-xylene		5213092940	5275876587	5368678560	5502731680	5343841400	5340844233
Aroclor-1242-1	(1)	128201947	133999579	140199986	151291416	151737820	141086150
Aroclor-1242-2	(2)	180116057	187990837	193950162	207443820	202594860	194419147
Aroclor-1242-3	(3)	98415916	102862087	106686518	114914080	114890880	107553896
Aroclor-1242-4	(4)	99499055	104564516	110212460	122453692	126942680	112734481
Aroclor-1242-5	(5)	120970385	126562719	132762996	145269744	140486480	133210465
Decachlorobiphenyl		2681588160	2805605107	2963333620	3194117160	3259185200	2980765849
Tetrachloro-m-xylene		5025574020	5171478640	5249343640	5411420520	4979995200	5167562404
Aroclor-1248-1	(1)	96991072	99650249	107360952	113149160	116291680	106688623
Aroclor-1248-2	(2)	134470113	139550641	150876132	162116232	173367240	152076072
Aroclor-1248-3	(3)	143621711	148800272	161120540	172009316	185838580	162278084
Aroclor-1248-4	(4)	169090611	174042627	187731984	198272620	205371660	186901900
Aroclor-1248-5	(5)	166537611	170537543	181997908	191074516	188929520	179815420
Decachlorobiphenyl		2758948620	2860350253	3062672360	3204082240	3305205400	3038251775
Tetrachloro-m-xylene		4985063440	5058850453	5283231640	5299839520	4943532600	5114103531
Aroclor-1254-1	(1)	258812283	261928736	275087630	291465840	309297220	279318342
Aroclor-1254-2	(2)	226599173	230656979	242348974	260549028	289659560	249962743
Aroclor-1254-3	(3)	367155524	370663624	385253628	402077168	407098000	386449589
Aroclor-1254-4	(4)	211061915	207363419	212342790	220544016	190805140	208423456
Aroclor-1254-5	(5)	304600854	307888127	321851642	335906840	345186360	323086765
Decachlorobiphenyl		2911755880	3004200760	3167779400	3323943760	3409494800	3163434920
Tetrachloro-m-xylene		5244001860	5240433640	5374931480	5434663200	4985288000	5255863636
Aroclor-1268-1	(1)	590846797	581634948	597230514	624849832	608144640	600541346

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	542437119	532276172	548178670	571646760	539255300	546758804	3
Aroclor-1268-3	(3)	433110112	424724803	437297710	459430228	441635260	439239623	3
Aroclor-1268-4	(4)	170526494	166707964	173722334	184577088	170298900	173166556	4
Aroclor-1268-5	(5)	1166175763	1139506068	1161503202	1197064764	1120826220	1157015203	2
Decachlorobiphenyl		5215914320	5162812320	5395934040	5697518440	5550218000	5404479424	4
Tetrachloro-m-xylene		5450783370	5333541467	5490349720	5602153360	4795336800	5334432943	6

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Instrument ID: ECD_O Date(s) Analyzed: 02/03/2025 02/04/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.91	3.81	4.01	97534400
		2	4.00	3.90	4.10	73947200
		3	4.08	3.98	4.18	207362000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.07	3.97	4.17	165633000
		2	4.57	4.47	4.67	90172400
		3	4.81	4.71	4.91	159712000
		4	4.99	4.89	5.09	88931200
		5	5.03	4.93	5.13	64613800
Aroclor-1262	500	1	6.85	6.75	6.95	530972000
		2	7.35	7.25	7.45	920474000
		3	7.64	7.54	7.74	360214000
		4	7.70	7.60	7.80	671730000
		5	8.20	8.10	8.30	285950000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Instrument ID: ECD_O Date(s) Analyzed: 02/03/2025 02/04/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.91	3.81	4.01	65094400
		2	3.99	3.89	4.09	49658400
		3	4.07	3.97	4.17	146489000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.07	3.97	4.17	114709000
		2	4.80	4.70	4.90	107451000
		3	4.97	4.87	5.07	59119000
		4	5.06	4.96	5.16	56134200
		5	5.23	5.13	5.33	60079000
Aroclor-1262	500	1	6.82	6.72	6.92	330774000
		2	7.32	7.22	7.42	540530000
		3	7.60	7.50	7.70	218438000
		4	7.66	7.56	7.76	379560000
		5	8.16	8.06	8.26	155383000

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 09:16 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.79	4.80	4.70	4.90	0.01
Aroclor-1016-2	(2)	4.81	4.81	4.71	4.91	0.00
Aroclor-1016-3	(3)	4.87	4.87	4.77	4.97	0.00
Aroclor-1016-4	(4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1016-5	(5)	5.25	5.25	5.15	5.35	0.00
Aroclor-1260-1	(1)	6.29	6.29	6.19	6.39	0.00
Aroclor-1260-2	(2)	6.48	6.48	6.38	6.58	0.00
Aroclor-1260-3	(3)	6.85	6.85	6.75	6.95	0.00
Aroclor-1260-4	(4)	7.11	7.11	7.01	7.21	0.00
Aroclor-1260-5	(5)	7.35	7.35	7.25	7.45	0.00
Tetrachloro-m-xylene		3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl		8.76	8.76	8.66	8.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 09:16 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.80	4.70	4.90	0.02
Aroclor-1016-2	(2)	4.80	4.82	4.72	4.92	0.02
Aroclor-1016-3	(3)	4.97	5.00	4.90	5.10	0.03
Aroclor-1016-4	(4)	5.01	5.04	4.94	5.14	0.03
Aroclor-1016-5	(5)	5.23	5.25	5.15	5.35	0.02
Aroclor-1260-1	(1)	6.26	6.29	6.19	6.39	0.03
Aroclor-1260-2	(2)	6.45	6.47	6.37	6.57	0.02
Aroclor-1260-3	(3)	6.60	6.63	6.53	6.73	0.03
Aroclor-1260-4	(4)	7.08	7.10	7.00	7.20	0.03
Aroclor-1260-5	(5)	7.32	7.34	7.24	7.44	0.02
Tetrachloro-m-xylene		3.69	3.72	3.62	3.82	0.03
Decachlorobiphenyl		8.71	8.73	8.63	8.83	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO109317.D Time Analyzed: 09:16

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.793	4.695	4.895	550.790	500.000	10.2
Aroclor-1016-2	4.813	4.714	4.914	550.890	500.000	10.2
Aroclor-1016-3	4.869	4.770	4.970	543.070	500.000	8.6
Aroclor-1016-4	4.990	4.891	5.091	548.160	500.000	9.6
Aroclor-1016-5	5.247	5.149	5.349	539.660	500.000	7.9
Aroclor-1260-1	6.290	6.190	6.390	523.710	500.000	4.7
Aroclor-1260-2	6.478	6.379	6.579	517.600	500.000	3.5
Aroclor-1260-3	6.847	6.747	6.947	511.430	500.000	2.3
Aroclor-1260-4	7.107	7.008	7.208	495.890	500.000	-0.8
Aroclor-1260-5	7.349	7.249	7.449	486.540	500.000	-2.7
Decachlorobiphenyl	8.759	8.658	8.858	45.000	50.000	-10.0
Tetrachloro-m-xylene	3.698	3.600	3.800	56.080	50.000	12.2

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO109317.D Time Analyzed: 09:16

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.778	4.703	4.903	559.060	500.000	11.8
Aroclor-1016-2	4.797	4.722	4.922	502.860	500.000	0.6
Aroclor-1016-3	4.973	4.898	5.098	524.970	500.000	5.0
Aroclor-1016-4	5.014	4.940	5.140	518.430	500.000	3.7
Aroclor-1016-5	5.229	5.153	5.353	510.340	500.000	2.1
Aroclor-1260-1	6.262	6.186	6.386	502.170	500.000	0.4
Aroclor-1260-2	6.450	6.373	6.573	497.440	500.000	-0.5
Aroclor-1260-3	6.603	6.527	6.727	495.150	500.000	-1.0
Aroclor-1260-4	7.075	6.999	7.199	499.130	500.000	-0.2
Aroclor-1260-5	7.315	7.239	7.439	499.000	500.000	-0.2
Decachlorobiphenyl	8.709	8.633	8.833	44.120	50.000	-11.8
Tetrachloro-m-xylene	3.694	3.618	3.818	54.310	50.000	8.6

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 17:22 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.79	4.80	4.70	4.90	0.01
Aroclor-1016-2	(2)	4.81	4.81	4.71	4.91	0.00
Aroclor-1016-3	(3)	4.87	4.87	4.77	4.97	0.00
Aroclor-1016-4	(4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1016-5	(5)	5.25	5.25	5.15	5.35	0.00
Aroclor-1260-1	(1)	6.29	6.29	6.19	6.39	0.00
Aroclor-1260-2	(2)	6.48	6.48	6.38	6.58	0.00
Aroclor-1260-3	(3)	6.85	6.85	6.75	6.95	0.00
Aroclor-1260-4	(4)	7.11	7.11	7.01	7.21	0.00
Aroclor-1260-5	(5)	7.35	7.35	7.25	7.45	0.00
Tetrachloro-m-xylene		3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl		8.76	8.76	8.66	8.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 17:22 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.80	4.70	4.90	0.02
Aroclor-1016-2	(2)	4.80	4.82	4.72	4.92	0.02
Aroclor-1016-3	(3)	4.98	5.00	4.90	5.10	0.02
Aroclor-1016-4	(4)	5.02	5.04	4.94	5.14	0.02
Aroclor-1016-5	(5)	5.23	5.25	5.15	5.35	0.02
Aroclor-1260-1	(1)	6.26	6.29	6.19	6.39	0.03
Aroclor-1260-2	(2)	6.45	6.47	6.37	6.57	0.02
Aroclor-1260-3	(3)	6.61	6.63	6.53	6.73	0.02
Aroclor-1260-4	(4)	7.08	7.10	7.00	7.20	0.02
Aroclor-1260-5	(5)	7.32	7.34	7.24	7.44	0.02
Tetrachloro-m-xylene		3.70	3.72	3.62	3.82	0.02
Decachlorobiphenyl		8.71	8.73	8.63	8.83	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO109331.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.794	4.695	4.895	529.170	500.000	5.8
Aroclor-1016-2	4.814	4.714	4.914	527.480	500.000	5.5
Aroclor-1016-3	4.870	4.770	4.970	517.980	500.000	3.6
Aroclor-1016-4	4.991	4.891	5.091	521.800	500.000	4.4
Aroclor-1016-5	5.249	5.149	5.349	546.460	500.000	9.3
Aroclor-1260-1	6.291	6.190	6.390	448.550	500.000	-10.3
Aroclor-1260-2	6.479	6.379	6.579	456.840	500.000	-8.6
Aroclor-1260-3	6.848	6.747	6.947	472.180	500.000	-5.6
Aroclor-1260-4	7.108	7.008	7.208	435.580	500.000	-12.9
Aroclor-1260-5	7.350	7.249	7.449	423.730	500.000	-15.3
Decachlorobiphenyl	8.759	8.658	8.858	39.540	50.000	-20.9
Tetrachloro-m-xylene	3.699	3.600	3.800	53.070	50.000	6.1

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO109331.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.780	4.703	4.903	573.010	500.000	14.6
Aroclor-1016-2	4.800	4.722	4.922	500.780	500.000	0.2
Aroclor-1016-3	4.976	4.898	5.098	538.260	500.000	7.7
Aroclor-1016-4	5.017	4.940	5.140	496.750	500.000	-0.7
Aroclor-1016-5	5.231	5.153	5.353	550.950	500.000	10.2
Aroclor-1260-1	6.264	6.186	6.386	483.980	500.000	-3.2
Aroclor-1260-2	6.452	6.373	6.573	475.980	500.000	-4.8
Aroclor-1260-3	6.606	6.527	6.727	473.890	500.000	-5.2
Aroclor-1260-4	7.077	6.999	7.199	438.870	500.000	-12.2
Aroclor-1260-5	7.318	7.239	7.439	441.630	500.000	-11.7
Decachlorobiphenyl	8.711	8.633	8.833	41.590	50.000	-16.8
Tetrachloro-m-xylene	3.697	3.618	3.818	54.770	50.000	9.5

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 22:43 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.79	4.80	4.70	4.90	0.01
Aroclor-1016-2	(2)	4.81	4.81	4.71	4.91	0.00
Aroclor-1016-3	(3)	4.87	4.87	4.77	4.97	0.00
Aroclor-1016-4	(4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1016-5	(5)	5.25	5.25	5.15	5.35	0.00
Aroclor-1260-1	(1)	6.29	6.29	6.19	6.39	0.00
Aroclor-1260-2	(2)	6.48	6.48	6.38	6.58	0.00
Aroclor-1260-3	(3)	6.85	6.85	6.75	6.95	0.00
Aroclor-1260-4	(4)	7.11	7.11	7.01	7.21	0.00
Aroclor-1260-5	(5)	7.35	7.35	7.25	7.45	0.00
Tetrachloro-m-xylene		3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl		8.76	8.76	8.66	8.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 22:43 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.80	4.70	4.90	0.02
Aroclor-1016-2	(2)	4.80	4.82	4.72	4.92	0.02
Aroclor-1016-3	(3)	4.98	5.00	4.90	5.10	0.03
Aroclor-1016-4	(4)	5.02	5.04	4.94	5.14	0.02
Aroclor-1016-5	(5)	5.23	5.25	5.15	5.35	0.02
Aroclor-1260-1	(1)	6.26	6.29	6.19	6.39	0.03
Aroclor-1260-2	(2)	6.45	6.47	6.37	6.57	0.02
Aroclor-1260-3	(3)	6.61	6.63	6.53	6.73	0.02
Aroclor-1260-4	(4)	7.08	7.10	7.00	7.20	0.02
Aroclor-1260-5	(5)	7.32	7.34	7.24	7.44	0.02
Tetrachloro-m-xylene		3.70	3.72	3.62	3.82	0.02
Decachlorobiphenyl		8.71	8.73	8.63	8.83	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO109344.D Time Analyzed: 22:43

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.792	4.695	4.895	531.740	500.000	6.3
Aroclor-1016-2	4.811	4.714	4.914	533.890	500.000	6.8
Aroclor-1016-3	4.867	4.770	4.970	522.950	500.000	4.6
Aroclor-1016-4	4.988	4.891	5.091	525.960	500.000	5.2
Aroclor-1016-5	5.246	5.149	5.349	557.920	500.000	11.6
Aroclor-1260-1	6.288	6.190	6.390	505.440	500.000	1.1
Aroclor-1260-2	6.477	6.379	6.579	501.420	500.000	0.3
Aroclor-1260-3	6.846	6.747	6.947	490.000	500.000	-2.0
Aroclor-1260-4	7.106	7.008	7.208	468.410	500.000	-6.3
Aroclor-1260-5	7.348	7.249	7.449	461.330	500.000	-7.7
Decachlorobiphenyl	8.757	8.658	8.858	42.570	50.000	-14.9
Tetrachloro-m-xylene	3.698	3.600	3.800	53.450	50.000	6.9

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO109344.D Time Analyzed: 22:43

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.780	4.703	4.903	566.450	500.000	13.3
Aroclor-1016-2	4.799	4.722	4.922	503.230	500.000	0.6
Aroclor-1016-3	4.975	4.898	5.098	532.710	500.000	6.5
Aroclor-1016-4	5.016	4.940	5.140	473.450	500.000	-5.3
Aroclor-1016-5	5.230	5.153	5.353	541.140	500.000	8.2
Aroclor-1260-1	6.264	6.186	6.386	528.710	500.000	5.7
Aroclor-1260-2	6.451	6.373	6.573	516.160	500.000	3.2
Aroclor-1260-3	6.605	6.527	6.727	505.880	500.000	1.2
Aroclor-1260-4	7.076	6.999	7.199	502.910	500.000	0.6
Aroclor-1260-5	7.316	7.239	7.439	497.040	500.000	-0.6
Decachlorobiphenyl	8.710	8.633	8.833	43.440	50.000	-13.1
Tetrachloro-m-xylene	3.696	3.618	3.818	55.700	50.000	11.4

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 02/01/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 06:43 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.79	4.80	4.70	4.90	0.01
Aroclor-1016-2	(2)	4.81	4.81	4.71	4.91	0.00
Aroclor-1016-3	(3)	4.87	4.87	4.77	4.97	0.00
Aroclor-1016-4	(4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1016-5	(5)	5.25	5.25	5.15	5.35	0.00
Aroclor-1260-1	(1)	6.29	6.29	6.19	6.39	0.00
Aroclor-1260-2	(2)	6.48	6.48	6.38	6.58	0.00
Aroclor-1260-3	(3)	6.85	6.85	6.75	6.95	0.00
Aroclor-1260-4	(4)	7.11	7.11	7.01	7.21	0.00
Aroclor-1260-5	(5)	7.35	7.35	7.25	7.45	0.00
Tetrachloro-m-xylene		3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl		8.76	8.76	8.66	8.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 02/01/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 06:43 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.80	4.70	4.90	0.02
Aroclor-1016-2	(2)	4.80	4.82	4.72	4.92	0.02
Aroclor-1016-3	(3)	4.97	5.00	4.90	5.10	0.03
Aroclor-1016-4	(4)	5.02	5.04	4.94	5.14	0.03
Aroclor-1016-5	(5)	5.23	5.25	5.15	5.35	0.02
Aroclor-1260-1	(1)	6.26	6.29	6.19	6.39	0.03
Aroclor-1260-2	(2)	6.45	6.47	6.37	6.57	0.02
Aroclor-1260-3	(3)	6.60	6.63	6.53	6.73	0.03
Aroclor-1260-4	(4)	7.08	7.10	7.00	7.20	0.03
Aroclor-1260-5	(5)	7.32	7.34	7.24	7.44	0.02
Tetrachloro-m-xylene		3.70	3.72	3.62	3.82	0.03
Decachlorobiphenyl		8.71	8.73	8.63	8.83	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO109364.D Time Analyzed: 06:43

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.793	4.695	4.895	524.500	500.000	4.9
Aroclor-1016-2	4.812	4.714	4.914	530.300	500.000	6.1
Aroclor-1016-3	4.868	4.770	4.970	516.520	500.000	3.3
Aroclor-1016-4	4.989	4.891	5.091	516.180	500.000	3.2
Aroclor-1016-5	5.247	5.149	5.349	540.580	500.000	8.1
Aroclor-1260-1	6.289	6.190	6.390	422.370	500.000	-15.5
Aroclor-1260-2	6.478	6.379	6.579	408.310	500.000	-18.3
Aroclor-1260-3	6.847	6.747	6.947	379.480	500.000	-24.1
Aroclor-1260-4	7.107	7.008	7.208	388.000	500.000	-22.4
Aroclor-1260-5	7.348	7.249	7.449	382.060	500.000	-23.6
Decachlorobiphenyl	8.757	8.658	8.858	33.880	50.000	-32.2
Tetrachloro-m-xylene	3.698	3.600	3.800	53.260	50.000	6.5

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO109364.D Time Analyzed: 06:43

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.778	4.703	4.903	566.230	500.000	13.2
Aroclor-1016-2	4.798	4.722	4.922	529.800	500.000	6.0
Aroclor-1016-3	4.973	4.898	5.098	549.850	500.000	10.0
Aroclor-1016-4	5.015	4.940	5.140	471.870	500.000	-5.6
Aroclor-1016-5	5.229	5.153	5.353	533.700	500.000	6.7
Aroclor-1260-1	6.262	6.186	6.386	485.550	500.000	-2.9
Aroclor-1260-2	6.450	6.373	6.573	458.810	500.000	-8.2
Aroclor-1260-3	6.604	6.527	6.727	431.070	500.000	-13.8
Aroclor-1260-4	7.075	6.999	7.199	405.070	500.000	-19.0
Aroclor-1260-5	7.316	7.239	7.439	399.550	500.000	-20.1
Decachlorobiphenyl	8.709	8.633	8.833	37.230	50.000	-25.5
Tetrachloro-m-xylene	3.695	3.618	3.818	55.990	50.000	12.0

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 02/04/2025 Initial Calibration Date(s): 02/03/2025 02/04/2025

Continuing Calib Time: 08:32 Initial Calibration Time(s): 18:17 02:30

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-2	(2)	4.81	4.81	4.71	4.91	0.00
Aroclor-1016-3	(3)	4.87	4.87	4.77	4.97	0.00
Aroclor-1016-4	(4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1016-5	(5)	5.25	5.25	5.15	5.35	0.00
Aroclor-1260-1	(1)	6.29	6.29	6.19	6.39	0.00
Aroclor-1260-2	(2)	6.48	6.48	6.38	6.58	0.00
Aroclor-1260-3	(3)	6.85	6.85	6.75	6.95	0.00
Aroclor-1260-4	(4)	7.11	7.11	7.01	7.21	0.00
Aroclor-1260-5	(5)	7.35	7.35	7.25	7.45	0.00
Tetrachloro-m-xylene		3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl		8.76	8.76	8.66	8.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 02/04/2025 Initial Calibration Date(s): 02/03/2025 02/04/2025

Continuing Calib Time: 08:32 Initial Calibration Time(s): 18:17 02:30

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.97	4.97	4.87	5.07	0.00
Aroclor-1016-4	(4)	5.02	5.02	4.92	5.12	0.00
Aroclor-1016-5	(5)	5.23	5.23	5.13	5.33	0.00
Aroclor-1260-1	(1)	6.26	6.26	6.16	6.36	0.00
Aroclor-1260-2	(2)	6.45	6.45	6.35	6.55	0.00
Aroclor-1260-3	(3)	6.60	6.60	6.50	6.70	0.00
Aroclor-1260-4	(4)	7.08	7.07	6.97	7.17	0.00
Aroclor-1260-5	(5)	7.32	7.31	7.21	7.41	-0.01
Tetrachloro-m-xylene		3.70	3.70	3.60	3.80	0.01
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/03/2025 02/03/2025

Client Sample No.: CCAL05 Date Analyzed: 02/04/2025

Lab Sample No.: AR1660CCC500 Data File : PO109405.D Time Analyzed: 08:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.793	4.693	4.893	537.250	500.000	7.5
Aroclor-1016-2	4.812	4.713	4.913	540.650	500.000	8.1
Aroclor-1016-3	4.869	4.769	4.969	539.300	500.000	7.9
Aroclor-1016-4	4.989	4.890	5.090	535.140	500.000	7.0
Aroclor-1016-5	5.247	5.148	5.348	541.600	500.000	8.3
Aroclor-1260-1	6.289	6.190	6.390	541.930	500.000	8.4
Aroclor-1260-2	6.478	6.378	6.578	536.860	500.000	7.4
Aroclor-1260-3	6.846	6.747	6.947	540.560	500.000	8.1
Aroclor-1260-4	7.107	7.007	7.207	536.910	500.000	7.4
Aroclor-1260-5	7.349	7.249	7.449	544.100	500.000	8.8
Decachlorobiphenyl	8.758	8.658	8.858	50.980	50.000	2.0
Tetrachloro-m-xylene	3.698	3.599	3.799	55.560	50.000	11.1

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/03/2025 02/03/2025

Client Sample No.: CCAL05 Date Analyzed: 02/04/2025

Lab Sample No.: AR1660CCC500 Data File : PO109405.D Time Analyzed: 08:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.779	4.679	4.879	514.180	500.000	2.8
Aroclor-1016-2	4.798	4.698	4.898	520.680	500.000	4.1
Aroclor-1016-3	4.974	4.874	5.074	516.610	500.000	3.3
Aroclor-1016-4	5.016	4.915	5.115	509.250	500.000	1.9
Aroclor-1016-5	5.229	5.129	5.329	521.880	500.000	4.4
Aroclor-1260-1	6.262	6.162	6.362	517.160	500.000	3.4
Aroclor-1260-2	6.449	6.349	6.549	519.610	500.000	3.9
Aroclor-1260-3	6.603	6.503	6.703	494.810	500.000	-1.0
Aroclor-1260-4	7.075	6.974	7.174	511.430	500.000	2.3
Aroclor-1260-5	7.315	7.214	7.414	516.330	500.000	3.3
Decachlorobiphenyl	8.708	8.607	8.807	54.160	50.000	8.3
Tetrachloro-m-xylene	3.695	3.595	3.795	52.840	50.000	5.7

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 02/04/2025 Initial Calibration Date(s): 02/03/2025 02/04/2025

Continuing Calib Time: 13:58 Initial Calibration Time(s): 18:17 02:30

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-2	(2)	4.81	4.81	4.71	4.91	0.00
Aroclor-1016-3	(3)	4.87	4.87	4.77	4.97	0.00
Aroclor-1016-4	(4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1016-5	(5)	5.25	5.25	5.15	5.35	0.01
Aroclor-1260-1	(1)	6.29	6.29	6.19	6.39	0.00
Aroclor-1260-2	(2)	6.48	6.48	6.38	6.58	0.00
Aroclor-1260-3	(3)	6.85	6.85	6.75	6.95	0.01
Aroclor-1260-4	(4)	7.11	7.11	7.01	7.21	0.00
Aroclor-1260-5	(5)	7.35	7.35	7.25	7.45	0.00
Tetrachloro-m-xylene		3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl		8.76	8.76	8.66	8.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 02/04/2025 Initial Calibration Date(s): 02/03/2025 02/04/2025

Continuing Calib Time: 13:58 Initial Calibration Time(s): 18:17 02:30

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.97	4.97	4.87	5.07	0.00
Aroclor-1016-4	(4)	5.01	5.02	4.92	5.12	0.01
Aroclor-1016-5	(5)	5.23	5.23	5.13	5.33	0.00
Aroclor-1260-1	(1)	6.26	6.26	6.16	6.36	0.00
Aroclor-1260-2	(2)	6.45	6.45	6.35	6.55	0.00
Aroclor-1260-3	(3)	6.60	6.60	6.50	6.70	0.00
Aroclor-1260-4	(4)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-5	(5)	7.31	7.31	7.21	7.41	0.00
Tetrachloro-m-xylene		3.69	3.70	3.60	3.80	0.01
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/03/2025 02/03/2025

Client Sample No.: CCAL06 Date Analyzed: 02/04/2025

Lab Sample No.: AR1660CCC500 Data File : PO109422.D Time Analyzed: 13:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.791	4.693	4.893	458.750	500.000	-8.3
Aroclor-1016-2	4.811	4.713	4.913	458.110	500.000	-8.4
Aroclor-1016-3	4.867	4.769	4.969	448.590	500.000	-10.3
Aroclor-1016-4	4.988	4.890	5.090	439.950	500.000	-12.0
Aroclor-1016-5	5.245	5.148	5.348	437.600	500.000	-12.5
Aroclor-1260-1	6.287	6.190	6.390	343.690	500.000	-31.3
Aroclor-1260-2	6.476	6.378	6.578	344.860	500.000	-31.0
Aroclor-1260-3	6.845	6.747	6.947	340.860	500.000	-31.8
Aroclor-1260-4	7.105	7.007	7.207	323.390	500.000	-35.3
Aroclor-1260-5	7.346	7.249	7.449	316.680	500.000	-36.7
Decachlorobiphenyl	8.755	8.658	8.858	35.320	50.000	-29.4
Tetrachloro-m-xylene	3.698	3.599	3.799	48.980	50.000	-2.0

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/03/2025 02/03/2025

Client Sample No.: CCAL06 Date Analyzed: 02/04/2025

Lab Sample No.: AR1660CCC500 Data File : PO109422.D Time Analyzed: 13:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.778	4.679	4.879	474.730	500.000	-5.1
Aroclor-1016-2	4.797	4.698	4.898	478.930	500.000	-4.2
Aroclor-1016-3	4.972	4.874	5.074	466.200	500.000	-6.8
Aroclor-1016-4	5.014	4.915	5.115	453.590	500.000	-9.3
Aroclor-1016-5	5.228	5.129	5.329	466.150	500.000	-6.8
Aroclor-1260-1	6.261	6.162	6.362	383.820	500.000	-23.2
Aroclor-1260-2	6.448	6.349	6.549	385.730	500.000	-22.9
Aroclor-1260-3	6.601	6.503	6.703	356.180	500.000	-28.8
Aroclor-1260-4	7.073	6.974	7.174	353.890	500.000	-29.2
Aroclor-1260-5	7.314	7.214	7.414	353.190	500.000	-29.4
Decachlorobiphenyl	8.707	8.607	8.807	38.440	50.000	-23.1
Tetrachloro-m-xylene	3.694	3.595	3.795	52.660	50.000	5.3

Analytical Sequence

Client: RU2 Engineering, LLC	SDG No.: Q1232	
Project: NYCDDC SANTWOBR Brooklyn Bridge BI	Instrument ID: ECD_O	
GC Column: ZB-MR1	ID: 0.32 (mm)	Inst. Calib. Date(s): 01/21/2025 01/21/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	01/21/2025	17:18	PO108981.D	8.76	3.70
AR1660ICC1000	AR1660ICC1000	01/21/2025	17:36	PO108982.D	8.76	3.70
AR1660ICC750	AR1660ICC750	01/21/2025	17:54	PO108983.D	8.76	3.70
AR1660ICC500	AR1660ICC500	01/21/2025	18:13	PO108984.D	8.76	3.70
AR1660ICC250	AR1660ICC250	01/21/2025	18:31	PO108985.D	8.76	3.70
AR1660ICC050	AR1660ICC050	01/21/2025	18:49	PO108986.D	8.76	3.70
AR1221ICC500	AR1221ICC500	01/21/2025	19:07	PO108987.D	8.76	3.70
AR1232ICC500	AR1232ICC500	01/21/2025	19:26	PO108988.D	8.76	3.70
AR1242ICC1000	AR1242ICC1000	01/21/2025	19:44	PO108989.D	8.76	3.70
AR1242ICC750	AR1242ICC750	01/21/2025	20:02	PO108990.D	8.76	3.70
AR1242ICC500	AR1242ICC500	01/21/2025	20:21	PO108991.D	8.76	3.70
AR1242ICC250	AR1242ICC250	01/21/2025	20:39	PO108992.D	8.76	3.70
AR1242ICC050	AR1242ICC050	01/21/2025	20:57	PO108993.D	8.76	3.70
AR1248ICC1000	AR1248ICC1000	01/21/2025	21:16	PO108994.D	8.76	3.70
AR1248ICC750	AR1248ICC750	01/21/2025	21:34	PO108995.D	8.76	3.70
AR1248ICC500	AR1248ICC500	01/21/2025	21:52	PO108996.D	8.76	3.70
AR1248ICC250	AR1248ICC250	01/21/2025	22:10	PO108997.D	8.76	3.70
AR1248ICC050	AR1248ICC050	01/21/2025	22:29	PO108998.D	8.76	3.70
AR1254ICC1000	AR1254ICC1000	01/21/2025	22:47	PO108999.D	8.76	3.70
AR1254ICC750	AR1254ICC750	01/21/2025	23:05	PO109000.D	8.76	3.70
AR1254ICC500	AR1254ICC500	01/21/2025	23:23	PO109001.D	8.76	3.70
AR1254ICC250	AR1254ICC250	01/21/2025	23:42	PO109002.D	8.76	3.70
AR1254ICC050	AR1254ICC050	01/22/2025	00:00	PO109003.D	8.76	3.70
AR1262ICC500	AR1262ICC500	01/22/2025	00:18	PO109004.D	8.76	3.70
AR1268ICC1000	AR1268ICC1000	01/22/2025	00:37	PO109005.D	8.76	3.70
AR1268ICC750	AR1268ICC750	01/22/2025	00:55	PO109006.D	8.76	3.70
AR1268ICC500	AR1268ICC500	01/22/2025	01:13	PO109007.D	8.76	3.70
AR1268ICC250	AR1268ICC250	01/22/2025	01:31	PO109008.D	8.76	3.70
AR1268ICC050	AR1268ICC050	01/22/2025	01:50	PO109009.D	8.76	3.70
AR1660CCC500	AR1660CCC500	01/31/2025	09:16	PO109317.D	8.76	3.70
IBLK	IBLK	01/31/2025	10:29	PO109321.D	8.76	3.70
PB166412BL	PB166412BL	01/31/2025	11:35	PO109323.D	8.76	3.70
PB166412BS	PB166412BS	01/31/2025	11:53	PO109324.D	8.76	3.70
AR1660CCC500	AR1660CCC500	01/31/2025	17:22	PO109331.D	8.76	3.70
IBLK	IBLK	01/31/2025	18:35	PO109335.D	8.76	3.70
AR1660CCC500	AR1660CCC500	01/31/2025	22:43	PO109344.D	8.76	3.70
IBLK	IBLK	01/31/2025	23:57	PO109348.D	8.76	3.70
JPP-46.2-012925	Q1232-03	02/01/2025	00:15	PO109349.D	8.76	3.70
JPP-46.2-012925MS	Q1232-03MS	02/01/2025	00:33	PO109350.D	8.76	3.70
JPP-46.2-012925MSD	Q1232-03MSD	02/01/2025	00:51	PO109351.D	8.76	3.70
JPP-46.1-012925	Q1232-07	02/01/2025	01:09	PO109352.D	8.76	3.70
JPP-42.1-012925	Q1232-11	02/01/2025	01:27	PO109353.D	8.76	3.70

Analytical Sequence

JPP-42.2-012925	Q1232-15	02/01/2025	01:46	PO109354.D	8.76	3.70
JPP-51.1-012925	Q1232-19	02/01/2025	02:04	PO109355.D	8.76	3.70
AR1660CCC500	AR1660CCC500	02/01/2025	06:43	PO109364.D	8.76	3.70
LBLK	LBLK	02/01/2025	07:01	PO109365.D	8.76	3.70
LBLK	LBLK	02/03/2025	17:58	PO109369.D	8.76	3.70
AR1660ICC1000	AR1660ICC1000	02/03/2025	18:17	PO109370.D	8.76	3.70
AR1660ICC750	AR1660ICC750	02/03/2025	18:35	PO109371.D	8.76	3.70
AR1660ICC500	AR1660ICC500	02/03/2025	18:53	PO109372.D	8.76	3.70
AR1660ICC250	AR1660ICC250	02/03/2025	19:12	PO109373.D	8.76	3.70
AR1660ICC050	AR1660ICC050	02/03/2025	19:30	PO109374.D	8.76	3.70
AR1221ICC500	AR1221ICC500	02/03/2025	19:49	PO109375.D	8.76	3.70
AR1232ICC500	AR1232ICC500	02/03/2025	20:07	PO109376.D	8.76	3.70
AR1242ICC1000	AR1242ICC1000	02/03/2025	20:25	PO109377.D	8.76	3.70
AR1242ICC750	AR1242ICC750	02/03/2025	20:44	PO109378.D	8.76	3.70
AR1242ICC500	AR1242ICC500	02/03/2025	21:02	PO109379.D	8.76	3.70
AR1242ICC250	AR1242ICC250	02/03/2025	21:19	PO109380.D	8.76	3.70
AR1242ICC050	AR1242ICC050	02/03/2025	21:38	PO109381.D	8.76	3.70
AR1248ICC1000	AR1248ICC1000	02/03/2025	21:56	PO109382.D	8.76	3.70
AR1248ICC750	AR1248ICC750	02/03/2025	22:14	PO109383.D	8.76	3.70
AR1248ICC500	AR1248ICC500	02/03/2025	22:32	PO109384.D	8.76	3.70
AR1248ICC250	AR1248ICC250	02/03/2025	22:50	PO109385.D	8.76	3.70
AR1248ICC050	AR1248ICC050	02/03/2025	23:08	PO109386.D	8.76	3.70
AR1254ICC1000	AR1254ICC1000	02/03/2025	23:27	PO109387.D	8.76	3.70
AR1254ICC750	AR1254ICC750	02/03/2025	23:45	PO109388.D	8.76	3.70
AR1254ICC500	AR1254ICC500	02/04/2025	00:03	PO109389.D	8.76	3.70
AR1254ICC250	AR1254ICC250	02/04/2025	00:22	PO109390.D	8.76	3.70
AR1254ICC050	AR1254ICC050	02/04/2025	00:40	PO109391.D	8.76	3.70
AR1262ICC500	AR1262ICC500	02/04/2025	00:58	PO109392.D	8.76	3.70
AR1268ICC1000	AR1268ICC1000	02/04/2025	01:17	PO109393.D	8.76	3.70
AR1268ICC750	AR1268ICC750	02/04/2025	01:35	PO109394.D	8.76	3.70
AR1268ICC500	AR1268ICC500	02/04/2025	01:54	PO109395.D	8.76	3.70
AR1268ICC250	AR1268ICC250	02/04/2025	02:12	PO109396.D	8.76	3.70
AR1268ICC050	AR1268ICC050	02/04/2025	02:30	PO109397.D	8.76	3.70
AR1660CCC500	AR1660CCC500	02/04/2025	08:32	PO109405.D	8.76	3.70
LBLK	LBLK	02/04/2025	08:51	PO109406.D	8.76	3.70
JPP-46.2-012925RE	Q1232-03RE	02/04/2025	09:09	PO109407.D	8.76	3.70
JPP-46.2-012925MSRE	Q1232-03MSRE	02/04/2025	09:27	PO109408.D	8.76	3.70
JPP-46.2-012925MSDRE	Q1232-03MSDRE	02/04/2025	09:46	PO109409.D	8.76	3.70
JPP-46.1-012925RE	Q1232-07RE	02/04/2025	10:04	PO109410.D	8.76	3.70
JPP-42.1-012925RE	Q1232-11RE	02/04/2025	10:22	PO109411.D	8.76	3.70
JPP-42.2-012925RE	Q1232-15RE	02/04/2025	10:40	PO109412.D	8.76	3.70
JPP-51.1-012925RE	Q1232-19RE	02/04/2025	10:59	PO109413.D	8.76	3.70
AR1660CCC500	AR1660CCC500	02/04/2025	13:58	PO109422.D	8.76	3.70
LBLK	LBLK	02/04/2025	14:16	PO109423.D	8.76	3.70

Analytical Sequence

Client: RU2 Engineering, LLC	SDG No.: Q1232
Project: NYCDDC SANTWOBR Brooklyn Bridge BI	Instrument ID: ECD_O
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 01/21/2025 01/21/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	01/21/2025	17:18	PO108981.D	8.71	3.70
AR1660ICC1000	AR1660ICC1000	01/21/2025	17:36	PO108982.D	8.71	3.70
AR1660ICC750	AR1660ICC750	01/21/2025	17:54	PO108983.D	8.71	3.70
AR1660ICC500	AR1660ICC500	01/21/2025	18:13	PO108984.D	8.73	3.72
AR1660ICC250	AR1660ICC250	01/21/2025	18:31	PO108985.D	8.71	3.70
AR1660ICC050	AR1660ICC050	01/21/2025	18:49	PO108986.D	8.71	3.70
AR1221ICC500	AR1221ICC500	01/21/2025	19:07	PO108987.D	8.71	3.70
AR1232ICC500	AR1232ICC500	01/21/2025	19:26	PO108988.D	8.71	3.70
AR1242ICC1000	AR1242ICC1000	01/21/2025	19:44	PO108989.D	8.71	3.70
AR1242ICC750	AR1242ICC750	01/21/2025	20:02	PO108990.D	8.71	3.70
AR1242ICC500	AR1242ICC500	01/21/2025	20:21	PO108991.D	8.71	3.70
AR1242ICC250	AR1242ICC250	01/21/2025	20:39	PO108992.D	8.71	3.70
AR1242ICC050	AR1242ICC050	01/21/2025	20:57	PO108993.D	8.71	3.70
AR1248ICC1000	AR1248ICC1000	01/21/2025	21:16	PO108994.D	8.71	3.70
AR1248ICC750	AR1248ICC750	01/21/2025	21:34	PO108995.D	8.71	3.70
AR1248ICC500	AR1248ICC500	01/21/2025	21:52	PO108996.D	8.71	3.70
AR1248ICC250	AR1248ICC250	01/21/2025	22:10	PO108997.D	8.71	3.70
AR1248ICC050	AR1248ICC050	01/21/2025	22:29	PO108998.D	8.71	3.70
AR1254ICC1000	AR1254ICC1000	01/21/2025	22:47	PO108999.D	8.71	3.70
AR1254ICC750	AR1254ICC750	01/21/2025	23:05	PO109000.D	8.71	3.70
AR1254ICC500	AR1254ICC500	01/21/2025	23:23	PO109001.D	8.71	3.70
AR1254ICC250	AR1254ICC250	01/21/2025	23:42	PO109002.D	8.71	3.70
AR1254ICC050	AR1254ICC050	01/22/2025	00:00	PO109003.D	8.71	3.70
AR1262ICC500	AR1262ICC500	01/22/2025	00:18	PO109004.D	8.71	3.70
AR1268ICC1000	AR1268ICC1000	01/22/2025	00:37	PO109005.D	8.71	3.70
AR1268ICC750	AR1268ICC750	01/22/2025	00:55	PO109006.D	8.71	3.70
AR1268ICC500	AR1268ICC500	01/22/2025	01:13	PO109007.D	8.71	3.70
AR1268ICC250	AR1268ICC250	01/22/2025	01:31	PO109008.D	8.71	3.70
AR1268ICC050	AR1268ICC050	01/22/2025	01:50	PO109009.D	8.71	3.70
AR1660CCC500	AR1660CCC500	01/31/2025	09:16	PO109317.D	8.71	3.69
IBLK	IBLK	01/31/2025	10:29	PO109321.D	8.71	3.70
PB166412BL	PB166412BL	01/31/2025	11:35	PO109323.D	8.71	3.70
PB166412BS	PB166412BS	01/31/2025	11:53	PO109324.D	8.71	3.70
AR1660CCC500	AR1660CCC500	01/31/2025	17:22	PO109331.D	8.71	3.70
IBLK	IBLK	01/31/2025	18:35	PO109335.D	8.71	3.70
AR1660CCC500	AR1660CCC500	01/31/2025	22:43	PO109344.D	8.71	3.70
IBLK	IBLK	01/31/2025	23:57	PO109348.D	8.71	3.70
JPP-46.2-012925	Q1232-03	02/01/2025	00:15	PO109349.D	8.71	3.70
JPP-46.2-012925MS	Q1232-03MS	02/01/2025	00:33	PO109350.D	8.71	3.70
JPP-46.2-012925MSD	Q1232-03MSD	02/01/2025	00:51	PO109351.D	8.71	3.70
JPP-46.1-012925	Q1232-07	02/01/2025	01:09	PO109352.D	8.71	3.70
JPP-42.1-012925	Q1232-11	02/01/2025	01:27	PO109353.D	8.71	3.70

Analytical Sequence

JPP-42.2-012925	Q1232-15	02/01/2025	01:46	PO109354.D	8.71	3.70
JPP-51.1-012925	Q1232-19	02/01/2025	02:04	PO109355.D	8.71	3.70
AR1660CCC500	AR1660CCC500	02/01/2025	06:43	PO109364.D	8.71	3.70
LBLK	LBLK	02/01/2025	07:01	PO109365.D	8.71	3.70
LBLK	LBLK	02/03/2025	17:58	PO109369.D	8.71	3.70
AR1660ICC1000	AR1660ICC1000	02/03/2025	18:17	PO109370.D	8.71	3.70
AR1660ICC750	AR1660ICC750	02/03/2025	18:35	PO109371.D	8.71	3.70
AR1660ICC500	AR1660ICC500	02/03/2025	18:53	PO109372.D	8.71	3.70
AR1660ICC250	AR1660ICC250	02/03/2025	19:12	PO109373.D	8.71	3.70
AR1660ICC050	AR1660ICC050	02/03/2025	19:30	PO109374.D	8.71	3.69
AR1221ICC500	AR1221ICC500	02/03/2025	19:49	PO109375.D	8.71	3.70
AR1232ICC500	AR1232ICC500	02/03/2025	20:07	PO109376.D	8.71	3.70
AR1242ICC1000	AR1242ICC1000	02/03/2025	20:25	PO109377.D	8.71	3.70
AR1242ICC750	AR1242ICC750	02/03/2025	20:44	PO109378.D	8.71	3.70
AR1242ICC500	AR1242ICC500	02/03/2025	21:02	PO109379.D	8.71	3.69
AR1242ICC250	AR1242ICC250	02/03/2025	21:19	PO109380.D	8.71	3.69
AR1242ICC050	AR1242ICC050	02/03/2025	21:38	PO109381.D	8.71	3.69
AR1248ICC1000	AR1248ICC1000	02/03/2025	21:56	PO109382.D	8.71	3.69
AR1248ICC750	AR1248ICC750	02/03/2025	22:14	PO109383.D	8.71	3.69
AR1248ICC500	AR1248ICC500	02/03/2025	22:32	PO109384.D	8.71	3.69
AR1248ICC250	AR1248ICC250	02/03/2025	22:50	PO109385.D	8.71	3.69
AR1248ICC050	AR1248ICC050	02/03/2025	23:08	PO109386.D	8.71	3.69
AR1254ICC1000	AR1254ICC1000	02/03/2025	23:27	PO109387.D	8.71	3.70
AR1254ICC750	AR1254ICC750	02/03/2025	23:45	PO109388.D	8.71	3.70
AR1254ICC500	AR1254ICC500	02/04/2025	00:03	PO109389.D	8.71	3.70
AR1254ICC250	AR1254ICC250	02/04/2025	00:22	PO109390.D	8.71	3.70
AR1254ICC050	AR1254ICC050	02/04/2025	00:40	PO109391.D	8.71	3.69
AR1262ICC500	AR1262ICC500	02/04/2025	00:58	PO109392.D	8.71	3.70
AR1268ICC1000	AR1268ICC1000	02/04/2025	01:17	PO109393.D	8.71	3.70
AR1268ICC750	AR1268ICC750	02/04/2025	01:35	PO109394.D	8.71	3.70
AR1268ICC500	AR1268ICC500	02/04/2025	01:54	PO109395.D	8.71	3.70
AR1268ICC250	AR1268ICC250	02/04/2025	02:12	PO109396.D	8.71	3.70
AR1268ICC050	AR1268ICC050	02/04/2025	02:30	PO109397.D	8.71	3.69
AR1660CCC500	AR1660CCC500	02/04/2025	08:32	PO109405.D	8.71	3.70
LBLK	LBLK	02/04/2025	08:51	PO109406.D	8.71	3.69
JPP-46.2-012925RE	Q1232-03RE	02/04/2025	09:09	PO109407.D	8.71	3.70
JPP-46.2-012925MSRE	Q1232-03MSRE	02/04/2025	09:27	PO109408.D	8.71	3.69
JPP-46.2-012925MSDRE	Q1232-03MSDRE	02/04/2025	09:46	PO109409.D	8.71	3.69
JPP-46.1-012925RE	Q1232-07RE	02/04/2025	10:04	PO109410.D	8.71	3.69
JPP-42.1-012925RE	Q1232-11RE	02/04/2025	10:22	PO109411.D	8.71	3.69
JPP-42.2-012925RE	Q1232-15RE	02/04/2025	10:40	PO109412.D	8.72	3.71
JPP-51.1-012925RE	Q1232-19RE	02/04/2025	10:59	PO109413.D	8.71	3.69
AR1660CCC500	AR1660CCC500	02/04/2025	13:58	PO109422.D	8.71	3.69
LBLK	LBLK	02/04/2025	14:16	PO109423.D	8.71	3.70



QC SAMPLE DATA

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	PB166412BL	SDG No.:	Q1232
Lab Sample ID:	PB166412BL	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109323.D	1	01/31/25 08:15	01/31/25 11:35	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	17.0	U	3.40	17.0	ug/kg
11104-28-2	Aroclor-1221	17.0	U	6.40	17.0	ug/kg
11141-16-5	Aroclor-1232	17.0	U	3.40	17.0	ug/kg
53469-21-9	Aroclor-1242	17.0	U	3.40	17.0	ug/kg
12672-29-6	Aroclor-1248	17.0	U	7.90	17.0	ug/kg
11097-69-1	Aroclor-1254	17.0	U	2.70	17.0	ug/kg
37324-23-5	Aroclor-1262	17.0	U	4.60	17.0	ug/kg
11100-14-4	Aroclor-1268	17.0	U	3.40	17.0	ug/kg
11096-82-5	Aroclor-1260	17.0	U	2.90	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	24.1		32 - 144	120%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.0		32 - 175	105%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/21/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/21/25
Client Sample ID:	PIBLK-PO108981.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PO108981.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO108981.D	1		01/21/25	PO012125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.6		60 - 140	103%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.6		60 - 140	108%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/31/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/31/25
Client Sample ID:	PIBLK-PO109321.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PO109321.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109321.D	1		01/31/25	PO013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.1		60 - 140	95%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.8		60 - 140	89%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/31/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/31/25
Client Sample ID:	PIBLK-PO109335.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PO109335.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109335.D	1		01/31/25	PO013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.7		60 - 140	99%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.1		60 - 140	81%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/31/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/31/25
Client Sample ID:	PIBLK-PO109348.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PO109348.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109348.D	1		01/31/25	PO013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.8		60 - 140	99%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.1		60 - 140	85%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	02/01/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	02/01/25
Client Sample ID:	PIBLK-PO109365.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PO109365.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109365.D	1		02/01/25	PO013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.8		60 - 140	99%	SPK: 20
2051-24-3	Decachlorobiphenyl	13.4		60 - 140	67%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	02/03/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	02/03/25
Client Sample ID:	PIBLK-PO109369.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PO109369.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109369.D	1		02/03/25	PO020325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.1		60 - 140	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.4		60 - 140	107%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	02/04/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	02/04/25
Client Sample ID:	PIBLK-PO109406.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PO109406.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109406.D	1		02/04/25	PO020425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.0		60 - 140	105%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.0		60 - 140	115%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	02/04/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	02/04/25
Client Sample ID:	PIBLK-PO109423.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PO109423.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109423.D	1		02/04/25	PO020425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.1		60 - 140	105%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.5		60 - 140	78%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	PB166412BS	SDG No.:	Q1232
Lab Sample ID:	PB166412BS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109324.D	1	01/31/25 08:15	01/31/25 11:53	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	162		3.40	17.0	ug/kg
11104-28-2	Aroclor-1221	17.0	U	6.40	17.0	ug/kg
11141-16-5	Aroclor-1232	17.0	U	3.40	17.0	ug/kg
53469-21-9	Aroclor-1242	17.0	U	3.40	17.0	ug/kg
12672-29-6	Aroclor-1248	17.0	U	7.90	17.0	ug/kg
11097-69-1	Aroclor-1254	17.0	U	2.70	17.0	ug/kg
37324-23-5	Aroclor-1262	17.0	U	4.60	17.0	ug/kg
11100-14-4	Aroclor-1268	17.0	U	3.40	17.0	ug/kg
11096-82-5	Aroclor-1260	149		2.90	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.1		32 - 144	116%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.8		32 - 175	109%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.2-012925MS	SDG No.:	Q1232
Lab Sample ID:	Q1232-03MS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	89.2
Sample Wt/Vol:	30.06 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109350.D	1	01/31/25 08:15	02/01/25 00:33	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	185		3.80	19.0	ug/kg
11104-28-2	Aroclor-1221	19.0	U	7.20	19.0	ug/kg
11141-16-5	Aroclor-1232	19.0	U	3.80	19.0	ug/kg
53469-21-9	Aroclor-1242	19.0	U	3.80	19.0	ug/kg
12672-29-6	Aroclor-1248	19.0	U	8.80	19.0	ug/kg
11097-69-1	Aroclor-1254	19.0	U	3.10	19.0	ug/kg
37324-23-5	Aroclor-1262	19.0	U	5.10	19.0	ug/kg
11100-14-4	Aroclor-1268	19.0	U	3.80	19.0	ug/kg
11096-82-5	Aroclor-1260	175		3.30	19.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.5		32 - 144	112%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.9		32 - 175	100%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.2-012925MSRE	SDG No.:	Q1232
Lab Sample ID:	Q1232-03MSRE	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	89.2
Sample Wt/Vol:	30.06 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109408.D	1	01/31/25 08:15	02/04/25 09:27	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	170		3.80	19.0	ug/kg
11104-28-2	Aroclor-1221	19.0	U	7.20	19.0	ug/kg
11141-16-5	Aroclor-1232	19.0	U	3.80	19.0	ug/kg
53469-21-9	Aroclor-1242	19.0	U	3.80	19.0	ug/kg
12672-29-6	Aroclor-1248	19.0	U	8.80	19.0	ug/kg
11097-69-1	Aroclor-1254	19.0	U	3.10	19.0	ug/kg
37324-23-5	Aroclor-1262	19.0	U	5.10	19.0	ug/kg
11100-14-4	Aroclor-1268	19.0	U	3.80	19.0	ug/kg
11096-82-5	Aroclor-1260	170		3.30	19.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.0		32 - 144	105%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.9		32 - 175	114%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.2-012925MSD	SDG No.:	Q1232
Lab Sample ID:	Q1232-03MSD	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	89.2
Sample Wt/Vol:	30.05 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109351.D	1	01/31/25 08:15	02/01/25 00:51	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	183		3.80	19.0	ug/kg
11104-28-2	Aroclor-1221	19.0	U	7.20	19.0	ug/kg
11141-16-5	Aroclor-1232	19.0	U	3.80	19.0	ug/kg
53469-21-9	Aroclor-1242	19.0	U	3.80	19.0	ug/kg
12672-29-6	Aroclor-1248	19.0	U	8.80	19.0	ug/kg
11097-69-1	Aroclor-1254	19.0	U	3.10	19.0	ug/kg
37324-23-5	Aroclor-1262	19.0	U	5.10	19.0	ug/kg
11100-14-4	Aroclor-1268	19.0	U	3.80	19.0	ug/kg
11096-82-5	Aroclor-1260	173		3.30	19.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.3		32 - 144	112%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.0		32 - 175	100%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.2-012925MSDRE	SDG No.:	Q1232
Lab Sample ID:	Q1232-03MSDRE	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	89.2
Sample Wt/Vol:	30.05 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109409.D	1	01/31/25 08:15	02/04/25 09:46	PB166412

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	167		3.80	19.0	ug/kg
11104-28-2	Aroclor-1221	19.0	U	7.20	19.0	ug/kg
11141-16-5	Aroclor-1232	19.0	U	3.80	19.0	ug/kg
53469-21-9	Aroclor-1242	19.0	U	3.80	19.0	ug/kg
12672-29-6	Aroclor-1248	19.0	U	8.80	19.0	ug/kg
11097-69-1	Aroclor-1254	19.0	U	3.10	19.0	ug/kg
37324-23-5	Aroclor-1262	19.0	U	5.10	19.0	ug/kg
11100-14-4	Aroclor-1268	19.0	U	3.80	19.0	ug/kg
11096-82-5	Aroclor-1260	167		3.30	19.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.8		32 - 144	104%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.8		32 - 175	114%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit