

DATA PACKAGE
GC SEMI-VOLATILES

PROJECT NAME : NYCDDC SANTWOBR BROOKLYN BRIDGE BBMCR

RU2 ENGINEERING, LLC
2 Melinda Drive

Monroe Township, NJ - 08831
Phone No: 732-261-2236

ORDER ID : Q1232
ATTENTION : Rutu Manani



Laboratory Certification ID # 20012



1) PESTICIDE-TCL Data	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	13
9) QC Data Summary For Pesticide-TCL	15
9.1) Deuterated Monitoring Compound Summary	16
9.2) MS/MSD Summary	18
9.3) LCS/LCSD Summary	20
9.4) Method Blank Summary	21
10) Sample Data	22
10.1) JPP-46.2-012925	23
10.2) JPP-46.1-012925	29
10.3) JPP-42.1-012925	38
10.4) JPP-42.2-012925	45
10.5) JPP-51.1-012925	53
11) Calibration Data Summary	62
11.1) Initial Calibration Data	63
11.1.1) PL012125	63
11.2) Continued Calibration Data	165
11.2.1) PL093930.D	165
11.2.2) PL093943.D	183
11.2.3) PL093959.D	201
11.2.4) PL093971.D	219
11.2.5) PEM Files	237
11.3) RESCHK Data	267
11.4) Analytical Seq	275
12) Compound Detection Summary	277
13) QC Sample Data	288
13.1) Method Blank Data	289
13.2) PIBLK Data	294
13.3) LCS Data	319

Table Of Contents for Q1232

13.4) MS Data	335
13.5) MSD Data	351
14) Manual Integration	367
15) Analytical Runlogs	375
16) Percent Solid	385
17) Extraction Logs	389
17.1) PB166413.pdf	389
17.2) PB166413IC.pdf	391
18) Standard Prep Logs	392
19) Shipping Document	438
19.1) Chain Of Custody	439
19.2) Lab Certificate	440
19.3) Internal COC	441

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Cover Page

Order ID : Q1232**Project ID :** NYCDDC SANTWOBR Brooklyn Bridge BBMCR**Client :** RU2 Engineering, LLC**Lab Sample Number**

Q1232-01
Q1232-02
Q1232-03
Q1232-04
Q1232-05
Q1232-06
Q1232-07
Q1232-08
Q1232-09
Q1232-10
Q1232-11
Q1232-12
Q1232-13
Q1232-14
Q1232-15
Q1232-16
Q1232-17
Q1232-18
Q1232-19
Q1232-20

Client Sample Number

JPP-46.2-012925
JPP-46.2-012925
JPP-46.2-012925
JPP-46.2-012925
JPP-46.1-012925
JPP-46.1-012925
JPP-46.1-012925
JPP-46.1-012925
JPP-42.1-012925
JPP-42.1-012925
JPP-42.1-012925
JPP-42.1-012925
JPP-42.2-012925
JPP-42.2-012925
JPP-42.2-012925
JPP-42.2-012925
JPP-51.1-012925
JPP-51.1-012925
JPP-51.1-012925
JPP-51.1-012925

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 2/8/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

RU2 Engineering, LLC

Project Name: NYCDDC SANTWOBR Brooklyn Bridge BBMCR

Project # N/A

Chemtech Project # Q1232

Test Name: Pesticide-TCL

A. Number of Samples and Date of Receipt:

20 Solid samples were received on 01/30/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Diesel Range Organics, Gasoline Range Organics, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, Paint Filter, PCB, Pesticide-TCL, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL and VOCMS Group1. This data package contains results for Pesticide-TCL.

C. Analytical Techniques:

The analysis was performed on instrument ECD_L. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df,; Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of Pesticide-TCLs was based on method 8081B and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for JPP-46.2-012925 [Decachlorobiphenyl(2) - 152%] as per method one surrogate is allowed to fail therefore no corrective action taken.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds .

The MSD recoveries met the acceptable requirements .

The RPD for {Q1239-10MSD} with File ID: PL093961.D met criteria except for Endrin aldehyde[30%] due to difference in results of MS and MSD.

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

For samples # JPP-46.1-012925 and JPP-51.1-012925 compound # delta-BHC, below Method detection limits, therefore it is not reported as Hit in Form-1.

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1232

MATRIX: Solid

METHOD: 8081B/3541

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified.			✓
2. Standard Summary Submitted.			✓
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD. The Initial Calibration met the requirements . The Continuous Calibration met the requirements .			✓
4. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. The Surrogate recoveries met the acceptable criteria except for JPP-46.2-012925 [Decachlorobiphenyl(2) - 152%] as per method one surrogate is allowed to fail therefore no corrective action taken.		✓	
6. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The MS recoveries met the requirements for all compounds . The MSD recoveries met the acceptable requirements . The Blank Spike met requirements for all samples . The RPD for {Q1239-10MSD} with File ID: PL093961.D met criteria except for Endrin aldehyde[30%] due to difference in results of MS and MSD.		✓	
7. Retention Time Shift Meet Criteria (if applicable) Comments:			✓
8. Extraction Holding Time Met If not met, list number of days exceeded for each sample:			✓



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

	NA	NO	YES
9. Analysis Holding Time Met			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			

ADDITIONAL COMMENTS:

For samples # JPP-46.1-012925 and JPP-51.1-012925 compound # delta-BHC, below Method detection limits, therefore it is not reported as Hit in Form-1.

The soil samples results are based on a dry weight basis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1232

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 02/08/2025

LAB CHRONICLE

OrderID:	Q1232	OrderDate:	1/30/2025 11:55:00 AM
Client:	RU2 Engineering, LLC	Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR
Contact:	Rutu Manani	Location:	E11,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1232-01	JPP-46.2-012925	SOIL	Diesel Range Organics Gasoline Range Organics	8015D 8015D	01/29/25	01/31/25	01/31/25 01/31/25	01/30/25
Q1232-03	JPP-46.2-012925	SOIL	Pesticide-TCL	8081B	01/29/25	01/31/25	01/31/25	01/30/25
Q1232-04	JPP-46.2-012925	TCLP	TCLP Pesticide	8081B	01/29/25	01/31/25	02/03/25	01/30/25
Q1232-05	JPP-46.1-012925	SOIL	Diesel Range Organics Gasoline Range Organics	8015D 8015D	01/29/25	01/31/25	01/31/25 02/03/25	01/30/25
Q1232-07	JPP-46.1-012925	SOIL	Pesticide-TCL	8081B	01/29/25	01/31/25	01/31/25	01/30/25
Q1232-08	JPP-46.1-012925	TCLP	TCLP Pesticide	8081B	01/29/25	01/31/25	02/03/25	01/30/25
Q1232-09	JPP-42.1-012925	SOIL	Diesel Range Organics Gasoline Range Organics	8015D 8015D	01/29/25	01/31/25	01/31/25 02/03/25	01/30/25
Q1232-11	JPP-42.1-012925	SOIL	Pesticide-TCL	8081B	01/29/25	01/31/25	01/31/25	01/30/25
Q1232-12	JPP-42.1-012925	TCLP	TCLP Pesticide	8081B	01/29/25	01/31/25	02/03/25	01/30/25
Q1232-13	JPP-42.2-012925	SOIL	Diesel Range Organics Gasoline Range Organics	8015D 8015D	01/29/25	01/31/25	01/31/25 02/03/25	01/30/25

LAB CHRONICLE

Q1232-15	JPP-42.2-012925	SOIL			01/29/25		01/30/25
			Pesticide-TCL	8081B		01/31/25	01/31/25
Q1232-16	JPP-42.2-012925	TCLP			01/29/25		01/30/25
			TCLP Pesticide	8081B		01/31/25	02/03/25
Q1232-17	JPP-51.1-012925	SOIL			01/29/25		01/30/25
			Diesel Range Organics	8015D		01/31/25	01/31/25
			Gasoline Range Organics	8015D			02/03/25
Q1232-19	JPP-51.1-012925	SOIL			01/29/25		01/30/25
			Pesticide-TCL	8081B		01/31/25	01/31/25
Q1232-20	JPP-51.1-012925	TCLP			01/29/25		01/30/25
			TCLP Pesticide	8081B		01/31/25	02/03/25

Hit Summary Sheet SW-846

SDG No.: Q1232

Order ID: Q1232

Client: RU2 Engineering, LLC

Project ID: NYCDDC SANTWOBR Brooklyn Bri

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : JPP-46.2-012925								
Q1232-03	JPP-46.2-012925	SOIL	Endosulfan II	2.90	P	0.34	1.90	ug/kg
Q1232-03	JPP-46.2-012925	SOIL	4,4-DDD	2.80	P	0.21	1.90	ug/kg
Total Concentration:				5.700				
Client ID : JPP-46.1-012925								
Q1232-07	JPP-46.1-012925	SOIL	Heptachlor	0.22	J	0.18	1.80	ug/kg
Q1232-07	JPP-46.1-012925	SOIL	Heptachlor epoxide	0.61	J	0.24	1.80	ug/kg
Q1232-07	JPP-46.1-012925	SOIL	4,4-DDE	0.43	J	0.14	1.80	ug/kg
Q1232-07	JPP-46.1-012925	SOIL	Endrin	0.60	JP	0.17	1.80	ug/kg
Q1232-07	JPP-46.1-012925	SOIL	4,4-DDT	1.60	JP	0.18	1.80	ug/kg
Q1232-07	JPP-46.1-012925	SOIL	alpha-Chlordane	1.20	JP	0.18	1.80	ug/kg
Total Concentration:				4.660				
Client ID : JPP-42.1-012925								
Q1232-11	JPP-42.1-012925	SOIL	Heptachlor epoxide	0.34	J	0.26	1.90	ug/kg
Q1232-11	JPP-42.1-012925	SOIL	Endrin	0.25	J	0.18	1.90	ug/kg
Q1232-11	JPP-42.1-012925	SOIL	4,4-DDT	1.10	JP	0.19	1.90	ug/kg
Q1232-11	JPP-42.1-012925	SOIL	alpha-Chlordane	1.10	JP	0.19	1.90	ug/kg
Total Concentration:				2.790				
Client ID : JPP-42.2-012925								
Q1232-15	JPP-42.2-012925	SOIL	Heptachlor	0.73	JP	0.18	1.80	ug/kg
Q1232-15	JPP-42.2-012925	SOIL	Heptachlor epoxide	0.28	J	0.25	1.80	ug/kg
Q1232-15	JPP-42.2-012925	SOIL	Endrin	0.20	J	0.17	1.80	ug/kg
Q1232-15	JPP-42.2-012925	SOIL	4,4-DDT	1.10	JP	0.18	1.80	ug/kg
Q1232-15	JPP-42.2-012925	SOIL	alpha-Chlordane	2.30	P	0.18	1.80	ug/kg
Q1232-15	JPP-42.2-012925	SOIL	gamma-Chlordane	1.20	JP	0.20	1.80	ug/kg
Total Concentration:				5.810				
Client ID : JPP-51.1-012925								
Q1232-19	JPP-51.1-012925	SOIL	Heptachlor	0.44	JP	0.18	1.80	ug/kg
Q1232-19	JPP-51.1-012925	SOIL	Heptachlor epoxide	0.78	JP	0.24	1.80	ug/kg
Q1232-19	JPP-51.1-012925	SOIL	Endrin	0.49	JP	0.17	1.80	ug/kg
Q1232-19	JPP-51.1-012925	SOIL	4,4-DDD	0.66	J	0.20	1.80	ug/kg
Q1232-19	JPP-51.1-012925	SOIL	4,4-DDT	2.00	P	0.18	1.80	ug/kg
Q1232-19	JPP-51.1-012925	SOIL	alpha-Chlordane	2.90	P	0.18	1.80	ug/kg
Total Concentration:				7.270				

Hit Summary Sheet
SW-846

SDG No.: Q1232

Order ID: Q1232

Client: RU2 Engineering, LLC

Project ID: NYCDDC SANTWOBR Brooklyn Bri

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------



QC SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Surrogate Summary

SDG No.: **Q1232**

Client: **RU2 Engineering, LLC**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL093928.D	PIBLK-PL093928.D	Decachlorobiphenyl	1	20	20.6	103		43	140
		Tetrachloro-m-xylene	1	20	21.1	106		77	126
		Decachlorobiphenyl	2	20	20.4	102		43	140
		Tetrachloro-m-xylene	2	20	19.8	99		77	126
PB166413BL	PB166413BL	Decachlorobiphenyl	1	20	24.6	123		10	148
		Tetrachloro-m-xylene	1	20	23.8	119		10	159
		Decachlorobiphenyl	2	20	24.1	121		10	148
		Tetrachloro-m-xylene	2	20	22.1	110		10	159
PB166413BS	PB166413BS	Decachlorobiphenyl	1	20	21.2	106		10	148
		Tetrachloro-m-xylene	1	20	20.0	100		10	159
		Decachlorobiphenyl	2	20	21.7	109		10	148
		Tetrachloro-m-xylene	2	20	18.6	93		10	159
I.BLK-PL093942.D	PIBLK-PL093942.D	Decachlorobiphenyl	1	20	22.6	113		43	140
		Tetrachloro-m-xylene	1	20	21.7	109		77	126
		Decachlorobiphenyl	2	20	22.3	111		43	140
		Tetrachloro-m-xylene	2	20	20.1	101		77	126
Q1232-03	JPP-46.2-012925	Decachlorobiphenyl	1	20	24.1	120		10	148
		Tetrachloro-m-xylene	1	20	18.1	91		10	159
		Decachlorobiphenyl	2	20	30.3	152	*	10	148
		Tetrachloro-m-xylene	2	20	22.0	110		10	159
Q1232-07	JPP-46.1-012925	Decachlorobiphenyl	1	20	20.4	102		10	148
		Tetrachloro-m-xylene	1	20	16.1	80		10	159
		Decachlorobiphenyl	2	20	12.8	64		10	148
		Tetrachloro-m-xylene	2	20	15.4	77		10	159
Q1232-11	JPP-42.1-012925	Decachlorobiphenyl	1	20	16.1	80		10	148
		Tetrachloro-m-xylene	1	20	14.7	74		10	159
		Decachlorobiphenyl	2	20	14.8	74		10	148
		Tetrachloro-m-xylene	2	20	15.6	78		10	159
Q1232-15	JPP-42.2-012925	Decachlorobiphenyl	1	20	16.6	83		10	148
		Tetrachloro-m-xylene	1	20	15.7	79		10	159
		Decachlorobiphenyl	2	20	13.3	66		10	148
		Tetrachloro-m-xylene	2	20	17.5	88		10	159
Q1232-19	JPP-51.1-012925	Decachlorobiphenyl	1	20	16.1	81		10	148
		Tetrachloro-m-xylene	1	20	16.3	81		10	159
		Decachlorobiphenyl	2	20	14.1	71		10	148
		Tetrachloro-m-xylene	2	20	16.4	82		10	159
I.BLK-PL093957.D	PIBLK-PL093957.D	Decachlorobiphenyl	1	20	20.4	102		43	140

Surrogate Summary

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093957.D	PIBLK-PL093957.D	Tetrachloro-m-xylene	1	20	22.0	110		77	126
		Decachlorobiphenyl	2	20	17.1	86		43	140
Q1239-10MS	357MS	Tetrachloro-m-xylene	2	20	20.8	104		77	126
		Decachlorobiphenyl	1	20	14.9	75		10	148
		Tetrachloro-m-xylene	1	20	19.1	95		10	159
		Decachlorobiphenyl	2	20	13.7	69		10	148
Q1239-10MSD	357MSD	Tetrachloro-m-xylene	2	20	18.0	90		10	159
		Decachlorobiphenyl	1	20	15.4	77		10	148
		Tetrachloro-m-xylene	1	20	18.7	93		10	159
		Decachlorobiphenyl	2	20	13.6	68		10	148
I.BLK-PL093970.D	PIBLK-PL093970.D	Tetrachloro-m-xylene	2	20	17.9	90		10	159
		Decachlorobiphenyl	1	20	19.3	96		43	140
		Tetrachloro-m-xylene	1	20	21.3	107		77	126
		Decachlorobiphenyl	2	20	16.1	81		43	140
		Tetrachloro-m-xylene	2	20	20.0	100		77	126

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8081B

DataFile : PL093960.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	357MS											
Q1239-10MS	alpha-BHC	19.63	0	19.6	ug/kg	100				60	144	
	beta-BHC	19.63	0	19.9	ug/kg	101				54	143	
	delta-BHC	19.63	0	19.1	ug/kg	97				47	144	
	gamma-BHC (Lindane)	19.63	0	19.2	ug/kg	98				61	140	
	Heptachlor	19.63	0	19.1	ug/kg	97				63	135	
	Aldrin	19.63	0	19.3	ug/kg	98				49	139	
	Heptachlor epoxide	19.63	0	20.0	ug/kg	102				32	180	
	Endosulfan I	19.63	0	19.0	ug/kg	97				56	142	
	Dieldrin	19.63	0	20.5	ug/kg	104				47	161	
	4,4'-DDE	19.63	0	21.4	ug/kg	109				55	136	
	Endrin	19.63	0	20.2	ug/kg	103				57	139	
	Endosulfan II	19.63	0	19.5	ug/kg	99				40	163	
	4,4'-DDD	19.63	0.28	22.0	ug/kg	111				37	192	
	Endosulfan sulfate	19.63	0	18.4	ug/kg	94				62	139	
	4,4'-DDT	19.63	0	18.9	ug/kg	96				51	146	
	Methoxychlor	19.63	0	17.5	ug/kg	89				54	136	
	Endrin ketone	19.63	0	17.5	ug/kg	89				60	129	
	Endrin aldehyde	19.63	0	13.7	ug/kg	70				59	132	
	alpha-Chlordane	19.63	0	20.4	ug/kg	104				30	192	
	gamma-Chlordane	19.63	0	21.0	ug/kg	107				44	175	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8081B

DataFile : PL093961.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q1239-10MSD	357MSD											
	alpha-BHC	19.65	0	19.8	ug/kg	101		1		60	144	20
	beta-BHC	19.65	0	20.1	ug/kg	102		1		54	143	20
	delta-BHC	19.65	0	20.0	ug/kg	102		5		47	144	20
	gamma-BHC (Lindane)	19.65	0	19.4	ug/kg	99		1		61	140	20
	Heptachlor	19.65	0	20.3	ug/kg	103		6		63	135	20
	Aldrin	19.65	0	19.5	ug/kg	99		1		49	139	20
	Heptachlor epoxide	19.65	0	20.0	ug/kg	102		0		32	180	20
	Endosulfan I	19.65	0	19.5	ug/kg	99		2		56	142	20
	Dieldrin	19.65	0	20.6	ug/kg	105		1		47	161	20
	4,4'-DDE	19.65	0	21.4	ug/kg	109		0		55	136	20
	Endrin	19.65	0	20.9	ug/kg	106		3		57	139	20
	Endosulfan II	19.65	0	20.1	ug/kg	102		3		40	163	20
	4,4'-DDD	19.65	0.28	22.2	ug/kg	112		1		37	192	20
	Endosulfan sulfate	19.65	0	19.7	ug/kg	100		6		62	139	20
	4,4'-DDT	19.65	0	20.5	ug/kg	104		8		51	146	20
	Methoxychlor	19.65	0	19.8	ug/kg	101		13		54	136	20
	Endrin ketone	19.65	0	18.8	ug/kg	96		8		60	129	20
	Endrin aldehyde	19.65	0	18.7	ug/kg	95		30	*	59	132	20
	alpha-Chlordane	19.65	0	20.6	ug/kg	105		1		30	192	20
	gamma-Chlordane	19.65	0	20.9	ug/kg	106		1		44	175	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Datafile : PL093941.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB166413BS	alpha-BHC	16.66	14.9	ug/kg	89				84	123	
	beta-BHC	16.66	15.1	ug/kg	91				82	123	
	delta-BHC	16.66	15.1	ug/kg	91				83	126	
	gamma-BHC (Lindane)	16.66	14.8	ug/kg	89				83	125	
	Heptachlor	16.66	16.0	ug/kg	96				83	122	
	Aldrin	16.66	15.0	ug/kg	90				82	124	
	Heptachlor epoxide	16.66	15.3	ug/kg	92				83	120	
	Endosulfan I	16.66	15.4	ug/kg	92				81	124	
	Dieldrin	16.66	15.5	ug/kg	93				85	121	
	4,4'-DDE	16.66	16.5	ug/kg	99				81	123	
	Endrin	16.66	16.8	ug/kg	101				76	130	
	Endosulfan II	16.66	16.2	ug/kg	97				80	125	
	4,4'-DDD	16.66	17.1	ug/kg	103				80	131	
	Endosulfan sulfate	16.66	16.2	ug/kg	97				81	122	
	4,4'-DDT	16.66	17.5	ug/kg	105				70	129	
	Methoxychlor	16.66	17.2	ug/kg	103				60	119	
	Endrin ketone	16.66	15.7	ug/kg	94				77	132	
	Endrin aldehyde	16.66	15.3	ug/kg	92				79	124	
	alpha-Chlordane	16.66	15.7	ug/kg	94				84	120	
	gamma-Chlordane	16.66	15.7	ug/kg	94				83	122	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB166413BL

Lab Name: CHEMTECH

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232

SAS No.: Q1232 SDG NO.: Q1232

Lab Sample ID: PB166413BL

Lab File ID: PL093940.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 01/31/2025

Date Analyzed (1): 01/31/2025

Date Analyzed (2): 01/31/2025

Time Analyzed (1): 14:28

Time Analyzed (2): 14:28

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB166413BS	PB166413BS	PL093941.D	01/31/2025	01/31/2025
JPP-46.2-012925	Q1232-03	PL093946.D	01/31/2025	01/31/2025
JPP-46.1-012925	Q1232-07	PL093947.D	01/31/2025	01/31/2025
JPP-42.1-012925	Q1232-11	PL093948.D	01/31/2025	01/31/2025
JPP-42.2-012925	Q1232-15	PL093949.D	01/31/2025	01/31/2025
JPP-51.1-012925	Q1232-19	PL093950.D	01/31/2025	01/31/2025
357MS	Q1239-10MS	PL093960.D	01/31/2025	01/31/2025
357MSD	Q1239-10MSD	PL093961.D	01/31/2025	01/31/2025

COMMENTS: _____



SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.2-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-03	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	89.2 Decanted:
Sample Wt/Vol:	30.05 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093946.D	1	01/31/25 08:15	01/31/25 17:01	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	1.90	U	0.20	1.90	ug/kg
319-85-7	beta-BHC	1.90	U	0.55	1.90	ug/kg
319-86-8	delta-BHC	1.90	U	0.53	1.90	ug/kg
58-89-9	gamma-BHC (Lindane)	1.90	U	0.21	1.90	ug/kg
76-44-8	Heptachlor	1.90	U	0.19	1.90	ug/kg
309-00-2	Aldrin	1.90	U	0.16	1.90	ug/kg
1024-57-3	Heptachlor epoxide	1.90	U	0.26	1.90	ug/kg
959-98-8	Endosulfan I	1.90	U	0.19	1.90	ug/kg
60-57-1	Dieldrin	1.90	U	0.17	1.90	ug/kg
72-55-9	4,4-DDE	1.90	U	0.15	1.90	ug/kg
72-20-8	Endrin	1.90	U	0.18	1.90	ug/kg
33213-65-9	Endosulfan II	2.90	P	0.34	1.90	ug/kg
72-54-8	4,4-DDD	2.80	P	0.21	1.90	ug/kg
1031-07-8	Endosulfan Sulfate	1.90	U	0.15	1.90	ug/kg
50-29-3	4,4-DDT	1.90	U	0.19	1.90	ug/kg
72-43-5	Methoxychlor	1.90	U	0.43	1.90	ug/kg
53494-70-5	Endrin ketone	1.90	U	0.25	1.90	ug/kg
7421-93-4	Endrin aldehyde	1.90	U	0.44	1.90	ug/kg
5103-71-9	alpha-Chlordane	1.90	U	0.19	1.90	ug/kg
5103-74-2	gamma-Chlordane	1.90	U	0.21	1.90	ug/kg
8001-35-2	Toxaphene	36.9	U	5.80	36.9	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	30.3	*	10 - 148	152%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.0		10 - 159	110%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.2-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-03	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	89.2
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093946.D	1	01/31/25 08:15	01/31/25 17:01	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093946.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 17:01
 Operator : AR\AJ
 Sample : Q1232-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 JPP-46.2-012925

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 02/03/2025
 Supervised By : Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:28:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.774	48784847	71802017	18.117m	21.997
28) SA Decachlor...	9.057	7.913	50355181	106.2E6	24.071m	30.320m#
Target Compounds						
15) B Endosulfa...	6.789	5.941	5783439	28730153	2.400m	7.757m#
16) A 4,4'-DDD	6.708	5.773	14182221	10058926	7.462	3.187m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093946.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 17:01
Operator : AR\AJ
Sample : Q1232-03
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

JPP-46.2-012925

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:28:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: 0.005 min
Response: 48784847
Conc: 18.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-46.2-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 71802017
Conc: 22.00 ng/ml




#15 Endosulfan II

R.T.: 6.789 min
Delta R.T.: -0.003 min
Response: 5783439
Conc: 2.40 ng/ml m



#15 Endosulfan II

R.T.: 5.941 min
Delta R.T.: 0.009 min
Response: 28730153
Conc: 7.76 ng/ml m




#16 4,4'-DDD

R.T.: 6.708 min
Delta R.T.: 0.000 min
Response: 14182221
Conc: 7.46 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-46.2-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#16 4,4'-DDD

R.T.: 5.773 min
Delta R.T.: -0.012 min
Response: 10058926
Conc: 3.19 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.057 min
Delta R.T.: 0.004 min
Response: 50355181
Conc: 24.07 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.913 min
Delta R.T.: 0.003 min
Response: 106243697
Conc: 30.32 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.1-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-07	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	94 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093947.D	1	01/31/25 08:15	01/31/25 17:14	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	1.80	U	0.19	1.80	ug/kg
319-85-7	beta-BHC	1.80	U	0.52	1.80	ug/kg
319-86-8	delta-BHC	1.80	U	0.50	1.80	ug/kg
58-89-9	gamma-BHC (Lindane)	1.80	U	0.20	1.80	ug/kg
76-44-8	Heptachlor	0.22	J	0.18	1.80	ug/kg
309-00-2	Aldrin	1.80	U	0.15	1.80	ug/kg
1024-57-3	Heptachlor epoxide	0.61	J	0.24	1.80	ug/kg
959-98-8	Endosulfan I	1.80	U	0.18	1.80	ug/kg
60-57-1	Dieldrin	1.80	U	0.16	1.80	ug/kg
72-55-9	4,4-DDE	0.43	J	0.14	1.80	ug/kg
72-20-8	Endrin	0.60	JP	0.17	1.80	ug/kg
33213-65-9	Endosulfan II	1.80	U	0.32	1.80	ug/kg
72-54-8	4,4-DDD	1.80	U	0.20	1.80	ug/kg
1031-07-8	Endosulfan Sulfate	1.80	U	0.14	1.80	ug/kg
50-29-3	4,4-DDT	1.60	JP	0.18	1.80	ug/kg
72-43-5	Methoxychlor	1.80	U	0.40	1.80	ug/kg
53494-70-5	Endrin ketone	1.80	U	0.23	1.80	ug/kg
7421-93-4	Endrin aldehyde	1.80	U	0.41	1.80	ug/kg
5103-71-9	alpha-Chlordane	1.20	JP	0.18	1.80	ug/kg
5103-74-2	gamma-Chlordane	1.80	U	0.20	1.80	ug/kg
8001-35-2	Toxaphene	35.1	U	5.50	35.1	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.4		10 - 148	102%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.1		10 - 159	80%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-46.1-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-07	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	94
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
Prep Method :	SW3541B	Test:	Pesticide-TCL
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093947.D	1	01/31/25 08:15	01/31/25 17:14	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093947.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 17:14
 Operator : AR\AJ
 Sample : Q1232-07
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

JPP-46.1-012925

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:28:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.773	43274564	50391528	16.071m	15.438
28) SA Decachlor...	9.058	7.910	42590400	44782358	20.360	12.780m#
Target Compounds						
4) MA Heptachlor	4.915	3.941	2040681	2171984	0.623	0.467m#
7) B delta-BHC	4.767	4.138	2444165	1817610	0.697m	0.383m#
8) B Heptachlo...	5.677	4.727	5143404	2377340	1.730m	0.569m#
11) B alpha-Chl...	6.021	5.037	9557852	7768517	3.428m	1.856m#
12) B 4,4'-DDE	6.179	5.227	2967867	4329717	1.219m	1.080m
14) MA Endrin	6.573	5.643	3994069	3021016	1.703m	0.818m#
17) MA 4,4'-DDT	7.024	6.028	1619569	14641858	0.821m	4.500m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093947.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 17:14
Operator : AR\AJ
Sample : Q1232-07
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

JPP-46.1-012925

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:28:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.537 min
Delta R.T.: -0.002 min
Response: 43274564
Conc: 16.07 ng/ml

Instrument :


ECD_L

ClientSampleId :

JPP-46.1-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: -0.001 min
Response: 50391528
Conc: 15.44 ng/ml




#4 Heptachlor

R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 2040681
Conc: 0.62 ng/ml



#4 Heptachlor

R.T.: 3.941 min
Delta R.T.: -0.004 min
Response: 2171984
Conc: 0.47 ng/ml m




#7 delta-BHC

R.T.: 4.767 min
Delta R.T.: -0.005 min
Response: 2444165
Conc: 0.70 ng/ml m

Instrument :
ECD_L
ClientSampleId :
JPP-46.1-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#7 delta-BHC

R.T.: 4.138 min
Delta R.T.: 0.002 min
Response: 1817610
Conc: 0.38 ng/ml m




#8 Heptachlor epoxide

R.T.: 5.677 min
Delta R.T.: -0.005 min
Response: 5143404
Conc: 1.73 ng/ml m



#8 Heptachlor epoxide

R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 2377340
Conc: 0.57 ng/ml m




#11 alpha-Chlordane

R.T.: 6.021 min
Delta R.T.: 0.003 min
Response: 9557852
Conc: 3.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-46.1-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#11 alpha-Chlordane

R.T.: 5.037 min
Delta R.T.: -0.004 min
Response: 7768517
Conc: 1.86 ng/ml




#12 4,4'-DDE

R.T.: 6.179 min
Delta R.T.: -0.012 min
Response: 2967867
Conc: 1.22 ng/ml



#12 4,4'-DDE

R.T.: 5.227 min
Delta R.T.: -0.003 min
Response: 4329717
Conc: 1.08 ng/ml




#14 Endrin

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 3994069
Conc: 1.70 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-46.1-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#14 Endrin

R.T.: 5.643 min
Delta R.T.: 0.006 min
Response: 3021016
Conc: 0.82 ng/ml




#17 4,4'-DDT

R.T.: 7.024 min
Delta R.T.: 0.001 min
Response: 1619569
Conc: 0.82 ng/ml



#17 4,4'-DDT

R.T.: 6.028 min
Delta R.T.: -0.006 min
Response: 14641858
Conc: 4.50 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.058 min
Delta R.T.: 0.005 min
Response: 42590400
Conc: 20.36 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-46.1-012925

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025



#28 Decachlorobiphenyl

R.T.: 7.910 min
Delta R.T.: 0.000 min
Response: 44782358
Conc: 12.78 ng/ml m

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-42.1-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-11	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	88.1
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093948.D	1	01/31/25 08:15	01/31/25 17:27	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	1.90	U	0.20	1.90	ug/kg
319-85-7	beta-BHC	1.90	U	0.55	1.90	ug/kg
319-86-8	delta-BHC	1.90	U	0.53	1.90	ug/kg
58-89-9	gamma-BHC (Lindane)	1.90	U	0.22	1.90	ug/kg
76-44-8	Heptachlor	1.90	U	0.19	1.90	ug/kg
309-00-2	Aldrin	1.90	U	0.16	1.90	ug/kg
1024-57-3	Heptachlor epoxide	0.34	J	0.26	1.90	ug/kg
959-98-8	Endosulfan I	1.90	U	0.19	1.90	ug/kg
60-57-1	Dieldrin	1.90	U	0.17	1.90	ug/kg
72-55-9	4,4-DDE	1.90	U	0.15	1.90	ug/kg
72-20-8	Endrin	0.25	J	0.18	1.90	ug/kg
33213-65-9	Endosulfan II	1.90	U	0.34	1.90	ug/kg
72-54-8	4,4-DDD	1.90	U	0.22	1.90	ug/kg
1031-07-8	Endosulfan Sulfate	1.90	U	0.15	1.90	ug/kg
50-29-3	4,4-DDT	1.10	JP	0.19	1.90	ug/kg
72-43-5	Methoxychlor	1.90	U	0.43	1.90	ug/kg
53494-70-5	Endrin ketone	1.90	U	0.25	1.90	ug/kg
7421-93-4	Endrin aldehyde	1.90	U	0.44	1.90	ug/kg
5103-71-9	alpha-Chlordane	1.10	JP	0.19	1.90	ug/kg
5103-74-2	gamma-Chlordane	1.90	U	0.22	1.90	ug/kg
8001-35-2	Toxaphene	37.4	U	5.90	37.4	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.1		10 - 148	80%	SPK: 20
877-09-8	Tetrachloro-m-xylene	15.6		10 - 159	78%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-42.1-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-11	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	88.1
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093948.D	1	01/31/25 08:15	01/31/25 17:27	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093948.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 17:27
 Operator : AR\AJ
 Sample : Q1232-11
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 JPP-42.1-012925

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 02/03/2025
 Supervised By : Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:29:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.536	2.772	39574609	50894754	14.697m	15.592m
28) SA Decachlor...	9.054	7.909	33628521	51685546	16.075m	14.750m
Target Compounds						
8) B Heptachlo...	5.678	4.728	1515664	3786246	0.510m	0.906m#
11) B alpha-Chl...	6.020	5.039	8306598	5812217	2.979m	1.388m#
14) MA Endrin	6.571	5.643	1355162	2403221	0.578m	0.651m
17) MA 4,4'-DDT	7.022	6.031	1832858	9362534	0.929m	2.877 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093948.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 17:27
Operator : AR\AJ
Sample : Q1232-11
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

JPP-42.1-012925

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:29:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.536 min
Delta R.T.: -0.003 min
Response: 39574609
Conc: 14.70 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-42.1-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.772 min
Delta R.T.: -0.002 min
Response: 50894754
Conc: 15.59 ng/ml




#8 Heptachlor epoxide

R.T.: 5.678 min
Delta R.T.: -0.004 min
Response: 1515664
Conc: 0.51 ng/ml



#8 Heptachlor epoxide

R.T.: 4.728 min
Delta R.T.: 0.001 min
Response: 3786246
Conc: 0.91 ng/ml




#11 alpha-Chlordane

R.T.: 6.020 min
Delta R.T.: 0.002 min
Response: 8306598
Conc: 2.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-42.1-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#11 alpha-Chlordane

R.T.: 5.039 min
Delta R.T.: -0.002 min
Response: 5812217
Conc: 1.39 ng/ml




#14 Endrin

R.T.: 6.571 min
Delta R.T.: -0.001 min
Response: 1355162
Conc: 0.58 ng/ml



#14 Endrin

R.T.: 5.643 min
Delta R.T.: 0.007 min
Response: 2403221
Conc: 0.65 ng/ml




#17 4,4'-DDT

R.T.: 7.022 min
Delta R.T.: 0.000 min
Response: 1832858
Conc: 0.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-42.1-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#17 4,4'-DDT

R.T.: 6.031 min
Delta R.T.: -0.003 min
Response: 9362534
Conc: 2.88 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.054 min
Delta R.T.: 0.001 min
Response: 33628521
Conc: 16.08 ng/ml m



#28 Decachlorobiphenyl

R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 51685546
Conc: 14.75 ng/ml m

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-42.2-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-15	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	93.2 Decanted:
Sample Wt/Vol:	30.06 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093949.D	1	01/31/25 08:15	01/31/25 17:40	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	1.80	U	0.19	1.80	ug/kg
319-85-7	beta-BHC	1.80	U	0.52	1.80	ug/kg
319-86-8	delta-BHC	1.80	U	0.50	1.80	ug/kg
58-89-9	gamma-BHC (Lindane)	1.80	U	0.20	1.80	ug/kg
76-44-8	Heptachlor	0.73	JP	0.18	1.80	ug/kg
309-00-2	Aldrin	1.80	U	0.15	1.80	ug/kg
1024-57-3	Heptachlor epoxide	0.28	J	0.25	1.80	ug/kg
959-98-8	Endosulfan I	1.80	U	0.18	1.80	ug/kg
60-57-1	Dieldrin	1.80	U	0.16	1.80	ug/kg
72-55-9	4,4-DDE	1.80	U	0.14	1.80	ug/kg
72-20-8	Endrin	0.20	J	0.17	1.80	ug/kg
33213-65-9	Endosulfan II	1.80	U	0.32	1.80	ug/kg
72-54-8	4,4-DDD	1.80	U	0.20	1.80	ug/kg
1031-07-8	Endosulfan Sulfate	1.80	U	0.14	1.80	ug/kg
50-29-3	4,4-DDT	1.10	JP	0.18	1.80	ug/kg
72-43-5	Methoxychlor	1.80	U	0.41	1.80	ug/kg
53494-70-5	Endrin ketone	1.80	U	0.24	1.80	ug/kg
7421-93-4	Endrin aldehyde	1.80	U	0.42	1.80	ug/kg
5103-71-9	alpha-Chlordane	2.30	P	0.18	1.80	ug/kg
5103-74-2	gamma-Chlordane	1.20	JP	0.20	1.80	ug/kg
8001-35-2	Toxaphene	35.3	U	5.60	35.3	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.6		10 - 148	83%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.5		10 - 159	88%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-42.2-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-15	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	93.2 Decanted:
Sample Wt/Vol:	30.06 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093949.D	1	01/31/25 08:15	01/31/25 17:40	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093949.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 17:40
 Operator : AR\AJ
 Sample : Q1232-15
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :

ECD_L

Client SampleId :

JPP-42.2-012925

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:29:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.536	2.773	42336240	57220614	15.722m	17.530
28) SA Decachlor...	9.055	7.911	34662205	46530408	16.570m	13.279
Target Compounds						
4) MA Heptachlor	4.914	3.941	6704825	5406053	2.046	1.161m#
8) B Heptachlor...	5.680	4.726	1403584	3248101	0.472m	0.777m#
10) B gamma-Chl...	5.938	4.975	9124545	4116551	3.274m	0.971m#
11) B alpha-Chl...	6.020	5.038	18118605	10765215	6.498m	2.571m#
14) MA Endrin	6.571	5.643	1243600	2034090	0.530m	0.551m
17) MA 4,4'-DDT	7.023	6.029	3120301	9633306	1.582	2.960m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093949.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 17:40
Operator : AR\AJ
Sample : Q1232-15
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

JPP-42.2-012925

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:29:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.536 min
Delta R.T.: -0.003 min
Response: 42336240
Conc: 15.72 ng/ml

Instrument :


ECD_L

ClientSampleId :

JPP-42.2-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: -0.001 min
Response: 57220614
Conc: 17.53 ng/ml




#4 Heptachlor

R.T.: 4.914 min
Delta R.T.: 0.000 min
Response: 6704825
Conc: 2.05 ng/ml



#4 Heptachlor

R.T.: 3.941 min
Delta R.T.: -0.004 min
Response: 5406053
Conc: 1.16 ng/ml m




#8 Heptachlor epoxide

R.T.: 5.680 min
Delta R.T.: -0.003 min
Response: 1403584
Conc: 0.47 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-42.2-012925


**Manual Integrations
APPROVED**

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#8 Heptachlor epoxide

R.T.: 4.726 min
Delta R.T.: 0.000 min
Response: 3248101
Conc: 0.78 ng/ml




#10 gamma-Chlordane

R.T.: 5.938 min
Delta R.T.: 0.000 min
Response: 9124545
Conc: 3.27 ng/ml



#10 gamma-Chlordane

R.T.: 4.975 min
Delta R.T.: -0.002 min
Response: 4116551
Conc: 0.97 ng/ml




#11 alpha-Chlordane

R.T.: 6.020 min
Delta R.T.: 0.002 min
Response: 18118605
Conc: 6.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-42.2-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#11 alpha-Chlordane

R.T.: 5.038 min
Delta R.T.: -0.002 min
Response: 10765215
Conc: 2.57 ng/ml




#14 Endrin

R.T.: 6.571 min
Delta R.T.: -0.001 min
Response: 1243600
Conc: 0.53 ng/ml



#14 Endrin

R.T.: 5.643 min
Delta R.T.: 0.006 min
Response: 2034090
Conc: 0.55 ng/ml




#17 4,4'-DDT

R.T.: 7.023 min
Delta R.T.: 0.001 min
Response: 3120301
Conc: 1.58 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-42.2-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#17 4,4'-DDT

R.T.: 6.029 min
Delta R.T.: -0.005 min
Response: 9633306
Conc: 2.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.002 min
Response: 34662205
Conc: 16.57 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: 0.001 min
Response: 46530408
Conc: 13.28 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-51.1-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-19	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	95.1 Decanted:
Sample Wt/Vol:	30.04 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093950.D	1	01/31/25 08:15	01/31/25 17:53	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	1.80	U	0.19	1.80	ug/kg
319-85-7	beta-BHC	1.80	U	0.51	1.80	ug/kg
319-86-8	delta-BHC	1.80	U	0.49	1.80	ug/kg
58-89-9	gamma-BHC (Lindane)	1.80	U	0.20	1.80	ug/kg
76-44-8	Heptachlor	0.44	JP	0.18	1.80	ug/kg
309-00-2	Aldrin	1.80	U	0.15	1.80	ug/kg
1024-57-3	Heptachlor epoxide	0.78	JP	0.24	1.80	ug/kg
959-98-8	Endosulfan I	1.80	U	0.18	1.80	ug/kg
60-57-1	Dieldrin	1.80	U	0.16	1.80	ug/kg
72-55-9	4,4-DDE	1.80	U	0.14	1.80	ug/kg
72-20-8	Endrin	0.49	JP	0.17	1.80	ug/kg
33213-65-9	Endosulfan II	1.80	U	0.32	1.80	ug/kg
72-54-8	4,4-DDD	0.66	J	0.20	1.80	ug/kg
1031-07-8	Endosulfan Sulfate	1.80	U	0.14	1.80	ug/kg
50-29-3	4,4-DDT	2.00	P	0.18	1.80	ug/kg
72-43-5	Methoxychlor	1.80	U	0.40	1.80	ug/kg
53494-70-5	Endrin ketone	1.80	U	0.23	1.80	ug/kg
7421-93-4	Endrin aldehyde	1.80	U	0.41	1.80	ug/kg
5103-71-9	alpha-Chlordane	2.90	P	0.18	1.80	ug/kg
5103-74-2	gamma-Chlordane	1.80	U	0.20	1.80	ug/kg
8001-35-2	Toxaphene	34.7	U	5.50	34.7	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.1		10 - 148	81%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.4		10 - 159	82%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/29/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	JPP-51.1-012925	SDG No.:	Q1232
Lab Sample ID:	Q1232-19	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	95.1 Decanted:
Sample Wt/Vol:	30.04 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093950.D	1	01/31/25 08:15	01/31/25 17:53	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093950.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 17:53
 Operator : AR\AJ
 Sample : Q1232-19
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 JPP-51.1-012925

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 02/03/2025
 Supervised By : Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:29:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.773	43874565	53536577	16.294	16.401
2) SA Decachlor...	9.056	7.911	33675208	49475561	16.098	14.119
Target Compounds						
4) MA Heptachlor	4.913	3.942	4091545	4074470	1.248m	0.875m#
7) B delta-BHC	4.770	4.139	2952260	3503895	0.842m	0.737m
8) B Heptachlo...	5.677	4.727	3117870	9279226	1.048m	2.220m#
11) B alpha-Chl...	6.020	5.039	22834894	16651498	8.189m	3.977m#
14) MA Endrin	6.571	5.643	2206276	5156319	0.941m	1.396m#
16) A 4,4'-DDD	6.717	5.772	3575083	1778442	1.881m	0.563m#
17) MA 4,4'-DDT	7.022	6.029	5686201	18262110	2.883m	5.612m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093950.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 17:53
Operator : AR\AJ
Sample : Q1232-19
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

JPP-51.1-012925

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:29:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.537 min
Delta R.T.: -0.002 min
Response: 43874565
Conc: 16.29 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-51.1-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: -0.001 min
Response: 53536577
Conc: 16.40 ng/ml




#4 Heptachlor

R.T.: 4.913 min
Delta R.T.: -0.002 min
Response: 4091545
Conc: 1.25 ng/ml m



#4 Heptachlor

R.T.: 3.942 min
Delta R.T.: -0.004 min
Response: 4074470
Conc: 0.88 ng/ml m




#7 delta-BHC

R.T.: 4.770 min
Delta R.T.: -0.002 min
Response: 2952260
Conc: 0.84 ng/ml m

Instrument :
ECD_L
ClientSampleId :
JPP-51.1-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#7 delta-BHC

R.T.: 4.139 min
Delta R.T.: 0.004 min
Response: 3503895
Conc: 0.74 ng/ml m




#8 Heptachlor epoxide

R.T.: 5.677 min
Delta R.T.: -0.005 min
Response: 3117870
Conc: 1.05 ng/ml m



#8 Heptachlor epoxide

R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 9279226
Conc: 2.22 ng/ml m




#11 alpha-Chlordane

R.T.: 6.020 min
Delta R.T.: 0.003 min
Response: 22834894
Conc: 8.19 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-51.1-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#11 alpha-Chlordane

R.T.: 5.039 min
Delta R.T.: -0.002 min
Response: 16651498
Conc: 3.98 ng/ml




#14 Endrin

R.T.: 6.571 min
Delta R.T.: -0.001 min
Response: 2206276
Conc: 0.94 ng/ml



#14 Endrin

R.T.: 5.643 min
Delta R.T.: 0.006 min
Response: 5156319
Conc: 1.40 ng/ml




#16 4,4'-DDD

R.T.: 6.717 min
Delta R.T.: 0.009 min
Response: 3575083
Conc: 1.88 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-51.1-012925


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#16 4,4'-DDD

R.T.: 5.772 min
Delta R.T.: -0.013 min
Response: 1778442
Conc: 0.56 ng/ml




#17 4,4'-DDT

R.T.: 7.022 min
Delta R.T.: 0.000 min
Response: 5686201
Conc: 2.88 ng/ml



#17 4,4'-DDT

R.T.: 6.029 min
Delta R.T.: -0.005 min
Response: 18262110
Conc: 5.61 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: 0.003 min
Response: 33675208
Conc: 16.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-51.1-012925

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: 0.000 min
Response: 49475561
Conc: 14.12 ng/ml



CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

RETENTION TIMES OF INITIAL CALIBRATION

Contract: RUTW01
Lab Code: CHEM **Case No.:** Q1232 **SAS No.:** Q1232 **SDG NO.:** Q1232
Instrument ID: ECD_L **Calibration Date(s):** 01/21/2025 01/21/2025
Calibration Times: 10:57 11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PL093728.D</u>	RT 075 = <u>PL093729.D</u>
RT 050 = <u>PL093730.D</u>	RT 025 = <u>PL093731.D</u>	RT 005 = <u>PL093732.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
4,4'-DDD	6.71	6.71	6.71	6.71	6.71	6.71	6.61	6.81
4,4'-DDE	6.19	6.19	6.19	6.19	6.19	6.19	6.09	6.29
4,4'-DDT	7.02	7.02	7.02	7.02	7.02	7.02	6.92	7.12
Aldrin	5.26	5.26	5.26	5.26	5.26	5.26	5.16	5.36
alpha-BHC	4.00	4.00	4.00	4.00	3.99	3.99	3.89	4.09
alpha-Chlordane	6.02	6.02	6.02	6.02	6.02	6.02	5.92	6.12
beta-BHC	4.53	4.53	4.53	4.53	4.53	4.53	4.43	4.63
Decachlorobiphenyl	9.05	9.05	9.05	9.05	9.05	9.05	8.95	9.15
delta-BHC	4.77	4.77	4.77	4.77	4.77	4.77	4.67	4.87
Dieldrin	6.34	6.34	6.34	6.34	6.34	6.34	6.24	6.44
Endosulfan I	6.07	6.07	6.07	6.07	6.07	6.07	5.97	6.17
Endosulfan II	6.79	6.79	6.79	6.79	6.79	6.79	6.69	6.89
Endosulfan sulfate	7.16	7.16	7.16	7.16	7.16	7.16	7.06	7.26
Endrin	6.57	6.57	6.57	6.57	6.57	6.57	6.47	6.67
Endrin aldehyde	6.92	6.92	6.92	6.92	6.92	6.92	6.82	7.02
Endrin ketone	7.64	7.64	7.64	7.64	7.64	7.64	7.54	7.74
gamma-BHC (Lindane)	4.33	4.33	4.33	4.33	4.33	4.33	4.23	4.43
gamma-Chlordane	5.94	5.94	5.94	5.94	5.94	5.94	5.84	6.04
Heptachlor	4.92	4.92	4.91	4.91	4.91	4.91	4.81	5.01
Heptachlor epoxide	5.68	5.68	5.68	5.68	5.68	5.68	5.58	5.78
Methoxychlor	7.50	7.50	7.50	7.50	7.50	7.50	7.40	7.60
Tetrachloro-m-xylene	3.54	3.54	3.54	3.54	3.54	3.54	3.44	3.64

RETENTION TIMES OF INITIAL CALIBRATION

Contract: RUTW01
Lab Code: CHEM **Case No.:** Q1232 **SAS No.:** Q1232 **SDG NO.:** Q1232
Instrument ID: ECD_L **Calibration Date(s):** 01/21/2025 01/21/2025
Calibration Times: 10:57 11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PL093728.D</u>	RT 075 = <u>PL093729.D</u>
RT 050 = <u>PL093730.D</u>	RT 025 = <u>PL093731.D</u>	RT 005 = <u>PL093732.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
4,4'-DDD	5.79	5.78	5.78	5.78	5.78	5.78	5.68	5.88
4,4'-DDE	5.23	5.23	5.23	5.23	5.23	5.23	5.13	5.33
4,4'-DDT	6.04	6.03	6.03	6.03	6.03	6.03	5.93	6.13
Aldrin	4.23	4.23	4.23	4.22	4.22	4.22	4.12	4.32
alpha-BHC	3.28	3.28	3.28	3.28	3.28	3.28	3.18	3.38
alpha-Chlordane	5.04	5.04	5.04	5.04	5.04	5.04	4.94	5.14
beta-BHC	3.91	3.91	3.91	3.91	3.91	3.91	3.81	4.01
Decachlorobiphenyl	7.91	7.91	7.91	7.91	7.91	7.91	7.81	8.01
delta-BHC	4.14	4.14	4.14	4.14	4.13	4.14	4.04	4.24
Dieldrin	5.36	5.36	5.36	5.36	5.36	5.36	5.26	5.46
Endosulfan I	5.10	5.10	5.10	5.10	5.10	5.10	5.00	5.20
Endosulfan II	5.93	5.93	5.93	5.93	5.93	5.93	5.83	6.03
Endosulfan sulfate	6.33	6.33	6.33	6.33	6.33	6.33	6.23	6.43
Endrin	5.64	5.64	5.64	5.64	5.64	5.64	5.54	5.74
Endrin aldehyde	6.11	6.11	6.11	6.11	6.11	6.11	6.01	6.21
Endrin ketone	6.84	6.84	6.84	6.84	6.84	6.84	6.74	6.94
gamma-BHC (Lindane)	3.61	3.61	3.61	3.61	3.61	3.61	3.51	3.71
gamma-Chlordane	4.98	4.98	4.98	4.98	4.98	4.98	4.88	5.08
Heptachlor	3.95	3.95	3.95	3.95	3.94	3.94	3.84	4.04
Heptachlor epoxide	4.73	4.73	4.73	4.73	4.73	4.73	4.63	4.83
Methoxychlor	6.61	6.61	6.61	6.61	6.61	6.61	6.51	6.71
Tetrachloro-m-xylene	2.78	2.77	2.77	2.77	2.77	2.77	2.67	2.87

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01
Lab Code: CHEM **Case No.:** Q1232 **SAS No.:** Q1232 **SDG NO.:** Q1232
Instrument ID: ECD_L
Calibration Date(s): 01/21/2025 01/21/2025
Calibration Times: 10:57 11:51

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL093728.D</u>		CF 075 = <u>PL093729.D</u>			
CF 050 = <u>PL093730.D</u>		CF 025 = <u>PL093731.D</u>		CF 005 = <u>PL093732.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	1660930000	1679650000	1932670000	1802720000	2426830000	1900560000	17
4,4'-DDE	2179870000	2169930000	2489080000	2321590000	3012520000	2434600000	14
4,4'-DDT	1755570000	1766710000	2016720000	1907120000	2414170000	1972060000	14
Aldrin	2924220000	2896750000	3292630000	3099660000	4146570000	3271970000	16
alpha-BHC	3537700000	3490280000	3918110000	3562830000	4660310000	3833850000	13
alpha-Chlordane	2458070000	2458490000	2788200000	2666580000	3570690000	2788400000	16
beta-BHC	1393460000	1394440000	1618290000	1508890000	2121530000	1607320000	19
Decachlorobiphenyl	1768480000	1816480000	2098320000	2018470000	2757820000	2091910000	19
delta-BHC	3233860000	3194550000	3605880000	3303370000	4188780000	3505290000	12
Dieldrin	2456580000	2440810000	2788190000	2639340000	3554340000	2775850000	17
Endosulfan I	2304400000	2298550000	2637060000	2528610000	3445650000	2642860000	18
Endosulfan II	2084130000	2100600000	2413950000	2287820000	3160260000	2409350000	18
Endosulfan sulfate	1923100000	1945070000	2248580000	2190510000	3011450000	2263740000	20
Endrin	2079430000	2060990000	2363220000	2218560000	3001890000	2344820000	17
Endrin aldehyde	1673120000	1696040000	1958970000	1896570000	2495580000	1944060000	17
Endrin ketone	2196850000	2205550000	2539700000	2413910000	3257130000	2522630000	17
gamma-BHC (Lindane)	3375960000	3339350000	3767250000	3460830000	4470850000	3682850000	13
gamma-Chlordane	2455830000	2471830000	2815630000	2678390000	3515170000	2787370000	16
Heptachlor	2922500000	2901690000	3325290000	3144100000	4093120000	3277340000	15
Heptachlor epoxide	2568680000	2575960000	2953630000	2835830000	3935020000	2973820000	19
Methoxychlor	907284000	922109000	1080370000	1020090000	1287130000	1043400000	15
Tetrachloro-m-xylene	2397870000	2402980000	2740040000	2595500000	3327420000	2692760000	14

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01
Lab Code: CHEM **Case No.:** Q1232 **SAS No.:** Q1232 **SDG NO.:** Q1232
Instrument ID: ECD_L
Calibration Date(s): 01/21/2025 01/21/2025
Calibration Times: 10:57 11:51

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL093728.D</u>		CF 075 = <u>PL093729.D</u>			
CF 050 = <u>PL093730.D</u>		CF 025 = <u>PL093731.D</u>		CF 005 = <u>PL093732.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	3134000000	3054730000	3379160000	2910470000	3304320000	3156540000	6
4,4'-DDE	3891920000	3807640000	4253650000	3749010000	4345130000	4009470000	7
4,4'-DDT	3270010000	3177800000	3542860000	3046890000	3232670000	3254050000	6
Aldrin	4482990000	4370810000	4856520000	4222470000	4876190000	4561800000	6
alpha-BHC	4914190000	4768640000	5271080000	4480730000	5010260000	4888980000	6
alpha-Chlordane	4056970000	3962110000	4424110000	3914810000	4574820000	4186560000	7
beta-BHC	1863440000	1842720000	2072180000	1889740000	2319100000	1997440000	10
Decachlorobiphenyl	3226690000	3193800000	3627020000	3320620000	4152210000	3504070000	11
delta-BHC	4741230000	4607910000	5098810000	4368820000	4939430000	4751240000	6
Dieldrin	4189300000	4076770000	4553570000	3958830000	4699760000	4295650000	7
Endosulfan I	3734100000	3661580000	4099030000	3635320000	4254550000	3876920000	7
Endosulfan II	3553260000	3487640000	3912960000	3484510000	4080760000	3703830000	7
Endosulfan sulfate	3408630000	3353240000	3757030000	3348270000	3963240000	3566080000	8
Endrin	3607760000	3481170000	3870730000	3406140000	4097610000	3692680000	8
Endrin aldehyde	2861460000	2820180000	3183430000	2892290000	3465840000	3044640000	9
Endrin ketone	3965120000	3881890000	4400080000	3907370000	4821740000	4195240000	10
gamma-BHC (Lindane)	4713370000	4597010000	5084610000	4384810000	4926270000	4741210000	6
gamma-Chlordane	4137240000	4016860000	4483010000	3935490000	4615500000	4237620000	7
Heptachlor	4505180000	4413750000	4924840000	4345980000	5084220000	4654790000	7
Heptachlor epoxide	4026840000	3946880000	4424170000	3927960000	4575440000	4180260000	7
Methoxychlor	1651870000	1634200000	1870410000	1643810000	2140390000	1788140000	12
Tetrachloro-m-xylene	3101220000	3058550000	3437230000	3066200000	3657590000	3264160000	8



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Instrument ID: ECD_L Date(s) Analyzed: 01/21/2025 01/21/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.24	6.14	6.34	23446000
		2	6.44	6.34	6.54	14767200
		3	7.06	6.96	7.16	75896000
		4	7.15	7.05	7.25	57345100
		5	7.93	7.83	8.03	43067100



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Instrument ID: ECD_L Date(s) Analyzed: 01/21/2025 01/21/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.00	4.90	5.10	27057100
		2	5.33	5.23	5.43	23947200
		3	5.68	5.58	5.78	24726400
		4	6.60	6.50	6.70	84987200
		5	7.04	6.94	7.14	80238300

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
 Data File : PL093728.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2025 10:57
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 13:55:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 13:52:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.775	239.8E6	310.1E6	93.340	94.861
28) SA Decachlor...	9.052	7.910	176.8E6	322.7E6	91.470	94.159
Target Compounds						
2) A alpha-BHC	3.995	3.277	353.8E6	491.4E6	94.898	96.496
3) MA gamma-BHC...	4.328	3.607	337.6E6	471.3E6	94.522	96.211
4) MA Heptachlor	4.915	3.946	292.3E6	450.5E6	93.553	95.550
5) MB Aldrin	5.257	4.225	292.4E6	448.3E6	94.074	96.001
6) B beta-BHC	4.526	3.907	139.3E6	186.3E6	92.535	94.696
7) B delta-BHC	4.773	4.136	323.4E6	474.1E6	94.561	96.366
8) B Heptachlo...	5.683	4.727	256.9E6	402.7E6	93.029	95.298
9) A Endosulfan I	6.069	5.097	230.4E6	373.4E6	93.268	95.341
10) B gamma-Chl...	5.940	4.977	245.6E6	413.7E6	93.175	95.989
11) B alpha-Chl...	6.018	5.041	245.8E6	405.7E6	93.707	95.671
12) B 4,4'-DDE	6.192	5.230	218.0E6	389.2E6	93.377	95.559
13) MA Dieldrin	6.344	5.361	245.7E6	418.9E6	93.677	95.834
14) MA Endrin	6.573	5.636	207.9E6	360.8E6	93.612	96.484
15) B Endosulfa...	6.793	5.932	208.4E6	355.3E6	92.668	95.182
16) A 4,4'-DDD	6.710	5.785	166.1E6	313.4E6	92.438	96.236
17) MA 4,4'-DDT	7.023	6.035	175.6E6	327.0E6	93.077	95.995
18) B Endrin al...	6.924	6.110	167.3E6	286.1E6	92.130	94.674
19) B Endosulfa...	7.158	6.333	192.3E6	340.9E6	92.198	95.138
20) A Methoxychlor	7.499	6.609	90728367	165.2E6	91.292	93.795
21) B Endrin ke...	7.643	6.838	219.7E6	396.5E6	92.761	94.800
22) Mirex	8.115	7.018	175.3E6	309.9E6	91.817	94.309



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093728.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 10:57
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 13:55:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 13:52:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.539 min
Delta R.T.: 0.000 min
Response: 239787086
Conc: 93.34 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#1 Tetrachloro-m-xylene

R.T.: 2.775 min
Delta R.T.: 0.000 min
Response: 310121914
Conc: 94.86 ng/ml




#2 alpha-BHC

R.T.: 3.995 min
Delta R.T.: 0.000 min
Response: 353769763
Conc: 94.90 ng/ml



#2 alpha-BHC


R.T.: 3.277 min
Delta R.T.: 0.000 min
Response: 491419278
Conc: 96.50 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.328 min
Delta R.T.: 0.000 min
Response: 337596128
Conc: 94.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#3 gamma-BHC (Lindane)

R.T.: 3.607 min
Delta R.T.: 0.000 min
Response: 471337352
Conc: 96.21 ng/ml




#4 Heptachlor

R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 292250336
Conc: 93.55 ng/ml



#4 Heptachlor


R.T.: 3.946 min
Delta R.T.: 0.000 min
Response: 450517764
Conc: 95.55 ng/ml



#5 Aldrin


R.T.: 5.257 min
Delta R.T.: 0.000 min
Response: 292421818
Conc: 94.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#5 Aldrin

R.T.: 4.225 min
Delta R.T.: 0.000 min
Response: 448299060
Conc: 96.00 ng/ml




#6 beta-BHC

R.T.: 4.526 min
Delta R.T.: 0.000 min
Response: 139346436
Conc: 92.54 ng/ml



#6 beta-BHC


R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 186343878
Conc: 94.70 ng/ml



#7 delta-BHC


R.T.: 4.773 min
Delta R.T.: 0.000 min
Response: 323385715
Conc: 94.56 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#7 delta-BHC

R.T.: 4.136 min
Delta R.T.: 0.000 min
Response: 474122692
Conc: 96.37 ng/ml




#8 Heptachlor epoxide

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 256867626
Conc: 93.03 ng/ml



#8 Heptachlor epoxide


R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 402684225
Conc: 95.30 ng/ml



#9 Endosulfan I


R.T.: 6.069 min
Delta R.T.: 0.000 min
Response: 230440433
Conc: 93.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#9 Endosulfan I

R.T.: 5.097 min
Delta R.T.: 0.000 min
Response: 373410327
Conc: 95.34 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 245583159
Conc: 93.17 ng/ml



#10 gamma-Chlordane


R.T.: 4.977 min
Delta R.T.: 0.000 min
Response: 413724271
Conc: 95.99 ng/ml



#11 alpha-Chlordane


R.T.: 6.018 min
Delta R.T.: 0.000 min
Response: 245806926
Conc: 93.71 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#11 alpha-Chlordane

R.T.: 5.041 min
Delta R.T.: 0.000 min
Response: 405697093
Conc: 95.67 ng/ml




#12 4,4'-DDE

R.T.: 6.192 min
Delta R.T.: 0.000 min
Response: 217987126
Conc: 93.38 ng/ml



#12 4,4'-DDE


R.T.: 5.230 min
Delta R.T.: 0.000 min
Response: 389192028
Conc: 95.56 ng/ml



#13 Dieldrin


R.T.: 6.344 min
Delta R.T.: 0.000 min
Response: 245657863
Conc: 93.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#13 Dieldrin

R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 418929941
Conc: 95.83 ng/ml




#14 Endrin

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 207942946
Conc: 93.61 ng/ml



#14 Endrin


R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 360776248
Conc: 96.48 ng/ml



#15 Endosulfan II


R.T.: 6.793 min
Delta R.T.: 0.000 min
Response: 208413423
Conc: 92.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#15 Endosulfan II

R.T.: 5.932 min
Delta R.T.: 0.000 min
Response: 355326342
Conc: 95.18 ng/ml




#16 4,4'-DDD

R.T.: 6.710 min
Delta R.T.: 0.000 min
Response: 166092692
Conc: 92.44 ng/ml



#16 4,4'-DDD


R.T.: 5.785 min
Delta R.T.: 0.000 min
Response: 313400411
Conc: 96.24 ng/ml



#17 4,4'-DDT


R.T.: 7.023 min
Delta R.T.: 0.000 min
Response: 175556551
Conc: 93.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#17 4,4'-DDT

R.T.: 6.035 min
Delta R.T.: 0.000 min
Response: 327000877
Conc: 96.00 ng/ml




#18 Endrin aldehyde

R.T.: 6.924 min
Delta R.T.: 0.000 min
Response: 167311887
Conc: 92.13 ng/ml



#18 Endrin aldehyde


R.T.: 6.110 min
Delta R.T.: 0.000 min
Response: 286145897
Conc: 94.67 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.158 min
Delta R.T.: 0.000 min
Response: 192310257
Conc: 92.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#19 Endosulfan Sulfate

R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 340862985
Conc: 95.14 ng/ml




#20 Methoxychlor

R.T.: 7.499 min
Delta R.T.: 0.000 min
Response: 90728367
Conc: 91.29 ng/ml



#20 Methoxychlor


R.T.: 6.609 min
Delta R.T.: 0.000 min
Response: 165186868
Conc: 93.80 ng/ml



#21 Endrin ketone


R.T.: 7.643 min
Delta R.T.: 0.000 min
Response: 219684904
Conc: 92.76 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#21 Endrin ketone

R.T.: 6.838 min
Delta R.T.: 0.000 min
Response: 396512430
Conc: 94.80 ng/ml




#22 Mirex

R.T.: 8.115 min
Delta R.T.: 0.000 min
Response: 175302981
Conc: 91.82 ng/ml



#22 Mirex


R.T.: 7.018 min
Delta R.T.: 0.000 min
Response: 309908166
Conc: 94.31 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.052 min
Delta R.T.: 0.000 min
Response: 176847808
Conc: 91.47 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100



#28 Decachlorobiphenyl

R.T.: 7.910 min
Delta R.T.: 0.000 min
Response: 322668553
Conc: 94.16 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
 Data File : PL093729.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2025 11:10
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 13:57:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 13:52:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.774	180.2E6	229.4E6	71.698	71.707
2) SA Decachlor...	9.053	7.910	136.2E6	239.5E6	71.914	71.521
Target Compounds						
2) A alpha-BHC	3.995	3.277	261.8E6	357.6E6	71.744	71.750
3) MA gamma-BHC...	4.327	3.607	250.5E6	344.8E6	71.676	71.853
4) MA Heptachlor	4.915	3.945	217.6E6	331.0E6	71.357	71.736
5) MB Aldrin	5.256	4.225	217.3E6	327.8E6	71.516	71.729
6) B beta-BHC	4.525	3.907	104.6E6	138.2E6	71.206	71.753
7) B delta-BHC	4.772	4.136	239.6E6	345.6E6	71.632	71.760
8) B Heptachlo...	5.683	4.727	193.2E6	296.0E6	71.570	71.629
9) A Endosulfan I	6.068	5.097	172.4E6	274.6E6	71.433	71.672
10) B gamma-Chl...	5.938	4.977	185.4E6	301.3E6	71.825	71.519
11) B alpha-Chl...	6.017	5.041	184.4E6	297.2E6	71.795	71.644
12) B 4,4'-DDE	6.191	5.230	162.7E6	285.6E6	71.391	71.673
13) MA Dieldrin	6.343	5.361	183.1E6	305.8E6	71.456	71.552
14) MA Endrin	6.573	5.637	154.6E6	261.1E6	71.302	71.468
15) B Endosulfa...	6.793	5.931	157.5E6	261.6E6	71.626	71.638
16) A 4,4'-DDD	6.709	5.784	126.0E6	229.1E6	71.668	71.836
17) MA 4,4'-DDT	7.022	6.034	132.5E6	238.3E6	71.766	71.567
18) B Endrin al...	6.923	6.110	127.2E6	211.5E6	71.622	71.578
19) B Endosulfa...	7.157	6.333	145.9E6	251.5E6	71.548	71.726
20) A Methoxychlor	7.498	6.609	69158182	122.6E6	71.303	71.307
21) B Endrin ke...	7.642	6.838	165.4E6	291.1E6	71.484	71.317
22) Mirex	8.115	7.018	133.6E6	229.7E6	71.590	71.522



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093729.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 11:10
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 13:57:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 13:52:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.539 min
Delta R.T.: 0.000 min
Response: 180223207
Conc: 71.70 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 229390960
Conc: 71.71 ng/ml




#2 alpha-BHC

R.T.: 3.995 min
Delta R.T.: 0.000 min
Response: 261771076
Conc: 71.74 ng/ml



#2 alpha-BHC


R.T.: 3.277 min
Delta R.T.: 0.000 min
Response: 357647885
Conc: 71.75 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.327 min
Delta R.T.: 0.000 min
Response: 250451076
Conc: 71.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#3 gamma-BHC (Lindane)

R.T.: 3.607 min
Delta R.T.: 0.000 min
Response: 344775667
Conc: 71.85 ng/ml




#4 Heptachlor

R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 217626800
Conc: 71.36 ng/ml



#4 Heptachlor


R.T.: 3.945 min
Delta R.T.: 0.000 min
Response: 331031082
Conc: 71.74 ng/ml



#5 Aldrin


R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 217256477
Conc: 71.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#5 Aldrin

R.T.: 4.225 min
Delta R.T.: 0.000 min
Response: 327810801
Conc: 71.73 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 104583271
Conc: 71.21 ng/ml



#6 beta-BHC


R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 138204189
Conc: 71.75 ng/ml



#7 delta-BHC


R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 239591328
Conc: 71.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#7 delta-BHC

R.T.: 4.136 min
Delta R.T.: 0.000 min
Response: 345593516
Conc: 71.76 ng/ml




#8 Heptachlor epoxide

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 193196976
Conc: 71.57 ng/ml



#8 Heptachlor epoxide


R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 296015685
Conc: 71.63 ng/ml



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 172391226
Conc: 71.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#9 Endosulfan I

R.T.: 5.097 min
Delta R.T.: 0.000 min
Response: 274618139
Conc: 71.67 ng/ml




#10 gamma-Chlordane

R.T.: 5.938 min
Delta R.T.: 0.000 min
Response: 185387384
Conc: 71.83 ng/ml



#10 gamma-Chlordane


R.T.: 4.977 min
Delta R.T.: 0.000 min
Response: 301264863
Conc: 71.52 ng/ml



#11 alpha-Chlordane


R.T.: 6.017 min
Delta R.T.: 0.000 min
Response: 184386873
Conc: 71.79 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#11 alpha-Chlordane

R.T.: 5.041 min
Delta R.T.: 0.000 min
Response: 297157887
Conc: 71.64 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 162744494
Conc: 71.39 ng/ml



#12 4,4'-DDE


R.T.: 5.230 min
Delta R.T.: 0.000 min
Response: 285573154
Conc: 71.67 ng/ml



#13 Dieldrin


R.T.: 6.343 min
Delta R.T.: 0.000 min
Response: 183061011
Conc: 71.46 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#13 Dieldrin

R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 305757664
Conc: 71.55 ng/ml




#14 Endrin

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 154574222
Conc: 71.30 ng/ml



#14 Endrin


R.T.: 5.637 min
Delta R.T.: 0.000 min
Response: 261087919
Conc: 71.47 ng/ml



#15 Endosulfan II


R.T.: 6.793 min
Delta R.T.: 0.000 min
Response: 157545345
Conc: 71.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#15 Endosulfan II

R.T.: 5.931 min
Delta R.T.: 0.000 min
Response: 261572754
Conc: 71.64 ng/ml




#16 4,4'-DDD

R.T.: 6.709 min
Delta R.T.: 0.000 min
Response: 125973987
Conc: 71.67 ng/ml



#16 4,4'-DDD


R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 229105062
Conc: 71.84 ng/ml



#17 4,4'-DDT


R.T.: 7.022 min
Delta R.T.: 0.000 min
Response: 132503378
Conc: 71.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#17 4,4'-DDT

R.T.: 6.034 min
Delta R.T.: 0.000 min
Response: 238334747
Conc: 71.57 ng/ml




#18 Endrin aldehyde

R.T.: 6.923 min
Delta R.T.: 0.000 min
Response: 127202927
Conc: 71.62 ng/ml



#18 Endrin aldehyde


R.T.: 6.110 min
Delta R.T.: 0.000 min
Response: 211513516
Conc: 71.58 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.157 min
Delta R.T.: 0.000 min
Response: 145880361
Conc: 71.55 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#19 Endosulfan Sulfate

R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 251492996
Conc: 71.73 ng/ml




#20 Methoxychlor

R.T.: 7.498 min
Delta R.T.: 0.000 min
Response: 69158182
Conc: 71.30 ng/ml



#20 Methoxychlor


R.T.: 6.609 min
Delta R.T.: 0.000 min
Response: 122564991
Conc: 71.31 ng/ml



#21 Endrin ketone


R.T.: 7.642 min
Delta R.T.: 0.000 min
Response: 165416371
Conc: 71.48 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#21 Endrin ketone

R.T.: 6.838 min
Delta R.T.: 0.000 min
Response: 291141748
Conc: 71.32 ng/ml




#22 Mirex

R.T.: 8.115 min
Delta R.T.: 0.000 min
Response: 133645314
Conc: 71.59 ng/ml



#22 Mirex


R.T.: 7.018 min
Delta R.T.: 0.000 min
Response: 229698971
Conc: 71.52 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: 0.000 min
Response: 136236362
Conc: 71.91 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075



#28 Decachlorobiphenyl

R.T.: 7.910 min
Delta R.T.: 0.000 min
Response: 239534634
Conc: 71.52 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093730.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 11:24
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 13:53:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 13:52:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.774	137.0E6	171.9E6	50.000	50.000
28) SA Decachlor...	9.053	7.910	104.9E6	181.4E6	50.000	50.000
Target Compounds						
2) A alpha-BHC	3.995	3.277	195.9E6	263.6E6	50.000	50.000
3) MA gamma-BHC...	4.327	3.607	188.4E6	254.2E6	50.000	50.000
4) MA Heptachlor	4.914	3.945	166.3E6	246.2E6	50.000	50.000
5) MB Aldrin	5.256	4.225	164.6E6	242.8E6	50.000	50.000
6) B beta-BHC	4.525	3.907	80914447	103.6E6	50.000	50.000
7) B delta-BHC	4.772	4.136	180.3E6	254.9E6	50.000	50.000
8) B Heptachlo...	5.682	4.727	147.7E6	221.2E6	50.000	50.000
9) A Endosulfan I	6.067	5.096	131.9E6	205.0E6	50.000	50.000
10) B gamma-Chl...	5.938	4.977	140.8E6	224.2E6	50.000	50.000
11) B alpha-Chl...	6.017	5.040	139.4E6	221.2E6	50.000	50.000
12) B 4,4'-DDE	6.191	5.230	124.5E6	212.7E6	50.000	50.000
13) MA Dieldrin	6.343	5.361	139.4E6	227.7E6	50.000	50.000
14) MA Endrin	6.572	5.636	118.2E6	193.5E6	50.000	50.000
15) B Endosulfa...	6.792	5.931	120.7E6	195.6E6	50.000	50.000
16) A 4,4'-DDD	6.708	5.784	96633741	169.0E6	50.000	50.000
17) MA 4,4'-DDT	7.022	6.034	100.8E6	177.1E6	50.000	50.000
18) B Endrin al...	6.923	6.110	97948460	159.2E6	50.000	50.000
19) B Endosulfa...	7.157	6.333	112.4E6	187.9E6	50.000	50.000
20) A Methoxychlor	7.498	6.609	54018493	93520516	50.000	50.000
21) B Endrin ke...	7.642	6.839	127.0E6	220.0E6	50.000	50.000
22) Mirex	8.115	7.018	103.3E6	173.7E6	50.000	50.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093730.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 11:24
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 13:53:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 13:52:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.539 min
Delta R.T.: 0.000 min
Response: 137001991
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 171861595
Conc: 50.00 ng/ml




#2 alpha-BHC

R.T.: 3.995 min
Delta R.T.: 0.000 min
Response: 195905536
Conc: 50.00 ng/ml



#2 alpha-BHC


R.T.: 3.277 min
Delta R.T.: 0.000 min
Response: 263554029
Conc: 50.00 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.327 min
Delta R.T.: 0.000 min
Response: 188362613
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#3 gamma-BHC (Lindane)

R.T.: 3.607 min
Delta R.T.: 0.000 min
Response: 254230505
Conc: 50.00 ng/ml




#4 Heptachlor

R.T.: 4.914 min
Delta R.T.: 0.000 min
Response: 166264355
Conc: 50.00 ng/ml



#4 Heptachlor


R.T.: 3.945 min
Delta R.T.: 0.000 min
Response: 246242003
Conc: 50.00 ng/ml



#5 Aldrin


R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 164631568
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#5 Aldrin

R.T.: 4.225 min
Delta R.T.: 0.000 min
Response: 242826175
Conc: 50.00 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 80914447
Conc: 50.00 ng/ml



#6 beta-BHC


R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 103608976
Conc: 50.00 ng/ml



#7 delta-BHC


R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 180293977
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#7 delta-BHC

R.T.: 4.136 min
Delta R.T.: 0.000 min
Response: 254940473
Conc: 50.00 ng/ml




#8 Heptachlor epoxide

R.T.: 5.682 min
Delta R.T.: 0.000 min
Response: 147681633
Conc: 50.00 ng/ml



#8 Heptachlor epoxide


R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 221208450
Conc: 50.00 ng/ml



#9 Endosulfan I


R.T.: 6.067 min
Delta R.T.: 0.000 min
Response: 131853108
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#9 Endosulfan I

R.T.: 5.096 min
Delta R.T.: 0.000 min
Response: 204951644
Conc: 50.00 ng/ml




#10 gamma-Chlordane

R.T.: 5.938 min
Delta R.T.: 0.000 min
Response: 140781558
Conc: 50.00 ng/ml



#10 gamma-Chlordane


R.T.: 4.977 min
Delta R.T.: 0.000 min
Response: 224150330
Conc: 50.00 ng/ml



#11 alpha-Chlordane


R.T.: 6.017 min
Delta R.T.: 0.000 min
Response: 139410024
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#11 alpha-Chlordane

R.T.: 5.040 min
Delta R.T.: 0.000 min
Response: 221205596
Conc: 50.00 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 124454141
Conc: 50.00 ng/ml



#12 4,4'-DDE


R.T.: 5.230 min
Delta R.T.: 0.000 min
Response: 212682331
Conc: 50.00 ng/ml



#13 Dieldrin


R.T.: 6.343 min
Delta R.T.: 0.000 min
Response: 139409444
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#13 Dieldrin

R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 227678389
Conc: 50.00 ng/ml




#14 Endrin

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 118161114
Conc: 50.00 ng/ml



#14 Endrin


R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 193536252
Conc: 50.00 ng/ml



#15 Endosulfan II


R.T.: 6.792 min
Delta R.T.: 0.000 min
Response: 120697329
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#15 Endosulfan II

R.T.: 5.931 min
Delta R.T.: 0.000 min
Response: 195648200
Conc: 50.00 ng/ml




#16 4,4'-DDD

R.T.: 6.708 min
Delta R.T.: 0.000 min
Response: 96633741
Conc: 50.00 ng/ml



#16 4,4'-DDD


R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 168957879
Conc: 50.00 ng/ml



#17 4,4'-DDT


R.T.: 7.022 min
Delta R.T.: 0.000 min
Response: 100835874
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#17 4,4'-DDT

R.T.: 6.034 min
Delta R.T.: 0.000 min
Response: 177142802
Conc: 50.00 ng/ml




#18 Endrin aldehyde

R.T.: 6.923 min
Delta R.T.: 0.000 min
Response: 97948460
Conc: 50.00 ng/ml



#18 Endrin aldehyde


R.T.: 6.110 min
Delta R.T.: 0.000 min
Response: 159171518
Conc: 50.00 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.157 min
Delta R.T.: 0.000 min
Response: 112428845
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#19 Endosulfan Sulfate

R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 187851270
Conc: 50.00 ng/ml




#20 Methoxychlor

R.T.: 7.498 min
Delta R.T.: 0.000 min
Response: 54018493
Conc: 50.00 ng/ml



#20 Methoxychlor


R.T.: 6.609 min
Delta R.T.: 0.000 min
Response: 93520516
Conc: 50.00 ng/ml



#21 Endrin ketone


R.T.: 7.642 min
Delta R.T.: 0.000 min
Response: 126985241
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#21 Endrin ketone

R.T.: 6.839 min
Delta R.T.: 0.000 min
Response: 220004203
Conc: 50.00 ng/ml




#22 Mirex

R.T.: 8.115 min
Delta R.T.: 0.000 min
Response: 103275089
Conc: 50.00 ng/ml



#22 Mirex


R.T.: 7.018 min
Delta R.T.: 0.000 min
Response: 173654301
Conc: 50.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: 0.000 min
Response: 104915987
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050



#28 Decachlorobiphenyl

R.T.: 7.910 min
Delta R.T.: 0.000 min
Response: 181351234
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
 Data File : PL093731.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2025 11:38
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 13:59:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 13:52:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.774	64887383	76654930	25.606	24.213
28) SA Decachlor...	9.052	7.909	50461717	83015469	26.208	24.840
Target Compounds						
2) A alpha-BHC	3.995	3.276	89070727	112.0E6	24.556	23.055
3) MA gamma-BHC...	4.327	3.607	86520755	109.6E6	24.821	23.349
4) MA Heptachlor	4.914	3.945	78602546	108.6E6	25.575	23.892
5) MB Aldrin	5.256	4.224	77491525	105.6E6	25.379	23.546
6) B beta-BHC	4.525	3.907	37722252	47243605	25.509	24.644
7) B delta-BHC	4.772	4.135	82584367	109.2E6	24.767	23.218
8) B Heptachlo...	5.683	4.726	70895850	98198954	25.936	24.060
9) A Endosulfan I	6.068	5.096	63215335	90882897	25.885	24.027
10) B gamma-Chl...	5.939	4.976	66959715	98387231	25.700	23.747
11) B alpha-Chl...	6.017	5.040	66664408	97870181	25.711	23.932
12) B 4,4'-DDE	6.191	5.229	58039701	93725344	25.344	23.876
13) MA Dieldrin	6.343	5.360	65983515	98970864	25.563	23.595
14) MA Endrin	6.572	5.636	55464000	85153560	25.436	23.710
15) B Endosulfa...	6.793	5.931	57195569	87112831	25.745	24.134
16) A 4,4'-DDD	6.708	5.784	45068086	72761634	25.477	23.324
17) MA 4,4'-DDT	7.022	6.034	47678056	76172310	25.612	23.370
18) B Endrin al...	6.923	6.110	47414192	72307343	26.251	24.600
19) B Endosulfa...	7.157	6.333	54762628	83706831	26.369	24.145
20) A Methoxychlor	7.498	6.609	25502321	41095325	25.958	24.173
21) B Endrin ke...	7.642	6.838	60347677	97684233	25.801	24.188
22) Mirex	8.115	7.018	50874505	81084696	26.652	25.185



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093731.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 11:38
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 13:59:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 13:52:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.539 min
Delta R.T.: 0.000 min
Response: 64887383
Conc: 25.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 76654930
Conc: 24.21 ng/ml




#2 alpha-BHC

R.T.: 3.995 min
Delta R.T.: 0.000 min
Response: 89070727
Conc: 24.56 ng/ml



#2 alpha-BHC


R.T.: 3.276 min
Delta R.T.: 0.000 min
Response: 112018157
Conc: 23.06 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.327 min
Delta R.T.: 0.000 min
Response: 86520755
Conc: 24.82 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#3 gamma-BHC (Lindane)

R.T.: 3.607 min
Delta R.T.: 0.000 min
Response: 109620260
Conc: 23.35 ng/ml




#4 Heptachlor

R.T.: 4.914 min
Delta R.T.: 0.000 min
Response: 78602546
Conc: 25.58 ng/ml



#4 Heptachlor


R.T.: 3.945 min
Delta R.T.: 0.000 min
Response: 108649410
Conc: 23.89 ng/ml



#5 Aldrin


R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 77491525
Conc: 25.38 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#5 Aldrin

R.T.: 4.224 min
Delta R.T.: 0.000 min
Response: 105561790
Conc: 23.55 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 37722252
Conc: 25.51 ng/ml



#6 beta-BHC


R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 47243605
Conc: 24.64 ng/ml



#7 delta-BHC


R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 82584367
Conc: 24.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#7 delta-BHC

R.T.: 4.135 min
Delta R.T.: 0.000 min
Response: 109220381
Conc: 23.22 ng/ml




#8 Heptachlor epoxide

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 70895850
Conc: 25.94 ng/ml



#8 Heptachlor epoxide


R.T.: 4.726 min
Delta R.T.: 0.000 min
Response: 98198954
Conc: 24.06 ng/ml



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 63215335
Conc: 25.89 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#9 Endosulfan I

R.T.: 5.096 min
Delta R.T.: 0.000 min
Response: 90882897
Conc: 24.03 ng/ml




#10 gamma-Chlordane

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 66959715
Conc: 25.70 ng/ml



#10 gamma-Chlordane


R.T.: 4.976 min
Delta R.T.: 0.000 min
Response: 98387231
Conc: 23.75 ng/ml



#11 alpha-Chlordane


R.T.: 6.017 min
Delta R.T.: 0.000 min
Response: 66664408
Conc: 25.71 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#11 alpha-Chlordane

R.T.: 5.040 min
Delta R.T.: 0.000 min
Response: 97870181
Conc: 23.93 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 58039701
Conc: 25.34 ng/ml



#12 4,4'-DDE


R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 93725344
Conc: 23.88 ng/ml



#13 Dieldrin


R.T.: 6.343 min
Delta R.T.: 0.000 min
Response: 65983515
Conc: 25.56 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#13 Dieldrin

R.T.: 5.360 min
Delta R.T.: 0.000 min
Response: 98970864
Conc: 23.59 ng/ml




#14 Endrin

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 55464000
Conc: 25.44 ng/ml



#14 Endrin


R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 85153560
Conc: 23.71 ng/ml



#15 Endosulfan II


R.T.: 6.793 min
Delta R.T.: 0.000 min
Response: 57195569
Conc: 25.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#15 Endosulfan II

R.T.: 5.931 min
Delta R.T.: 0.000 min
Response: 87112831
Conc: 24.13 ng/ml




#16 4,4'-DDD

R.T.: 6.708 min
Delta R.T.: 0.000 min
Response: 45068086
Conc: 25.48 ng/ml



#16 4,4'-DDD


R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 72761634
Conc: 23.32 ng/ml



#17 4,4'-DDT


R.T.: 7.022 min
Delta R.T.: 0.000 min
Response: 47678056
Conc: 25.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#17 4,4'-DDT

R.T.: 6.034 min
Delta R.T.: 0.000 min
Response: 76172310
Conc: 23.37 ng/ml




#18 Endrin aldehyde

R.T.: 6.923 min
Delta R.T.: 0.000 min
Response: 47414192
Conc: 26.25 ng/ml



#18 Endrin aldehyde


R.T.: 6.110 min
Delta R.T.: 0.000 min
Response: 72307343
Conc: 24.60 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.157 min
Delta R.T.: 0.000 min
Response: 54762628
Conc: 26.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#19 Endosulfan Sulfate

R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 83706831
Conc: 24.15 ng/ml




#20 Methoxychlor

R.T.: 7.498 min
Delta R.T.: 0.000 min
Response: 25502321
Conc: 25.96 ng/ml



#20 Methoxychlor


R.T.: 6.609 min
Delta R.T.: 0.000 min
Response: 41095325
Conc: 24.17 ng/ml



#21 Endrin ketone


R.T.: 7.642 min
Delta R.T.: 0.000 min
Response: 60347677
Conc: 25.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#21 Endrin ketone

R.T.: 6.838 min
Delta R.T.: 0.000 min
Response: 97684233
Conc: 24.19 ng/ml




#22 Mirex

R.T.: 8.115 min
Delta R.T.: 0.000 min
Response: 50874505
Conc: 26.65 ng/ml



#22 Mirex


R.T.: 7.018 min
Delta R.T.: 0.000 min
Response: 81084696
Conc: 25.19 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.052 min
Delta R.T.: 0.000 min
Response: 50461717
Conc: 26.21 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025



#28 Decachlorobiphenyl

R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 83015469
Conc: 24.84 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
 Data File : PL093732.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2025 11:51
 Operator : AR\AJ
 Sample : PSTDICC005
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC005

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 14:01:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 13:52:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.774	16637105	18287931	6.178	5.603
28) SA Decachlor...	9.052	7.909	13789093	20761045	6.592	5.925
Target Compounds						
2) A alpha-BHC	3.994	3.276	23301548	25051289	6.078	5.124
3) MA gamma-BHC...	4.326	3.606	22354233	24631359	6.070	5.195
4) MA Heptachlor	4.914	3.944	20465600	25421102	6.245	5.461
5) MB Aldrin	5.255	4.223	20732862	24380948	6.337	5.345
6) B beta-BHC	4.525	3.906	10607660	11595524	6.600	5.805
7) B delta-BHC	4.771	4.134	20943898	24697126	5.975	5.198
8) B Heptachlo...	5.682	4.726	19675106	22877181	6.616	5.473
9) A Endosulfan I	6.068	5.096	17228246	21272747	6.519	5.487
10) B gamma-Chl...	5.938	4.976	17575834	23077513	6.306	5.446
11) B alpha-Chl...	6.017	5.040	17853432	22874114	6.403	5.464
12) B 4,4'-DDE	6.191	5.229	15062588	21725638	6.187	5.419
13) MA Dieldrin	6.343	5.360	17771692	23498784	6.402	5.470
14) MA Endrin	6.573	5.635	15009439	20488065	6.401	5.548
15) B Endosulfa...	6.793	5.930	15801314	20403798	6.558	5.509
16) A 4,4'-DDD	6.708	5.783	12134151	16521614	6.385	5.234
17) MA 4,4'-DDT	7.022	6.034	12070833	16163358	6.121	4.967
18) B Endrin al...	6.922	6.110	12477919	17329206	6.418	5.692
19) B Endosulfa...	7.156	6.332	15057236	19816189	6.651	5.557
20) A Methoxychlor	7.498	6.609	6435643	10701964	6.168	5.985
21) B Endrin ke...	7.642	6.837	16285626	24108712	6.456	5.747
22) Mirex	8.114	7.018	13884960	20156166	6.667	5.960



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093732.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 11:51
Operator : AR\AJ
Sample : PSTDICC005
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC005

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 14:01:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 13:52:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 16637105
Conc: 6.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 18287931
Conc: 5.60 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 23301548
Conc: 6.08 ng/ml



#2 alpha-BHC


R.T.: 3.276 min
Delta R.T.: 0.000 min
Response: 25051289
Conc: 5.12 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.326 min
Delta R.T.: 0.000 min
Response: 22354233
Conc: 6.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#3 gamma-BHC (Lindane)

R.T.: 3.606 min
Delta R.T.: 0.000 min
Response: 24631359
Conc: 5.20 ng/ml




#4 Heptachlor

R.T.: 4.914 min
Delta R.T.: 0.000 min
Response: 20465600
Conc: 6.24 ng/ml



#4 Heptachlor


R.T.: 3.944 min
Delta R.T.: 0.000 min
Response: 25421102
Conc: 5.46 ng/ml



#5 Aldrin


R.T.: 5.255 min
Delta R.T.: 0.000 min
Response: 20732862
Conc: 6.34 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#5 Aldrin

R.T.: 4.223 min
Delta R.T.: 0.000 min
Response: 24380948
Conc: 5.34 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 10607660
Conc: 6.60 ng/ml



#6 beta-BHC


R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 11595524
Conc: 5.81 ng/ml



#7 delta-BHC


R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 20943898
Conc: 5.97 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#7 delta-BHC

R.T.: 4.134 min
Delta R.T.: 0.000 min
Response: 24697126
Conc: 5.20 ng/ml




#8 Heptachlor epoxide

R.T.: 5.682 min
Delta R.T.: 0.000 min
Response: 19675106
Conc: 6.62 ng/ml



#8 Heptachlor epoxide


R.T.: 4.726 min
Delta R.T.: 0.000 min
Response: 22877181
Conc: 5.47 ng/ml



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 17228246
Conc: 6.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#9 Endosulfan I

R.T.: 5.096 min
Delta R.T.: 0.000 min
Response: 21272747
Conc: 5.49 ng/ml




#10 gamma-Chlordane

R.T.: 5.938 min
Delta R.T.: 0.000 min
Response: 17575834
Conc: 6.31 ng/ml



#10 gamma-Chlordane


R.T.: 4.976 min
Delta R.T.: 0.000 min
Response: 23077513
Conc: 5.45 ng/ml



#11 alpha-Chlordane


R.T.: 6.017 min
Delta R.T.: 0.000 min
Response: 17853432
Conc: 6.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#11 alpha-Chlordane

R.T.: 5.040 min
Delta R.T.: 0.000 min
Response: 22874114
Conc: 5.46 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 15062588
Conc: 6.19 ng/ml



#12 4,4'-DDE


R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 21725638
Conc: 5.42 ng/ml



#13 Dieldrin


R.T.: 6.343 min
Delta R.T.: 0.000 min
Response: 17771692
Conc: 6.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#13 Dieldrin

R.T.: 5.360 min
Delta R.T.: 0.000 min
Response: 23498784
Conc: 5.47 ng/ml




#14 Endrin

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 15009439
Conc: 6.40 ng/ml



#14 Endrin


R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 20488065
Conc: 5.55 ng/ml



#15 Endosulfan II


R.T.: 6.793 min
Delta R.T.: 0.000 min
Response: 15801314
Conc: 6.56 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#15 Endosulfan II

R.T.: 5.930 min
Delta R.T.: 0.000 min
Response: 20403798
Conc: 5.51 ng/ml




#16 4,4'-DDD

R.T.: 6.708 min
Delta R.T.: 0.000 min
Response: 12134151
Conc: 6.38 ng/ml



#16 4,4'-DDD


R.T.: 5.783 min
Delta R.T.: 0.000 min
Response: 16521614
Conc: 5.23 ng/ml



#17 4,4'-DDT


R.T.: 7.022 min
Delta R.T.: 0.000 min
Response: 12070833
Conc: 6.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#17 4,4'-DDT

R.T.: 6.034 min
Delta R.T.: 0.000 min
Response: 16163358
Conc: 4.97 ng/ml




#18 Endrin aldehyde

R.T.: 6.922 min
Delta R.T.: 0.000 min
Response: 12477919
Conc: 6.42 ng/ml



#18 Endrin aldehyde


R.T.: 6.110 min
Delta R.T.: 0.000 min
Response: 17329206
Conc: 5.69 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.156 min
Delta R.T.: 0.000 min
Response: 15057236
Conc: 6.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#19 Endosulfan Sulfate

R.T.: 6.332 min
Delta R.T.: 0.000 min
Response: 19816189
Conc: 5.56 ng/ml




#20 Methoxychlor

R.T.: 7.498 min
Delta R.T.: 0.000 min
Response: 6435643
Conc: 6.17 ng/ml



#20 Methoxychlor


R.T.: 6.609 min
Delta R.T.: 0.000 min
Response: 10701964
Conc: 5.98 ng/ml



#21 Endrin ketone


R.T.: 7.642 min
Delta R.T.: 0.000 min
Response: 16285626
Conc: 6.46 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#21 Endrin ketone

R.T.: 6.837 min
Delta R.T.: 0.000 min
Response: 24108712
Conc: 5.75 ng/ml




#22 Mirex

R.T.: 8.114 min
Delta R.T.: 0.000 min
Response: 13884960
Conc: 6.67 ng/ml



#22 Mirex

R.T.: 7.018 min
Delta R.T.: 0.000 min
Response: 20156166
Conc: 5.96 ng/ml




#28 Decachlorobiphenyl

ICAL Form

R.T.: 9.052 min
Delta R.T.: 0.000 min
Response: 13789093
Conc: 6.59 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005



#28 Decachlorobiphenyl

R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 20761045
Conc: 5.92 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093735.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 12:32
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 13:40:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 13:40:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.773	118.0E6	178.8E6	50.000	50.000
28) SA Decachlor...	9.053	7.909	91356144	160.3E6	50.000	50.000
Target Compounds						
23) Chlordane-1	4.700	3.771	55335446	61106259	500.000	500.000
24) Chlordane-2	5.229	4.347	55911116	70304921	500.000	500.000
25) Chlordane-3	5.939	4.977	183.8E6	213.9E6	500.000	500.000
26) Chlordane-4	6.021	5.039	220.6E6	206.1E6	500.000	500.000
27) Chlordane-5	6.870	5.935	42155882	74355315	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093735.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 12:32
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 13:40:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 13:40:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 118044809
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: 0.000 min
Response: 178786091
Conc: 50.00 ng/ml




#23 Chlordane-1

R.T.: 4.700 min
Delta R.T.: 0.000 min
Response: 55335446
Conc: 500.00 ng/ml



#23 Chlordane-1


R.T.: 3.771 min
Delta R.T.: 0.000 min
Response: 61106259
Conc: 500.00 ng/ml



#24 Chlordane-2


R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 55911116
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#24 Chlordane-2

R.T.: 4.347 min
Delta R.T.: 0.000 min
Response: 70304921
Conc: 500.00 ng/ml




#25 Chlordane-3

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 183782045
Conc: 500.00 ng/ml



#25 Chlordane-3


R.T.: 4.977 min
Delta R.T.: 0.000 min
Response: 213941021
Conc: 500.00 ng/ml



#26 Chlordane-4


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 220583333
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#26 Chlordane-4

R.T.: 5.039 min
Delta R.T.: 0.000 min
Response: 206126766
Conc: 500.00 ng/ml




#27 Chlordane-5

R.T.: 6.870 min
Delta R.T.: 0.000 min
Response: 42155882
Conc: 500.00 ng/ml



#27 Chlordane-5


R.T.: 5.935 min
Delta R.T.: 0.000 min
Response: 74355315
Conc: 500.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: 0.000 min
Response: 91356144
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500



#28 Decachlorobiphenyl

R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 160259410
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093740.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 13:39
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PTOXICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 14:13:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:13:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.774	124.5E6	157.4E6	50.000	50.000
7) SA Decachlor...	9.053	7.909	96684586	169.0E6	50.000	50.000
Target Compounds						
2) Toxaphene-1	6.235	5.002	11722975	13528528	500.000	500.000
3) Toxaphene-2	6.440	5.326	7383579	11973587	500.000	500.000
4) Toxaphene-3	7.058	5.684	37947977	12363221	500.000	500.000
5) Toxaphene-4	7.147	6.599	28672538	42493596	500.000	500.000
6) Toxaphene-5	7.932	7.039	21533557	40119156	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093740.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 13:39
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PTOXICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 14:13:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:13:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
Signal #1 Info : 30M x 0.32mm x 0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.539 min
Delta R.T.: 0.000 min
Response: 124524341
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PTOXICC500




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 157364468
Conc: 50.00 ng/ml




#2 Toxaphene-1

R.T.: 6.235 min
Delta R.T.: 0.000 min
Response: 11722975
Conc: 500.00 ng/ml



#2 Toxaphene-1


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 13528528
Conc: 500.00 ng/ml



#3 Toxaphene-2


R.T.: 6.440 min
Delta R.T.: 0.000 min
Response: 7383579
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PTOXICC500




#3 Toxaphene-2

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 11973587
Conc: 500.00 ng/ml




#4 Toxaphene-3

R.T.: 7.058 min
Delta R.T.: 0.000 min
Response: 37947977
Conc: 500.00 ng/ml



#4 Toxaphene-3


R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 12363221
Conc: 500.00 ng/ml



#5 Toxaphene-4


R.T.: 7.147 min
Delta R.T.: 0.000 min
Response: 28672538
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PTOXICC500




#5 Toxaphene-4

R.T.: 6.599 min
Delta R.T.: 0.000 min
Response: 42493596
Conc: 500.00 ng/ml




#6 Toxaphene-5

R.T.: 7.932 min
Delta R.T.: 0.000 min
Response: 21533557
Conc: 500.00 ng/ml



#6 Toxaphene-5


R.T.: 7.039 min
Delta R.T.: 0.000 min
Response: 40119156
Conc: 500.00 ng/ml



#7 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: 0.000 min
Response: 96684586
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PTOXICC500



#7 Decachlorobiphenyl

R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 169042393
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
 Data File : PL093743.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2025 14:20
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL012125

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 14:28:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:02:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.774	141.2E6	176.6E6	52.434	54.093
28) SA Decachlor...	9.054	7.910	108.7E6	190.8E6	51.975	54.460
Target Compounds						
2) A alpha-BHC	3.995	3.277	200.6E6	269.7E6	52.316	55.168
3) MA gamma-BHC...	4.327	3.607	192.5E6	260.8E6	52.258	55.000
4) MA Heptachlor	4.915	3.945	166.4E6	253.0E6	50.760	54.351
5) MB Aldrin	5.256	4.225	167.1E6	250.6E6	51.084	54.939
6) B beta-BHC	4.525	3.907	81194319	107.5E6	50.515	53.798
7) B delta-BHC	4.772	4.135	181.5E6	260.8E6	51.768	54.892
8) B Heptachlo...	5.683	4.727	150.9E6	228.4E6	50.731	54.639
9) A Endosulfan I	6.068	5.096	134.8E6	212.0E6	51.009	54.694
10) B gamma-Chl...	5.938	4.977	144.6E6	231.3E6	51.881	54.579
11) B alpha-Chl...	6.017	5.041	143.8E6	228.4E6	51.572	54.561
12) B 4,4'-DDE	6.191	5.229	129.1E6	221.4E6	53.039	55.209
13) MA Dieldrin	6.343	5.361	143.4E6	234.2E6	51.649	54.528
14) MA Endrin	6.573	5.636	118.4E6	194.0E6	50.501	52.537
15) B Endosulfa...	6.793	5.931	124.2E6	202.0E6	51.563	54.540
16) A 4,4'-DDD	6.709	5.785	101.9E6	178.4E6	53.599	56.515
17) MA 4,4'-DDT	7.023	6.035	104.2E6	181.3E6	52.862	55.711
18) B Endrin al...	6.923	6.111	101.8E6	165.3E6	52.345	54.299
19) B Endosulfa...	7.158	6.333	115.9E6	193.8E6	51.182	54.342
20) A Methoxychlor	7.499	6.610	56233890	95758805	53.895	53.552
21) B Endrin ke...	7.642	6.838	132.2E6	231.2E6	52.424	55.120
22) Mirex	8.115	7.019	106.9E6	182.6E6	51.338	54.005



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093743.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 14:20
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL012125

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 14:28:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:02:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.539 min
Delta R.T.: 0.000 min
Response: 141191268
Conc: 52.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 176566835
Conc: 54.09 ng/ml




#2 alpha-BHC

R.T.: 3.995 min
Delta R.T.: 0.000 min
Response: 200569905
Conc: 52.32 ng/ml



#2 alpha-BHC


R.T.: 3.277 min
Delta R.T.: 0.000 min
Response: 269716716
Conc: 55.17 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.327 min
Delta R.T.: 0.000 min
Response: 192459165
Conc: 52.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125




#3 gamma-BHC (Lindane)

R.T.: 3.607 min
Delta R.T.: 0.000 min
Response: 260765963
Conc: 55.00 ng/ml




#4 Heptachlor

R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 166357805
Conc: 50.76 ng/ml



#4 Heptachlor


R.T.: 3.945 min
Delta R.T.: 0.000 min
Response: 252994081
Conc: 54.35 ng/ml



#5 Aldrin


R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 167145833
Conc: 51.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125




#5 Aldrin

R.T.: 4.225 min
Delta R.T.: 0.001 min
Response: 250621610
Conc: 54.94 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 81194319
Conc: 50.52 ng/ml



#6 beta-BHC


R.T.: 3.907 min
Delta R.T.: 0.001 min
Response: 107457324
Conc: 53.80 ng/ml



#7 delta-BHC


R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 181463119
Conc: 51.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125




#7 delta-BHC

R.T.: 4.135 min
Delta R.T.: 0.000 min
Response: 260806257
Conc: 54.89 ng/ml




#8 Heptachlor epoxide

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 150865872
Conc: 50.73 ng/ml



#8 Heptachlor epoxide


R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 228406178
Conc: 54.64 ng/ml



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 134808929
Conc: 51.01 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125




#9 Endosulfan I

R.T.: 5.096 min
Delta R.T.: 0.000 min
Response: 212043680
Conc: 54.69 ng/ml




#10 gamma-Chlordane

R.T.: 5.938 min
Delta R.T.: 0.000 min
Response: 144611164
Conc: 51.88 ng/ml



#10 gamma-Chlordane


R.T.: 4.977 min
Delta R.T.: 0.000 min
Response: 231285603
Conc: 54.58 ng/ml



#11 alpha-Chlordane


R.T.: 6.017 min
Delta R.T.: 0.000 min
Response: 143803294
Conc: 51.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125




#11 alpha-Chlordane

R.T.: 5.041 min
Delta R.T.: 0.001 min
Response: 228422304
Conc: 54.56 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 129128935
Conc: 53.04 ng/ml



#12 4,4'-DDE


R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 221359531
Conc: 55.21 ng/ml



#13 Dieldrin


R.T.: 6.343 min
Delta R.T.: 0.000 min
Response: 143369662
Conc: 51.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125




#13 Dieldrin

R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 234235081
Conc: 54.53 ng/ml




#14 Endrin

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 118416700
Conc: 50.50 ng/ml



#14 Endrin


R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 194001333
Conc: 52.54 ng/ml



#15 Endosulfan II


R.T.: 6.793 min
Delta R.T.: 0.000 min
Response: 124234027
Conc: 51.56 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125




#15 Endosulfan II

R.T.: 5.931 min
Delta R.T.: 0.001 min
Response: 202005610
Conc: 54.54 ng/ml




#16 4,4'-DDD

R.T.: 6.709 min
Delta R.T.: 0.000 min
Response: 101868364
Conc: 53.60 ng/ml



#16 4,4'-DDD


R.T.: 5.785 min
Delta R.T.: 0.001 min
Response: 178390286
Conc: 56.51 ng/ml



#17 4,4'-DDT


R.T.: 7.023 min
Delta R.T.: 0.000 min
Response: 104247092
Conc: 52.86 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125




#17 4,4'-DDT

R.T.: 6.035 min
Delta R.T.: 0.000 min
Response: 181287401
Conc: 55.71 ng/ml




#18 Endrin aldehyde

R.T.: 6.923 min
Delta R.T.: 0.000 min
Response: 101761385
Conc: 52.34 ng/ml



#18 Endrin aldehyde


R.T.: 6.111 min
Delta R.T.: 0.001 min
Response: 165320369
Conc: 54.30 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.158 min
Delta R.T.: 0.002 min
Response: 115863257
Conc: 51.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125




#19 Endosulfan Sulfate

R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 193788608
Conc: 54.34 ng/ml




#20 Methoxychlor

R.T.: 7.499 min
Delta R.T.: 0.000 min
Response: 56233890
Conc: 53.90 ng/ml



#20 Methoxychlor


R.T.: 6.610 min
Delta R.T.: 0.000 min
Response: 95758805
Conc: 53.55 ng/ml



#21 Endrin ketone


R.T.: 7.642 min
Delta R.T.: 0.000 min
Response: 132246340
Conc: 52.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125




#21 Endrin ketone

R.T.: 6.838 min
Delta R.T.: 0.000 min
Response: 231241910
Conc: 55.12 ng/ml




#22 Mirex

R.T.: 8.115 min
Delta R.T.: 0.000 min
Response: 106911145
Conc: 51.34 ng/ml



#22 Mirex


R.T.: 7.019 min
Delta R.T.: 0.000 min
Response: 182637497
Conc: 54.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.054 min
Delta R.T.: 0.001 min
Response: 108726317
Conc: 51.97 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL012125



#28 Decachlorobiphenyl

R.T.: 7.910 min
Delta R.T.: 0.001 min
Response: 190831116
Conc: 54.46 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 11:17 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	4.00	3.90	4.10	0.01
beta-BHC	4.53	4.53	4.43	4.63	0.00
delta-BHC	4.77	4.77	4.67	4.87	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.91	4.91	4.81	5.01	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.79	6.79	6.69	6.89	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.64	7.64	7.54	7.74	0.00
Endrin aldehyde	6.92	6.92	6.82	7.02	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00

CALIBRATION VERIFICATION SUMMARY

 Contract: RUTW01

 Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

 Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

 Continuing Calib Time: 11:17 Initial Calibration Time(s): 10:57 11:51

 GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
alpha-BHC	3.28	3.28	3.18	3.38	0.00
beta-BHC	3.91	3.91	3.81	4.01	0.00
delta-BHC	4.14	4.14	4.04	4.24	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.95	3.95	3.85	4.05	0.01
Aldrin	4.22	4.23	4.13	4.33	0.01
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endosulfan I	5.10	5.10	5.00	5.20	0.00
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Endosulfan II	5.93	5.93	5.83	6.03	0.00
4,4'-DDD	5.78	5.78	5.68	5.88	0.00
Endosulfan sulfate	6.33	6.33	6.23	6.43	0.00
4,4'-DDT	6.03	6.03	5.93	6.13	0.00
Methoxychlor	6.61	6.61	6.51	6.71	0.00
Endrin ketone	6.84	6.84	6.74	6.94	0.00
Endrin aldehyde	6.11	6.11	6.01	6.21	0.00
alpha-Chlordane	5.04	5.04	4.94	5.14	0.00
gamma-Chlordane	4.98	4.98	4.88	5.08	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/31/2025

Lab Sample No.: PSTDCCC050 Data File : PL093930.D Time Analyzed: 11:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.709	6.608	6.808	50.910	50.000	1.8
4,4'-DDE	6.192	6.091	6.291	50.300	50.000	0.6
4,4'-DDT	7.023	6.922	7.122	52.710	50.000	5.4
Aldrin	5.256	5.156	5.356	45.640	50.000	-8.7
alpha-BHC	3.993	3.895	4.095	47.930	50.000	-4.1
alpha-Chlordane	6.018	5.917	6.117	46.290	50.000	-7.4
beta-BHC	4.525	4.425	4.625	47.920	50.000	-4.2
Decachlorobiphenyl	9.055	8.953	9.153	45.000	50.000	-10.0
delta-BHC	4.772	4.672	4.872	44.790	50.000	-10.4
Dieldrin	6.344	6.243	6.443	45.550	50.000	-8.9
Endosulfan I	6.068	5.967	6.167	45.360	50.000	-9.3
Endosulfan II	6.793	6.692	6.892	45.890	50.000	-8.2
Endosulfan sulfate	7.158	7.057	7.257	44.420	50.000	-11.2
Endrin	6.572	6.472	6.672	46.850	50.000	-6.3
Endrin aldehyde	6.923	6.823	7.023	43.790	50.000	-12.4
Endrin ketone	7.644	7.542	7.742	44.060	50.000	-11.9
gamma-BHC (Lindane)	4.326	4.227	4.427	47.240	50.000	-5.5
gamma-Chlordane	5.939	5.838	6.038	46.300	50.000	-7.4
Heptachlor	4.914	4.814	5.014	49.110	50.000	-1.8
Heptachlor epoxide	5.683	5.582	5.782	44.870	50.000	-10.3
Methoxychlor	7.500	7.398	7.598	51.410	50.000	2.8
Tetrachloro-m-xylene	3.538	3.439	3.639	47.330	50.000	-5.3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/31/2025

Lab Sample No.: PSTDCCC050 Data File : PL093930.D Time Analyzed: 11:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.784	5.684	5.884	51.360	50.000	2.7
4,4'-DDE	5.229	5.130	5.330	51.230	50.000	2.5
4,4'-DDT	6.034	5.934	6.134	53.970	50.000	7.9
Aldrin	4.224	4.125	4.325	46.010	50.000	-8.0
alpha-BHC	3.276	3.177	3.377	48.050	50.000	-3.9
alpha-Chlordane	5.040	4.940	5.140	47.780	50.000	-4.4
beta-BHC	3.906	3.807	4.007	47.900	50.000	-4.2
Decachlorobiphenyl	7.910	7.810	8.010	47.390	50.000	-5.2
delta-BHC	4.135	4.036	4.236	46.170	50.000	-7.7
Dieldrin	5.361	5.261	5.461	47.240	50.000	-5.5
Endosulfan I	5.096	4.996	5.196	44.020	50.000	-12.0
Endosulfan II	5.931	5.831	6.031	47.980	50.000	-4.0
Endosulfan sulfate	6.333	6.233	6.433	47.140	50.000	-5.7
Endrin	5.636	5.536	5.736	51.340	50.000	2.7
Endrin aldehyde	6.110	6.010	6.210	44.640	50.000	-10.7
Endrin ketone	6.839	6.739	6.939	45.680	50.000	-8.6
gamma-BHC (Lindane)	3.606	3.507	3.707	45.940	50.000	-8.1
gamma-Chlordane	4.977	4.877	5.077	48.730	50.000	-2.5
Heptachlor	3.945	3.845	4.045	48.380	50.000	-3.2
Heptachlor epoxide	4.727	4.627	4.827	46.190	50.000	-7.6
Methoxychlor	6.610	6.509	6.709	52.540	50.000	5.1
Tetrachloro-m-xylene	2.774	2.674	2.874	47.580	50.000	-4.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093930.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 11:17
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:23:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.774	127.4E6	155.3E6	47.327	47.575
28) SA Decachlor...	9.055	7.910	94144637	166.0E6	45.004	47.385
Target Compounds						
2) A alpha-BHC	3.993	3.276	183.7E6	234.9E6	47.926	48.046
3) MA gamma-BHC...	4.326	3.606	174.0E6	217.8E6	47.236	45.944
4) MA Heptachlor	4.914	3.945	160.9E6	225.2E6	49.109	48.385
5) MB Aldrin	5.256	4.224	149.3E6	209.9E6	45.636	46.013
6) B beta-BHC	4.525	3.906	77014994	95671218	47.915	47.897
7) B delta-BHC	4.772	4.135	157.0E6	219.4E6	44.786	46.173
8) B Heptachlo...	5.683	4.727	133.4E6	193.1E6	44.870	46.194
9) A Endosulfan I	6.068	5.096	119.9E6	170.7E6	45.357	44.018
10) B gamma-Chl...	5.939	4.977	129.1E6	206.5E6	46.299	48.732
11) B alpha-Chl...	6.018	5.040	129.1E6	200.0E6	46.293	47.776
12) B 4,4'-DDE	6.192	5.229	122.5E6	205.4E6	50.304	51.231
13) MA Dieldrin	6.344	5.361	126.4E6	202.9E6	45.545	47.243
14) MA Endrin	6.572	5.636	109.9E6	189.6E6	46.848m	51.340
15) B Endosulfa...	6.793	5.931	110.6E6	177.7E6	45.886	47.978
16) A 4,4'-DDD	6.709	5.784	96757725	162.1E6	50.910	51.357
17) MA 4,4'-DDT	7.023	6.034	103.9E6	175.6E6	52.706	53.968
18) B Endrin al...	6.923	6.110	85137438	135.9E6	43.794	44.644
19) B Endosulfa...	7.158	6.333	100.5E6	168.1E6	44.416	47.138
20) A Methoxychlor	7.500	6.610	53646135	93951668	51.415	52.542
21) B Endrin ke...	7.644	6.839	111.1E6	191.6E6	44.056	45.678
22) Mirex	8.116	7.019	86631474	145.6E6	41.600	43.065

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093930.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 11:17
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:23:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.538 min
Delta R.T.: -0.001 min
Response: 127440428
Conc: 47.33 ng/ml

Instrument :

ECD_L


ClientSampleId :

PSTDCCC050

Manual Integrations


APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 155293788
Conc: 47.58 ng/ml




#2 alpha-BHC

R.T.: 3.993 min
Delta R.T.: -0.001 min
Response: 183740365
Conc: 47.93 ng/ml



#2 alpha-BHC

R.T.: 3.276 min
Delta R.T.: 0.000 min
Response: 234896985
Conc: 48.05 ng/ml



#3 gamma-BHC (Lindane)

R.T.: 4.326 min
Delta R.T.: 0.000 min
Response: 173961937
Conc: 47.24 ng/ml

Instrument :


ECD_L

ClientSampleId :

PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#3 gamma-BHC (Lindane)

R.T.: 3.606 min
Delta R.T.: 0.000 min
Response: 217830151
Conc: 45.94 ng/ml




#4 Heptachlor

R.T.: 4.914 min
Delta R.T.: 0.000 min
Response: 160948171
Conc: 49.11 ng/ml



#4 Heptachlor

R.T.: 3.945 min
Delta R.T.: 0.000 min
Response: 225221625
Conc: 48.38 ng/ml




#5 Aldrin

R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 149319432
Conc: 45.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#5 Aldrin

R.T.: 4.224 min
Delta R.T.: 0.000 min
Response: 209902281
Conc: 46.01 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 77014994
Conc: 47.92 ng/ml



#6 beta-BHC

R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 95671218
Conc: 47.90 ng/ml




#7 delta-BHC

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 156988220
Conc: 44.79 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#7 delta-BHC

R.T.: 4.135 min
Delta R.T.: 0.000 min
Response: 219380269
Conc: 46.17 ng/ml




#8 Heptachlor epoxide

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 133436773
Conc: 44.87 ng/ml



#8 Heptachlor epoxide

R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 193104475
Conc: 46.19 ng/ml




#9 Endosulfan I

R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 119873285
Conc: 45.36 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#9 Endosulfan I

R.T.: 5.096 min
Delta R.T.: 0.000 min
Response: 170652524
Conc: 44.02 ng/ml




#10 gamma-Chlordane

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 129053291
Conc: 46.30 ng/ml



#10 gamma-Chlordane

R.T.: 4.977 min
Delta R.T.: 0.000 min
Response: 206509157
Conc: 48.73 ng/ml




#11 alpha-Chlordane

R.T.: 6.018 min
Delta R.T.: 0.000 min
Response: 129083034
Conc: 46.29 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#11 alpha-Chlordane

R.T.: 5.040 min
Delta R.T.: 0.000 min
Response: 200018297
Conc: 47.78 ng/ml




#12 4,4'-DDE

R.T.: 6.192 min
Delta R.T.: 0.000 min
Response: 122470412
Conc: 50.30 ng/ml



#12 4,4'-DDE

R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 205409127
Conc: 51.23 ng/ml




#13 Dieldrin

R.T.: 6.344 min
Delta R.T.: 0.000 min
Response: 126427219
Conc: 45.55 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#13 Dieldrin

R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 202939963
Conc: 47.24 ng/ml




#14 Endrin

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 109850042
Conc: 46.85 ng/ml m



#14 Endrin

R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 189582652
Conc: 51.34 ng/ml




#15 Endosulfan II

R.T.: 6.793 min
Delta R.T.: 0.001 min
Response: 110556362
Conc: 45.89 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#15 Endosulfan II

R.T.: 5.931 min
Delta R.T.: 0.000 min
Response: 177701011
Conc: 47.98 ng/ml




#16 4,4'-DDD

R.T.: 6.709 min
Delta R.T.: 0.000 min
Response: 96757725
Conc: 50.91 ng/ml



#16 4,4'-DDD

R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 162110968
Conc: 51.36 ng/ml




#17 4,4'-DDT

R.T.: 7.023 min
Delta R.T.: 0.000 min
Response: 103939183
Conc: 52.71 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#17 4,4'-DDT

R.T.: 6.034 min
Delta R.T.: 0.000 min
Response: 175614385
Conc: 53.97 ng/ml




#18 Endrin aldehyde

R.T.: 6.923 min
Delta R.T.: 0.000 min
Response: 85137438
Conc: 43.79 ng/ml



#18 Endrin aldehyde

R.T.: 6.110 min
Delta R.T.: 0.000 min
Response: 135925969
Conc: 44.64 ng/ml



#19 Endosulfan Sulfate

R.T.: 7.158 min
Delta R.T.: 0.000 min
Response: 100546133
Conc: 44.42 ng/ml

Instrument :


ECD_L

ClientSampleId :

PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#19 Endosulfan Sulfate

R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 168097604
Conc: 47.14 ng/ml




#20 Methoxychlor

R.T.: 7.500 min
Delta R.T.: 0.002 min
Response: 53646135
Conc: 51.41 ng/ml



#20 Methoxychlor

R.T.: 6.610 min
Delta R.T.: 0.000 min
Response: 93951668
Conc: 52.54 ng/ml




#21 Endrin ketone

R.T.: 7.644 min
Delta R.T.: 0.002 min
Response: 111136103
Conc: 44.06 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#21 Endrin ketone

R.T.: 6.839 min
Delta R.T.: 0.000 min
Response: 191629552
Conc: 45.68 ng/ml




#22 Mirex

R.T.: 8.116 min
Delta R.T.: 0.001 min
Response: 86631474
Conc: 41.60 ng/ml



#22 Mirex

R.T.: 7.019 min
Delta R.T.: 0.001 min
Response: 145642498
Conc: 43.07 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.002 min
Response: 94144637
Conc: 45.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025



#28 Decachlorobiphenyl

R.T.: 7.910 min
Delta R.T.: 0.000 min
Response: 166040514
Conc: 47.39 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 15:10 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	4.00	3.90	4.10	0.01
beta-BHC	4.53	4.53	4.43	4.63	0.00
delta-BHC	4.77	4.77	4.67	4.87	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.91	4.91	4.81	5.01	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.79	6.79	6.69	6.89	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.64	7.64	7.54	7.74	0.00
Endrin aldehyde	6.92	6.92	6.82	7.02	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 15:10 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
alpha-BHC	3.28	3.28	3.18	3.38	0.00
beta-BHC	3.91	3.91	3.81	4.01	0.00
delta-BHC	4.14	4.14	4.04	4.24	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.94	3.95	3.85	4.05	0.01
Aldrin	4.22	4.23	4.13	4.33	0.01
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endosulfan I	5.10	5.10	5.00	5.20	0.00
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Endosulfan II	5.93	5.93	5.83	6.03	0.00
4,4'-DDD	5.78	5.78	5.68	5.88	0.00
Endosulfan sulfate	6.33	6.33	6.23	6.43	0.00
4,4'-DDT	6.03	6.03	5.93	6.13	0.00
Methoxychlor	6.61	6.61	6.51	6.71	0.00
Endrin ketone	6.84	6.84	6.74	6.94	0.00
Endrin aldehyde	6.11	6.11	6.01	6.21	0.00
alpha-Chlordane	5.04	5.04	4.94	5.14	0.00
gamma-Chlordane	4.98	4.98	4.88	5.08	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/31/2025

Lab Sample No.: PSTDCCC050 Data File : PL093943.D Time Analyzed: 15:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.710	6.608	6.808	54.880	50.000	9.8
4,4'-DDE	6.192	6.091	6.291	53.640	50.000	7.3
4,4'-DDT	7.023	6.922	7.122	53.570	50.000	7.1
Aldrin	5.256	5.156	5.356	48.820	50.000	-2.4
alpha-BHC	3.994	3.895	4.095	50.190	50.000	0.4
alpha-Chlordane	6.018	5.917	6.117	50.080	50.000	0.2
beta-BHC	4.525	4.425	4.625	49.660	50.000	-0.7
Decachlorobiphenyl	9.056	8.953	9.153	48.430	50.000	-3.1
delta-BHC	4.772	4.672	4.872	47.390	50.000	-5.2
Dieldrin	6.344	6.243	6.443	49.090	50.000	-1.8
Endosulfan I	6.069	5.967	6.167	49.170	50.000	-1.7
Endosulfan II	6.794	6.692	6.892	48.340	50.000	-3.3
Endosulfan sulfate	7.159	7.057	7.257	46.610	50.000	-6.8
Endrin	6.572	6.472	6.672	49.260	50.000	-1.5
Endrin aldehyde	6.924	6.823	7.023	46.580	50.000	-6.8
Endrin ketone	7.644	7.542	7.742	46.800	50.000	-6.4
gamma-BHC (Lindane)	4.326	4.227	4.427	49.360	50.000	-1.3
gamma-Chlordane	5.939	5.838	6.038	49.950	50.000	-0.1
Heptachlor	4.914	4.814	5.014	51.680	50.000	3.4
Heptachlor epoxide	5.683	5.582	5.782	47.960	50.000	-4.1
Methoxychlor	7.500	7.398	7.598	53.190	50.000	6.4
Tetrachloro-m-xylene	3.537	3.439	3.639	49.370	50.000	-1.3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/31/2025

Lab Sample No.: PSTDCCC050 Data File : PL093943.D Time Analyzed: 15:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.784	5.684	5.884	53.890	50.000	7.8
4,4'-DDE	5.229	5.130	5.330	51.950	50.000	3.9
4,4'-DDT	6.034	5.934	6.134	53.920	50.000	7.8
Aldrin	4.224	4.125	4.325	47.070	50.000	-5.9
alpha-BHC	3.276	3.177	3.377	49.480	50.000	-1.0
alpha-Chlordane	5.040	4.940	5.140	49.320	50.000	-1.4
beta-BHC	3.906	3.807	4.007	48.590	50.000	-2.8
Decachlorobiphenyl	7.911	7.810	8.010	49.210	50.000	-1.6
delta-BHC	4.135	4.036	4.236	46.340	50.000	-7.3
Dieldrin	5.361	5.261	5.461	48.660	50.000	-2.7
Endosulfan I	5.096	4.996	5.196	45.000	50.000	-10.0
Endosulfan II	5.931	5.831	6.031	49.800	50.000	-0.4
Endosulfan sulfate	6.333	6.233	6.433	48.390	50.000	-3.2
Endrin	5.636	5.536	5.736	51.550	50.000	3.1
Endrin aldehyde	6.110	6.010	6.210	45.680	50.000	-8.6
Endrin ketone	6.839	6.739	6.939	48.250	50.000	-3.5
gamma-BHC (Lindane)	3.606	3.507	3.707	47.750	50.000	-4.5
gamma-Chlordane	4.976	4.877	5.077	50.160	50.000	0.3
Heptachlor	3.944	3.845	4.045	49.210	50.000	-1.6
Heptachlor epoxide	4.726	4.627	4.827	47.640	50.000	-4.7
Methoxychlor	6.609	6.509	6.709	52.440	50.000	4.9
Tetrachloro-m-xylene	2.773	2.674	2.874	48.600	50.000	-2.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093943.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 15:10
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:27:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.773	132.9E6	158.6E6	49.373	48.598
28) SA Decachlor...	9.056	7.911	101.3E6	172.5E6	48.428	49.215
Target Compounds						
2) A alpha-BHC	3.994	3.276	192.4E6	241.9E6	50.186	49.477
3) MA gamma-BHC...	4.326	3.606	181.8E6	226.4E6	49.362	47.751
4) MA Heptachlor	4.914	3.944	169.4E6	229.0E6	51.679	49.207
5) MB Aldrin	5.256	4.224	159.7E6	214.7E6	48.818	47.070
6) B beta-BHC	4.525	3.906	79824408	97064464	49.663	48.594
7) B delta-BHC	4.772	4.135	166.1E6	220.2E6	47.394	46.342
8) B Heptachlo...	5.683	4.726	142.6E6	199.2E6	47.965	47.644
9) A Endosulfan I	6.069	5.096	129.9E6	174.5E6	49.170	45.002
10) B gamma-Chl...	5.939	4.976	139.2E6	212.6E6	49.947	50.159
11) B alpha-Chl...	6.018	5.040	139.6E6	206.5E6	50.080	49.315
12) B 4,4'-DDE	6.192	5.229	130.6E6	208.3E6	53.645	51.946
13) MA Dieldrin	6.344	5.361	136.3E6	209.0E6	49.087	48.660
14) MA Endrin	6.572	5.636	115.5E6	190.4E6	49.261m	51.548
15) B Endosulfa...	6.794	5.931	116.5E6	184.5E6	48.343	49.801
16) A 4,4'-DDD	6.710	5.784	104.3E6	170.1E6	54.876	53.888
17) MA 4,4'-DDT	7.023	6.034	105.6E6	175.4E6	53.568	53.917
18) B Endrin al...	6.924	6.110	90562659	139.1E6	46.584	45.676
19) B Endosulfa...	7.159	6.333	105.5E6	172.6E6	46.606	48.393
20) A Methoxychlor	7.500	6.609	55495800	93773372	53.188	52.442
21) B Endrin ke...	7.644	6.839	118.1E6	202.4E6	46.801	48.249
22) Mirex	8.117	7.019	93662867	155.8E6	44.976	46.076

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093943.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 15:10
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:27:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.537 min
Delta R.T.: -0.002 min
Response: 132948418
Conc: 49.37 ng/ml

Instrument :


ECD_L

ClientSampleId :

PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: -0.001 min
Response: 158630081
Conc: 48.60 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: -0.001 min
Response: 192404145
Conc: 50.19 ng/ml



#2 alpha-BHC

R.T.: 3.276 min
Delta R.T.: -0.001 min
Response: 241892491
Conc: 49.48 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.326 min
Delta R.T.: 0.000 min
Response: 181790929
Conc: 49.36 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#3 gamma-BHC (Lindane)

R.T.: 3.606 min
Delta R.T.: -0.001 min
Response: 226396596
Conc: 47.75 ng/ml




#4 Heptachlor

R.T.: 4.914 min
Delta R.T.: 0.000 min
Response: 169368307
Conc: 51.68 ng/ml



#4 Heptachlor

R.T.: 3.944 min
Delta R.T.: 0.000 min
Response: 229049342
Conc: 49.21 ng/ml




#5 Aldrin

R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 159729603
Conc: 48.82 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#5 Aldrin

R.T.: 4.224 min
Delta R.T.: -0.001 min
Response: 214722350
Conc: 47.07 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 79824408
Conc: 49.66 ng/ml



#6 beta-BHC

R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 97064464
Conc: 48.59 ng/ml




#7 delta-BHC

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 166130357
Conc: 47.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#7 delta-BHC

R.T.: 4.135 min
Delta R.T.: 0.000 min
Response: 220181787
Conc: 46.34 ng/ml




#8 Heptachlor epoxide

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 142638245
Conc: 47.96 ng/ml



#8 Heptachlor epoxide

R.T.: 4.726 min
Delta R.T.: 0.000 min
Response: 199163293
Conc: 47.64 ng/ml




#9 Endosulfan I

R.T.: 6.069 min
Delta R.T.: 0.001 min
Response: 129949655
Conc: 49.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#9 Endosulfan I

R.T.: 5.096 min
Delta R.T.: 0.000 min
Response: 174469159
Conc: 45.00 ng/ml




#10 gamma-Chlordane

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 139220700
Conc: 49.95 ng/ml



#10 gamma-Chlordane

R.T.: 4.976 min
Delta R.T.: 0.000 min
Response: 212552875
Conc: 50.16 ng/ml




#11 alpha-Chlordane

R.T.: 6.018 min
Delta R.T.: 0.000 min
Response: 139642336
Conc: 50.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#11 alpha-Chlordane

R.T.: 5.040 min
Delta R.T.: 0.000 min
Response: 206461185
Conc: 49.32 ng/ml




#12 4,4'-DDE

R.T.: 6.192 min
Delta R.T.: 0.000 min
Response: 130603013
Conc: 53.64 ng/ml



#12 4,4'-DDE

R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 208276185
Conc: 51.95 ng/ml




#13 Dieldrin

R.T.: 6.344 min
Delta R.T.: 0.001 min
Response: 136258977
Conc: 49.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#13 Dieldrin

R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 209025626
Conc: 48.66 ng/ml




#14 Endrin

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 115508037
Conc: 49.26 ng/ml m



#14 Endrin

R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 190350317
Conc: 51.55 ng/ml




#15 Endosulfan II

R.T.: 6.794 min
Delta R.T.: 0.002 min
Response: 116474797
Conc: 48.34 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#15 Endosulfan II

R.T.: 5.931 min
Delta R.T.: 0.000 min
Response: 184453995
Conc: 49.80 ng/ml




#16 4,4'-DDD

R.T.: 6.710 min
Delta R.T.: 0.002 min
Response: 104295691
Conc: 54.88 ng/ml



#16 4,4'-DDD

R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 170099239
Conc: 53.89 ng/ml




#17 4,4'-DDT

R.T.: 7.023 min
Delta R.T.: 0.001 min
Response: 105638372
Conc: 53.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#17 4,4'-DDT

R.T.: 6.034 min
Delta R.T.: 0.000 min
Response: 175446797
Conc: 53.92 ng/ml




#18 Endrin aldehyde

R.T.: 6.924 min
Delta R.T.: 0.001 min
Response: 90562659
Conc: 46.58 ng/ml



#18 Endrin aldehyde

R.T.: 6.110 min
Delta R.T.: 0.000 min
Response: 139067043
Conc: 45.68 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.159 min
Delta R.T.: 0.002 min
Response: 105503247
Conc: 46.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#19 Endosulfan Sulfate

R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 172573421
Conc: 48.39 ng/ml




#20 Methoxychlor

R.T.: 7.500 min
Delta R.T.: 0.002 min
Response: 55495800
Conc: 53.19 ng/ml



#20 Methoxychlor

R.T.: 6.609 min
Delta R.T.: 0.000 min
Response: 93773372
Conc: 52.44 ng/ml




#21 Endrin ketone

R.T.: 7.644 min
Delta R.T.: 0.002 min
Response: 118061517
Conc: 46.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#21 Endrin ketone

R.T.: 6.839 min
Delta R.T.: 0.000 min
Response: 202414678
Conc: 48.25 ng/ml




#22 Mirex

R.T.: 8.117 min
Delta R.T.: 0.002 min
Response: 93662867
Conc: 44.98 ng/ml



#22 Mirex

R.T.: 7.019 min
Delta R.T.: 0.001 min
Response: 155823162
Conc: 46.08 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: 0.004 min
Response: 101308174
Conc: 48.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: 0.000 min
Response: 172452418
Conc: 49.21 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 20:18 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	4.00	3.90	4.10	0.01
beta-BHC	4.52	4.53	4.43	4.63	0.01
delta-BHC	4.77	4.77	4.67	4.87	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.91	4.91	4.81	5.01	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.79	6.79	6.69	6.89	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.64	7.64	7.54	7.74	0.00
Endrin aldehyde	6.92	6.92	6.82	7.02	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 20:18 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
alpha-BHC	3.28	3.28	3.18	3.38	0.00
beta-BHC	3.91	3.91	3.81	4.01	0.00
delta-BHC	4.14	4.14	4.04	4.24	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.95	3.95	3.85	4.05	0.01
Aldrin	4.22	4.23	4.13	4.33	0.01
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endosulfan I	5.10	5.10	5.00	5.20	0.00
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Endosulfan II	5.93	5.93	5.83	6.03	0.00
4,4'-DDD	5.78	5.78	5.68	5.88	0.00
Endosulfan sulfate	6.33	6.33	6.23	6.43	0.00
4,4'-DDT	6.04	6.03	5.93	6.13	0.00
Methoxychlor	6.61	6.61	6.51	6.71	0.00
Endrin ketone	6.84	6.84	6.74	6.94	0.00
Endrin aldehyde	6.11	6.11	6.01	6.21	0.00
alpha-Chlordane	5.04	5.04	4.94	5.14	0.00
gamma-Chlordane	4.98	4.98	4.88	5.08	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/31/2025

Lab Sample No.: PSTDCCC050 Data File : PL093959.D Time Analyzed: 20:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.710	6.608	6.808	54.500	50.000	9.0
4,4'-DDE	6.192	6.091	6.291	51.300	50.000	2.6
4,4'-DDT	7.023	6.922	7.122	48.330	50.000	-3.3
Aldrin	5.256	5.156	5.356	47.800	50.000	-4.4
alpha-BHC	3.994	3.895	4.095	50.690	50.000	1.4
alpha-Chlordane	6.018	5.917	6.117	48.070	50.000	-3.9
beta-BHC	4.524	4.425	4.625	50.600	50.000	1.2
Decachlorobiphenyl	9.056	8.953	9.153	45.870	50.000	-8.3
delta-BHC	4.772	4.672	4.872	49.100	50.000	-1.8
Dieldrin	6.343	6.243	6.443	47.370	50.000	-5.3
Endosulfan I	6.068	5.967	6.167	46.980	50.000	-6.0
Endosulfan II	6.794	6.692	6.892	46.660	50.000	-6.7
Endosulfan sulfate	7.159	7.057	7.257	45.370	50.000	-9.3
Endrin	6.572	6.472	6.672	45.760	50.000	-8.5
Endrin aldehyde	6.924	6.823	7.023	45.260	50.000	-9.5
Endrin ketone	7.644	7.542	7.742	45.780	50.000	-8.4
gamma-BHC (Lindane)	4.326	4.227	4.427	49.510	50.000	-1.0
gamma-Chlordane	5.939	5.838	6.038	48.430	50.000	-3.1
Heptachlor	4.914	4.814	5.014	49.860	50.000	-0.3
Heptachlor epoxide	5.683	5.582	5.782	46.670	50.000	-6.7
Methoxychlor	7.500	7.398	7.598	47.640	50.000	-4.7
Tetrachloro-m-xylene	3.538	3.439	3.639	49.950	50.000	-0.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 01/31/2025

Lab Sample No.: PSTDCCC050 Data File : PL093959.D Time Analyzed: 20:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.784	5.684	5.884	53.950	50.000	7.9
4,4'-DDE	5.229	5.130	5.330	53.020	50.000	6.0
4,4'-DDT	6.035	5.934	6.134	48.230	50.000	-3.5
Aldrin	4.224	4.125	4.325	48.560	50.000	-2.9
alpha-BHC	3.276	3.177	3.377	50.850	50.000	1.7
alpha-Chlordane	5.040	4.940	5.140	50.410	50.000	0.8
beta-BHC	3.907	3.807	4.007	50.650	50.000	1.3
Decachlorobiphenyl	7.911	7.810	8.010	44.510	50.000	-11.0
delta-BHC	4.135	4.036	4.236	48.490	50.000	-3.0
Dieldrin	5.361	5.261	5.461	48.940	50.000	-2.1
Endosulfan I	5.096	4.996	5.196	45.790	50.000	-8.4
Endosulfan II	5.931	5.831	6.031	48.780	50.000	-2.4
Endosulfan sulfate	6.334	6.233	6.433	46.930	50.000	-6.1
Endrin	5.636	5.536	5.736	48.830	50.000	-2.3
Endrin aldehyde	6.111	6.010	6.210	44.850	50.000	-10.3
Endrin ketone	6.839	6.739	6.939	46.120	50.000	-7.8
gamma-BHC (Lindane)	3.606	3.507	3.707	49.220	50.000	-1.6
gamma-Chlordane	4.975	4.877	5.077	50.890	50.000	1.8
Heptachlor	3.945	3.845	4.045	49.210	50.000	-1.6
Heptachlor epoxide	4.727	4.627	4.827	49.060	50.000	-1.9
Methoxychlor	6.610	6.509	6.709	46.040	50.000	-7.9
Tetrachloro-m-xylene	2.773	2.674	2.874	50.250	50.000	0.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093959.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 20:18
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:33:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.773	134.5E6	164.0E6	49.947	50.255
28) SA Decachlor...	9.056	7.911	95956611	156.0E6	45.870	44.506
Target Compounds						
2) A alpha-BHC	3.994	3.276	194.3E6	248.6E6	50.689	50.852
3) MA gamma-BHC...	4.326	3.606	182.3E6	233.4E6	49.509	49.224
4) MA Heptachlor	4.914	3.945	163.4E6	229.0E6	49.864	49.205
5) MB Aldrin	5.256	4.224	156.4E6	221.5E6	47.802	48.559
6) B beta-BHC	4.524	3.907	81325359	101.2E6	50.597	50.650
7) B delta-BHC	4.772	4.135	172.1E6	230.4E6	49.105	48.494
8) B Heptachlo...	5.683	4.727	138.8E6	205.1E6	46.671	49.057
9) A Endosulfan I	6.068	5.096	124.2E6	177.5E6	46.979	45.788
10) B gamma-Chl...	5.939	4.975	135.0E6	215.7E6	48.432	50.894m
11) B alpha-Chl...	6.018	5.040	134.0E6	211.0E6	48.071	50.409
12) B 4,4'-DDE	6.192	5.229	124.9E6	212.6E6	51.304	53.022
13) MA Dieldrin	6.343	5.361	131.5E6	210.2E6	47.371	48.940
14) MA Endrin	6.572	5.636	107.3E6	180.3E6	45.759m	48.833
15) B Endosulfa...	6.794	5.931	112.4E6	180.7E6	46.658	48.777
16) A 4,4'-DDD	6.710	5.784	103.6E6	170.3E6	54.501	53.947
17) MA 4,4'-DDT	7.023	6.035	95307062	157.0E6	48.329	48.233
18) B Endrin al...	6.924	6.111	87990160	136.5E6	45.261	44.845
19) B Endosulfa...	7.159	6.334	102.7E6	167.4E6	45.367	46.929
20) A Methoxychlor	7.500	6.610	49709035	82322200	47.642	46.038
21) B Endrin ke...	7.644	6.839	115.5E6	193.5E6	45.777	46.117
22) Mirex	8.117	7.019	85395308	144.1E6	41.006	42.598

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093959.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 20:18
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:33:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.538 min
Delta R.T.: -0.001 min
Response: 134495166
Conc: 49.95 ng/ml

Instrument :


ECD_L

ClientSampleId :

PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: 0.000 min
Response: 164039017
Conc: 50.25 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 194333908
Conc: 50.69 ng/ml



#2 alpha-BHC

R.T.: 3.276 min
Delta R.T.: 0.000 min
Response: 248614691
Conc: 50.85 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.326 min
Delta R.T.: 0.000 min
Response: 182334965
Conc: 49.51 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#3 gamma-BHC (Lindane)

R.T.: 3.606 min
Delta R.T.: 0.000 min
Response: 233380358
Conc: 49.22 ng/ml




#4 Heptachlor

R.T.: 4.914 min
Delta R.T.: 0.000 min
Response: 163422577
Conc: 49.86 ng/ml



#4 Heptachlor

R.T.: 3.945 min
Delta R.T.: 0.000 min
Response: 229039951
Conc: 49.21 ng/ml




#5 Aldrin

R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 156405208
Conc: 47.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#5 Aldrin

R.T.: 4.224 min
Delta R.T.: 0.000 min
Response: 221518046
Conc: 48.56 ng/ml




#6 beta-BHC

R.T.: 4.524 min
Delta R.T.: 0.000 min
Response: 81325359
Conc: 50.60 ng/ml



#6 beta-BHC

R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 101169833
Conc: 50.65 ng/ml




#7 delta-BHC

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 172125561
Conc: 49.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#7 delta-BHC

R.T.: 4.135 min
Delta R.T.: 0.000 min
Response: 230407963
Conc: 48.49 ng/ml




#8 Heptachlor epoxide

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 138792617
Conc: 46.67 ng/ml



#8 Heptachlor epoxide

R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 205068803
Conc: 49.06 ng/ml




#9 Endosulfan I

R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 124157471
Conc: 46.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#9 Endosulfan I

R.T.: 5.096 min
Delta R.T.: 0.000 min
Response: 177517018
Conc: 45.79 ng/ml




#10 gamma-Chlordane

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 134996839
Conc: 48.43 ng/ml



#10 gamma-Chlordane

R.T.: 4.975 min
Delta R.T.: -0.002 min
Response: 215667901
Conc: 50.89 ng/ml m




#11 alpha-Chlordane

R.T.: 6.018 min
Delta R.T.: 0.000 min
Response: 134040951
Conc: 48.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#11 alpha-Chlordane

R.T.: 5.040 min
Delta R.T.: 0.000 min
Response: 211040317
Conc: 50.41 ng/ml




#12 4,4'-DDE

R.T.: 6.192 min
Delta R.T.: 0.000 min
Response: 124903447
Conc: 51.30 ng/ml



#12 4,4'-DDE

R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 212589791
Conc: 53.02 ng/ml




#13 Dieldrin

R.T.: 6.343 min
Delta R.T.: 0.000 min
Response: 131494756
Conc: 47.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#13 Dieldrin

R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 210228997
Conc: 48.94 ng/ml




#14 Endrin

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 107296542
Conc: 45.76 ng/ml m



#14 Endrin

R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 180323797
Conc: 48.83 ng/ml




#15 Endosulfan II

R.T.: 6.794 min
Delta R.T.: 0.002 min
Response: 112415926
Conc: 46.66 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#15 Endosulfan II

R.T.: 5.931 min
Delta R.T.: 0.000 min
Response: 180661726
Conc: 48.78 ng/ml




#16 4,4'-DDD

R.T.: 6.710 min
Delta R.T.: 0.001 min
Response: 103581747
Conc: 54.50 ng/ml



#16 4,4'-DDD

R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 170285310
Conc: 53.95 ng/ml




#17 4,4'-DDT

R.T.: 7.023 min
Delta R.T.: 0.001 min
Response: 95307062
Conc: 48.33 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#17 4,4'-DDT

R.T.: 6.035 min
Delta R.T.: 0.000 min
Response: 156951923
Conc: 48.23 ng/ml




#18 Endrin aldehyde

R.T.: 6.924 min
Delta R.T.: 0.001 min
Response: 87990160
Conc: 45.26 ng/ml



#18 Endrin aldehyde

R.T.: 6.111 min
Delta R.T.: 0.000 min
Response: 136537004
Conc: 44.85 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.159 min
Delta R.T.: 0.001 min
Response: 102700044
Conc: 45.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#19 Endosulfan Sulfate

R.T.: 6.334 min
Delta R.T.: 0.000 min
Response: 167352654
Conc: 46.93 ng/ml




#20 Methoxychlor

R.T.: 7.500 min
Delta R.T.: 0.002 min
Response: 49709035
Conc: 47.64 ng/ml



#20 Methoxychlor

R.T.: 6.610 min
Delta R.T.: 0.000 min
Response: 82322200
Conc: 46.04 ng/ml



#21 Endrin ketone

R.T.: 7.644 min
Delta R.T.: 0.002 min
Response: 115478712
Conc: 45.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025

#21 Endrin ketone


R.T.: 6.839 min
Delta R.T.: 0.000 min
Response: 193472127
Conc: 46.12 ng/ml

#22 Mirex

R.T.: 8.117 min
Delta R.T.: 0.002 min
Response: 85395308
Conc: 41.01 ng/ml

#22 Mirex

R.T.: 7.019 min
Delta R.T.: 0.002 min
Response: 144060209
Conc: 42.60 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: 0.003 min
Response: 95956611
Conc: 45.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: 0.000 min
Response: 155951173
Conc: 44.51 ng/ml

CALIBRATION VERIFICATION SUMMARY

 Contract: RUTW01

 Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

 Continuing Calib Date: 02/01/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

 Continuing Calib Time: 00:57 Initial Calibration Time(s): 10:57 11:51

 GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	4.00	3.90	4.10	0.01
beta-BHC	4.53	4.53	4.43	4.63	0.00
delta-BHC	4.77	4.77	4.67	4.87	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.91	4.81	5.01	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.35	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.80	6.79	6.69	6.89	-0.01
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.03	7.02	6.92	7.12	-0.01
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.65	7.64	7.54	7.74	0.00
Endrin aldehyde	6.93	6.92	6.82	7.02	-0.01
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

Continuing Calib Date: 02/01/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 00:57 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
alpha-BHC	3.28	3.28	3.18	3.38	0.00
beta-BHC	3.91	3.91	3.81	4.01	0.00
delta-BHC	4.14	4.14	4.04	4.24	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.95	3.95	3.85	4.05	0.01
Aldrin	4.22	4.23	4.13	4.33	0.01
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endosulfan I	5.10	5.10	5.00	5.20	0.00
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Endosulfan II	5.93	5.93	5.83	6.03	0.00
4,4'-DDD	5.79	5.78	5.68	5.88	0.00
Endosulfan sulfate	6.33	6.33	6.23	6.43	0.00
4,4'-DDT	6.04	6.03	5.93	6.13	0.00
Methoxychlor	6.61	6.61	6.51	6.71	0.00
Endrin ketone	6.84	6.84	6.74	6.94	0.00
Endrin aldehyde	6.11	6.11	6.01	6.21	0.00
alpha-Chlordane	5.04	5.04	4.94	5.14	0.00
gamma-Chlordane	4.98	4.98	4.88	5.08	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/01/2025

Lab Sample No.: PSTDCCC050 Data File : PL093971.D Time Analyzed: 00:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.711	6.608	6.808	54.730	50.000	9.5
4,4'-DDE	6.192	6.091	6.291	51.870	50.000	3.7
4,4'-DDT	7.025	6.922	7.122	44.830	50.000	-10.3
Aldrin	5.257	5.156	5.356	49.150	50.000	-1.7
alpha-BHC	3.994	3.895	4.095	51.630	50.000	3.3
alpha-Chlordane	6.018	5.917	6.117	49.050	50.000	-1.9
beta-BHC	4.526	4.425	4.625	50.470	50.000	0.9
Decachlorobiphenyl	9.057	8.953	9.153	44.060	50.000	-11.9
delta-BHC	4.772	4.672	4.872	49.410	50.000	-1.2
Dieldrin	6.345	6.243	6.443	47.520	50.000	-5.0
Endosulfan I	6.069	5.967	6.167	48.060	50.000	-3.9
Endosulfan II	6.795	6.692	6.892	46.330	50.000	-7.3
Endosulfan sulfate	7.160	7.057	7.257	44.270	50.000	-11.5
Endrin	6.573	6.472	6.672	44.470	50.000	-11.1
Endrin aldehyde	6.925	6.823	7.023	44.530	50.000	-10.9
Endrin ketone	7.645	7.542	7.742	44.660	50.000	-10.7
gamma-BHC (Lindane)	4.327	4.227	4.427	50.040	50.000	0.1
gamma-Chlordane	5.940	5.838	6.038	49.550	50.000	-0.9
Heptachlor	4.915	4.814	5.014	50.120	50.000	0.2
Heptachlor epoxide	5.683	5.582	5.782	47.790	50.000	-4.4
Methoxychlor	7.500	7.398	7.598	43.640	50.000	-12.7
Tetrachloro-m-xylene	3.538	3.439	3.639	51.110	50.000	2.2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/01/2025

Lab Sample No.: PSTDCCC050 Data File : PL093971.D Time Analyzed: 00:57

COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
4,4'-DDD	5.785	5.684	5.884	56.200	50.000	12.4
4,4'-DDE	5.230	5.130	5.330	54.610	50.000	9.2
4,4'-DDT	6.035	5.934	6.134	46.460	50.000	-7.1
Aldrin	4.224	4.125	4.325	50.030	50.000	0.1
alpha-BHC	3.276	3.177	3.377	53.120	50.000	6.2
alpha-Chlordane	5.041	4.940	5.140	51.720	50.000	3.4
beta-BHC	3.906	3.807	4.007	52.880	50.000	5.8
Decachlorobiphenyl	7.911	7.810	8.010	42.340	50.000	-15.3
delta-BHC	4.135	4.036	4.236	50.050	50.000	0.1
Dieldrin	5.361	5.261	5.461	50.490	50.000	1.0
Endosulfan I	5.097	4.996	5.196	45.740	50.000	-8.5
Endosulfan II	5.932	5.831	6.031	49.620	50.000	-0.8
Endosulfan sulfate	6.334	6.233	6.433	47.400	50.000	-5.2
Endrin	5.637	5.536	5.736	49.760	50.000	-0.5
Endrin aldehyde	6.111	6.010	6.210	45.370	50.000	-9.3
Endrin ketone	6.840	6.739	6.939	45.870	50.000	-8.3
gamma-BHC (Lindane)	3.606	3.507	3.707	51.100	50.000	2.2
gamma-Chlordane	4.977	4.877	5.077	52.470	50.000	4.9
Heptachlor	3.945	3.845	4.045	50.220	50.000	0.4
Heptachlor epoxide	4.727	4.627	4.827	50.070	50.000	0.1
Methoxychlor	6.611	6.509	6.709	45.230	50.000	-9.5
Tetrachloro-m-xylene	2.774	2.674	2.874	52.270	50.000	4.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093971.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2025 00:57
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 01:09:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.774	137.6E6	170.6E6	51.107	52.269
28) SA Decachlor...	9.057	7.911	92166830	148.4E6	44.059	42.337
Target Compounds						
2) A alpha-BHC	3.994	3.276	197.9E6	259.7E6	51.627	53.118
3) MA gamma-BHC...	4.327	3.606	184.3E6	242.3E6	50.044	51.101
4) MA Heptachlor	4.915	3.945	164.3E6	233.8E6	50.120	50.222
5) MB Aldrin	5.257	4.224	160.8E6	228.2E6	49.147	50.029
6) B beta-BHC	4.526	3.906	81125313	105.6E6	50.472	52.878
7) B delta-BHC	4.772	4.135	173.2E6	237.8E6	49.409	50.051
8) B Heptachlo...	5.683	4.727	142.1E6	209.3E6	47.787	50.066
9) A Endosulfan I	6.069	5.097	127.0E6	177.3E6	48.064	45.738
10) B gamma-Chl...	5.940	4.977	138.1E6	222.3E6	49.551	52.466
11) B alpha-Chl...	6.018	5.041	136.8E6	216.5E6	49.052	51.718
12) B 4,4'-DDE	6.192	5.230	126.3E6	218.9E6	51.875	54.608
13) MA Dieldrin	6.345	5.361	131.9E6	216.9E6	47.522	50.491
14) MA Endrin	6.573	5.637	104.3E6	183.8E6	44.467m	49.762
15) B Endosulfa...	6.795	5.932	111.6E6	183.8E6	46.333	49.617
16) A 4,4'-DDD	6.711	5.785	104.0E6	177.4E6	54.726	56.198
17) MA 4,4'-DDT	7.025	6.035	88412487	151.2E6	44.833	46.456
18) B Endrin al...	6.925	6.111	86568776	138.1E6	44.530	45.374
19) B Endosulfa...	7.160	6.334	100.2E6	169.0E6	44.272	47.401
20) A Methoxychlor	7.500	6.611	45529670	80874445	43.636	45.228
21) B Endrin ke...	7.645	6.840	112.7E6	192.4E6	44.661	45.870
22) Mirex	8.118	7.020	82894240	140.6E6	39.805	41.587

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093971.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2025 00:57
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 01:09:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.538 min
Delta R.T.: -0.001 min
Response: 137617679
Conc: 51.11 ng/ml

Instrument :


ECD_L

ClientSampleId :

PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 170615450
Conc: 52.27 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 197929895
Conc: 51.63 ng/ml



#2 alpha-BHC

R.T.: 3.276 min
Delta R.T.: 0.000 min
Response: 259690513
Conc: 53.12 ng/ml



#3 gamma-BHC (Lindane)

R.T.: 4.327 min
Delta R.T.: 0.000 min
Response: 184303795
Conc: 50.04 ng/ml

Instrument :


ECD_L

ClientSampleId :

PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#3 gamma-BHC (Lindane)

R.T.: 3.606 min
Delta R.T.: -0.001 min
Response: 242279064
Conc: 51.10 ng/ml




#4 Heptachlor

R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 164258733
Conc: 50.12 ng/ml



#4 Heptachlor

R.T.: 3.945 min
Delta R.T.: 0.000 min
Response: 233773504
Conc: 50.22 ng/ml




#5 Aldrin

R.T.: 5.257 min
Delta R.T.: 0.001 min
Response: 160808163
Conc: 49.15 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#5 Aldrin

R.T.: 4.224 min
Delta R.T.: 0.000 min
Response: 228224277
Conc: 50.03 ng/ml




#6 beta-BHC

R.T.: 4.526 min
Delta R.T.: 0.000 min
Response: 81125313
Conc: 50.47 ng/ml



#6 beta-BHC

R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 105619555
Conc: 52.88 ng/ml




#7 delta-BHC

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 173193791
Conc: 49.41 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#7 delta-BHC

R.T.: 4.135 min
Delta R.T.: 0.000 min
Response: 237802685
Conc: 50.05 ng/ml




#8 Heptachlor epoxide

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 142110295
Conc: 47.79 ng/ml



#8 Heptachlor epoxide

R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 209288866
Conc: 50.07 ng/ml




#9 Endosulfan I

R.T.: 6.069 min
Delta R.T.: 0.002 min
Response: 127025462
Conc: 48.06 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#9 Endosulfan I

R.T.: 5.097 min
Delta R.T.: 0.000 min
Response: 177320788
Conc: 45.74 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.002 min
Response: 138116385
Conc: 49.55 ng/ml



#10 gamma-Chlordane

R.T.: 4.977 min
Delta R.T.: 0.000 min
Response: 222332589
Conc: 52.47 ng/ml




#11 alpha-Chlordane

R.T.: 6.018 min
Delta R.T.: 0.001 min
Response: 136776971
Conc: 49.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#11 alpha-Chlordane

R.T.: 5.041 min
Delta R.T.: 0.000 min
Response: 216520758
Conc: 51.72 ng/ml




#12 4,4'-DDE

R.T.: 6.192 min
Delta R.T.: 0.000 min
Response: 126294058
Conc: 51.87 ng/ml



#12 4,4'-DDE

R.T.: 5.230 min
Delta R.T.: 0.000 min
Response: 218947610
Conc: 54.61 ng/ml




#13 Dieldrin

R.T.: 6.345 min
Delta R.T.: 0.002 min
Response: 131914872
Conc: 47.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#13 Dieldrin

R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 216891243
Conc: 50.49 ng/ml




#14 Endrin

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 104266266
Conc: 44.47 ng/ml m



#14 Endrin

R.T.: 5.637 min
Delta R.T.: 0.000 min
Response: 183753883
Conc: 49.76 ng/ml




#15 Endosulfan II

R.T.: 6.795 min
Delta R.T.: 0.003 min
Response: 111631719
Conc: 46.33 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#15 Endosulfan II

R.T.: 5.932 min
Delta R.T.: 0.000 min
Response: 183774349
Conc: 49.62 ng/ml




#16 4,4'-DDD

R.T.: 6.711 min
Delta R.T.: 0.002 min
Response: 104011017
Conc: 54.73 ng/ml



#16 4,4'-DDD

R.T.: 5.785 min
Delta R.T.: 0.000 min
Response: 177390621
Conc: 56.20 ng/ml




#17 4,4'-DDT

R.T.: 7.025 min
Delta R.T.: 0.003 min
Response: 88412487
Conc: 44.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#17 4,4'-DDT

R.T.: 6.035 min
Delta R.T.: 0.000 min
Response: 151169658
Conc: 46.46 ng/ml




#18 Endrin aldehyde

R.T.: 6.925 min
Delta R.T.: 0.002 min
Response: 86568776
Conc: 44.53 ng/ml



#18 Endrin aldehyde

R.T.: 6.111 min
Delta R.T.: 0.000 min
Response: 138148088
Conc: 45.37 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.160 min
Delta R.T.: 0.003 min
Response: 100219540
Conc: 44.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#19 Endosulfan Sulfate

R.T.: 6.334 min
Delta R.T.: 0.000 min
Response: 169037546
Conc: 47.40 ng/ml




#20 Methoxychlor

R.T.: 7.500 min
Delta R.T.: 0.003 min
Response: 45529670
Conc: 43.64 ng/ml



#20 Methoxychlor

R.T.: 6.611 min
Delta R.T.: 0.001 min
Response: 80874445
Conc: 45.23 ng/ml




#21 Endrin ketone

R.T.: 7.645 min
Delta R.T.: 0.003 min
Response: 112662187
Conc: 44.66 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#21 Endrin ketone

R.T.: 6.840 min
Delta R.T.: 0.000 min
Response: 192435988
Conc: 45.87 ng/ml




#22 Mirex

R.T.: 8.118 min
Delta R.T.: 0.004 min
Response: 82894240
Conc: 39.81 ng/ml



#22 Mirex

R.T.: 7.020 min
Delta R.T.: 0.002 min
Response: 140640985
Conc: 41.59 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.057 min
Delta R.T.: 0.004 min
Response: 92166830
Conc: 44.06 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: 0.001 min
Response: 148352395
Conc: 42.34 ng/ml

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093726.D Date Analyzed: 01/21/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:30

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.052	8.950	9.150	18.070	20.000	-9.7
Tetrachloro-m-xylene	3.538	3.490	3.590	18.530	20.000	-7.4
alpha-BHC	3.994	3.940	4.040	9.490	10.000	-5.1
beta-BHC	4.525	4.470	4.580	9.790	10.000	-2.1
gamma-BHC (Lindane)	4.326	4.280	4.380	9.300	10.000	-7.0
Endrin	6.572	6.500	6.640	41.270	50.000	-17.5
4,4'-DDT	7.022	6.950	7.090	82.410	100.000	-17.6
Methoxychlor	7.498	7.430	7.570	190.380	250.000	-23.8

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093726.D Date Analyzed: 01/21/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:30

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.909	7.810	8.010	17.950	20.000	-10.3
Tetrachloro-m-xylene	2.775	2.720	2.830	17.900	20.000	-10.5
alpha-BHC	3.277	3.230	3.330	8.620	10.000	-13.8
beta-BHC	3.907	3.860	3.960	9.800	10.000	-2.0
gamma-BHC (Lindane)	3.607	3.560	3.660	8.300	10.000	-17.0
Endrin	5.636	5.570	5.710	42.700	50.000	-14.6
4,4'-DDT	6.034	5.960	6.100	96.510	100.000	-3.5
Methoxychlor	6.609	6.540	6.680	209.940	250.000	-16.0

Data File: PEM
 PL093726.D **Date Acquired** 1/21/2025 10:30
Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.57	96765136.66	105215770.7	8450634.06	8.03
Endrin aldehyde	6.92	3175682.472			
Endrin ketone	7.64	5274951.584			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.64	157695791.8	174071574	16375782.2	9.41
Endrin aldehyde #2	6.11	6776503.08			
Endrin ketone #2	6.84	9599279.119			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.02	162509369.8	166424298.6	3914928.76	2.35
4,4'-DDE	6.19	560248.444			
4,4'-DDD	6.71	3354680.315			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.03	314041690	320417432.8	6375742.79	1.99
4,4'-DDE #2	5.23	775353.914			
4,4'-DDD #2	5.78	5600388.877			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
 Data File : PL093726.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2025 10:30
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 01/22/2025
 Supervised By : Ankita Jodhani 01/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 14:04:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:02:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.775	49897579	58438387	18.530	17.903
2) SA Decachlor...	9.052	7.909	37808316	62882920	18.074	17.946
Target Compounds						
2) A alpha-BHC	3.994	3.277	36373358	42163610	9.487	8.624
3) MA gamma-BHC...	4.326	3.607	34234012	39348781	9.296	8.299
6) B beta-BHC	4.525	3.907	15730216	19569860	9.787	9.797
12) B 4,4'-DDE	6.193	5.230	560248	775354	0.230m	0.193
14) MA Endrin	6.572	5.636	96765137	157.7E6	41.268	42.705
16) A 4,4'-DDD	6.707	5.785	3354680	5600389	1.765m	1.774
17) MA 4,4'-DDT	7.022	6.034	162.5E6	314.0E6	82.406	96.508
18) B Endrin al...	6.921	6.109	3175682	6776503	1.634m	2.226 #
20) A Methoxychlor	7.498	6.609	198.6E6	375.4E6	190.379	209.937
21) B Endrin ke...	7.640	6.838	5274952	9599279	2.091	2.288

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093726.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 10:30
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/22/2025
Supervised By : Ankita Jodhani 01/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 14:04:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:02:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 49897579
Conc: 18.53 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/22/2025
Supervised By :Ankita Jodhani 01/22/2025




#1 Tetrachloro-m-xylene

R.T.: 2.775 min
Delta R.T.: 0.000 min
Response: 58438387
Conc: 17.90 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 36373358
Conc: 9.49 ng/ml



#2 alpha-BHC

R.T.: 3.277 min
Delta R.T.: 0.000 min
Response: 42163610
Conc: 8.62 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.326 min
Delta R.T.: 0.000 min
Response: 34234012
Conc: 9.30 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/22/2025
Supervised By :Ankita Jodhani 01/22/2025




#3 gamma-BHC (Lindane)

R.T.: 3.607 min
Delta R.T.: 0.000 min
Response: 39348781
Conc: 8.30 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 15730216
Conc: 9.79 ng/ml



#6 beta-BHC

R.T.: 3.907 min
Delta R.T.: 0.001 min
Response: 19569860
Conc: 9.80 ng/ml




#12 4,4'-DDE

R.T.: 6.193 min
Delta R.T.: 0.002 min
Response: 560248
Conc: 0.23 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/22/2025
Supervised By :Ankita Jodhani 01/22/2025




#12 4,4'-DDE

R.T.: 5.230 min
Delta R.T.: 0.002 min
Response: 775354
Conc: 0.19 ng/ml




#14 Endrin

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 96765137
Conc: 41.27 ng/ml



#14 Endrin

R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 157695792
Conc: 42.70 ng/ml




#16 4,4'-DDD

R.T.: 6.707 min
Delta R.T.: 0.000 min
Response: 3354680
Conc: 1.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/22/2025
Supervised By :Ankita Jodhani 01/22/2025




#16 4,4'-DDD

R.T.: 5.785 min
Delta R.T.: 0.001 min
Response: 5600389
Conc: 1.77 ng/ml




#17 4,4'-DDT

R.T.: 7.022 min
Delta R.T.: 0.000 min
Response: 162509370
Conc: 82.41 ng/ml



#17 4,4'-DDT

R.T.: 6.034 min
Delta R.T.: 0.000 min
Response: 314041690
Conc: 96.51 ng/ml




#18 Endrin aldehyde

R.T.: 6.921 min
Delta R.T.: -0.001 min
Response: 3175682
Conc: 1.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/22/2025
Supervised By :Ankita Jodhani 01/22/2025




#18 Endrin aldehyde

R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 6776503
Conc: 2.23 ng/ml




#20 Methoxychlor

R.T.: 7.498 min
Delta R.T.: 0.000 min
Response: 198641245
Conc: 190.38 ng/ml



#20 Methoxychlor

R.T.: 6.609 min
Delta R.T.: 0.000 min
Response: 375396697
Conc: 209.94 ng/ml




#21 Endrin ketone

R.T.: 7.640 min
Delta R.T.: -0.001 min
Response: 5274952
Conc: 2.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/22/2025
Supervised By :Ankita Jodhani 01/22/2025




#21 Endrin ketone

R.T.: 6.838 min
Delta R.T.: 0.000 min
Response: 9599279
Conc: 2.29 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.052 min
Delta R.T.: 0.000 min
Response: 37808316
Conc: 18.07 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 62882920
Conc: 17.95 ng/ml

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093929.D Date Analyzed: 01/31/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.055	8.950	9.160	20.810	20.000	4.1
Tetrachloro-m-xylene	3.536	3.490	3.590	21.970	20.000	9.9
alpha-BHC	3.993	3.940	4.040	11.710	10.000	17.1
beta-BHC	4.524	4.470	4.570	11.950	10.000	19.5
gamma-BHC (Lindane)	4.325	4.270	4.380	11.550	10.000	15.5
Endrin	6.571	6.500	6.640	48.650	50.000	-2.7
4,4'-DDT	7.023	6.950	7.090	111.680	100.000	11.7
Methoxychlor	7.499	7.430	7.570	263.710	250.000	5.5

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093929.D Date Analyzed: 01/31/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.909	7.810	8.010	20.550	20.000	2.8
Tetrachloro-m-xylene	2.773	2.720	2.820	21.090	20.000	5.5
alpha-BHC	3.276	3.230	3.330	10.040	10.000	0.4
beta-BHC	3.906	3.860	3.960	10.760	10.000	7.6
gamma-BHC (Lindane)	3.606	3.560	3.660	9.410	10.000	-5.9
Endrin	5.636	5.570	5.710	52.600	50.000	5.2
4,4'-DDT	6.034	5.960	6.100	118.390	100.000	18.4
Methoxychlor	6.609	6.540	6.680	266.850	250.000	6.7

Data File: PEM
 PL093929.D **Date Acquired** 1/31/2025 11:04
Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.57	114069036.6	123405513.4	9336476.86	Down 7.57
Endrin aldehyde	6.92	3204039.84			
Endrin ketone	7.64	6132437.023			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.64	194240646	206778735.2	12538089.3	6.06
Endrin aldehyde #2	6.11	5122985.811			
Endrin ketone #2	6.84	7415103.441			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.02	220234745.2	221974836.9	1740091.71	0.78
4,4'-DDE	0.00	0			
4,4'-DDD	6.71	1740091.71			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.03	385239381.2	387063377.5	1823996.3	0.47
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.78	1823996.297			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093929.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 11:04
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 02/03/2025
 Supervised By : Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:23:18 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.536	2.773	59160870	68856501	21.970	21.095
28) SA Decachlor...	9.055	7.909	43538365	71991432	20.813	20.545
Target Compounds						
2) A alpha-BHC	3.993	3.276	44899644	49070780	11.711	10.037
3) MA gamma-BHC...	4.325	3.606	42528053	44605929	11.548	9.408
6) B beta-BHC	4.524	3.906	19203907	21501109	11.948	10.764
14) MA Endrin	6.571	5.636	114.1E6	194.2E6	48.647m	52.601
16) A 4,4'-DDD	6.708	5.783	1740092	1823996	0.916m	0.578m#
17) MA 4,4'-DDT	7.023	6.034	220.2E6	385.2E6	111.678	118.388
18) B Endrin al...	6.923	6.110	3204040	5122986	1.648	1.683
20) A Methoxychlor	7.499	6.609	275.2E6	477.2E6	263.715	266.852
21) B Endrin ke...	7.642	6.838	6132437	7415103	2.431	1.768 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093929.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 11:04
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 02/03/2025
Supervised By : Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:23:18 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.536 min
Delta R.T.: -0.003 min
Response: 59160870
Conc: 21.97 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: -0.002 min
Response: 68856501
Conc: 21.09 ng/ml




#2 alpha-BHC

R.T.: 3.993 min
Delta R.T.: -0.002 min
Response: 44899644
Conc: 11.71 ng/ml



#2 alpha-BHC

R.T.: 3.276 min
Delta R.T.: -0.001 min
Response: 49070780
Conc: 10.04 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.325 min
Delta R.T.: -0.002 min
Response: 42528053
Conc: 11.55 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#3 gamma-BHC (Lindane)

R.T.: 3.606 min
Delta R.T.: -0.001 min
Response: 44605929
Conc: 9.41 ng/ml




#6 beta-BHC

R.T.: 4.524 min
Delta R.T.: -0.001 min
Response: 19203907
Conc: 11.95 ng/ml



#6 beta-BHC

R.T.: 3.906 min
Delta R.T.: -0.001 min
Response: 21501109
Conc: 10.76 ng/ml




#14 Endrin

R.T.: 6.571 min
Delta R.T.: -0.001 min
Response: 114069037
Conc: 48.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#14 Endrin

R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 194240646
Conc: 52.60 ng/ml




#16 4,4'-DDD

R.T.: 6.708 min
Delta R.T.: 0.000 min
Response: 1740092
Conc: 0.92 ng/ml m



#16 4,4'-DDD

R.T.: 5.783 min
Delta R.T.: -0.001 min
Response: 1823996
Conc: 0.58 ng/ml m




#17 4,4'-DDT

R.T.: 7.023 min
Delta R.T.: 0.000 min
Response: 220234745
Conc: 111.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#17 4,4'-DDT

R.T.: 6.034 min
Delta R.T.: 0.000 min
Response: 385239381
Conc: 118.39 ng/ml




#18 Endrin aldehyde

R.T.: 6.923 min
Delta R.T.: 0.000 min
Response: 3204040
Conc: 1.65 ng/ml



#18 Endrin aldehyde

R.T.: 6.110 min
Delta R.T.: 0.000 min
Response: 5122986
Conc: 1.68 ng/ml




#20 Methoxychlor

R.T.: 7.499 min
Delta R.T.: 0.001 min
Response: 275158888
Conc: 263.71 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#20 Methoxychlor

R.T.: 6.609 min
Delta R.T.: 0.000 min
Response: 477168648
Conc: 266.85 ng/ml




#21 Endrin ketone

R.T.: 7.642 min
Delta R.T.: 0.000 min
Response: 6132437
Conc: 2.43 ng/ml



#21 Endrin ketone

R.T.: 6.838 min
Delta R.T.: 0.000 min
Response: 7415103
Conc: 1.77 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.002 min
Response: 43538365
Conc: 20.81 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025



#28 Decachlorobiphenyl

R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 71991432
Conc: 20.55 ng/ml

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093958.D Date Analyzed: 01/31/2025

Lab Sample No.(PEM): PEM Time Analyzed: 19:39

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.056	8.960	9.160	20.910	20.000	4.6
Tetrachloro-m-xylene	3.537	3.490	3.590	23.840	20.000	19.2
alpha-BHC	3.993	3.940	4.040	12.420	10.000	24.2
beta-BHC	4.525	4.470	4.580	12.480	10.000	24.8
gamma-BHC (Lindane)	4.326	4.280	4.380	11.910	10.000	19.1
Endrin	6.572	6.500	6.640	47.160	50.000	-5.7
4,4'-DDT	7.024	6.950	7.090	98.200	100.000	-1.8
Methoxychlor	7.501	7.430	7.570	231.470	250.000	-7.4

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093958.D Date Analyzed: 01/31/2025

Lab Sample No.(PEM): PEM Time Analyzed: 19:39

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.911	7.810	8.010	18.550	20.000	-7.3
Tetrachloro-m-xylene	2.774	2.720	2.820	22.890	20.000	14.5
alpha-BHC	3.276	3.230	3.330	10.850	10.000	8.5
beta-BHC	3.907	3.860	3.960	12.090	10.000	20.9
gamma-BHC (Lindane)	3.606	3.560	3.660	10.490	10.000	4.9
Endrin	5.637	5.570	5.710	50.900	50.000	1.8
4,4'-DDT	6.034	5.960	6.100	108.940	100.000	8.9
Methoxychlor	6.611	6.540	6.680	235.360	250.000	-5.9

Data File: PEM
 PL093958.D **Date Acquired** 1/31/2025 19:39
Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.57	110571791.9	124023498.6	13451706.7	Down 10.85
Endrin aldehyde	6.92	3858593.612			
Endrin ketone	7.64	9593113.101			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.64	187947389.1	211853935	23906545.9	11.28
Endrin aldehyde #2	6.11	7418834.795			
Endrin ketone #2	6.84	16487711.06			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.02	193658434.1	205489483	11831048.9	5.76
4,4'-DDE	6.19	884644.732			
4,4'-DDD	6.71	10946404.2			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.03	354508166	369800573.4	15292407.4	4.14
4,4'-DDE #2	5.23	640806.411			
4,4'-DDD #2	5.78	14651600.97			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093958.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 19:39
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 02/03/2025
 Supervised By : Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:32:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.774	64208122	74728930	23.845	22.894
28) SA Decachlor...	9.056	7.911	43746756	65001026	20.912	18.550
Target Compounds						
2) A alpha-BHC	3.993	3.276	47630507	53034269	12.424	10.848
3) MA gamma-BHC...	4.326	3.606	43873273	49725275	11.913	10.488
6) B beta-BHC	4.525	3.907	20063671	24147956	12.483	12.089
12) B 4,4'-DDE	6.187	5.228	884645	640806	0.363m	0.160m#
14) MA Endrin	6.572	5.637	110.6E6	187.9E6	47.156m	50.897
16) A 4,4'-DDD	6.709	5.785	10946404	14651601	5.760	4.642
17) MA 4,4'-DDT	7.024	6.034	193.7E6	354.5E6	98.201	108.944
18) B Endrin al...	6.924	6.110	3858594	7418835	1.985	2.437
20) A Methoxychlor	7.501	6.611	241.5E6	420.9E6	231.467	235.364
21) B Endrin ke...	7.644	6.839	9593113	16487711	3.803	3.930

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093958.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 19:39
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 02/03/2025
Supervised By : Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:32:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.537 min
Delta R.T.: -0.002 min
Response: 64208122
Conc: 23.84 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 74728930
Conc: 22.89 ng/ml




#2 alpha-BHC

R.T.: 3.993 min
Delta R.T.: -0.001 min
Response: 47630507
Conc: 12.42 ng/ml



#2 alpha-BHC

R.T.: 3.276 min
Delta R.T.: 0.000 min
Response: 53034269
Conc: 10.85 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.326 min
Delta R.T.: -0.001 min
Response: 43873273
Conc: 11.91 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#3 gamma-BHC (Lindane)

R.T.: 3.606 min
Delta R.T.: 0.000 min
Response: 49725275
Conc: 10.49 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 20063671
Conc: 12.48 ng/ml



#6 beta-BHC

R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 24147956
Conc: 12.09 ng/ml



#12 4,4'-DDE

R.T.: 6.187 min
Delta R.T.: -0.004 min
Response: 884645
Conc: 0.36 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025

#12 4,4'-DDE


R.T.: 5.228 min
Delta R.T.: -0.002 min
Response: 640806
Conc: 0.16 ng/ml

#14 Endrin

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 110571792
Conc: 47.16 ng/ml

#14 Endrin

R.T.: 5.637 min
Delta R.T.: 0.000 min
Response: 187947389
Conc: 50.90 ng/ml




#16 4,4'-DDD

R.T.: 6.709 min
Delta R.T.: 0.001 min
Response: 10946404
Conc: 5.76 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#16 4,4'-DDD

R.T.: 5.785 min
Delta R.T.: 0.000 min
Response: 14651601
Conc: 4.64 ng/ml




#17 4,4'-DDT

R.T.: 7.024 min
Delta R.T.: 0.002 min
Response: 193658434
Conc: 98.20 ng/ml



#17 4,4'-DDT

R.T.: 6.034 min
Delta R.T.: 0.000 min
Response: 354508166
Conc: 108.94 ng/ml




#18 Endrin aldehyde

R.T.: 6.924 min
Delta R.T.: 0.001 min
Response: 3858594
Conc: 1.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#18 Endrin aldehyde

R.T.: 6.110 min
Delta R.T.: 0.000 min
Response: 7418835
Conc: 2.44 ng/ml




#20 Methoxychlor

R.T.: 7.501 min
Delta R.T.: 0.003 min
Response: 241512375
Conc: 231.47 ng/ml



#20 Methoxychlor

R.T.: 6.611 min
Delta R.T.: 0.001 min
Response: 420862172
Conc: 235.36 ng/ml




#21 Endrin ketone

R.T.: 7.644 min
Delta R.T.: 0.002 min
Response: 9593113
Conc: 3.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#21 Endrin ketone

R.T.: 6.839 min
Delta R.T.: 0.000 min
Response: 16487711
Conc: 3.93 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: 0.004 min
Response: 43746756
Conc: 20.91 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: 0.001 min
Response: 65001026
Conc: 18.55 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093727.D
Acq On : 21 Jan 2025 10:43
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e

Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Title : GC Extractables
Last Update : Tue Jan 21 14:02:23 2025
Integrator: ChemStation

RT#1	RT#2	Resolution
3.539	5.939	100.00%
5.939	6.068	100.00%
6.068	6.191	100.00%
6.191	6.343	100.00%
6.343	7.157	100.00%
7.157	7.499	100.00%
7.499	7.642	100.00%
7.642	9.053	100.00%

Signal #2

2.774	4.977	100.00%
4.977	5.097	100.00%
5.097	5.230	100.00%
5.230	5.361	100.00%
5.361	6.333	100.00%
6.333	6.609	100.00%
6.609	6.838	100.00%
6.838	7.910	100.00%

PL012125.M Tue Jan 21 14:11:38 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
 Data File : PL093727.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2025 10:43
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 RESCHK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 14:04:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:02:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.774	48739758	56764042	18.100	17.390
28) SA Decachlor...	9.053	7.910	37826748	61983547	18.082	17.689
Target Compounds						
9) A Endosulfan I	6.068	5.097	24505533	31106405	9.272	8.023
10) B gamma-Chl...	5.939	4.977	26686746	36721665	9.574	8.666
12) B 4,4'-DDE	6.191	5.230	45047667	71134812	18.503	17.742
13) MA Dieldrin	6.343	5.361	50553851	73721045	18.212	17.162
19) B Endosulfa...	7.157	6.333	42219467	62907773	18.650	17.641
20) A Methoxychlor	7.499	6.609	83993166	151.6E6	80.500	84.806
21) B Endrin ke...	7.642	6.838	44318803	70010295	17.569	16.688



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093727.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 10:43
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
RESCHK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 14:04:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:02:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.539 min
Delta R.T.: 0.001 min
Response: 48739758
Conc: 18.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 56764042
Conc: 17.39 ng/ml




#9 Endosulfan I

R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 24505533
Conc: 9.27 ng/ml



#9 Endosulfan I


R.T.: 5.097 min
Delta R.T.: 0.000 min
Response: 31106405
Conc: 8.02 ng/ml



#10 gamma-Chlordane


R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 26686746
Conc: 9.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#10 gamma-Chlordane

R.T.: 4.977 min
Delta R.T.: 0.000 min
Response: 36721665
Conc: 8.67 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 45047667
Conc: 18.50 ng/ml



#12 4,4'-DDE


R.T.: 5.230 min
Delta R.T.: 0.001 min
Response: 71134812
Conc: 17.74 ng/ml



#13 Dieldrin


R.T.: 6.343 min
Delta R.T.: 0.000 min
Response: 50553851
Conc: 18.21 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#13 Dieldrin

R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 73721045
Conc: 17.16 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.157 min
Delta R.T.: 0.001 min
Response: 42219467
Conc: 18.65 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 62907773
Conc: 17.64 ng/ml



#20 Methoxychlor


R.T.: 7.499 min
Delta R.T.: 0.000 min
Response: 83993166
Conc: 80.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#20 Methoxychlor

R.T.: 6.609 min
Delta R.T.: 0.000 min
Response: 151645256
Conc: 84.81 ng/ml




#21 Endrin ketone

R.T.: 7.642 min
Delta R.T.: 0.000 min
Response: 44318803
Conc: 17.57 ng/ml



#21 Endrin ketone


R.T.: 6.838 min
Delta R.T.: 0.000 min
Response: 70010295
Conc: 16.69 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: 0.000 min
Response: 37826748
Conc: 18.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK



#28 Decachlorobiphenyl

R.T.: 7.910 min
Delta R.T.: 0.001 min
Response: 61983547
Conc: 17.69 ng/ml

Analytical Sequence

Client: RU2 Engineering, LLC	SDG No.: Q1232
Project: NYCDDC SANTWOBR Brooklyn Bridge BI	Instrument ID: ECD_L
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 01/21/2025 01/21/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
LBLK	LBLK	01/21/2025	10:16	PL093725.D	9.05	3.54
PEM	PEM	01/21/2025	10:30	PL093726.D	9.05	3.54
RESCHK	RESCHK	01/21/2025	10:43	PL093727.D	9.05	3.54
PSTDICC100	PSTDICC100	01/21/2025	10:57	PL093728.D	9.05	3.54
PSTDICC075	PSTDICC075	01/21/2025	11:10	PL093729.D	9.05	3.54
PSTDICC050	PSTDICC050	01/21/2025	11:24	PL093730.D	9.05	3.54
PSTDICC025	PSTDICC025	01/21/2025	11:38	PL093731.D	9.05	3.54
PSTDICC005	PSTDICC005	01/21/2025	11:51	PL093732.D	9.05	3.54
PCHLORICC500	PCHLORICC500	01/21/2025	12:32	PL093735.D	9.05	3.54
PTOXICC500	PTOXICC500	01/21/2025	13:39	PL093740.D	9.05	3.54
LBLK	LBLK	01/31/2025	10:51	PL093928.D	9.05	3.54
PEM	PEM	01/31/2025	11:04	PL093929.D	9.06	3.54
PSTDCCC050	PSTDCCC050	01/31/2025	11:17	PL093930.D	9.06	3.54
PB166413BL	PB166413BL	01/31/2025	14:28	PL093940.D	9.06	3.54
PB166413BS	PB166413BS	01/31/2025	14:43	PL093941.D	9.06	3.54
LBLK	LBLK	01/31/2025	14:57	PL093942.D	9.06	3.54
PSTDCCC050	PSTDCCC050	01/31/2025	15:10	PL093943.D	9.06	3.54
JPP-46.2-012925	Q1232-03	01/31/2025	17:01	PL093946.D	9.06	3.54
JPP-46.1-012925	Q1232-07	01/31/2025	17:14	PL093947.D	9.06	3.54
JPP-42.1-012925	Q1232-11	01/31/2025	17:27	PL093948.D	9.05	3.54
JPP-42.2-012925	Q1232-15	01/31/2025	17:40	PL093949.D	9.06	3.54
JPP-51.1-012925	Q1232-19	01/31/2025	17:53	PL093950.D	9.06	3.54
LBLK	LBLK	01/31/2025	19:26	PL093957.D	9.06	3.54
PEM	PEM	01/31/2025	19:39	PL093958.D	9.06	3.54
PSTDCCC050	PSTDCCC050	01/31/2025	20:18	PL093959.D	9.06	3.54
357MS	Q1239-10MS	01/31/2025	20:31	PL093960.D	9.06	3.54
357MSD	Q1239-10MSD	01/31/2025	20:45	PL093961.D	9.06	3.54
LBLK	LBLK	01/31/2025	22:43	PL093970.D	9.06	3.54
PSTDCCC050	PSTDCCC050	02/01/2025	00:57	PL093971.D	9.06	3.54

Analytical Sequence

Client: RU2 Engineering, LLC	SDG No.: Q1232
Project: NYCDDC SANTWOBR Brooklyn Bridge BI	Instrument ID: ECD_L
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 01/21/2025 01/21/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	01/21/2025	10:16	PL093725.D	7.91	2.78
PEM	PEM	01/21/2025	10:30	PL093726.D	7.91	2.78
RESCHK	RESCHK	01/21/2025	10:43	PL093727.D	7.91	2.77
PSTDICC100	PSTDICC100	01/21/2025	10:57	PL093728.D	7.91	2.78
PSTDICC075	PSTDICC075	01/21/2025	11:10	PL093729.D	7.91	2.77
PSTDICC050	PSTDICC050	01/21/2025	11:24	PL093730.D	7.91	2.77
PSTDICC025	PSTDICC025	01/21/2025	11:38	PL093731.D	7.91	2.77
PSTDICC005	PSTDICC005	01/21/2025	11:51	PL093732.D	7.91	2.77
PCHLORICC500	PCHLORICC500	01/21/2025	12:32	PL093735.D	7.91	2.77
PTOXICC500	PTOXICC500	01/21/2025	13:39	PL093740.D	7.91	2.77
IBLK	IBLK	01/31/2025	10:51	PL093928.D	7.91	2.77
PEM	PEM	01/31/2025	11:04	PL093929.D	7.91	2.77
PSTDCCC050	PSTDCCC050	01/31/2025	11:17	PL093930.D	7.91	2.77
PB166413BL	PB166413BL	01/31/2025	14:28	PL093940.D	7.91	2.77
PB166413BS	PB166413BS	01/31/2025	14:43	PL093941.D	7.91	2.77
IBLK	IBLK	01/31/2025	14:57	PL093942.D	7.91	2.77
PSTDCCC050	PSTDCCC050	01/31/2025	15:10	PL093943.D	7.91	2.77
JPP-46.2-012925	Q1232-03	01/31/2025	17:01	PL093946.D	7.91	2.77
JPP-46.1-012925	Q1232-07	01/31/2025	17:14	PL093947.D	7.91	2.77
JPP-42.1-012925	Q1232-11	01/31/2025	17:27	PL093948.D	7.91	2.77
JPP-42.2-012925	Q1232-15	01/31/2025	17:40	PL093949.D	7.91	2.77
JPP-51.1-012925	Q1232-19	01/31/2025	17:53	PL093950.D	7.91	2.77
IBLK	IBLK	01/31/2025	19:26	PL093957.D	7.91	2.77
PEM	PEM	01/31/2025	19:39	PL093958.D	7.91	2.77
PSTDCCC050	PSTDCCC050	01/31/2025	20:18	PL093959.D	7.91	2.77
357MS	Q1239-10MS	01/31/2025	20:31	PL093960.D	7.91	2.77
357MSD	Q1239-10MSD	01/31/2025	20:45	PL093961.D	7.91	2.77
IBLK	IBLK	01/31/2025	22:43	PL093970.D	7.91	2.77
PSTDCCC050	PSTDCCC050	02/01/2025	00:57	PL093971.D	7.91	2.77

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

357MS

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1232

SAS No.: Q1232

SDG NO.: Q1232

Lab Sample ID: Q1239-10MS

Date(s) Analyzed: 01/31/2025

01/31/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.79	6.74	6.84	18.7	4.2
	2	5.93	5.88	5.98	19.5	
4,4'-DDD	1	6.71	6.66	6.76	22.0	2.8
	2	5.79	5.74	5.84	21.4	
4,4'-DDT	1	7.02	6.97	7.07	18.9	0.5
	2	6.04	5.99	6.09	18.8	
Endrin aldehyde	1	6.92	6.87	6.97	13.7	3.7
	2	6.11	6.06	6.16	13.2	
Endosulfan sulfate	1	7.16	7.11	7.21	18.2	1.1
	2	6.33	6.28	6.38	18.4	
Methoxychlor	1	7.50	7.45	7.55	17.4	0.6
	2	6.61	6.56	6.66	17.5	
Endrin ketone	1	7.65	7.60	7.70	17.5	0
	2	6.84	6.79	6.89	17.5	
alpha-BHC	1	3.99	3.94	4.04	19.6	2.1
	2	3.28	3.23	3.33	19.2	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	19.2	1.6
	2	3.61	3.56	3.66	18.9	
Heptachlor	1	4.92	4.87	4.97	19.1	3.7
	2	3.95	3.90	4.00	18.4	
Aldrin	1	5.26	5.21	5.31	19.3	0
	2	4.22	4.17	4.27	19.3	
beta-BHC	1	4.53	4.48	4.58	19.8	0.5
	2	3.91	3.86	3.96	19.9	
delta-BHC	1	4.77	4.72	4.82	19.1	1.1
	2	4.14	4.09	4.19	18.9	
Heptachlor epoxide	1	5.68	5.63	5.73	18.6	7.3
	2	4.73	4.68	4.78	20.0	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

357MS

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232

SAS No.: Q1232 SDG NO.: Q1232

Lab Sample ID: Q1239-10MS

Date(s) Analyzed: 01/31/2025 01/31/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.07	6.02	6.12	19.0	0.5
	2	5.10	5.05	5.15	18.9	
gamma-Chlordane	1	5.94	5.89	5.99	20.2	3.9
	2	4.98	4.93	5.03	21.0	
alpha-Chlordane	1	6.02	5.97	6.07	19.8	3
	2	5.04	4.99	5.09	20.4	
4,4'-DDE	1	6.19	6.14	6.24	19.9	7.3
	2	5.23	5.18	5.28	21.4	
Dieldrin	1	6.35	6.30	6.40	19.2	6.5
	2	5.36	5.31	5.41	20.5	
Endrin	1	6.57	6.52	6.62	18.6	8.2
	2	5.64	5.59	5.69	20.2	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

357MSD

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232

SAS No.: Q1232 SDG NO.: Q1232

Lab Sample ID: Q1239-10MSD

Date(s) Analyzed: 01/31/2025 01/31/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.79	6.74	6.84	19.2	4.6
	2	5.93	5.88	5.98	20.1	
4,4'-DDD	1	6.71	6.66	6.76	22.2	2.7
	2	5.78	5.73	5.83	21.6	
alpha-Chlordane	1	6.02	5.97	6.07	20.0	3
	2	5.04	4.99	5.09	20.6	
4,4'-DDE	1	6.19	6.14	6.24	20.0	6.8
	2	5.23	5.18	5.28	21.4	
Dieldrin	1	6.34	6.29	6.39	19.2	7
	2	5.36	5.31	5.41	20.6	
Endrin	1	6.57	6.52	6.62	19.0	9.5
	2	5.64	5.59	5.69	20.9	
4,4'-DDT	1	7.02	6.97	7.07	20.5	0.5
	2	6.03	5.98	6.08	20.4	
Endrin aldehyde	1	6.92	6.87	6.97	18.7	0
	2	6.11	6.06	6.16	18.7	
Endosulfan sulfate	1	7.16	7.11	7.21	18.9	4.1
	2	6.33	6.28	6.38	19.7	
Methoxychlor	1	7.50	7.45	7.55	19.8	1
	2	6.61	6.56	6.66	19.6	
Endrin ketone	1	7.64	7.59	7.69	18.8	0
	2	6.84	6.79	6.89	18.8	
alpha-BHC	1	3.99	3.94	4.04	19.8	2
	2	3.28	3.23	3.33	19.4	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	19.4	1
	2	3.61	3.56	3.66	19.2	
Heptachlor	1	4.91	4.86	4.96	20.3	3.5
	2	3.94	3.89	3.99	19.6	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

357MSD

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232

SAS No.: Q1232 SDG NO.: Q1232

Lab Sample ID: Q1239-10MSD

Date(s) Analyzed: 01/31/2025 01/31/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Aldrin	1	5.26	5.21	5.31	19.5	1
	2	4.22	4.17	4.27	19.3	
beta-BHC	1	4.53	4.48	4.58	20.1	0.5
	2	3.91	3.86	3.96	20.0	
delta-BHC	1	4.77	4.72	4.82	20.0	3
	2	4.13	4.08	4.18	19.4	
Heptachlor epoxide	1	5.68	5.63	5.73	18.9	5.7
	2	4.73	4.68	4.78	20.0	
Endosulfan I	1	6.07	6.02	6.12	19.5	1
	2	5.10	5.05	5.15	19.3	
gamma-Chlordane	1	5.94	5.89	5.99	20.1	3.9
	2	4.98	4.93	5.03	20.9	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

JPP-42.1-012925

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1232

SAS No.: Q1232

SDG NO.: Q1232

Lab Sample ID: Q1232-11

Date(s) Analyzed: 01/31/2025 01/31/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column:(2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDT	1	7.02	6.97	7.07	0.35	103.3
	2	6.03	5.98	6.08	1.10	
alpha-Chlordane	1	6.02	5.97	6.07	1.10	70.9
	2	5.04	4.99	5.09	0.52	
Endrin	1	6.57	6.52	6.62	0.22	11.4
	2	5.64	5.59	5.69	0.25	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

JPP-42.2-012925

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1232

SAS No.: Q1232

SDG NO.: Q1232

Lab Sample ID: Q1232-15

Date(s) Analyzed: 01/31/2025 01/31/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDT	1	7.02	6.97	7.07	0.56	64.4
	2	6.03	5.98	6.08	1.10	
Heptachlor	1	4.91	4.86	4.96	0.73	55.4
	2	3.94	3.89	3.99	0.41	
gamma-Chlordane	1	5.94	5.89	5.99	1.20	110.4
	2	4.98	4.93	5.03	0.35	
alpha-Chlordane	1	6.02	5.97	6.07	2.30	86
	2	5.04	4.99	5.09	0.92	
Endrin	1	6.57	6.52	6.62	0.19	3.7
	2	5.64	5.59	5.69	0.20	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

JPP-46.1-012925

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1232

SAS No.: Q1232

SDG NO.: Q1232

Lab Sample ID: Q1232-07

Date(s) Analyzed: 01/31/2025 01/31/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDT	1	7.02	6.97	7.07	0.29	138.5
	2	6.03	5.98	6.08	1.60	
alpha-Chlordane	1	6.02	5.97	6.07	1.20	58.2
	2	5.04	4.99	5.09	0.66	
4,4'-DDE	1	6.18	6.13	6.23	0.43	12.2
	2	5.23	5.18	5.28	0.38	
Endrin	1	6.57	6.52	6.62	0.60	69.8
	2	5.64	5.59	5.69	0.29	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

JPP-46.2-012925

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1232

SAS No.: Q1232

SDG NO.: Q1232

Lab Sample ID: Q1232-03

Date(s) Analyzed: 01/31/2025

01/31/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column:(2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.79	6.74	6.84	0.90	105.6
	2	5.94	5.89	5.99	2.90	
4,4'-DDD	1	6.71	6.66	6.76	2.80	80
	2	5.77	5.72	5.82	1.20	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

JPP-51.1-012925

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232

SAS No.: Q1232 SDG NO.: Q1232

Lab Sample ID: Q1232-19

Date(s) Analyzed: 01/31/2025 01/31/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDT	1	7.02	6.97	7.07	1.00	66.7
	2	6.03	5.98	6.08	2.00	
Heptachlor	1	4.91	4.86	4.96	0.44	34.8
	2	3.94	3.89	3.99	0.31	
Heptachlor epoxide	1	5.68	5.63	5.73	0.37	71.6
	2	4.73	4.68	4.78	0.78	
alpha-Chlordane	1	6.02	5.97	6.07	2.90	69.8
	2	5.04	4.99	5.09	1.40	
Endrin	1	6.57	6.52	6.62	0.33	39.3
	2	5.64	5.59	5.69	0.49	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB166413BS

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1232

SAS No.: Q1232

SDG NO.: Q1232

Lab Sample ID: PB166413BS

Date(s) Analyzed: 01/31/2025

01/31/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.71	6.66	6.76	17.0	0.6
	2	5.79	5.74	5.84	17.1	
4,4'-DDT	1	7.03	6.98	7.08	17.5	0
	2	6.04	5.99	6.09	17.5	
Aldrin	1	5.26	5.21	5.31	15.0	3.4
	2	4.23	4.18	4.28	14.5	
4,4'-DDE	1	6.19	6.14	6.24	16.5	0.6
	2	5.23	5.18	5.28	16.4	
Endosulfan II	1	6.80	6.75	6.85	15.5	4.4
	2	5.93	5.88	5.98	16.2	
Endrin aldehyde	1	6.93	6.88	6.98	15.0	2
	2	6.11	6.06	6.16	15.3	
Endosulfan sulfate	1	7.16	7.11	7.21	15.1	7
	2	6.33	6.28	6.38	16.2	
Methoxychlor	1	7.50	7.45	7.55	16.6	3.6
	2	6.61	6.56	6.66	17.2	
Endrin ketone	1	7.65	7.60	7.70	15.2	3.2
	2	6.84	6.79	6.89	15.7	
alpha-BHC	1	4.00	3.95	4.05	14.9	2
	2	3.28	3.23	3.33	14.6	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	14.8	3.4
	2	3.61	3.56	3.66	14.3	
Heptachlor	1	4.92	4.87	4.97	16.0	4.5
	2	3.95	3.90	4.00	15.3	
beta-BHC	1	4.53	4.48	4.58	15.0	0.7
	2	3.91	3.86	3.96	15.1	
delta-BHC	1	4.78	4.73	4.83	15.1	4.1
	2	4.14	4.09	4.19	14.5	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB166413BS

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232

SAS No.: Q1232 SDG NO.: Q1232

Lab Sample ID: PB166413BS

Date(s) Analyzed: 01/31/2025 01/31/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Heptachlor epoxide	1	5.69	5.64	5.74	14.8	3.3
	2	4.73	4.68	4.78	15.3	
Endosulfan I	1	6.07	6.02	6.12	15.4	2
	2	5.10	5.05	5.15	15.1	
gamma-Chlordane	1	5.94	5.89	5.99	15.4	1.9
	2	4.98	4.93	5.03	15.7	
alpha-Chlordane	1	6.02	5.97	6.07	15.6	0.6
	2	5.04	4.99	5.09	15.7	
Dieldrin	1	6.35	6.30	6.40	15.4	0.6
	2	5.36	5.31	5.41	15.5	
Endrin	1	6.58	6.53	6.63	15.7	6.8
	2	5.64	5.59	5.69	16.8	



QC SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	PB166413BL	SDG No.:	Q1232
Lab Sample ID:	PB166413BL	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	100 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093940.D	1	01/31/25 08:15	01/31/25 14:28	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	1.70	U	0.18	1.70	ug/kg
319-85-7	beta-BHC	1.70	U	0.49	1.70	ug/kg
319-86-8	delta-BHC	1.70	U	0.47	1.70	ug/kg
58-89-9	gamma-BHC (Lindane)	1.70	U	0.19	1.70	ug/kg
76-44-8	Heptachlor	1.70	U	0.17	1.70	ug/kg
309-00-2	Aldrin	1.70	U	0.14	1.70	ug/kg
1024-57-3	Heptachlor epoxide	1.70	U	0.23	1.70	ug/kg
959-98-8	Endosulfan I	1.70	U	0.17	1.70	ug/kg
60-57-1	Dieldrin	1.70	U	0.15	1.70	ug/kg
72-55-9	4,4-DDE	1.70	U	0.13	1.70	ug/kg
72-20-8	Endrin	1.70	U	0.16	1.70	ug/kg
33213-65-9	Endosulfan II	1.70	U	0.30	1.70	ug/kg
72-54-8	4,4-DDD	1.70	U	0.19	1.70	ug/kg
1031-07-8	Endosulfan Sulfate	1.70	U	0.13	1.70	ug/kg
50-29-3	4,4-DDT	1.70	U	0.17	1.70	ug/kg
72-43-5	Methoxychlor	1.70	U	0.38	1.70	ug/kg
53494-70-5	Endrin ketone	1.70	U	0.22	1.70	ug/kg
7421-93-4	Endrin aldehyde	1.70	U	0.39	1.70	ug/kg
5103-71-9	alpha-Chlordane	1.70	U	0.17	1.70	ug/kg
5103-74-2	gamma-Chlordane	1.70	U	0.19	1.70	ug/kg
8001-35-2	Toxaphene	33.0	U	5.20	33.0	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	24.6		10 - 148	123%	SPK: 20
877-09-8	Tetrachloro-m-xylene	23.8		10 - 159	119%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	PB166413BL	SDG No.:	Q1232
Lab Sample ID:	PB166413BL	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	100
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
Prep Method :	SW3541B	Test:	Pesticide-TCL
		PH :	
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093940.D	1	01/31/25 08:15	01/31/25 14:28	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093940.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 14:28
Operator : AR\AJ
Sample : PB166413BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB166413BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:26:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.774	64156821	72110298	23.826	22.092
28) SA Decachlor...	9.056	7.910	51373198	84523384	24.558	24.122

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093940.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 14:28
Operator : AR\AJ
Sample : PB166413BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB166413BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:26:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.537 min
Delta R.T.: -0.002 min
Response: 64156821
Conc: 23.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB166413BL




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 72110298
Conc: 22.09 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: 0.003 min
Response: 51373198
Conc: 24.56 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.910 min
Delta R.T.: 0.000 min
Response: 84523384
Conc: 24.12 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/21/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/21/25
Client Sample ID:	PIBLK-PL093725.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PL093725.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093725.D	1		01/21/25	PL012125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.050	U	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.050	U	0.014	0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0054	0.050	ug/L
309-00-2	Aldrin	0.050	U	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.050	U	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.050	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.050	U	0.0045	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.050	U	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.050	U	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.050	U	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.050	U	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.050	U	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.050	U	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.050	U	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.050	U	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.1		43 - 140	111%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.8		77 - 126	104%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/21/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/21/25
Client Sample ID:	PIBLK-PL093725.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PL093725.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
Prep Method :	3510C	Test:	Pesticide-TCL
		PH :	
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093725.D	1		01/21/25	PL012125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
 Data File : PL093725.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2025 10:16
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 14:04:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:02:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.775	55919553	66932258	20.767	20.505
28) SA Decachlor...	9.052	7.909	46293108	76642664	22.130	21.872

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012125\
Data File : PL093725.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 10:16
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 14:04:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:02:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.539 min
Delta R.T.: 0.000 min
Response: 55919553
Conc: 20.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.775 min
Delta R.T.: 0.000 min
Response: 66932258
Conc: 20.51 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.052 min
Delta R.T.: 0.000 min
Response: 46293108
Conc: 22.13 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 76642664
Conc: 21.87 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/31/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/31/25
Client Sample ID:	PIBLK-PL093928.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PL093928.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093928.D	1		01/31/25	pl013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.050	U	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.050	U	0.014	0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0054	0.050	ug/L
309-00-2	Aldrin	0.050	U	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.050	U	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.050	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.050	U	0.0045	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.050	U	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.050	U	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.050	U	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.050	U	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.050	U	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.050	U	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.050	U	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.050	U	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.6		43 - 140	103%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.1		77 - 126	106%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/31/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/31/25
Client Sample ID:	PIBLK-PL093928.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PL093928.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Decanted:	
		Final Vol:	10000
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093928.D	1		01/31/25	pl013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093928.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 10:51
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:23:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.773	56869031	64496193	21.119	19.759
28) SA Decachlor...	9.053	7.909	43185846	71542365	20.644	20.417

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093928.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 10:51
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:23:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.537 min
Delta R.T.: -0.002 min
Response: 56869031
Conc: 21.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: -0.001 min
Response: 64496193
Conc: 19.76 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: 0.000 min
Response: 43185846
Conc: 20.64 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 71542365
Conc: 20.42 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/31/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/31/25
Client Sample ID:	PIBLK-PL093942.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PL093942.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093942.D	1		01/31/25	pl013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.050	U	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.050	U	0.014	0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0054	0.050	ug/L
309-00-2	Aldrin	0.050	U	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.050	U	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.050	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.050	U	0.0045	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.050	U	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.050	U	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.050	U	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.050	U	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.050	U	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.050	U	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.050	U	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.050	U	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.6		43 - 140	113%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.7		77 - 126	109%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/31/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/31/25
Client Sample ID:	PIBLK-PL093942.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PL093942.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
Prep Method :	3510C	PH :	
		Test:	Pesticide-TCL
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093942.D	1		01/31/25	pl013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 14:57
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:27:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.773	58455331	65748686	21.708	20.143
28) SA Decachlor...	9.056	7.910	47390121	78106190	22.654	22.290

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 14:57
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:27:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.537 min
Delta R.T.: -0.002 min
Response: 58455331
Conc: 21.71 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: -0.002 min
Response: 65748686
Conc: 20.14 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: 0.004 min
Response: 47390121
Conc: 22.65 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.910 min
Delta R.T.: 0.000 min
Response: 78106190
Conc: 22.29 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/31/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/31/25
Client Sample ID:	PIBLK-PL093957.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PL093957.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093957.D	1		01/31/25	pl013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.050	U	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.050	U	0.014	0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0054	0.050	ug/L
309-00-2	Aldrin	0.050	U	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.050	U	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.050	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.050	U	0.0045	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.050	U	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.050	U	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.050	U	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.050	U	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.050	U	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.050	U	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.050	U	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.050	U	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.4		43 - 140	102%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.0		77 - 126	110%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/31/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/31/25
Client Sample ID:	PIBLK-PL093957.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PL093957.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093957.D	1		01/31/25	pl013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093957.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 19:26
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:32:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.774	59144195	67936809	21.964	20.813
28) SA Decachlor...	9.056	7.911	42654762	59975666	20.390	17.116

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093957.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 19:26
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:32:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.538 min
Delta R.T.: -0.001 min
Response: 59144195
Conc: 21.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 67936809
Conc: 20.81 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: 0.003 min
Response: 42654762
Conc: 20.39 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: 0.001 min
Response: 59975666
Conc: 17.12 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/31/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/31/25
Client Sample ID:	PIBLK-PL093970.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PL093970.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093970.D	1		01/31/25	pl013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.050	U	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.050	U	0.014	0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.050	U	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.050	U	0.0054	0.050	ug/L
309-00-2	Aldrin	0.050	U	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.050	U	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.050	U	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.050	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.050	U	0.0045	0.050	ug/L
72-20-8	Endrin	0.050	U	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.050	U	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.050	U	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.050	U	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.050	U	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.050	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.050	U	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.050	U	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.050	U	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.050	U	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	19.3		43 - 140	96%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.3		77 - 126	107%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/31/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/31/25
Client Sample ID:	PIBLK-PL093970.D	SDG No.:	Q1232
Lab Sample ID:	I.BLK-PL093970.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
Prep Method :	3510C	PH :	
		Test:	Pesticide-TCL
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093970.D	1		01/31/25	pl013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093970.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 22:43
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:37:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.773	57423032	65419079	21.325	20.042
28) SA Decachlor...	9.057	7.911	40262923	56527597	19.247	16.132

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093970.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 22:43
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:37:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.537 min
Delta R.T.: -0.002 min
Response: 57423032
Conc: 21.32 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: 0.000 min
Response: 65419079
Conc: 20.04 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.057 min
Delta R.T.: 0.004 min
Response: 40262923
Conc: 19.25 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: 0.001 min
Response: 56527597
Conc: 16.13 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	PB166413BS	SDG No.:	Q1232
Lab Sample ID:	PB166413BS	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	100 Decanted:
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093941.D	1	01/31/25 08:15	01/31/25 14:43	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	14.9		0.18	1.70	ug/kg
319-85-7	beta-BHC	15.1		0.49	1.70	ug/kg
319-86-8	delta-BHC	15.1		0.47	1.70	ug/kg
58-89-9	gamma-BHC (Lindane)	14.8		0.19	1.70	ug/kg
76-44-8	Heptachlor	16.0		0.17	1.70	ug/kg
309-00-2	Aldrin	15.0		0.14	1.70	ug/kg
1024-57-3	Heptachlor epoxide	15.3		0.23	1.70	ug/kg
959-98-8	Endosulfan I	15.4		0.17	1.70	ug/kg
60-57-1	Dieldrin	15.5		0.15	1.70	ug/kg
72-55-9	4,4-DDE	16.5		0.13	1.70	ug/kg
72-20-8	Endrin	16.8		0.16	1.70	ug/kg
33213-65-9	Endosulfan II	16.2		0.30	1.70	ug/kg
72-54-8	4,4-DDD	17.1		0.19	1.70	ug/kg
1031-07-8	Endosulfan Sulfate	16.2		0.13	1.70	ug/kg
50-29-3	4,4-DDT	17.5		0.17	1.70	ug/kg
72-43-5	Methoxychlor	17.2		0.38	1.70	ug/kg
53494-70-5	Endrin ketone	15.7		0.22	1.70	ug/kg
7421-93-4	Endrin aldehyde	15.3		0.39	1.70	ug/kg
5103-71-9	alpha-Chlordane	15.7		0.17	1.70	ug/kg
5103-74-2	gamma-Chlordane	15.7		0.19	1.70	ug/kg
8001-35-2	Toxaphene	33.0	U	5.20	33.0	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.7		10 - 148	109%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.0		10 - 159	100%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	PB166413BS	SDG No.:	Q1232
Lab Sample ID:	PB166413BS	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	100
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
Prep Method :	SW3541B	Test:	Pesticide-TCL
		PH :	
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093941.D	1	01/31/25 08:15	01/31/25 14:43	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093941.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 14:43
 Operator : AR\AJ
 Sample : PB166413BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PB166413BS

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:26:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.541	2.774	53863775	60606861	20.003	18.567
28) SA Decachlor...	9.059	7.911	44293906	76038120	21.174	21.700
Target Compounds						
2) A alpha-BHC	3.997	3.276	171.2E6	213.9E6	44.668	43.753
3) MA gamma-BHC...	4.330	3.607	163.2E6	203.3E6	44.307	42.882
4) MA Heptachlor	4.917	3.945	157.6E6	213.8E6	48.094	45.939
5) MB Aldrin	5.259	4.225	146.8E6	199.2E6	44.877	43.660
6) B beta-BHC	4.528	3.907	72120927	90605027	44.870	45.361
7) B delta-BHC	4.775	4.135	159.2E6	207.2E6	45.405	43.618
8) B Heptachlo...	5.686	4.727	132.1E6	192.4E6	44.421	46.029
9) A Endosulfan I	6.072	5.098	121.9E6	175.3E6	46.106	45.217
10) B gamma-Chl...	5.940	4.976	129.1E6	199.9E6	46.303m	47.166m
11) B alpha-Chl...	6.021	5.041	130.6E6	196.9E6	46.837	47.038
12) B 4,4'-DDE	6.194	5.230	120.3E6	196.8E6	49.415	49.095
13) MA Dieldrin	6.346	5.362	128.0E6	200.1E6	46.118	46.582
14) MA Endrin	6.575	5.637	110.2E6	186.1E6	47.003m	50.392
15) B Endosulfa...	6.796	5.932	111.7E6	179.8E6	46.371	48.551
16) A 4,4'-DDD	6.712	5.785	96870322	161.6E6	50.969	51.185
17) MA 4,4'-DDT	7.025	6.035	103.5E6	170.6E6	52.489	52.414
18) B Endrin al...	6.926	6.111	87351319	139.8E6	44.933	45.924
19) B Endosulfa...	7.161	6.334	102.7E6	173.3E6	45.361	48.596
20) A Methoxychlor	7.502	6.610	51975133	92217194	49.813	51.572
21) B Endrin ke...	7.646	6.839	115.1E6	197.6E6	45.617	47.107
22) Mirex	8.119	7.020	87775989	147.9E6	42.150	43.722

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093941.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 14:43
Operator : AR\AJ
Sample : PB166413BS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PB166413BS



Manual Integrations


APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:26:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.541 min
Delta R.T.: 0.002 min
Response: 53863775
Conc: 20.00 ng/ml

Instrument :


ECD_L

ClientSampleId :

PB166413BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 60606861
Conc: 18.57 ng/ml




#2 alpha-BHC

R.T.: 3.997 min
Delta R.T.: 0.003 min
Response: 171248506
Conc: 44.67 ng/ml



#2 alpha-BHC

R.T.: 3.276 min
Delta R.T.: 0.000 min
Response: 213909521
Conc: 43.75 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.330 min
Delta R.T.: 0.003 min
Response: 163177663
Conc: 44.31 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB166413BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#3 gamma-BHC (Lindane)

R.T.: 3.607 min
Delta R.T.: 0.000 min
Response: 203313828
Conc: 42.88 ng/ml




#4 Heptachlor

R.T.: 4.917 min
Delta R.T.: 0.003 min
Response: 157620122
Conc: 48.09 ng/ml



#4 Heptachlor

R.T.: 3.945 min
Delta R.T.: 0.000 min
Response: 213838193
Conc: 45.94 ng/ml




#5 Aldrin

R.T.: 5.259 min
Delta R.T.: 0.003 min
Response: 146835897
Conc: 44.88 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB166413BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#5 Aldrin

R.T.: 4.225 min
Delta R.T.: 0.000 min
Response: 199167275
Conc: 43.66 ng/ml




#6 beta-BHC

R.T.: 4.528 min
Delta R.T.: 0.003 min
Response: 72120927
Conc: 44.87 ng/ml



#6 beta-BHC

R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 90605027
Conc: 45.36 ng/ml




#7 delta-BHC

R.T.: 4.775 min
Delta R.T.: 0.003 min
Response: 159158634
Conc: 45.41 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB166413BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#7 delta-BHC

R.T.: 4.135 min
Delta R.T.: 0.000 min
Response: 207239446
Conc: 43.62 ng/ml




#8 Heptachlor epoxide

R.T.: 5.686 min
Delta R.T.: 0.003 min
Response: 132098839
Conc: 44.42 ng/ml



#8 Heptachlor epoxide

R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 192412348
Conc: 46.03 ng/ml




#9 Endosulfan I

R.T.: 6.072 min
Delta R.T.: 0.004 min
Response: 121851374
Conc: 46.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB166413BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#9 Endosulfan I

R.T.: 5.098 min
Delta R.T.: 0.001 min
Response: 175304199
Conc: 45.22 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.002 min
Response: 129062469
Conc: 46.30 ng/ml m



#10 gamma-Chlordane

R.T.: 4.976 min
Delta R.T.: -0.001 min
Response: 199870170
Conc: 47.17 ng/ml m




#11 alpha-Chlordane

R.T.: 6.021 min
Delta R.T.: 0.004 min
Response: 130599407
Conc: 46.84 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB166413BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#11 alpha-Chlordane

R.T.: 5.041 min
Delta R.T.: 0.000 min
Response: 196927472
Conc: 47.04 ng/ml




#12 4,4'-DDE

R.T.: 6.194 min
Delta R.T.: 0.003 min
Response: 120305149
Conc: 49.41 ng/ml



#12 4,4'-DDE

R.T.: 5.230 min
Delta R.T.: 0.000 min
Response: 196843014
Conc: 49.09 ng/ml




#13 Dieldrin

R.T.: 6.346 min
Delta R.T.: 0.003 min
Response: 128016367
Conc: 46.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB166413BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#13 Dieldrin

R.T.: 5.362 min
Delta R.T.: 0.000 min
Response: 200100466
Conc: 46.58 ng/ml




#14 Endrin

R.T.: 6.575 min
Delta R.T.: 0.002 min
Response: 110213851
Conc: 47.00 ng/ml m



#14 Endrin

R.T.: 5.637 min
Delta R.T.: 0.000 min
Response: 186080914
Conc: 50.39 ng/ml




#15 Endosulfan II

R.T.: 6.796 min
Delta R.T.: 0.004 min
Response: 111724724
Conc: 46.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB166413BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#15 Endosulfan II

R.T.: 5.932 min
Delta R.T.: 0.000 min
Response: 179824512
Conc: 48.55 ng/ml




#16 4,4'-DDD

R.T.: 6.712 min
Delta R.T.: 0.004 min
Response: 96870322
Conc: 50.97 ng/ml



#16 4,4'-DDD

R.T.: 5.785 min
Delta R.T.: 0.000 min
Response: 161568406
Conc: 51.19 ng/ml




#17 4,4'-DDT

R.T.: 7.025 min
Delta R.T.: 0.003 min
Response: 103511557
Conc: 52.49 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB166413BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#17 4,4'-DDT

R.T.: 6.035 min
Delta R.T.: 0.001 min
Response: 170558141
Conc: 52.41 ng/ml




#18 Endrin aldehyde

R.T.: 6.926 min
Delta R.T.: 0.004 min
Response: 87351319
Conc: 44.93 ng/ml



#18 Endrin aldehyde

R.T.: 6.111 min
Delta R.T.: 0.001 min
Response: 139823266
Conc: 45.92 ng/ml



#19 Endosulfan Sulfate

R.T.: 7.161 min
Delta R.T.: 0.003 min
Response: 102686507
Conc: 45.36 ng/ml

Instrument :


ECD_L

ClientSampleId :

PB166413BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#19 Endosulfan Sulfate

R.T.: 6.334 min
Delta R.T.: 0.000 min
Response: 173296175
Conc: 48.60 ng/ml




#20 Methoxychlor

R.T.: 7.502 min
Delta R.T.: 0.004 min
Response: 51975133
Conc: 49.81 ng/ml



#20 Methoxychlor

R.T.: 6.610 min
Delta R.T.: 0.000 min
Response: 92217194
Conc: 51.57 ng/ml




#21 Endrin ketone

R.T.: 7.646 min
Delta R.T.: 0.004 min
Response: 115075104
Conc: 45.62 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB166413BS


**Manual Integrations
APPROVED**

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#21 Endrin ketone

R.T.: 6.839 min
Delta R.T.: 0.000 min
Response: 197623235
Conc: 47.11 ng/ml




#22 Mirex

R.T.: 8.119 min
Delta R.T.: 0.004 min
Response: 87775989
Conc: 42.15 ng/ml



#22 Mirex

R.T.: 7.020 min
Delta R.T.: 0.002 min
Response: 147863707
Conc: 43.72 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.059 min
Delta R.T.: 0.006 min
Response: 44293906
Conc: 21.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB166413BS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: 0.002 min
Response: 76038120
Conc: 21.70 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	357MS	SDG No.:	Q1232
Lab Sample ID:	Q1239-10MS	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	84.8 Decanted:
Sample Wt/Vol:	30.04 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093960.D	1	01/31/25 08:15	01/31/25 20:31	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	19.6		0.21	2.00	ug/kg
319-85-7	beta-BHC	19.9		0.58	2.00	ug/kg
319-86-8	delta-BHC	19.1		0.55	2.00	ug/kg
58-89-9	gamma-BHC (Lindane)	19.2		0.22	2.00	ug/kg
76-44-8	Heptachlor	19.1		0.20	2.00	ug/kg
309-00-2	Aldrin	19.3		0.16	2.00	ug/kg
1024-57-3	Heptachlor epoxide	20.0		0.27	2.00	ug/kg
959-98-8	Endosulfan I	19.0		0.20	2.00	ug/kg
60-57-1	Dieldrin	20.5		0.18	2.00	ug/kg
72-55-9	4,4-DDE	21.4		0.15	2.00	ug/kg
72-20-8	Endrin	20.2		0.19	2.00	ug/kg
33213-65-9	Endosulfan II	19.5		0.35	2.00	ug/kg
72-54-8	4,4-DDD	22.0		0.22	2.00	ug/kg
1031-07-8	Endosulfan Sulfate	18.4		0.15	2.00	ug/kg
50-29-3	4,4-DDT	18.9		0.20	2.00	ug/kg
72-43-5	Methoxychlor	17.5		0.45	2.00	ug/kg
53494-70-5	Endrin ketone	17.5		0.26	2.00	ug/kg
7421-93-4	Endrin aldehyde	13.7		0.46	2.00	ug/kg
5103-71-9	alpha-Chlordane	20.4		0.20	2.00	ug/kg
5103-74-2	gamma-Chlordane	21.0		0.22	2.00	ug/kg
8001-35-2	Toxaphene	38.9	U	6.10	38.9	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	14.9		10 - 148	75%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.1		10 - 159	95%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	357MS	SDG No.:	Q1232
Lab Sample ID:	Q1239-10MS	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	84.8
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093960.D	1	01/31/25 08:15	01/31/25 20:31	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093960.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 20:31
 Operator : AR\AJ
 Sample : Q1239-10MS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 357MS

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 02/03/2025
 Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:33:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.536	2.774	51414596	58832338	19.094m	18.024
28) SA Decachlor...	9.057	7.911	31279196	48142245	14.952	13.739
Target Compounds						
2) A alpha-BHC	3.994	3.276	191.6E6	239.2E6	49.969	48.930
3) MA gamma-BHC...	4.327	3.606	180.4E6	228.4E6	48.988	48.175
4) MA Heptachlor	4.915	3.945	159.2E6	218.3E6	48.569	46.904
5) MB Aldrin	5.257	4.224	161.0E6	224.1E6	49.201	49.116
6) B beta-BHC	4.525	3.906	81072574	101.5E6	50.439	50.802
7) B delta-BHC	4.772	4.135	170.9E6	228.3E6	48.751	48.042
8) B Heptachlo...	5.684	4.727	140.7E6	212.8E6	47.306	50.902
9) A Endosulfan I	6.069	5.097	128.2E6	186.2E6	48.520	48.027
10) B gamma-Chl...	5.940	4.977	143.7E6	226.2E6	51.550	53.387
11) B alpha-Chl...	6.019	5.041	140.8E6	218.0E6	50.489	52.061
12) B 4,4'-DDE	6.192	5.230	123.7E6	218.4E6	50.794	54.462
13) MA Dieldrin	6.345	5.361	135.5E6	224.8E6	48.827	52.339
14) MA Endrin	6.574	5.637	111.4E6	190.4E6	47.496	51.554
15) B Endosulfa...	6.794	5.932	114.6E6	184.5E6	47.558	49.804
16) A 4,4'-DDD	6.710	5.785	106.5E6	171.9E6	56.051	54.443
17) MA 4,4'-DDT	7.024	6.035	94770168	155.6E6	48.057	47.825
18) B Endrin al...	6.924	6.111	67631389	102.3E6	34.789	33.595
19) B Endosulfa...	7.159	6.334	105.0E6	167.0E6	46.401	46.818
20) A Methoxychlor	7.500	6.610	46323353	79508850	44.397m	44.465
21) B Endrin ke...	7.645	6.839	112.2E6	187.1E6	44.489	44.597
22) Mirex	8.119	7.019	88253104	150.1E6	42.379	44.370

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093960.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 20:31
Operator : AR\AJ
Sample : Q1239-10MS
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

357MS

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:33:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.536 min
Delta R.T.: -0.003 min
Response: 51414596
Conc: 19.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 58832338
Conc: 18.02 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 191573060
Conc: 49.97 ng/ml



#2 alpha-BHC

R.T.: 3.276 min
Delta R.T.: 0.000 min
Response: 239216365
Conc: 48.93 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.327 min
Delta R.T.: 0.000 min
Response: 180414145
Conc: 48.99 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#3 gamma-BHC (Lindane)

R.T.: 3.606 min
Delta R.T.: 0.000 min
Response: 228406050
Conc: 48.17 ng/ml




#4 Heptachlor

R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 159176402
Conc: 48.57 ng/ml



#4 Heptachlor

R.T.: 3.945 min
Delta R.T.: 0.000 min
Response: 218328245
Conc: 46.90 ng/ml




#5 Aldrin

R.T.: 5.257 min
Delta R.T.: 0.001 min
Response: 160983220
Conc: 49.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#5 Aldrin

R.T.: 4.224 min
Delta R.T.: 0.000 min
Response: 224056312
Conc: 49.12 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 81072574
Conc: 50.44 ng/ml



#6 beta-BHC

R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 101474057
Conc: 50.80 ng/ml




#7 delta-BHC

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 170885702
Conc: 48.75 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#7 delta-BHC

R.T.: 4.135 min
Delta R.T.: 0.000 min
Response: 228261078
Conc: 48.04 ng/ml




#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.001 min
Response: 140679215
Conc: 47.31 ng/ml



#8 Heptachlor epoxide

R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 212784536
Conc: 50.90 ng/ml




#9 Endosulfan I

R.T.: 6.069 min
Delta R.T.: 0.002 min
Response: 128231596
Conc: 48.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#9 Endosulfan I

R.T.: 5.097 min
Delta R.T.: 0.000 min
Response: 186197238
Conc: 48.03 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.001 min
Response: 143688480
Conc: 51.55 ng/ml



#10 gamma-Chlordane

R.T.: 4.977 min
Delta R.T.: 0.000 min
Response: 226233417
Conc: 53.39 ng/ml




#11 alpha-Chlordane

R.T.: 6.019 min
Delta R.T.: 0.002 min
Response: 140783127
Conc: 50.49 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#11 alpha-Chlordane

R.T.: 5.041 min
Delta R.T.: 0.000 min
Response: 217957882
Conc: 52.06 ng/ml




#12 4,4'-DDE

R.T.: 6.192 min
Delta R.T.: 0.000 min
Response: 123663019
Conc: 50.79 ng/ml



#12 4,4'-DDE

R.T.: 5.230 min
Delta R.T.: 0.000 min
Response: 218364523
Conc: 54.46 ng/ml




#13 Dieldrin

R.T.: 6.345 min
Delta R.T.: 0.002 min
Response: 135537422
Conc: 48.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#13 Dieldrin

R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 224830219
Conc: 52.34 ng/ml




#14 Endrin

R.T.: 6.574 min
Delta R.T.: 0.002 min
Response: 111370119
Conc: 47.50 ng/ml



#14 Endrin

R.T.: 5.637 min
Delta R.T.: 0.000 min
Response: 190373930
Conc: 51.55 ng/ml




#15 Endosulfan II

R.T.: 6.794 min
Delta R.T.: 0.002 min
Response: 114583644
Conc: 47.56 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#15 Endosulfan II

R.T.: 5.932 min
Delta R.T.: 0.000 min
Response: 184464053
Conc: 49.80 ng/ml




#16 4,4'-DDD

R.T.: 6.710 min
Delta R.T.: 0.002 min
Response: 106527786
Conc: 56.05 ng/ml



#16 4,4'-DDD

R.T.: 5.785 min
Delta R.T.: 0.000 min
Response: 171850479
Conc: 54.44 ng/ml




#17 4,4'-DDT

R.T.: 7.024 min
Delta R.T.: 0.002 min
Response: 94770168
Conc: 48.06 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#17 4,4'-DDT

R.T.: 6.035 min
Delta R.T.: 0.000 min
Response: 155624195
Conc: 47.82 ng/ml




#18 Endrin aldehyde

R.T.: 6.924 min
Delta R.T.: 0.002 min
Response: 67631389
Conc: 34.79 ng/ml



#18 Endrin aldehyde

R.T.: 6.111 min
Delta R.T.: 0.001 min
Response: 102285489
Conc: 33.60 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.159 min
Delta R.T.: 0.002 min
Response: 105040812
Conc: 46.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#19 Endosulfan Sulfate

R.T.: 6.334 min
Delta R.T.: 0.001 min
Response: 166956037
Conc: 46.82 ng/ml




#20 Methoxychlor

R.T.: 7.500 min
Delta R.T.: 0.002 min
Response: 46323353
Conc: 44.40 ng/ml m



#20 Methoxychlor

R.T.: 6.610 min
Delta R.T.: 0.001 min
Response: 79508850
Conc: 44.46 ng/ml




#21 Endrin ketone

R.T.: 7.645 min
Delta R.T.: 0.003 min
Response: 112229195
Conc: 44.49 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#21 Endrin ketone

R.T.: 6.839 min
Delta R.T.: 0.000 min
Response: 187096384
Conc: 44.60 ng/ml




#22 Mirex

R.T.: 8.119 min
Delta R.T.: 0.004 min
Response: 88253104
Conc: 42.38 ng/ml



#22 Mirex

R.T.: 7.019 min
Delta R.T.: 0.002 min
Response: 150055609
Conc: 44.37 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.057 min
Delta R.T.: 0.004 min
Response: 31279196
Conc: 14.95 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: 0.001 min
Response: 48142245
Conc: 13.74 ng/ml

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	357MSD	SDG No.:	Q1232
Lab Sample ID:	Q1239-10MSD	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	84.8 Decanted:
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093961.D	1	01/31/25 08:15	01/31/25 20:45	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	19.8		0.21	2.00	ug/kg
319-85-7	beta-BHC	20.1		0.58	2.00	ug/kg
319-86-8	delta-BHC	20.0		0.55	2.00	ug/kg
58-89-9	gamma-BHC (Lindane)	19.4		0.22	2.00	ug/kg
76-44-8	Heptachlor	20.3		0.20	2.00	ug/kg
309-00-2	Aldrin	19.5		0.17	2.00	ug/kg
1024-57-3	Heptachlor epoxide	20.0		0.27	2.00	ug/kg
959-98-8	Endosulfan I	19.5		0.20	2.00	ug/kg
60-57-1	Dieldrin	20.6		0.18	2.00	ug/kg
72-55-9	4,4-DDE	21.4		0.15	2.00	ug/kg
72-20-8	Endrin	20.9		0.19	2.00	ug/kg
33213-65-9	Endosulfan II	20.1		0.35	2.00	ug/kg
72-54-8	4,4-DDD	22.2		0.22	2.00	ug/kg
1031-07-8	Endosulfan Sulfate	19.7		0.15	2.00	ug/kg
50-29-3	4,4-DDT	20.5		0.20	2.00	ug/kg
72-43-5	Methoxychlor	19.8		0.45	2.00	ug/kg
53494-70-5	Endrin ketone	18.8		0.26	2.00	ug/kg
7421-93-4	Endrin aldehyde	18.7		0.46	2.00	ug/kg
5103-71-9	alpha-Chlordane	20.6		0.20	2.00	ug/kg
5103-74-2	gamma-Chlordane	20.9		0.22	2.00	ug/kg
8001-35-2	Toxaphene	38.9	U	6.20	38.9	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	15.4		10 - 148	77%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.7		10 - 159	93%	SPK: 20

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/30/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/30/25
Client Sample ID:	357MSD	SDG No.:	Q1232
Lab Sample ID:	Q1239-10MSD	Matrix:	SOIL
Analytical Method:	SW8081	% Solid:	84.8 Decanted:
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093961.D	1	01/31/25 08:15	01/31/25 20:45	PB166413

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
 Data File : PL093961.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2025 20:45
 Operator : AR\AJ
 Sample : Q1239-10MSD
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 357MSD

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 02/03/2025
 Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 00:34:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.536	2.773	50300528	58521224	18.680m	17.928
28) SA Decachlor...	9.056	7.911	32207548	47799485	15.396	13.641
Target Compounds						
2) A alpha-BHC	3.993	3.276	193.1E6	241.4E6	50.362	49.377
3) MA gamma-BHC...	4.326	3.606	181.6E6	231.1E6	49.311	48.750
4) MA Heptachlor	4.914	3.944	169.1E6	232.1E6	51.608	49.869
5) MB Aldrin	5.256	4.223	162.2E6	223.9E6	49.585	49.085
6) B beta-BHC	4.525	3.906	82285429	101.8E6	51.194	50.971
7) B delta-BHC	4.772	4.134	178.5E6	234.6E6	50.937	49.368
8) B Heptachlo...	5.683	4.726	142.9E6	212.5E6	48.061	50.825
9) A Endosulfan I	6.069	5.096	131.1E6	190.7E6	49.609	49.182
10) B gamma-Chl...	5.939	4.975	142.5E6	225.5E6	51.108	53.213m
11) B alpha-Chl...	6.018	5.040	142.0E6	219.5E6	50.916	52.436
12) B 4,4'-DDE	6.192	5.229	124.0E6	218.9E6	50.915	54.583
13) MA Dieldrin	6.344	5.360	135.7E6	225.5E6	48.886	52.484
14) MA Endrin	6.572	5.635	113.5E6	196.2E6	48.400m	53.138m
15) B Endosulfa...	6.794	5.931	118.0E6	189.3E6	48.982	51.113
16) A 4,4'-DDD	6.708	5.784	107.3E6	173.5E6	56.444m	54.951
17) MA 4,4'-DDT	7.023	6.034	102.7E6	169.1E6	52.056	51.970
18) B Endrin al...	6.924	6.110	92653355	144.8E6	47.660	47.566
19) B Endosulfa...	7.159	6.333	109.0E6	178.8E6	48.131	50.138
20) A Methoxychlor	7.500	6.609	52631768	89126412	50.443	49.843
21) B Endrin ke...	7.644	6.838	121.0E6	200.4E6	47.957	47.780
22) Mirex	8.117	7.019	88327251	149.6E6	42.414	44.228

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013125\
Data File : PL093961.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2025 20:45
Operator : AR\AJ
Sample : Q1239-10MSD
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

357MSD

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 02/03/2025

Supervised By :Ankita Jodhani 02/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 00:34:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.536 min
Delta R.T.: -0.003 min
Response: 50300528
Conc: 18.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: -0.001 min
Response: 58521224
Conc: 17.93 ng/ml




#2 alpha-BHC

R.T.: 3.993 min
Delta R.T.: -0.001 min
Response: 193081300
Conc: 50.36 ng/ml



#2 alpha-BHC

R.T.: 3.276 min
Delta R.T.: -0.001 min
Response: 241401784
Conc: 49.38 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.326 min
Delta R.T.: 0.000 min
Response: 181606158
Conc: 49.31 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#3 gamma-BHC (Lindane)

R.T.: 3.606 min
Delta R.T.: -0.001 min
Response: 231135430
Conc: 48.75 ng/ml




#4 Heptachlor

R.T.: 4.914 min
Delta R.T.: 0.000 min
Response: 169137974
Conc: 51.61 ng/ml



#4 Heptachlor

R.T.: 3.944 min
Delta R.T.: -0.001 min
Response: 232129771
Conc: 49.87 ng/ml




#5 Aldrin

R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 162239210
Conc: 49.58 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#5 Aldrin

R.T.: 4.223 min
Delta R.T.: -0.001 min
Response: 223914541
Conc: 49.08 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 82285429
Conc: 51.19 ng/ml



#6 beta-BHC

R.T.: 3.906 min
Delta R.T.: -0.001 min
Response: 101811998
Conc: 50.97 ng/ml




#7 delta-BHC

R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 178547228
Conc: 50.94 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#7 delta-BHC

R.T.: 4.134 min
Delta R.T.: -0.001 min
Response: 234560895
Conc: 49.37 ng/ml




#8 Heptachlor epoxide

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 142925303
Conc: 48.06 ng/ml



#8 Heptachlor epoxide

R.T.: 4.726 min
Delta R.T.: 0.000 min
Response: 212460541
Conc: 50.82 ng/ml




#9 Endosulfan I

R.T.: 6.069 min
Delta R.T.: 0.002 min
Response: 131110254
Conc: 49.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#9 Endosulfan I

R.T.: 5.096 min
Delta R.T.: 0.000 min
Response: 190673728
Conc: 49.18 ng/ml




#10 gamma-Chlordane

R.T.: 5.939 min
Delta R.T.: 0.001 min
Response: 142457749
Conc: 51.11 ng/ml



#10 gamma-Chlordane

R.T.: 4.975 min
Delta R.T.: -0.002 min
Response: 225497434
Conc: 53.21 ng/ml m




#11 alpha-Chlordane

R.T.: 6.018 min
Delta R.T.: 0.001 min
Response: 141973735
Conc: 50.92 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#11 alpha-Chlordane

R.T.: 5.040 min
Delta R.T.: 0.000 min
Response: 219526442
Conc: 52.44 ng/ml




#12 4,4'-DDE

R.T.: 6.192 min
Delta R.T.: 0.001 min
Response: 123956358
Conc: 50.91 ng/ml



#12 4,4'-DDE

R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 218850428
Conc: 54.58 ng/ml




#13 Dieldrin

R.T.: 6.344 min
Delta R.T.: 0.002 min
Response: 135700909
Conc: 48.89 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#13 Dieldrin

R.T.: 5.360 min
Delta R.T.: 0.000 min
Response: 225450604
Conc: 52.48 ng/ml




#14 Endrin

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 113489417
Conc: 48.40 ng/ml m



#14 Endrin

R.T.: 5.635 min
Delta R.T.: -0.002 min
Response: 196220745
Conc: 53.14 ng/ml m




#15 Endosulfan II

R.T.: 6.794 min
Delta R.T.: 0.002 min
Response: 118015882
Conc: 48.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#15 Endosulfan II

R.T.: 5.931 min
Delta R.T.: 0.000 min
Response: 189315492
Conc: 51.11 ng/ml




#16 4,4'-DDD

R.T.: 6.708 min
Delta R.T.: 0.000 min
Response: 107274421
Conc: 56.44 ng/ml m



#16 4,4'-DDD

R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 173455259
Conc: 54.95 ng/ml




#17 4,4'-DDT

R.T.: 7.023 min
Delta R.T.: 0.000 min
Response: 102656848
Conc: 52.06 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025



#17 4,4'-DDT


R.T.: 6.034 min
Delta R.T.: 0.000 min
Response: 169112807
Conc: 51.97 ng/ml

#18 Endrin aldehyde

R.T.: 6.924 min
Delta R.T.: 0.001 min
Response: 92653355
Conc: 47.66 ng/ml

#18 Endrin aldehyde

R.T.: 6.110 min
Delta R.T.: 0.000 min
Response: 144821876
Conc: 47.57 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.159 min
Delta R.T.: 0.001 min
Response: 108956035
Conc: 48.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#19 Endosulfan Sulfate

R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 178795532
Conc: 50.14 ng/ml




#20 Methoxychlor

R.T.: 7.500 min
Delta R.T.: 0.002 min
Response: 52631768
Conc: 50.44 ng/ml



#20 Methoxychlor

R.T.: 6.609 min
Delta R.T.: 0.000 min
Response: 89126412
Conc: 49.84 ng/ml




#21 Endrin ketone

R.T.: 7.644 min
Delta R.T.: 0.002 min
Response: 120977141
Conc: 47.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025




#21 Endrin ketone

R.T.: 6.838 min
Delta R.T.: 0.000 min
Response: 200448431
Conc: 47.78 ng/ml




#22 Mirex

R.T.: 8.117 min
Delta R.T.: 0.003 min
Response: 88327251
Conc: 42.41 ng/ml



#22 Mirex

R.T.: 7.019 min
Delta R.T.: 0.000 min
Response: 149573926
Conc: 44.23 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: 0.003 min
Response: 32207548
Conc: 15.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
357MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: 0.001 min
Response: 47799485
Conc: 13.64 ng/ml

Manual Integration Report

Sequence:	PL012125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL093726.D	4,4"-DDD	Abdul	1/22/2025 8:17:27 AM	Ankita	1/22/2025 8:30:27	Peak Integrated by Software
PEM	PL093726.D	4,4"-DDE	Abdul	1/22/2025 8:17:27 AM	Ankita	1/22/2025 8:30:27	Peak Integrated by Software
PEM	PL093726.D	Endrin aldehyde	Abdul	1/22/2025 8:17:27 AM	Ankita	1/22/2025 8:30:27	Peak Integrated by Software
PCHLORICV500	PL093744.D	Chlordane-2	Abdul	1/22/2025 8:17:34 AM	Ankita	1/22/2025 8:30:28	Peak Integrated by Software
PCHLORICV500	PL093744.D	Chlordane-5	Abdul	1/22/2025 8:17:34 AM	Ankita	1/22/2025 8:30:28	Peak Integrated by Software
PCHLORICV500	PL093744.D	Chlordane-5 #2	Abdul	1/22/2025 8:17:34 AM	Ankita	1/22/2025 8:30:28	Peak Integrated by Software
PEM	PL093747.D	4,4"-DDE	Abdul	1/22/2025 8:17:08 AM	Ankita	1/22/2025 8:30:30	Peak Integrated by Software
PEM	PL093747.D	4,4"-DDE #2	Abdul	1/22/2025 8:17:08 AM	Ankita	1/22/2025 8:30:30	Peak Integrated by Software
PSTDCCC050	PL093748.D	Heptachlor	Abdul	1/22/2025 8:17:12 AM	Ankita	1/22/2025 8:30:32	Peak Integrated by Software

Manual Integration Report

Sequence:	pl013125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL093929.D	4,4"-DDD	Abdul	2/3/2025 9:48:09 AM	Ankita	2/3/2025 1:12:24	Peak Integrated by Software
PEM	PL093929.D	4,4"-DDD #2	Abdul	2/3/2025 9:48:09 AM	Ankita	2/3/2025 1:12:24	Peak Integrated by Software
PEM	PL093929.D	Endrin	Abdul	2/3/2025 9:48:09 AM	Ankita	2/3/2025 1:12:24	Peak Integrated by Software
PSTDCCC050	PL093930.D	Endrin	Abdul	2/3/2025 9:48:14 AM	Ankita	2/3/2025 1:12:25	Peak Integrated by Software
PCHLORCCC500	PL093931.D	Chlordane-2	Abdul	2/3/2025 9:48:17 AM	Ankita	2/3/2025 1:12:27	Peak Integrated by Software
PCHLORCCC500	PL093931.D	Chlordane-5	Abdul	2/3/2025 9:48:17 AM	Ankita	2/3/2025 1:12:27	Peak Integrated by Software
PCHLORCCC500	PL093931.D	Chlordane-5 #2	Abdul	2/3/2025 9:48:17 AM	Ankita	2/3/2025 1:12:27	Peak Integrated by Software
PB166413BS	PL093941.D	Endrin	Abdul	2/3/2025 9:48:34 AM	Ankita	2/3/2025 1:12:35	Peak Integrated by Software
PB166413BS	PL093941.D	gamma-Chlordane	Abdul	2/3/2025 9:48:34 AM	Ankita	2/3/2025 1:12:35	Peak Integrated by Software
PB166413BS	PL093941.D	gamma-Chlordane #2	Abdul	2/3/2025 9:48:34 AM	Ankita	2/3/2025 1:12:35	Peak Integrated by Software
PSTDCCC050	PL093943.D	Endrin	Abdul	2/3/2025 9:48:38 AM	Ankita	2/3/2025 1:12:36	Peak Integrated by Software
PCHLORCCC500	PL093944.D	Chlordane-2	Abdul	2/3/2025 9:48:41 AM	Ankita	2/3/2025 1:12:38	Peak Integrated by Software
PCHLORCCC500	PL093944.D	Chlordane-5	Abdul	2/3/2025 9:48:41 AM	Ankita	2/3/2025 1:12:38	Peak Integrated by Software

Manual Integration Report

Sequence:	pl013125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PCHLORCCC500	PL093944.D	Chlordane-5 #2	Abdul	2/3/2025 9:48:41 AM	Ankita	2/3/2025 1:12:38	Peak Integrated by Software
PTOXCCC500	PL093945.D	Toxaphene-2	Abdul	2/3/2025 9:48:44 AM	Ankita	2/3/2025 1:12:40	Peak Integrated by Software
PTOXCCC500	PL093945.D	Toxaphene-5 #2	Abdul	2/3/2025 9:48:44 AM	Ankita	2/3/2025 1:12:40	Peak Integrated by Software
Q1232-03	PL093946.D	4,4"-DDD #2	Abdul	2/3/2025 9:48:47 AM	Ankita	2/3/2025 1:12:41	Peak Integrated by Software
Q1232-03	PL093946.D	Decachlorobiphenyl	Abdul	2/3/2025 9:48:47 AM	Ankita	2/3/2025 1:12:41	Peak Integrated by Software
Q1232-03	PL093946.D	Decachlorobiphenyl #2	Abdul	2/3/2025 9:48:47 AM	Ankita	2/3/2025 1:12:41	Peak Integrated by Software
Q1232-03	PL093946.D	Endosulfan II	Abdul	2/3/2025 9:48:47 AM	Ankita	2/3/2025 1:12:41	Peak Integrated by Software
Q1232-03	PL093946.D	Endosulfan II #2	Abdul	2/3/2025 9:48:47 AM	Ankita	2/3/2025 1:12:41	Peak Integrated by Software
Q1232-03	PL093946.D	Tetrachloro-m-xylene	Abdul	2/3/2025 9:48:47 AM	Ankita	2/3/2025 1:12:41	Peak Integrated by Software
Q1232-07	PL093947.D	4,4"-DDE	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	4,4"-DDE #2	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	4,4"-DDT	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	4,4"-DDT #2	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software

Manual Integration Report

Sequence:	pl013125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q1232-07	PL093947.D	alpha-Chlordane	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	alpha-Chlordane #2	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	Decachlorobiphenyl #2	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	delta-BHC	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	delta-BHC #2	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	Endrin	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	Endrin #2	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	Heptachlor #2	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	Heptachlor epoxide	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	Heptachlor epoxide #2	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-07	PL093947.D	Tetrachloro-m-xylene	Abdul	2/3/2025 9:48:50 AM	Ankita	2/3/2025 1:12:43	Peak Integrated by Software
Q1232-11	PL093948.D	4,4"-DDT	Abdul	2/3/2025 9:48:54 AM	Ankita	2/3/2025 1:12:45	Peak Integrated by Software
Q1232-11	PL093948.D	alpha-Chlordane	Abdul	2/3/2025 9:48:54 AM	Ankita	2/3/2025 1:12:45	Peak Integrated by Software

Manual Integration Report

Sequence:	pl013125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q1232-11	PL093948.D	alpha-Chlordane #2	Abdul	2/3/2025 9:48:54 AM	Ankita	2/3/2025 1:12:45	Peak Integrated by Software
Q1232-11	PL093948.D	Decachlorobiphenyl	Abdul	2/3/2025 9:48:54 AM	Ankita	2/3/2025 1:12:45	Peak Integrated by Software
Q1232-11	PL093948.D	Decachlorobiphenyl #2	Abdul	2/3/2025 9:48:54 AM	Ankita	2/3/2025 1:12:45	Peak Integrated by Software
Q1232-11	PL093948.D	Endrin	Abdul	2/3/2025 9:48:54 AM	Ankita	2/3/2025 1:12:45	Peak Integrated by Software
Q1232-11	PL093948.D	Endrin #2	Abdul	2/3/2025 9:48:54 AM	Ankita	2/3/2025 1:12:45	Peak Integrated by Software
Q1232-11	PL093948.D	Heptachlor epoxide	Abdul	2/3/2025 9:48:54 AM	Ankita	2/3/2025 1:12:45	Peak Integrated by Software
Q1232-11	PL093948.D	Heptachlor epoxide #2	Abdul	2/3/2025 9:48:54 AM	Ankita	2/3/2025 1:12:45	Peak Integrated by Software
Q1232-11	PL093948.D	Tetrachloro-m-xylene	Abdul	2/3/2025 9:48:54 AM	Ankita	2/3/2025 1:12:45	Peak Integrated by Software
Q1232-11	PL093948.D	Tetrachloro-m-xylene #2	Abdul	2/3/2025 9:48:54 AM	Ankita	2/3/2025 1:12:45	Peak Integrated by Software
Q1232-15	PL093949.D	4,4"-DDT #2	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software
Q1232-15	PL093949.D	alpha-Chlordane	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software
Q1232-15	PL093949.D	alpha-Chlordane #2	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software
Q1232-15	PL093949.D	Decachlorobiphenyl	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software

Manual Integration Report

Sequence:	pl013125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q1232-15	PL093949.D	Endrin	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software
Q1232-15	PL093949.D	Endrin #2	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software
Q1232-15	PL093949.D	gamma-Chlordane	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software
Q1232-15	PL093949.D	gamma-Chlordane #2	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software
Q1232-15	PL093949.D	Heptachlor #2	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software
Q1232-15	PL093949.D	Heptachlor epoxide	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software
Q1232-15	PL093949.D	Heptachlor epoxide #2	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software
Q1232-15	PL093949.D	Tetrachloro-m-xylene	Abdul	2/3/2025 9:48:58 AM	Ankita	2/3/2025 1:12:47	Peak Integrated by Software
Q1232-19	PL093950.D	4,4"-DDD	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	4,4"-DDD #2	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	4,4"-DDT	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	4,4"-DDT #2	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	alpha-Chlordane	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software

Manual Integration Report

Sequence:	pl013125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q1232-19	PL093950.D	alpha-Chlordane #2	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	delta-BHC	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	delta-BHC #2	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	Endrin	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	Endrin #2	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	Heptachlor	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	Heptachlor #2	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	Heptachlor epoxide	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
Q1232-19	PL093950.D	Heptachlor epoxide #2	Abdul	2/3/2025 9:49:02 AM	Ankita	2/3/2025 1:12:48	Peak Integrated by Software
PEM	PL093958.D	4,4"-DDE	Abdul	2/3/2025 9:49:29 AM	Ankita	2/3/2025 1:13:27	Peak Integrated by Software
PEM	PL093958.D	4,4"-DDE #2	Abdul	2/3/2025 9:49:29 AM	Ankita	2/3/2025 1:13:27	Peak Integrated by Software
PEM	PL093958.D	Endrin	Abdul	2/3/2025 9:49:29 AM	Ankita	2/3/2025 1:13:27	Peak Integrated by Software
PSTDCCC050	PL093959.D	Endrin	Abdul	2/3/2025 9:54:37 AM	Ankita	2/3/2025 1:13:28	Peak Integrated by Software

Manual Integration Report

Sequence:	pl013125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PL093959.D	gamma-Chlordane #2	Abdul	2/3/2025 9:54:37 AM	Ankita	2/3/2025 1:13:28	Peak Integrated by Software
Q1239-10MS	PL093960.D	Methoxychlor	Abdul	2/3/2025 9:54:42 AM	Ankita	2/3/2025 1:13:30	Peak Integrated by Software
Q1239-10MS	PL093960.D	Tetrachloro-m-xylene	Abdul	2/3/2025 9:54:42 AM	Ankita	2/3/2025 1:13:30	Peak Integrated by Software
Q1239-10MSD	PL093961.D	4,4"-DDD	Abdul	2/3/2025 9:54:46 AM	Ankita	2/3/2025 1:13:32	Peak Integrated by Software
Q1239-10MSD	PL093961.D	Endrin	Abdul	2/3/2025 9:54:46 AM	Ankita	2/3/2025 1:13:32	Peak Integrated by Software
Q1239-10MSD	PL093961.D	Endrin #2	Abdul	2/3/2025 9:54:46 AM	Ankita	2/3/2025 1:13:32	Peak Integrated by Software
Q1239-10MSD	PL093961.D	gamma-Chlordane #2	Abdul	2/3/2025 9:54:46 AM	Ankita	2/3/2025 1:13:32	Peak Integrated by Software
Q1239-10MSD	PL093961.D	Tetrachloro-m-xylene	Abdul	2/3/2025 9:54:46 AM	Ankita	2/3/2025 1:13:32	Peak Integrated by Software
PSTDCCC050	PL093971.D	Endrin	Abdul	2/3/2025 9:49:56 AM	Ankita	2/3/2025 1:14:06	Peak Integrated by Software
PSTDCCC050	PL093978.D	Endrin	Abdul	2/3/2025 9:54:09 AM	Ankita	2/3/2025 1:14:16	Peak Integrated by Software

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL012125

Review By	Abdul	Review On	1/22/2025 8:17:54 AM
Supervise By	Ankita	Supervise On	1/22/2025 8:30:39 AM
SubDirectory	PL012125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL093724.D	21 Jan 2025 10:03	AR\AJ	Ok
2	I.BLK	PL093725.D	21 Jan 2025 10:16	AR\AJ	Ok
3	PEM	PL093726.D	21 Jan 2025 10:30	AR\AJ	Ok,M
4	RESCHK	PL093727.D	21 Jan 2025 10:43	AR\AJ	Ok
5	PSTDICC100	PL093728.D	21 Jan 2025 10:57	AR\AJ	Ok
6	PSTDICC075	PL093729.D	21 Jan 2025 11:10	AR\AJ	Ok
7	PSTDICC050	PL093730.D	21 Jan 2025 11:24	AR\AJ	Ok
8	PSTDICC025	PL093731.D	21 Jan 2025 11:38	AR\AJ	Ok
9	PSTDICC005	PL093732.D	21 Jan 2025 11:51	AR\AJ	Ok
10	PCHLORICC1000	PL093733.D	21 Jan 2025 12:05	AR\AJ	Ok
11	PCHLORICC750	PL093734.D	21 Jan 2025 12:18	AR\AJ	Ok
12	PCHLORICC500	PL093735.D	21 Jan 2025 12:32	AR\AJ	Ok
13	PCHLORICC250	PL093736.D	21 Jan 2025 12:45	AR\AJ	Ok
14	PCHLORICC050	PL093737.D	21 Jan 2025 12:59	AR\AJ	Ok
15	PTOXICC1000	PL093738.D	21 Jan 2025 13:12	AR\AJ	Ok
16	PTOXICC750	PL093739.D	21 Jan 2025 13:26	AR\AJ	Ok
17	PTOXICC500	PL093740.D	21 Jan 2025 13:39	AR\AJ	Ok
18	PTOXICC250	PL093741.D	21 Jan 2025 13:53	AR\AJ	Ok
19	PTOXICC100	PL093742.D	21 Jan 2025 14:07	AR\AJ	Ok
20	PSTDICV050	PL093743.D	21 Jan 2025 14:20	AR\AJ	Ok
21	PCHLORICV500	PL093744.D	21 Jan 2025 14:47	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL012125

Review By	Abdul	Review On	1/22/2025 8:17:54 AM
Supervise By	Ankita	Supervise On	1/22/2025 8:30:39 AM
SubDirectory	PL012125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PL093745.D	21 Jan 2025 15:14	AR\AJ	Ok
23	I.BLK	PL093746.D	21 Jan 2025 15:41	AR\AJ	Ok
24	PEM	PL093747.D	21 Jan 2025 15:54	AR\AJ	Ok,M
25	PSTDCCC050	PL093748.D	21 Jan 2025 16:07	AR\AJ	Ok,M
26	Q1093-01RE	PL093749.D	21 Jan 2025 16:21	AR\AJ	Confirms
27	I.BLK	PL093750.D	21 Jan 2025 16:34	AR\AJ	Ok
28	PSTDCCC050	PL093751.D	21 Jan 2025 16:48	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL013125

Review By	Abdul	Review On	2/3/2025 9:55:31 AM
Supervise By	Ankita	Supervise On	2/3/2025 1:14:23 PM
SubDirectory	PL013125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL093927.D	31 Jan 2025 10:38	AR\AJ	Ok
2	I.BLK	PL093928.D	31 Jan 2025 10:51	AR\AJ	Ok
3	PEM	PL093929.D	31 Jan 2025 11:04	AR\AJ	Ok,M
4	PSTDCCC050	PL093930.D	31 Jan 2025 11:17	AR\AJ	Ok,M
5	PCHLORCCC500	PL093931.D	31 Jan 2025 11:54	AR\AJ	Ok,M
6	PTOXCCC500	PL093932.D	31 Jan 2025 12:07	AR\AJ	Ok
7	PB166365BL	PL093933.D	31 Jan 2025 12:29	AR\AJ	Ok
8	PB166365BS	PL093934.D	31 Jan 2025 13:08	AR\AJ	Ok
9	PB166365BSD	PL093935.D	31 Jan 2025 13:22	AR\AJ	Ok,M
10	PB166365BS	PL093936.D	31 Jan 2025 13:35	AR\AJ	Ok,M
11	PB166365BS	PL093937.D	31 Jan 2025 13:49	AR\AJ	Ok,M
12	Q1211-01	PL093938.D	31 Jan 2025 14:02	AR\AJ	Ok
13	Q1211-02	PL093939.D	31 Jan 2025 14:15	AR\AJ	Ok,M
14	PB166413BL	PL093940.D	31 Jan 2025 14:28	AR\AJ	Ok
15	PB166413BS	PL093941.D	31 Jan 2025 14:43	AR\AJ	Ok,M
16	I.BLK	PL093942.D	31 Jan 2025 14:57	AR\AJ	Ok
17	PSTDCCC050	PL093943.D	31 Jan 2025 15:10	AR\AJ	Ok,M
18	PCHLORCCC500	PL093944.D	31 Jan 2025 15:23	AR\AJ	Ok,M
19	PTOXCCC500	PL093945.D	31 Jan 2025 16:34	AR\AJ	Ok,M
20	Q1232-03	PL093946.D	31 Jan 2025 17:01	AR\AJ	Ok,M
21	Q1232-07	PL093947.D	31 Jan 2025 17:14	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL013125

Review By	Abdul	Review On	2/3/2025 9:55:31 AM
Supervise By	Ankita	Supervise On	2/3/2025 1:14:23 PM
SubDirectory	PL013125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1232-11	PL093948.D	31 Jan 2025 17:27	AR\AJ	Ok,M
23	Q1232-15	PL093949.D	31 Jan 2025 17:40	AR\AJ	Ok,M
24	Q1232-19	PL093950.D	31 Jan 2025 17:53	AR\AJ	Ok,M
25	Q1235-03	PL093951.D	31 Jan 2025 18:07	AR\AJ	Ok,M
26	Q1235-07	PL093952.D	31 Jan 2025 18:20	AR\AJ	Ok,M
27	Q1239-01	PL093953.D	31 Jan 2025 18:33	AR\AJ	Ok,M
28	Q1239-04	PL093954.D	31 Jan 2025 18:46	AR\AJ	Ok,M
29	Q1239-07	PL093955.D	31 Jan 2025 18:59	AR\AJ	Ok,M
30	Q1239-10	PL093956.D	31 Jan 2025 19:12	AR\AJ	Ok,M
31	I.BLK	PL093957.D	31 Jan 2025 19:26	AR\AJ	Ok
32	PEM	PL093958.D	31 Jan 2025 19:39	AR\AJ	Ok,M
33	PSTDCCC050	PL093959.D	31 Jan 2025 20:18	AR\AJ	Ok,M
34	Q1239-10MS	PL093960.D	31 Jan 2025 20:31	AR\AJ	Ok,M
35	Q1239-10MSD	PL093961.D	31 Jan 2025 20:45	AR\AJ	Ok,M
36	Q1241-03	PL093962.D	31 Jan 2025 20:58	AR\AJ	Ok,M
37	Q1241-07	PL093963.D	31 Jan 2025 21:11	AR\AJ	Ok,M
38	Q1241-11	PL093964.D	31 Jan 2025 21:24	AR\AJ	Ok,M
39	Q1241-15	PL093965.D	31 Jan 2025 21:37	AR\AJ	Ok,M
40	Q1241-19	PL093966.D	31 Jan 2025 21:50	AR\AJ	Ok,M
41	Q1242-03	PL093967.D	31 Jan 2025 22:04	AR\AJ	Ok,M
42	Q1243-01	PL093968.D	31 Jan 2025 22:17	AR\AJ	Ok
43	Q1244-01	PL093969.D	31 Jan 2025 22:30	AR\AJ	Ok,M
44	I.BLK	PL093970.D	31 Jan 2025 22:43	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL013125

Review By	Abdul	Review On	2/3/2025 9:55:31 AM
Supervise By	Ankita	Supervise On	2/3/2025 1:14:23 PM
SubDirectory	PL013125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	PSTDCCC050	PL093971.D	01 Feb 2025 00:57	AR\AJ	Ok,M
46	Q1215-03	PL093972.D	01 Feb 2025 01:10	AR\AJ	Ok,M
47	Q1215-03MS	PL093973.D	01 Feb 2025 01:24	AR\AJ	Ok,M
48	Q1215-03MSD	PL093974.D	01 Feb 2025 01:37	AR\AJ	Ok,M
49	Q1216-15	PL093975.D	01 Feb 2025 01:50	AR\AJ	ReRun
50	Q1219-01	PL093976.D	01 Feb 2025 02:03	AR\AJ	Ok,M
51	I.BLK	PL093977.D	01 Feb 2025 02:43	AR\AJ	Ok
52	PSTDCCC050	PL093978.D	01 Feb 2025 02:56	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL012125

Review By	Abdul	Review On	1/22/2025 8:17:54 AM
Supervise By	Ankita	Supervise On	1/22/2025 8:30:39 AM
SubDirectory	PL012125	HP Acquire Method	HP Processing Method pl012125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL093724.D	21 Jan 2025 10:03		AR\AJ	Ok
2	I.BLK	I.BLK	PL093725.D	21 Jan 2025 10:16		AR\AJ	Ok
3	PEM	PEM	PL093726.D	21 Jan 2025 10:30		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL093727.D	21 Jan 2025 10:43		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PL093728.D	21 Jan 2025 10:57		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL093729.D	21 Jan 2025 11:10		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL093730.D	21 Jan 2025 11:24		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL093731.D	21 Jan 2025 11:38		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL093732.D	21 Jan 2025 11:51		AR\AJ	Ok
10	PCHLORICC1000	PCHLORICC1000	PL093733.D	21 Jan 2025 12:05		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL093734.D	21 Jan 2025 12:18		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL093735.D	21 Jan 2025 12:32		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL093736.D	21 Jan 2025 12:45		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL093737.D	21 Jan 2025 12:59		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL093738.D	21 Jan 2025 13:12		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL093739.D	21 Jan 2025 13:26		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL093740.D	21 Jan 2025 13:39		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL093741.D	21 Jan 2025 13:53		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL012125

Review By	Abdul	Review On	1/22/2025 8:17:54 AM
Supervise By	Ankita	Supervise On	1/22/2025 8:30:39 AM
SubDirectory	PL012125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL093742.D	21 Jan 2025 14:07		AR\AJ	Ok
20	PSTDICV050	ICVPL012125	PL093743.D	21 Jan 2025 14:20		AR\AJ	Ok
21	PCHLORICV500	ICVPL012125CHLOR	PL093744.D	21 Jan 2025 14:47		AR\AJ	Ok,M
22	PTOXICV500	ICVPL012125TOX	PL093745.D	21 Jan 2025 15:14		AR\AJ	Ok
23	I.BLK	I.BLK	PL093746.D	21 Jan 2025 15:41		AR\AJ	Ok
24	PEM	PEM	PL093747.D	21 Jan 2025 15:54		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PL093748.D	21 Jan 2025 16:07		AR\AJ	Ok,M
26	Q1093-01RE	RT-3407RE	PL093749.D	21 Jan 2025 16:21	F Flag coming , DCB high in 2nd column	AR\AJ	Confirms
27	I.BLK	I.BLK	PL093750.D	21 Jan 2025 16:34		AR\AJ	Ok
28	PSTDCCC050	PSTDCCC050	PL093751.D	21 Jan 2025 16:48		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL013125

Review By	Abdul	Review On	2/3/2025 9:55:31 AM
Supervise By	Ankita	Supervise On	2/3/2025 1:14:23 PM
SubDirectory	PL013125	HP Acquire Method	HP Processing Method pl012125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL093927.D	31 Jan 2025 10:38		AR\AJ	Ok
2	I.BLK	I.BLK	PL093928.D	31 Jan 2025 10:51		AR\AJ	Ok
3	PEM	PEM	PL093929.D	31 Jan 2025 11:04		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL093930.D	31 Jan 2025 11:17		AR\AJ	Ok,M
5	PCHLORCCC500	PCHLORCCC500	PL093931.D	31 Jan 2025 11:54		AR\AJ	Ok,M
6	PTOXCCC500	PTOXCCC500	PL093932.D	31 Jan 2025 12:07		AR\AJ	Ok
7	PB166365BL	PB166365BL	PL093933.D	31 Jan 2025 12:29		AR\AJ	Ok
8	PB166365BS	PB166365BS	PL093934.D	31 Jan 2025 13:08		AR\AJ	Ok
9	PB166365BSD	PB166365BSD	PL093935.D	31 Jan 2025 13:22		AR\AJ	Ok,M
10	PB166365BS	PB166365BS	PL093936.D	31 Jan 2025 13:35		AR\AJ	Ok,M
11	PB166365BS	PB166365BS	PL093937.D	31 Jan 2025 13:49		AR\AJ	Ok,M
12	Q1211-01	TAPHHA-MW01-01282	PL093938.D	31 Jan 2025 14:02	TCMX high in 1st column	AR\AJ	Ok
13	Q1211-02	TAPIAL2-MW03-01282	PL093939.D	31 Jan 2025 14:15		AR\AJ	Ok,M
14	PB166413BL	PB166413BL	PL093940.D	31 Jan 2025 14:28		AR\AJ	Ok
15	PB166413BS	PB166413BS	PL093941.D	31 Jan 2025 14:43		AR\AJ	Ok,M
16	I.BLK	I.BLK	PL093942.D	31 Jan 2025 14:57		AR\AJ	Ok
17	PSTDCCC050	PSTDCCC050	PL093943.D	31 Jan 2025 15:10		AR\AJ	Ok,M
18	PCHLORCCC500	PCHLORCCC500	PL093944.D	31 Jan 2025 15:23		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL013125

Review By	Abdul	Review On	2/3/2025 9:55:31 AM
Supervise By	Ankita	Supervise On	2/3/2025 1:14:23 PM
SubDirectory	PL013125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXC500	PTOXC500	PL093945.D	31 Jan 2025 16:34		AR\AJ	Ok,M
20	Q1232-03	JPP-46.2-012925	PL093946.D	31 Jan 2025 17:01	DCB high in 2nd column	AR\AJ	Ok,M
21	Q1232-07	JPP-46.1-012925	PL093947.D	31 Jan 2025 17:14		AR\AJ	Ok,M
22	Q1232-11	JPP-42.1-012925	PL093948.D	31 Jan 2025 17:27		AR\AJ	Ok,M
23	Q1232-15	JPP-42.2-012925	PL093949.D	31 Jan 2025 17:40		AR\AJ	Ok,M
24	Q1232-19	JPP-51.1-012925	PL093950.D	31 Jan 2025 17:53		AR\AJ	Ok,M
25	Q1235-03	JPP-51.2-012925	PL093951.D	31 Jan 2025 18:07		AR\AJ	Ok,M
26	Q1235-07	JPP-16.1-012925	PL093952.D	31 Jan 2025 18:20		AR\AJ	Ok,M
27	Q1239-01	286	PL093953.D	31 Jan 2025 18:33		AR\AJ	Ok,M
28	Q1239-04	348	PL093954.D	31 Jan 2025 18:46		AR\AJ	Ok,M
29	Q1239-07	RBR22266	PL093955.D	31 Jan 2025 18:59		AR\AJ	Ok,M
30	Q1239-10	357	PL093956.D	31 Jan 2025 19:12		AR\AJ	Ok,M
31	I.BLK	I.BLK	PL093957.D	31 Jan 2025 19:26		AR\AJ	Ok
32	PEM	PEM	PL093958.D	31 Jan 2025 19:39		AR\AJ	Ok,M
33	PSTDCCC050	PSTDCCC050	PL093959.D	31 Jan 2025 20:18		AR\AJ	Ok,M
34	Q1239-10MS	357MS	PL093960.D	31 Jan 2025 20:31		AR\AJ	Ok,M
35	Q1239-10MSD	357MSD	PL093961.D	31 Jan 2025 20:45	RPD Fail	AR\AJ	Ok,M
36	Q1241-03	JPP-3.5-013025	PL093962.D	31 Jan 2025 20:58		AR\AJ	Ok,M
37	Q1241-07	JPP-5.3-013025	PL093963.D	31 Jan 2025 21:11		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL013125

Review By	Abdul	Review On	2/3/2025 9:55:31 AM
Supervise By	Ankita	Supervise On	2/3/2025 1:14:23 PM
SubDirectory	PL013125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q1241-11	JPP-5.2-013025	PL093964.D	31 Jan 2025 21:24		AR\AJ	Ok,M
39	Q1241-15	JPP-5.4-013025	PL093965.D	31 Jan 2025 21:37		AR\AJ	Ok,M
40	Q1241-19	JPP-51.4-013025	PL093966.D	31 Jan 2025 21:50		AR\AJ	Ok,M
41	Q1242-03	JPP-6.2-013025	PL093967.D	31 Jan 2025 22:04		AR\AJ	Ok,M
42	Q1243-01	CL-01-01302025	PL093968.D	31 Jan 2025 22:17		AR\AJ	Ok
43	Q1244-01	EO-02-01302025	PL093969.D	31 Jan 2025 22:30		AR\AJ	Ok,M
44	I.BLK	I.BLK	PL093970.D	31 Jan 2025 22:43		AR\AJ	Ok
45	PSTDCCC050	PSTDCCC050	PL093971.D	01 Feb 2025 00:57	Comp#22 recovery low	AR\AJ	Ok,M
46	Q1215-03	JPP-29.1-012825	PL093972.D	01 Feb 2025 01:10		AR\AJ	Ok,M
47	Q1215-03MS	JPP-29.1-012825MS	PL093973.D	01 Feb 2025 01:24		AR\AJ	Ok,M
48	Q1215-03MSD	JPP-29.1-012825MSD	PL093974.D	01 Feb 2025 01:37		AR\AJ	Ok,M
49	Q1216-15	JPP-26.1-012825	PL093975.D	01 Feb 2025 01:50	DCB high in both column	AR\AJ	ReRun
50	Q1219-01	LAW-25-0015	PL093976.D	01 Feb 2025 02:03	DCB high in 1st column	AR\AJ	Ok,M
51	I.BLK	I.BLK	PL093977.D	01 Feb 2025 02:43		AR\AJ	Ok
52	PSTDCCC050	PSTDCCC050	PL093978.D	01 Feb 2025 02:56	Comp#22 recovery low	AR\AJ	Ok,M

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/31/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 16:45
In Date: 01/30/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:14
Out Date: 01/31/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB134481

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1215-01	JPP-29.1-012825	1	1.15	8.54	9.69	8.75	89.0	
Q1215-03	JPP-29.1-012825	2	1.16	8.48	9.64	8.69	88.8	
Q1215-05	JPP-29.2-012825	3	1.19	8.70	9.89	8.77	87.1	
Q1215-07	JPP-29.2-012825	4	1.15	8.63	9.78	8.81	88.8	
Q1216-01	JPP-18.1-012825	5	1.19	8.45	9.64	8.05	81.2	
Q1216-03	JPP-18.1-012825	6	1.16	8.82	9.98	8.51	83.3	
Q1216-05	JPP-21.1-012825	7	1.15	8.40	9.55	8.83	91.4	
Q1216-07	JPP-21.1-012825	8	1.15	8.75	9.9	9.06	90.4	
Q1216-09	JPP-21.2-012825	9	1.19	8.42	9.61	8.29	84.3	
Q1216-11	JPP-21.2-012825	10	1.15	8.36	9.51	8.2	84.3	
Q1216-13	JPP-26.1-012825	11	1.19	8.46	9.65	7.87	79.0	
Q1216-15	JPP-26.1-012825	12	1.17	8.76	9.93	8.42	82.8	
Q1216-17	JPP-26.2-012825	13	1.16	8.63	9.79	8.52	85.3	
Q1216-19	JPP-26.2-012825	14	1.17	8.51	9.68	8.47	85.8	
Q1232-01	JPP-46.2-012925	15	1.12	8.77	9.89	8.99	89.7	
Q1232-03	JPP-46.2-012925	16	1.15	8.37	9.52	8.62	89.2	
Q1232-05	JPP-46.1-012925	17	1.17	8.50	9.67	9.14	93.8	
Q1232-07	JPP-46.1-012925	18	1.15	8.72	9.87	9.35	94.0	
Q1232-09	JPP-42.1-012925	19	1.14	8.37	9.51	8.56	88.6	
Q1232-11	JPP-42.1-012925	20	1.19	8.43	9.62	8.62	88.1	
Q1232-13	JPP-42.2-012925	21	1.15	8.50	9.65	8.98	92.1	
Q1232-15	JPP-42.2-012925	22	1.15	8.37	9.52	8.95	93.2	
Q1232-17	JPP-51.1-012925	23	1.19	8.42	9.61	9.14	94.4	
Q1232-19	JPP-51.1-012925	24	1.12	8.75	9.87	9.44	95.1	
Q1233-01	WIPE-1	25	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q1233-02	WIPE-2	26	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q1235-01	JPP-51.2-012925	27	1.15	8.60	9.75	8.99	91.2	
Q1235-03	JPP-51.2-012925	28	1.15	8.51	9.66	8.96	91.8	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/31/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 16:45
In Date: 01/30/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:14
Out Date: 01/31/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB134481

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1235-05	JPP-16.1-012925	29	1.15	8.75	9.9	8.94	89.0	
Q1235-07	JPP-16.1-012925	30	1.12	8.77	9.89	8.94	89.2	
Q1237-01	HL6PX1	31	1.16	8.53	9.69	9.27	95.1	
Q1237-02	HL6PX2	32	1.16	8.70	9.86	9.28	93.3	
Q1237-03	HL6PX3	33	1.15	8.82	9.97	9.27	92.1	
Q1237-04	HL6PX4	34	1.15	8.78	9.93	9.43	94.3	
Q1237-05	HL6PX5	35	1.17	8.54	9.71	9.33	95.6	
Q1237-06	HL6PX6	36	1.17	8.57	9.74	9.07	92.2	
Q1239-01	286	37	1.14	8.49	9.63	8.68	88.8	
Q1239-04	348	38	1.14	8.83	9.97	9.00	89.0	
Q1239-07	RBR22266	39	1.17	8.74	9.91	9.00	89.6	
Q1239-10	357	40	1.16	8.80	9.96	8.62	84.8	
Q1240-01	MEG-OIL	41	1.00	1.00	2.00	2.00	100.0	oil sample

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

13448

WorkList Name : %1-013025

WorkList ID : 187270

Department : Wet-Chemistry

Date : 01-30-2025 07:55:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1215-01	JPP-29.1-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1215-03	JPP-29.1-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1215-05	JPP-29.2-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1215-07	JPP-29.2-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1216-01	JPP-18.1-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1216-03	JPP-18.1-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1216-05	JPP-21.1-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1216-07	JPP-21.1-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1216-09	JPP-21.2-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1216-11	JPP-21.2-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1216-13	JPP-26.1-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1216-15	JPP-26.1-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1216-17	JPP-26.2-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1216-19	JPP-26.2-012825	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1232-01	JPP-46.2-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/28/2025	Chemtech -SO
Q1232-03	JPP-46.2-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1232-05	JPP-46.1-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1232-07	JPP-46.1-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1232-09	JPP-42.1-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1232-11	JPP-42.1-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1232-13	JPP-42.2-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO

Date/Time 01/30/25 15:20

Raw Sample Received by: SD WCI

Raw Sample Relinquished by: CSM

Date/Time 01/30/25

Raw Sample Received by:

Raw Sample Relinquished by:

387 of 442

WORKLIST(Hardcopy Internal Chain)

134481

WorkList Name : %1-013025

WorkList ID : 187270

Department : Wet-Chemistry

Date : 01-30-2025 07:55:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1232-15	JPP-42.2-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1232-17	JPP-51.1-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1232-19	JPP-51.1-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1233-01	WIPE-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	N41	01/30/2025	Chemtech -SO
Q1233-02	WIPE-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	N41	01/30/2025	Chemtech -SO
Q1235-01	JPP-51.2-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1235-03	JPP-51.2-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1235-05	JPP-16.1-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1235-07	JPP-16.1-012925	Solid	Percent Solids	Cool 4 deg C	RUTW01	E11	01/29/2025	Chemtech -SO
Q1237-01	HL6PX1	Solid	Percent Solids	Cool 4 deg C	GENV01	N31	01/30/2025	Chemtech -SO
Q1237-02	HL6PX2	Solid	Percent Solids	Cool 4 deg C	GENV01	N31	01/30/2025	Chemtech -SO
Q1237-03	HL6PX3	Solid	Percent Solids	Cool 4 deg C	GENV01	N31	01/30/2025	Chemtech -SO
Q1237-04	HL6PX4	Solid	Percent Solids	Cool 4 deg C	GENV01	N31	01/30/2025	Chemtech -SO
Q1237-05	HL6PX5	Solid	Percent Solids	Cool 4 deg C	GENV01	N31	01/30/2025	Chemtech -SO
Q1237-06	HL6PX6	Solid	Percent Solids	Cool 4 deg C	GENV01	N31	01/30/2025	Chemtech -SO
Q1239-01	286	Solid	Percent Solids	Cool 4 deg C	PSEG03	N31	01/30/2025	Chemtech -SO
Q1239-04	348	Solid	Percent Solids	Cool 4 deg C	PSEG03	N31	01/30/2025	Chemtech -SO
Q1239-07	RBR22266	Solid	Percent Solids	Cool 4 deg C	PSEG03	N31	01/30/2025	Chemtech -SO
Q1239-10	357	Solid	Percent Solids	Cool 4 deg C	PSEG03	N31	01/30/2025	Chemtech -SO
Q1240-01	MEG-OIL	Solid	Percent Solids	Cool 4 deg C	PSEG03	N41	01/30/2025	Chemtech -SO

Date/Time 01/30/25 15:20
 Raw Sample Received by: JB WLC
 Raw Sample Relinquished by: cf gm

Date/Time 01/30/25 17:10
 Raw Sample Received by: cf sm
 Raw Sample Relinquished by: JB WLC

SOP ID: M3541-ASE Extraction-14

Clean Up SOP #: Florisil

Matrix : Solid

Weigh By: EH

Balance check: RJ

Balance ID: EX-SC-2

pH Strip Lot#: N/A

Extraction By: RJ

Filter By: RJ

pH Meter ID: N/A

Hood ID: 3,7

Extraction Start Date : 01/31/2025

Extraction Start Time : 08:15

Extraction End Date : 01/31/2025

Extraction End Time : 11:15

Concentration By: EH

Supervisor By : rajesh

Extraction Method: ☐ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP24091
Surrogate	1.0ML	200 PPB	PP24123
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2579
Baked Na2SO4	N/A	EP2580
Sand	N/A	E2865
Hexane	N/A	E3872
Florisil	N/A	E3806
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40 BTS721.

KD Bath ID: N/A

KD Bath Temperature: N/A

Envap ID: NEVAP-02

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
01/31/25	RJ (Sat 10:15)	Y.P. PESTHUB
11:20	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 01/31/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB166413BL	PBLK413	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10			U7-1
PB166413BS	PLCS413	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10			2
Q1232-03	JPP-46.2-012925	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	C		3
Q1232-07	JPP-46.1-012925	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	C		4
Q1232-11	JPP-42.1-012925	Pesticide-TCL	30.08	N/A	ritesh	Evelyn	10	C		5
Q1232-15	JPP-42.2-012925	Pesticide-TCL	30.06	N/A	ritesh	Evelyn	10	C		6
Q1232-19	JPP-51.1-012925	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	C		U6-1
Q1235-03	JPP-51.2-012925	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10	C		2
Q1235-07	JPP-16.1-012925	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	C		3
Q1239-01	286	Pesticide-TCL	30.08	N/A	ritesh	Evelyn	10	C		4
Q1239-04	348	Pesticide-TCL	30.09	N/A	ritesh	Evelyn	10	C		5
Q1239-07	RBR22266	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	C		6
Q1239-10	357	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	C		U5-1
Q1239-10MS	357MS	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	C		2
Q1239-10MS D	357MSD	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10	C		3
Q1241-03	JPP-3.5-013025	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10	C		4
Q1241-07	JPP-5.3-013025	Pesticide-TCL	30.07	N/A	ritesh	Evelyn	10	C		5
Q1241-11	JPP-5.2-013025	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	C		6
Q1241-15	JPP-5.4-013025	Pesticide-TCL	30.06	N/A	ritesh	Evelyn	10	C		U4-1
Q1241-19	JPP-51.4-013025	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	C		2
Q1242-03	JPP-6.2-013025	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10	C		3
Q1243-01	CL-01-01302025	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	C		4
Q1244-01	EO-02-01302025	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10	C		5

* Extracts relinquished on the same date as received.

166413
8-15

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1235

WorkList ID : 187330

Department : Extraction

Date : 01-31-2025 08:11:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1232-03	JPP-46.2-012925	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/29/2025	8081B
Q1232-07	JPP-46.1-012925	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/29/2025	8081B
Q1232-11	JPP-42.1-012925	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/29/2025	8081B
Q1232-15	JPP-42.2-012925	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/29/2025	8081B
Q1232-19	JPP-51.1-012925	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/29/2025	8081B
Q1235-03	JPP-51.2-012925	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/29/2025	8081B
Q1235-07	JPP-16.1-012925	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/29/2025	8081B
Q1239-01	286	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	N31	01/30/2025	8081B
Q1239-04	348	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	N31	01/30/2025	8081B
Q1239-07	RBR22266	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	N31	01/30/2025	8081B
Q1239-10	357	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	N31	01/30/2025	8081B
Q1241-03	JPP-3.5-013025	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/30/2025	8081B
Q1241-07	JPP-5.3-013025	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/30/2025	8081B
Q1241-11	JPP-5.2-013025	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/30/2025	8081B
Q1241-15	JPP-5.4-013025	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/30/2025	8081B
Q1241-19	JPP-51.4-013025	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/30/2025	8081B
Q1242-03	JPP-6.2-013025	Solid	Pesticide-TCL	Cool 4 deg C	RUTW01	E11	01/30/2025	8081B
Q1243-01	CL-01-01302025	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	N41	01/30/2025	8081B
Q1244-01	EO-02-01302025	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	N51	01/30/2025	8081B

Date/Time 01/31/25 8:12
 Raw Sample Received by: RJ (Ext 104)
 Raw Sample Relinquished by: CR SM

Date/Time 01/31/25 8:35
 Raw Sample Received by: CR SM
 Raw Sample Relinquished by: RJ (Ext 104)

Prep Standard - Chemical Standard Summary

Order ID : Q1232
Test : Pesticide-TCL
Prepbatch ID : PB166413,
Sequence ID/Qc Batch ID: pl013125,

Standard ID :
EP2579,EP2580,PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683,PP23686,PP23687,PP23690,PP23693,PP23695,PP23698,PP23733,PP23793,PP24091,PP24095,PP24123,

Chemical ID :
E2865,E3551,E3792,E3805,E3806,E3843,E3846,E3847,E3872,P11146,P11896,P13036,P13039,P13245,P13349,P13350,P13353,P13359,P13402,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2579	01/06/2025	06/16/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 01/06/2025

FROM 8000.00000ml of E3846 + 8000.00000ml of E3847 = Final Quantity: 8000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2580	01/17/2025	07/01/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 01/17/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP23673	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
10/01/2024								

FROM 1.00000ml of P13349 + 9.00000ml of E3792 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3629	20 PPM PEST stock Solution 1st source(RESTEK)	PP23674	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
10/01/2024								

FROM 1.00000ml of P13036 + 9.00000ml of E3792 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1472	20 PPM Pest Stock Solution 2nd Source	PP23675	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
10/01/2024								

FROM 1.00000ml of P13039 + 9.00000ml of E3792 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1273	20 PPM Mirex Stock (Primary Source)	PP23676	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
10/01/2024								

FROM 0.20000ml of P11146 + 9.80000ml of E3792 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3663	20 PPM MIREX Stock STD (Secondary source)	PP23677	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani 10/01/2024

FROM 0.20000ml of P11146 + 9.80000ml of E3792 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3630	100/100 PPB PEST Working std.1st Source(RESTEK)	PP23678	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani 10/01/2024

FROM 98.50000ml of E3792 + 0.50000ml of PP23673 + 0.50000ml of PP23674 + 0.50000ml of PP23676 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
80	100/100 PPB Pesticide Working Solution 2nd Source	PP23679	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
10/01/2024								

FROM 98.50000ml of E3792 + 0.50000ml of PP23673 + 0.50000ml of PP23675 + 0.50000ml of PP23677 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
386	1000/100 PPB Chlordane STD (Restek)	PP23680	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
10/01/2024								

FROM 0.10000ml of P11896 + 99.40000ml of E3792 + 0.50000ml of PP23673 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3746	1000/100 ppb Chlordane STD-RESTEK 2ND SOURCE	PP23681	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
								10/01/2024

FROM 0.10000ml of P11896 + 99.40000ml of E3792 + 0.50000ml of PP23673 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
383	1000/100 PPB Toxaphene STD (Restek)	PP23682	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
								10/01/2024

FROM 0.10000ml of P13359 + 99.40000ml of E3792 + 0.50000ml of PP23673 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3669	1000/100 PPB TOXAPHENE STD 2nd source (RESTEK)	PP23683	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani 10/01/2024

FROM 0.10000ml of P13402 + 99.40000ml of E3792 + 0.50000ml of PP23673 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3632	50 PPB ICAL PEST STD(RESTEK)	PP23686	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani 10/01/2024

FROM 0.50000ml of E3792 + 0.50000ml of PP23678 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3988	50 PPB PEST ICV STD(RESTEK)	PP23687	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
								10/01/2024

FROM 0.50000ml of E3792 + 0.50000ml of PP23679 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
529	CHLOR 500 PPB STD	PP23690	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
								10/01/2024

FROM 0.50000ml of E3792 + 0.50000ml of PP23680 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
532	CHLOR 500 PPB ICV STD	PP23693	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
10/01/2024								

FROM 0.50000ml of E3792 + 0.50000ml of PP23681 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
534	TOX 500 PPB STD	PP23695	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
10/01/2024								

FROM 0.50000ml of E3792 + 0.50000ml of PP23682 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3670	TOX 500 PPB ICV std (RESTEK)	PP23698	09/21/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
								10/01/2024

FROM 0.50000ml of E3792 + 0.50000ml of PP23683 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP23733	10/03/2024	03/30/2025	Ankita Jodhani	None	None	Yogesh Patel
								10/03/2024

FROM 1.00000ml of P13350 + 9.00000ml of E3805 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
518	Pest/PCB I.BLK 20 PPB	PP23793	10/03/2024	03/30/2025	Ankita Jodhani	None	None	Yogesh Patel
								10/03/2024

FROM 99.90000ml of E3805 + 0.10000ml of PP23733 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
79	500 PPB Pesticide Spike Solution	PP24091	12/17/2024	03/11/2025	Abdul Mirza	None	None	Ankita Jodhani
								12/18/2024

FROM 95.00000ml of E3843 + 2.50000ml of PP23675 + 2.50000ml of PP23677 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4027	Pesticide resolution Check Mixture 8081	PP24095	12/23/2024	06/16/2025	Abdul Mirza	None	None	Ankita Jodhani
								12/30/2024

FROM 1.00000ml of P13245 + 99.00000ml of E3847 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24123	01/20/2025	06/26/2025	Abdul Mirza	None	None	Ankita Jodhani
								01/20/2025

FROM 1.00000ml of P13353 + 999.00000ml of E3846 = Final Quantity: 1000.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	0000243821	06/30/2025	04/30/2020 / RAJESH	04/28/2020 / RAJESH	E2865

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24C1862008	03/11/2025	09/12/2024 / Rajesh	09/11/2024 / Rajesh	E3792

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24C1862008	03/30/2025	09/30/2024 / Rajesh	09/25/2024 / Rajesh	E3805

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agela Technologies Inc.	FS0006 / Cleanert Florisil cartridge	M06518	03/25/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3806

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	06/05/2025	12/05/2024 / Rajesh	12/05/2024 / Rajesh	E3843

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	06/26/2025	12/26/2024 / Rajesh	12/13/2024 / Rajesh	E3846

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	06/16/2025	12/16/2024 / Rajesh	12/13/2024 / Rajesh	E3847

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	07/29/2025	01/29/2025 / Rajesh	01/29/2025 / Rajesh	E3872

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	102821	03/21/2025	09/21/2024 / Abdul	10/29/2021 / Abdul	P11146

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0181737	03/21/2025	09/21/2024 / Abdul	06/17/2022 / Abdul	P11896

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0200423	03/21/2025	09/21/2024 / Abdul	12/26/2023 / Abdul	P13036

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0199099	03/21/2025	09/21/2024 / Abdul	12/26/2023 / Abdul	P13039

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	19161 / 8081 pesticide resolution check mixture	013124	06/23/2025	12/23/2024 / Abdul	02/09/2024 / Abdul	P13245

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	03/21/2025	09/21/2024 / Abdul	04/22/2024 / Abdul	P13349

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	04/03/2025	10/03/2024 / Ankita	04/22/2024 / Abdul	P13350

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	07/20/2025	01/20/2025 / Abdul	04/22/2024 / Abdul	P13353

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0203830	03/21/2025	09/21/2024 / Abdul	05/03/2024 / Abdul	P13359

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0203038	03/21/2025	09/21/2024 / Abdul	05/15/2024 / Abdul	P13402

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Sand
Purified
Washed and Ignited



Material No.: 3382-05
Batch No.: 0000243821
Manufactured Date: 2018/04/09
Retest Date: 2025/04/07
Revision No: 1


Certificate of Analysis

Test	Specification	Result
Substances Soluble in HCl	$\leq 0.16\%$	0.01

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC

E 2865


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %
COMMENTS		
QC: PhC Irma Belmares		

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3

Hexanes (95% n-hexane)
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9262-03
Batch No.: 24C1862008
Manufactured Date: 2024-01-30
Expiration Date: 2025-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.4 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 09/11/24

E 3192

Jamie Croak
Director Quality Operations, Bioscience Production

Hexanes (95% n-hexane)
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9262-03
Batch No.: 24C1862008
Manufactured Date: 2024-01-30
Expiration Date: 2025-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.4 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 9/25/24

E 3805

Jamie Croak
Director Quality Operations, Bioscience Production

Cleanert Florisil

1g/6ml 30/pkg

固相萃取产品

LOT#:M06518

MFG#:F04074



Made in China

CAT# FS0006

Agela Technologies

E 3806



Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 12/5/24

E 3843

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Rec'd by RP On 12/13/24

E 3846

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 12/13/24

E3847

J. Croak

Jamie Croak
Director Quality Operations, Bioscience Production

Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₈ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 1/29/25

E 3872

J. Croak

Jamie Croak
Director Quality Operations, Bioscience Production



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0181737

Description : Chlordane Standard

Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Chlordane CAS # 57-74-9 Purity ----%	1,006.0 µg/mL (Lot 978545)	+/- 5.9753 µg/mL Gravimetric +/- 31.8975 µg/mL Unstressed +/- 41.6615 µg/mL Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

P 11892
↓
P 11896
5

06/17/2022

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)


Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Josh McCloskey
Josh McCloskey - Operations Technician I

Date Mixed: 11-Feb-2022

Balance: B442140311

Marlina Cowan
Marlina Cowan - Operations Tech I

Date Passed: 24-Feb-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 11892
↓
P 11896 / (5)

AK
06/17/2022



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 Lot No.: A0199099

Description : Organochlorine Pesticide Mix AB #1
Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL Pkg Amt: > 1 mL

Expiration Date : June 30, 2027 Storage: 10°C or colder

Ship: Ambient

P130397 5
↓
P13043 1
✓
12-26-2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.0 µg/mL	+/- 8.9732
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	200.1 µg/mL	+/- 8.9762
3	beta-BHC	319-85-7	BCCC6425	99%	200.3 µg/mL	+/- 8.9844
4	delta-BHC	319-86-8	14450800	98%	200.0 µg/mL	+/- 8.9740
5	Heptachlor	76-44-8	813251	99%	200.1 µg/mL	+/- 8.9754
6	Aldrin	309-00-2	14389400	98%	200.0 µg/mL	+/- 8.9718
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.1 µg/mL	+/- 8.9754
8	trans-Chlordane	5103-74-2	32943	98%	199.9 µg/mL	+/- 8.9696
9	cis-Chlordane	5103-71-9	31766	98%	200.1 µg/mL	+/- 8.9762
10	Endosulfan I	959-98-8	BCCF4060	99%	200.1 µg/mL	+/- 8.9754
11	4,4'-DDE	72-55-9	GHYQG	99%	200.1 µg/mL	+/- 8.9777
12	Dieldrin	60-57-1	11129900	98%	200.0 µg/mL	+/- 8.9718
13	Endrin	72-20-8	14123200	98%	199.9 µg/mL	+/- 8.9696
14	4,4'-DDD	72-54-8	HAN02	99%	200.1 µg/mL	+/- 8.9777
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	200.0 µg/mL	+/- 8.9718

17	Endrin aldehyde	7421-93-4	30720	98%	200.1	µg/mL	+/- 8.9784
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.0	µg/mL	+/- 8.9732
19	Methoxychlor	72-43-5	13668200	99%	200.1	µg/mL	+/- 8.9777
20	Endrin ketone	53494-70-5	1-ABS-16-7	98%	200.0	µg/mL	+/- 8.9740

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P 13039
 ↓
 P13043
 1
 5
 12/26/23

Quality Confirmation Test

Column:
 30m x .25mm x .2µm
 Rtx-CLP II (cat.# 11323)

Carrier Gas:
 helium-constant pressure 20 psi.

Temp. Program:
 150°C to 300°C
 @ 4°C/min. (hold 5 min.)


Inj. Temp:
 200°C

Det. Temp:
 300°C

Det. Type:
 ECD

Split Vent:
 Split ratio 50:1

Inj. Vol
 1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Josh McCloskey
 Josh McCloskey - Operations Technician I

Date Mixed: 19-Jun-2023

Balance Serial # 1128360905

Jennifer Pollino
 Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 23-Jun-2023

Manufactured under Restek's ISO 9001:2015
 Registered Quality System
 Certificate #FM 80397



CERTIFIED WEIGHT REPORT

Part Number: **79136**
 Lot Number: **102821**
 Description: **Mirex**

Solvent(s): **Acetone**
 Lot#: **81025**




Expiration Date: **102826**
 Recommended Storage: **Refrigerate (4 °C)**
 Nominal Concentration (µg/mL): **1000**
 NIST Test ID#: **6UTB**

Weight(s) shown below were combined and diluted to (mL): **50.0**
 5E-05 Balance Uncertainty
 0.006 Flask Uncertainty

<i>Eli Aliaga</i>		102821
Formulated By:	Eli Aliaga	DATE
<i>Pedro L. Rentas</i>		102821
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Mirex	437	9492400	1000	99.4	0.5	0.05034	0.05039	1000.9	10.3	2385-85-5	N/A	ori-rat 306mg/kg

Method GC7MSD-1.M: Column: SPB-608 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 150°C (4min.), Temp 2 = 290°C (13.5 min.), Rate = 8°C/min., Injector B= 200°C, Detector B = 290°C. Split Ratio = 100:1, Scan Rate = 2. Analysis performed by Candice Warren.



P11142
 To
 P11146
 AR
 11/02/21

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 **Lot No.:** A0200423

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 10°C or colder

Ship: Ambient

P 13034
↓
P 13038
12.26.2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.5 µg/mL	+/- 8.9956
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	199.9 µg/mL	+/- 8.9696
3	beta-BHC	319-85-7	BCCC6425	99%	200.0 µg/mL	+/- 8.9732
4	delta-BHC	319-86-8	14450800	98%	199.9 µg/mL	+/- 8.9696
5	Heptachlor	76-44-8	813251	99%	202.0 µg/mL	+/- 9.0629
6	Aldrin	309-00-2	14389400	98%	200.9 µg/mL	+/- 9.0136
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.0 µg/mL	+/- 8.9732
8	trans-Chlordane	5103-74-2	34616	99%	200.5 µg/mL	+/- 8.9956
9	cis-Chlordane	5103-71-9	31766	98%	201.4 µg/mL	+/- 9.0356
10	Endosulfan I	959-98-8	BCCF4060	99%	200.0 µg/mL	+/- 8.9732
11	4,4'-DDE	72-55-9	GHYQG	99%	201.5 µg/mL	+/- 9.0405
12	Dieldrin	60-57-1	14515000	98%	199.9 µg/mL	+/- 8.9696
13	Endrin	72-20-8	14485300	98%	200.4 µg/mL	+/- 8.9916
14	4,4'-DDD	72-54-8	HAN02	99%	200.5 µg/mL	+/- 8.9956
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	201.9 µg/mL	+/- 9.0575

17	Endrin aldehyde	7421-93-4	30720	98%	201.4 µg/mL	+/- 9.0356
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.5 µg/mL	+/- 8.9956
19	Methoxychlor	72-43-5	14563200	98%	200.9 µg/mL	+/- 9.0136
20	Endrin ketone	53494-70-5	14537700	98%	199.9 µg/mL	+/- 8.9696

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13034
P13038
1
5
12/26/2023

Quality Confirmation Test

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 50:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 31-Jul-2023 **Balance Serial #** B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 03-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED WEIGHT REPORT

Part Number: **19161**
 Lot Number: **013124**
 Description: **CLP Pesticides & PCB's Resolution Check Standard**
 9 components
 Expiration Date: **013129**
 Recommended Storage: **Refrigerate (4 °C)**
 Nominal Concentration (µg/mL): **Varied**
 NIST Test ID#: **6UTB**

Solvent(s):
 Hexane 273615 (50%)
 Toluene 28508 (50%)

Volume(s) shown below were combined and diluted to (mL): 100.0 5E-05 Balance Uncertainty
 0.021 Flask Uncertainty

		013124
Formulated By:	Lawrence Barry	DATE
		013124
Reviewed By:	Pedro L. Rentas	DATE

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc.(ug/mL)	Final Conc.(ug/mL)	Expanded Uncertainty (+/-) µg/mL	SDS Information (Solvent Safety Info. On Attached pg.)		
									CAS#	OSHA PEL (TWA)	LD50
1. trans-Chlordane	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	5103-74-2	0.5mg/m3 (skin)	ori-rat 500mg/kg
2. Endosulfan I	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	959-98-8	0.1mg/m3 (skin)	ori-rat 18mg/kg
3. 4,4'-DDE	19361	013124	0.010	1.00	0.004	201.6	2.0	0.03	72-55-9	N/A	ori-rat 880mg/kg
4. Dieldrin	19361	013124	0.010	1.00	0.004	202.8	2.0	0.03	60-57-1	0.25mg/m3 (skin)	ori-rat 38300ug/kg
5. Endosulfan sulfate	19361	013124	0.010	1.00	0.004	204.2	2.0	0.03	1031-07-8	N/A	ori-rat 18mg/kg
6. Endrin ketone	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	53494-70-5	N/A	N/A
7. 4,4'-Methoxychlor	19361	013124	0.010	1.00	0.004	1000.7	10.0	0.09	72-43-5	10mg/m3	ori-rat 6000mg/kg
8. 2,4,5,6-Tetrachloro-m-xylene	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	877-09-8	N/A	N/A
9. Decachlorobiphenyl (209)	19361	013124	0.010	1.00	0.004	202.0	2.0	0.03	2051-24-3	N/A	N/A

• The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 • Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
 • Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 • All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 • Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

P13243 } (5)
 ↓
 P13247 }
 ↓
 02/9/2024



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0203038

Description : Toxaphene Standard

Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	----%	1,009.0 µg/mL	+/- 55.9920

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13358
↓
P 13369
(12)

[Signature]
05-06-2024

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 10-Oct-2023

Balance Serial # 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13358 } (12)
↓
P13369
|


05-06-2024

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0203038

Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2028 **Storage:** 10°C or colder
Ship: Ambient

P13402 } (5)
P13406 }
5/22/2021

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	---%	1,009.0 µg/mL	+/- 55.9920

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I


Date Mixed: 10-Oct-2023

Balance Serial # 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13402
↓
P13406 } (5)

5/22/2024



SHIPPING DOCUMENTS

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. **Q1232**
QUOTE NO.
COC Number **2041691**

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: RU2 Engineering LLC
ADDRESS: 2 Melinda Drive
CITY: Monroe Twp, NJ 08831

CITY: ZIP:

ATTENTION: Rutu Manani

PHONE: 609-409-4564

FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: SANDTWOBR BMCR Project

PROJECT NO.: LOCATION: Brooklyn, NYC

PROJECT MANAGER: Rutu Manani

e-mail: Rmanani@Ru2eng.com

PHONE:

FAX:

CLIENT BILLING INFORMATION

BILL TO: Same as company address

ADDRESS:

CITY:

STATE:

ZIP:

ATTENTION:

PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) Standard 10 days DAYS*
HARDCOPY (DATA PACKAGE): Standard 10 days DAYS*
EDD: Standard 10 days DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☒ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

1. TCL VOC+TICS
2. TCLP VOCs
3. TPH GRO-DRO
4. TCL SVOCs+TICS
5. TAL Metals
6. Pesticides PCBs
7. RCRA Characterization
8. Paint filter
9. full TCLP

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	← Specify Preservatives A-HCl B-HNO3 C-H2SO4 D-NaOH E-IGP F-OTHER	
1.	JPP-46.2-012925	Soil		G	01/29/25	9:00	3	X	X	X								
2.	JPP-46.2-012925	Soil	L		1/29/25	9:05	8			X	X	X	X	X	X	X		
3.	JPP-46.1-012925	Soil		G	1/29/25	10:10	3	X	X	X								
4.	JPP-46.1-012925	Soil	L		1/29/25	10:18	8			X	X	X	X	X	X	X		
5.	JPP-42.1-012925	Soil		G	1/29/25	11:24	3	X	X	X								
6.	JPP-42.1-012925	Soil	L		1/29/25	11:30	8			X	X	X	X	X	X	X		
7.	JPP-42.2-012925	Soil		G	1/29/25	12:24	3	X	X	X								
8.	JPP-42.2-012925	Soil	L		1/29/25	12:30	8			X	X	X	X	X	X	X		
9.	JPP-51.1-012925	Soil		G	1/29/25	14:14	3	X	X	X								
10.	JPP-51.1-012925	Soil	L		1/29/25	14:20	8			X	X	X	X	X	X	X		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. RA	DATE/TIME: 1/30/2025	RECEIVED BY: 1. [Signature] 1045 1-30-25	Conditions of bottles or coolers at receipt: ☒ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.16 °C
RELINQUISHED BY SAMPLER: 2. [Signature]	DATE/TIME:	RECEIVED BY: 2. [Signature]	Comments: Preserve extra Sample Jar for additional analysis is required.
RELINQUISHED BY SAMPLER: 3. [Signature]	DATE/TIME: 1152 1-30-25	RECEIVED BY: 3. [Signature]	Page 1 of 2

CLIENT: ☐ Hand Delivered ☐ Other
CHEMTECH: ☐ Picked Up ☐ Field Sampling

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488


LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1232	RUTW01	Order Date : 1/30/2025 11:55:00 AM	Project Mgr : Kiran
Client Name : RU2 Engineering, LLC		Project Name : NYCDDC SANTWOBR B	Report Type : NYS ASP B
Client Contact : Rutu Manani		Receive DateTime : 1/30/2025 11:52:00 AM	EDD Type : Excel NY
Invoice Name : RU2 Engineering, LLC		Purchase Order :	Hard Copy Date :
Invoice Contact : Rutu Manani			Date Signoff : 1/30/2025 1:25:39 PM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1232-01	JPP-46.2-012925	Solid	01/29/2025	09:00					
					VOCMS Group1		8260D	10 Bus. Days	
Q1232-05	JPP-46.1-012925	Solid	01/29/2025	10:10					
					VOCMS Group1		8260D	10 Bus. Days	
Q1232-09	JPP-42.1-012925	Solid	01/29/2025	11:24					
					VOCMS Group1		8260D	10 Bus. Days	
Q1232-13	JPP-42.2-012925	Solid	01/29/2025	12:24					
					VOCMS Group1		8260D	10 Bus. Days	
Q1232-17	JPP-51.1-012925	Solid	01/29/2025	14:14					
					VOCMS Group1		8260D	10 Bus. Days	

Relinquished By :

Date / Time :


1/30/25 1700

Received By :

Date / Time :


1.30.25. 17:00

Storage Area : VOA Refridgerator Room

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1232	RUTW01	Order Date : 1/30/2025 11:55:00 AM	Project Mgr : Kiran
Client Name : RU2 Engineering, LLC		Project Name : NYCDDC SANTWOBR B	Report Type : NYS ASP B
Client Contact : Rutu Manani		Receive DateTime : 1/30/2025 11:52:00 AM	EDD Type : Excel NY
Invoice Name : RU2 Engineering, LLC		Purchase Order :	Hard Copy Date :
Invoice Contact : Rutu Manani			Date Signoff : 1/30/2025 1:25:39 PM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1232-01	JPP-46.2-012925	Solid	01/29/2025	09:00					
					Gasoline Range Organics		8015D	10 Bus. Days	
Q1232-05	JPP-46.1-012925	Solid	01/29/2025	10:10					
					Gasoline Range Organics		8015D	10 Bus. Days	
Q1232-09	JPP-42.1-012925	Solid	01/29/2025	11:24					
					Gasoline Range Organics		8015D	10 Bus. Days	
Q1232-13	JPP-42.2-012925	Solid	01/29/2025	12:24					
					Gasoline Range Organics		8015D	10 Bus. Days	
Q1232-17	JPP-51.1-012925	Solid	01/29/2025	14:14					
					Gasoline Range Organics		8015D	10 Bus. Days	

Relinquished By :

Date / Time : 1/30/25 17:00

Received By :

Date / Time : 1-30-25 17:00

Storage Area : VOA Refridgerator Room