

Surrogate Summary

SDG No.: **Q1232**

Client: **RU2 Engineering, LLC**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL093928.D	PIBLK-PL093928.D	Decachlorobiphenyl	1	20	20.6	103		43	140
		Tetrachloro-m-xylene	1	20	21.1	106		77	126
		Decachlorobiphenyl	2	20	20.4	102		43	140
		Tetrachloro-m-xylene	2	20	19.8	99		77	126
PB166413BL	PB166413BL	Decachlorobiphenyl	1	20	24.6	123		10	148
		Tetrachloro-m-xylene	1	20	23.8	119		10	159
		Decachlorobiphenyl	2	20	24.1	121		10	148
		Tetrachloro-m-xylene	2	20	22.1	110		10	159
PB166413BS	PB166413BS	Decachlorobiphenyl	1	20	21.2	106		10	148
		Tetrachloro-m-xylene	1	20	20.0	100		10	159
		Decachlorobiphenyl	2	20	21.7	109		10	148
		Tetrachloro-m-xylene	2	20	18.6	93		10	159
I.BLK-PL093942.D	PIBLK-PL093942.D	Decachlorobiphenyl	1	20	22.6	113		43	140
		Tetrachloro-m-xylene	1	20	21.7	109		77	126
		Decachlorobiphenyl	2	20	22.3	111		43	140
		Tetrachloro-m-xylene	2	20	20.1	101		77	126
Q1232-03	JPP-46.2-012925	Decachlorobiphenyl	1	20	24.1	120		10	148
		Tetrachloro-m-xylene	1	20	18.1	91		10	159
		Decachlorobiphenyl	2	20	30.3	152	*	10	148
		Tetrachloro-m-xylene	2	20	22.0	110		10	159
Q1232-07	JPP-46.1-012925	Decachlorobiphenyl	1	20	20.4	102		10	148
		Tetrachloro-m-xylene	1	20	16.1	80		10	159
		Decachlorobiphenyl	2	20	12.8	64		10	148
		Tetrachloro-m-xylene	2	20	15.4	77		10	159
Q1232-11	JPP-42.1-012925	Decachlorobiphenyl	1	20	16.1	80		10	148
		Tetrachloro-m-xylene	1	20	14.7	74		10	159
		Decachlorobiphenyl	2	20	14.8	74		10	148
		Tetrachloro-m-xylene	2	20	15.6	78		10	159
Q1232-15	JPP-42.2-012925	Decachlorobiphenyl	1	20	16.6	83		10	148
		Tetrachloro-m-xylene	1	20	15.7	79		10	159
		Decachlorobiphenyl	2	20	13.3	66		10	148
		Tetrachloro-m-xylene	2	20	17.5	88		10	159
Q1232-19	JPP-51.1-012925	Decachlorobiphenyl	1	20	16.1	81		10	148
		Tetrachloro-m-xylene	1	20	16.3	81		10	159
		Decachlorobiphenyl	2	20	14.1	71		10	148
		Tetrachloro-m-xylene	2	20	16.4	82		10	159
I.BLK-PL093957.D	PIBLK-PL093957.D	Decachlorobiphenyl	1	20	20.4	102		43	140

Surrogate Summary

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093957.D	PIBLK-PL093957.D	Tetrachloro-m-xylene	1	20	22.0	110		77	126
		Decachlorobiphenyl	2	20	17.1	86		43	140
Q1239-10MS	357MS	Tetrachloro-m-xylene	2	20	20.8	104		77	126
		Decachlorobiphenyl	1	20	14.9	75		10	148
		Tetrachloro-m-xylene	1	20	19.1	95		10	159
		Decachlorobiphenyl	2	20	13.7	69		10	148
		Tetrachloro-m-xylene	2	20	18.0	90		10	159
		Decachlorobiphenyl	1	20	15.4	77		10	148
Q1239-10MSD	357MSD	Tetrachloro-m-xylene	1	20	18.7	93		10	159
		Decachlorobiphenyl	2	20	13.6	68		10	148
		Tetrachloro-m-xylene	2	20	17.9	90		10	159
		Decachlorobiphenyl	1	20	19.3	96		43	140
		Tetrachloro-m-xylene	1	20	21.3	107		77	126
		Decachlorobiphenyl	2	20	16.1	81		43	140
I.BLK-PL093970.D	PIBLK-PL093970.D	Tetrachloro-m-xylene	2	20	20.0	100		77	126