

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL094022.D Time Analyzed: 22:41

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.056	8.953	9.153	48.960	50.000	-2.1
Endrin	6.572	6.472	6.672	50.650	50.000	1.3
gamma-BHC (Lindane)	4.327	4.227	4.427	51.980	50.000	4.0
Heptachlor	4.915	4.814	5.014	51.640	50.000	3.3
Heptachlor epoxide	5.683	5.582	5.782	50.990	50.000	2.0
Methoxychlor	7.500	7.398	7.598	53.980	50.000	8.0
Tetrachloro-m-xylene	3.538	3.439	3.639	51.210	50.000	2.4

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL094022.D Time Analyzed: 22:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.910	7.810	8.010	53.520	50.000	7.0
Endrin	5.636	5.536	5.736	56.890	50.000	13.8
gamma-BHC (Lindane)	3.607	3.507	3.707	52.380	50.000	4.8
Heptachlor	3.945	3.845	4.045	51.290	50.000	2.6
Heptachlor epoxide	4.727	4.627	4.827	52.980	50.000	6.0
Methoxychlor	6.610	6.509	6.709	56.660	50.000	13.3
Tetrachloro-m-xylene	2.774	2.674	2.874	51.590	50.000	3.2