Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG No.: Q1232

Instrument ID: BNA_F Calibration Date(s): 02/06/2025 02/06/2025

Calibration Time(s): 11:07 14:14

LAB FILE ID: RRF2.5 = BF141472.D RRF005 = BF141473.D RRF010 = BF141474.D
RRF020 = BF141475.D RRF040 = BF141476.D RRF050 = BF141477.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.235	1.290	1.365	1.289	1.273	1.276	3.8
Benzaldehyde		0.892	0.993	0.981	0.821	0.747	0.857	13.9
Phenol-d6		1.619	1.629	1.704	1.613	1.599	1.612	3.1
Phenol		1.693	1.681	1.817	1.685	1.659	1.678	4.4
bis(2-Chloroethyl)ether		1.262	1.229	1.296	1.266	1.252	1.261	1.6
2-Chlorophenol		1.399	1.410	1.446	1.379	1.368	1.378	3.4
2-Methylphenol		1.116	1.089	1.121	1.077	1.059	1.079	3.1
2,2-oxybis(1-Chloropropane)		2.222	2.245	2.313	2.165	2.119	2.158	5.5
Acetophenone		0.517	0.502	0.521	0.493	0.483	0.498	3.4
3+4-Methylphenols		1.442	1.417	1.459	1.372	1.344	1.371	5.5
n-Nitroso-di-n-propylamine	0.950	0.978	0.985	1.039	0.969	0.947	0.964	4.0
Nitrobenzene-d5		0.380	0.379	0.396	0.385	0.378	0.383	1.6
Hexachloroethane		0.539	0.545	0.589	0.553	0.556	0.550	3.6
Nitrobenzene		0.370	0.371	0.383	0.374	0.369	0.374	1.3
Isophorone		0.611	0.594	0.633	0.611	0.601	0.611	2.0
2-Nitrophenol		0.171	0.175	0.190	0.193	0.190	0.186	4.9
2,4-Dimethylphenol		0.208	0.212	0.221	0.223	0.218	0.217	2.5
bis(2-Chloroethoxy)methane		0.384	0.390	0.408	0.394	0.387	0.392	2.0
2,4-Dichlorophenol		0.286	0.287	0.312	0.296	0.290	0.294	2.9
Naphthalene		1.067	1.041	1.096	1.037	1.012	1.041	3.0
4-Chloroaniline		0.354	0.352	0.382	0.364	0.360	0.363	2.8
Hexachlorobutadiene		0.193	0.191	0.200	0.192	0.188	0.192	2.0
Caprolactam		0.083	0.085	0.094	0.093	0.088	0.089	4.8
4-Chloro-3-methylphenol		0.310	0.324	0.342	0.329	0.324	0.325	2.9
2-Methylnaphthalene		0.688	0.689	0.719	0.669	0.659	0.677	3.4
Hexachlorocyclopentadiene			0.138	0.180	0.204	0.205	0.192	15.4
2,4,6-Trichlorophenol		0.364	0.365	0.389	0.377	0.373	0.374	2.3
2-Fluorobiphenyl		1.373	1.360	1.385	1.269	1.241	1.298	5.6
2,4,5-Trichlorophenol		0.385	0.407	0.430	0.410	0.394	0.405	3.5
1,1-Biphenyl		1.582	1.564	1.632	1.545	1.517	1.551	3.2
2-Chloronaphthalene		1.167	1.161	1.211	1.135	1.114	1.147	3.2
2-Nitroaniline		0.371	0.371	0.394	0.390	0.382	0.385	2.9
Dimethylphthalate		1.382	1.359	1.408	1.345	1.324	1.358	2.1
Acenaphthylene		1.739	1.736	1.791	1.696	1.660	1.705	3.2

All other compounds must meet a minimum RRF of 0.010.

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG No.: Q1232

Instrument ID: BNA_F Calibration Date(s): 02/06/2025 02/06/2025

Calibration Time(s): 11:07 14:14

LAB FILE ID: RRF2.5 = BF141472.D RRF005 = BF141473.D RRF010 = BF141474.D
RRF020 = BF141475.D RRF040 = BF141476.D RRF050 = BF141477.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.289	0.290	0.314	0.301	0.295	0.297	3.1
3-Nitroaniline		0.312	0.313	0.335	0.319	0.320	0.319	2.4
Acenaphthene		1.186	1.164	1.188	1.152	1.108	1.147	3.1
2,4-Dinitrophenol			0.133	0.167	0.177	0.185	0.176	13.8
4-Nitrophenol		0.190	0.216	0.247	0.249	0.249	0.237	10.4
Dibenzofuran		1.768	1.722	1.774	1.673	1.633	1.683	4.4
2,4-Dinitrotoluene		0.383	0.403	0.414	0.398	0.390	0.395	2.8
Diethylphthalate		1.400	1.346	1.398	1.330	1.300	1.336	3.7
4-Chlorophenyl-phenylether		0.661	0.643	0.659	0.622	0.602	0.626	4.7
Fluorene		1.359	1.361	1.377	1.290	1.239	1.301	4.9
4-Nitroaniline		0.315	0.319	0.333	0.335	0.319	0.324	2.3
4,6-Dinitro-2-methylphenol			0.116	0.140	0.142	0.142	0.139	8.5
n-Nitrosodiphenylamine		0.650	0.653	0.674	0.627	0.625	0.640	3.0
2,4,6-Tribromophenol		0.201	0.200	0.210	0.200	0.199	0.201	2.1
4-Bromophenyl-phenylether		0.228	0.220	0.234	0.221	0.220	0.224	2.5
Hexachlorobenzene		0.229	0.234	0.241	0.227	0.224	0.230	2.5
Atrazine		0.200	0.203	0.214	0.185	0.185	0.192	7.1
Pentachlorophenol		0.116	0.131	0.151	0.154	0.158	0.147	11.7
Phenanthrene		1.141	1.132	1.161	1.084	1.071	1.101	3.9
Anthracene		1.168	1.126	1.158	1.082	1.084	1.107	4.0
Carbazole		1.022	1.026	1.076	0.986	0.966	0.996	4.9
Di-n-butylphthalate		1.165	1.156	1.223	1.141	1.132	1.150	3.3
Fluoranthene		1.251	1.239	1.243	1.151	1.129	1.167	6.6
Pyrene		1.528	1.557	1.717	1.745	1.774	1.703	6.7
Terphenyl-d14		1.129	1.104	1.211	1.198	1.209	1.185	4.1
Butylbenzylphthalate		0.525	0.540	0.605	0.609	0.603	0.590	6.8
3,3-Dichlorobenzidine		0.397	0.411	0.407	0.377	0.363	0.381	6.3
Benzo(a)anthracene		1.381	1.332	1.350	1.345	1.309	1.329	2.5
Chrysene		1.187	1.223	1.314	1.201	1.215	1.228	3.3
Bis(2-ethylhexyl)phthalate		0.593	0.610	0.685	0.689	0.680	0.665	6.7
Di-n-octyl phthalate		0.802	0.814	0.910	0.965	1.010	0.949	11.8
Benzo(b)fluoranthene		1.286	1.407	1.398	1.336	1.350	1.343	3.9
Benzo(k)fluoranthene		1.113	1.216	1.247	1.142	1.074	1.147	5.5
Benzo(a)pyrene		1.085	1.074	1.115	1.080	1.069	1.087	1.4
Indeno(1,2,3-cd)pyrene		0.987	1.029	1.186	1.288	1.341	1.236	14.1

All other compounds must meet a minimum RRF of 0.010.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG No.: Q1232

Instrument ID: BNA_F Calibration Date(s): 02/06/2025 02/06/2025

Calibration Time(s): 11:07 14:14

LAB FILE ID: RRF2.5 = BF141472.D RRF005 = BF141473.D RRF010 = BF141474.D
RRF020 = BF141475.D RRF040 = BF141476.D RRF050 = BF141477.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo(a,h)anthracene		0.796	0.828	0.970	1.051	1.092	1.001	14.1
Benzo(g,h,i)perylene		0.805	0.864	1.001	1.100	1.144	1.048	15.6
1,2,4,5-Tetrachlorobenzene		0.582	0.588	0.600	0.584	0.560	0.579	2.5
1,4-Dioxane		0.511	0.519	0.523	0.517	0.502	0.513	1.7
2,3,4,6-Tetrachlorophenol		0.332	0.361	0.359	0.348	0.348	0.349	3.0

All other compounds must meet a minimum RRF of 0.010.