

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bf021025

SequenceID : bf021025

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	<u>✓</u>
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	<u>✓</u>
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	<u>✓</u>
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	<u>✓</u>
5. GC/MS Calibration Met:	_____	_____	<u>✓</u>
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	<u>✓</u>
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. Compound #85 is biased high in SSTDCCC (BF141531.D). If any samples are found with Hit of this compound they will be analyzed again with proper passing CCC.	_____	<u>✓</u>	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	<u>✓</u>	_____
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria	_____	_____	<u>✓</u>
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
a. B/N Fraction			
d. Acid Fraction			
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria	_____	<u>✓</u>	_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
a. B/N Fraction	<u>The Recovery and RPD fail for some compound in Q1299-04 MS/MSD due to matrix interference. Hence, no corrective action is required.</u>		
d. Acid Fraction			
9. Internal Standard Area/Retention Time Shift Meet Criteria	_____	_____	<u>✓</u>
Comments:			
10. Extraction Holding Time Met	_____	_____	<u>✓</u>
If not met, list number of days exceeded for each sample:			
11. Analysis Holding Time Met	_____	<u>✓</u>	_____
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

Recovery and RPD fail high for some compound in PB166619BS/BSD. This data will be used for hardcopies.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bf021025

SequenceID : bf021025 NA NO YES