

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1239-10MS	Client Sample ID:	357MS					DataFile:	BF141345.D		
Benzaldehyde	1200	0	1000	ug/Kg	83				10	86	
Phenol	1200	0	880	ug/Kg	73				67	126	
bis(2-Chloroethyl)ether	1200	0	900	ug/Kg	75				54	125	
2-Chlorophenol	1200	0	900	ug/Kg	75	*			79	107	
2-Methylphenol	1200	0	860	ug/Kg	72				66	122	
2,2-oxybis(1-Chloropropane)	1200	0	880	ug/Kg	73				65	110	
Acetophenone	1200	0	990	ug/Kg	83				75	111	
3+4-Methylphenols	1200	0	920	ug/Kg	77				66	104	
N-Nitroso-di-n-propylamine	1200	0	900	ug/Kg	75				59	119	
Hexachloroethane	1200	0	860	ug/Kg	72				65	117	
Nitrobenzene	1200	0	840	ug/Kg	70				70	119	
Isophorone	1200	0	900	ug/Kg	75	*			76	122	
2-Nitrophenol	1200	0	900	ug/Kg	75				54	145	
2,4-Dimethylphenol	1200	0	1100	ug/Kg	92				44	135	
bis(2-Chloroethoxy)methane	1200	0	880	ug/Kg	73				68	112	
2,4-Dichlorophenol	1200	0	900	ug/Kg	75				72	118	
Naphthalene	1200	0	850	ug/Kg	71	*			72	110	
4-Chloroaniline	1200	0	210	ug/Kg	18				10	91	
Hexachlorobutadiene	1200	0	840	ug/Kg	70				66	114	
Caprolactam	1200	0	990	ug/Kg	83				51	134	
4-Chloro-3-methylphenol	1200	0	870	ug/Kg	73				57	132	
2-Methylnaphthalene	1200	0	840	ug/Kg	70				59	123	
Hexachlorocyclopentadiene	2400	0	3200	ug/Kg	133				10	175	
2,4,6-Trichlorophenol	1200	0	920	ug/Kg	77				72	117	
2,4,5-Trichlorophenol	1200	0	890	ug/Kg	74				72	117	
1,1-Biphenyl	1200	0	990	ug/Kg	83				75	113	
2-Chloronaphthalene	1200	0	850	ug/Kg	71				67	118	
2-Nitroaniline	1200	0	900	ug/Kg	75				69	127	
Dimethylphthalate	1200	0	910	ug/Kg	76				70	113	
Acenaphthylene	1200	0	940	ug/Kg	78	*			79	118	
2,6-Dinitrotoluene	1200	0	860	ug/Kg	72				70	125	
3-Nitroaniline	1200	0	530	ug/Kg	44				30	99	
Acenaphthene	1200	0	980	ug/Kg	82				70	121	
2,4-Dinitrophenol	2400	0	2200	ug/Kg	92				10	155	
4-Nitrophenol	2400	0	1600	ug/Kg	67				45	133	
Dibenzofuran	1200	0	870	ug/Kg	73				72	110	
2,4-Dinitrotoluene	1200	0	880	ug/Kg	73				55	128	
Diethylphthalate	1200	0	900	ug/Kg	75				70	112	
4-Chlorophenyl-phenylether	1200	0	850	ug/Kg	71				71	108	
Fluorene	1200	0	850	ug/Kg	71				68	116	
4-Nitroaniline	1200	0	800	ug/Kg	67				55	120	
4,6-Dinitro-2-methylphenol	1200	0	1100	ug/Kg	92				10	160	
N-Nitrosodiphenylamine	1200	0	990	ug/Kg	83				73	118	
4-Bromophenyl-phenylether	1200	0	970	ug/Kg	81				65	121	
Hexachlorobenzene	1200	0	940	ug/Kg	78				67	118	
Atrazine	1200	0	1000	ug/Kg	83				79	127	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Pentachlorophenol	2400	0	1800	ug/Kg	75				47	128	
Phenanthrene	1200	0	940	ug/Kg	78				52	128	
Anthracene	1200	0	950	ug/Kg	79				62	124	
Carbazole	1200	0	850	ug/Kg	71				59	119	
Di-n-butylphthalate	1200	0	960	ug/Kg	80				69	118	
Fluoranthene	1200	0	840	ug/Kg	70				44	125	
Pyrene	1200	0	740	ug/Kg	62				26	142	
Butylbenzylphthalate	1200	0	860	ug/Kg	72				64	126	
3,3-Dichlorobenzidine	1200	0	670	ug/Kg	56				33	116	
Benzo(a)anthracene	1200	0	860	ug/Kg	72				71	114	
Chrysene	1200	0	910	ug/Kg	76				57	121	
bis(2-Ethylhexyl)phthalate	1200	0	890	ug/Kg	74				42	169	
Di-n-octyl phthalate	1200	0	1000	ug/Kg	83				23	175	
Benzo(b)fluoranthene	1200	0	1000	ug/Kg	83				67	121	
Benzo(k)fluoranthene	1200	0	890	ug/Kg	74				57	134	
Benzo(a)pyrene	1200	0	990	ug/Kg	83				70	142	
Indeno(1,2,3-cd)pyrene	1200	0	720	ug/Kg	60				40	129	
Dibenz(a,h)anthracene	1200	0	730	ug/Kg	61				43	123	
Benzo(g,h,i)perylene	1200	0	640	ug/Kg	53				24	125	
1,2,4,5-Tetrachlorobenzene	1200	0	960	ug/Kg	80				69	124	
1,4-Dioxane	1200	0	900	ug/Kg	75				46	112	
2,3,4,6-Tetrachlorophenol	1200	0	930	ug/Kg	78				69	112	