

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1239-10MSD	Client Sample ID:	357MSD					DataFile:	BF141346.D		
Benzaldehyde	1200	0	1100	ug/Kg	92	*	10		10	86	20
Phenol	1200	0	920	ug/Kg	77		5		67	126	20
bis(2-Chloroethyl)ether	1200	0	940	ug/Kg	78		4		54	125	20
2-Chlorophenol	1200	0	950	ug/Kg	79		5		79	107	20
2-Methylphenol	1200	0	900	ug/Kg	75		4		66	122	20
2,2-oxybis(1-Chloropropane)	1200	0	920	ug/Kg	77		5		65	110	20
Acetophenone	1200	0	1100	ug/Kg	92		10		75	111	20
3+4-Methylphenols	1200	0	970	ug/Kg	81		5		66	104	20
N-Nitroso-di-n-propylamine	1200	0	940	ug/Kg	78		4		59	119	20
Hexachloroethane	1200	0	910	ug/Kg	76		5		65	117	20
Nitrobenzene	1200	0	910	ug/Kg	76		8		70	119	20
Isophorone	1200	0	960	ug/Kg	80		6		76	122	20
2-Nitrophenol	1200	0	970	ug/Kg	81		8		54	145	20
2,4-Dimethylphenol	1200	0	1200	ug/Kg	100		8		44	135	20
bis(2-Chloroethoxy)methane	1200	0	950	ug/Kg	79		8		68	112	20
2,4-Dichlorophenol	1200	0	950	ug/Kg	79		5		72	118	20
Naphthalene	1200	0	900	ug/Kg	75		5		72	110	20
4-Chloroaniline	1200	0	240	ug/Kg	20		11		10	91	20
Hexachlorobutadiene	1200	0	890	ug/Kg	74		6		66	114	20
Caprolactam	1200	0	1100	ug/Kg	92		10		51	134	20
4-Chloro-3-methylphenol	1200	0	930	ug/Kg	78		7		57	132	20
2-Methylnaphthalene	1200	0	890	ug/Kg	74		6		59	123	20
Hexachlorocyclopentadiene	2400	0	3400	ug/Kg	142		7		10	175	20
2,4,6-Trichlorophenol	1200	0	970	ug/Kg	81		5		72	117	20
2,4,5-Trichlorophenol	1200	0	950	ug/Kg	79		7		72	117	20
1,1-Biphenyl	1200	0	1000	ug/Kg	83		0		75	113	20
2-Chloronaphthalene	1200	0	900	ug/Kg	75		5		67	118	20
2-Nitroaniline	1200	0	950	ug/Kg	79		5		69	127	20
Dimethylphthalate	1200	0	970	ug/Kg	81		6		70	113	20
Acenaphthylene	1200	0	990	ug/Kg	83		6		79	118	20
2,6-Dinitrotoluene	1200	0	910	ug/Kg	76		5		70	125	20
3-Nitroaniline	1200	0	550	ug/Kg	46		4		30	99	20
Acenaphthene	1200	0	1000	ug/Kg	83		1		70	121	20
2,4-Dinitrophenol	2400	0	2300	ug/Kg	96		4		10	155	20
4-Nitrophenol	2400	0	1700	ug/Kg	71		6		45	133	20
Dibenzofuran	1200	0	920	ug/Kg	77		5		72	110	20
2,4-Dinitrotoluene	1200	0	930	ug/Kg	78		7		55	128	20
Diethylphthalate	1200	0	950	ug/Kg	79		5		70	112	20
4-Chlorophenyl-phenylether	1200	0	900	ug/Kg	75		5		71	108	20
Fluorene	1200	0	900	ug/Kg	75		5		68	116	20
4-Nitroaniline	1200	0	870	ug/Kg	73		9		55	120	20
4,6-Dinitro-2-methylphenol	1200	0	1200	ug/Kg	100		8		10	160	20
N-Nitrosodiphenylamine	1200	0	1100	ug/Kg	92		10		73	118	20
4-Bromophenyl-phenylether	1200	0	1000	ug/Kg	83		2		65	121	20
Hexachlorobenzene	1200	0	990	ug/Kg	83		6		67	118	20
Atrazine	1200	0	1100	ug/Kg	92		10		79	127	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Pentachlorophenol	2400	0	2000	ug/Kg	83		10		47	128	20
Phenanthrene	1200	0	1000	ug/Kg	83		6		52	128	20
Anthracene	1200	0	1000	ug/Kg	83		5		62	124	20
Carbazole	1200	0	920	ug/Kg	77		8		59	119	20
Di-n-butylphthalate	1200	0	1000	ug/Kg	83		4		69	118	20
Fluoranthene	1200	0	920	ug/Kg	77		10		44	125	20
Pyrene	1200	0	810	ug/Kg	68		9		26	142	20
Butylbenzylphthalate	1200	0	930	ug/Kg	78		8		64	126	20
3,3-Dichlorobenzidine	1200	0	750	ug/Kg	63		12		33	116	20
Benzo(a)anthracene	1200	0	930	ug/Kg	78		8		71	114	20
Chrysene	1200	0	980	ug/Kg	82		8		57	121	20
bis(2-Ethylhexyl)phthalate	1200	0	970	ug/Kg	81		9		42	169	20
Di-n-octyl phthalate	1200	0	1100	ug/Kg	92		10		23	175	20
Benzo(b)fluoranthene	1200	0	990	ug/Kg	83		0		67	121	20
Benzo(k)fluoranthene	1200	0	1100	ug/Kg	92		22	*	57	134	20
Benzo(a)pyrene	1200	0	1100	ug/Kg	92		10		70	142	20
Indeno(1,2,3-cd)pyrene	1200	0	770	ug/Kg	64		6		40	129	20
Dibenz(a,h)anthracene	1200	0	780	ug/Kg	65		6		43	123	20
Benzo(g,h,i)perylene	1200	0	690	ug/Kg	58		9		24	125	20
1,2,4,5-Tetrachlorobenzene	1200	0	1000	ug/Kg	83		4		69	124	20
1,4-Dioxane	1200	0	950	ug/Kg	79		5		46	112	20
2,3,4,6-Tetrachlorophenol	1200	0	1000	ug/Kg	83		6		69	112	20