

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/07/2025

Lab File ID: BF141506.D Time Analyzed: 14:33

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	91880	6.787	360290	8.08	198348	9.83
UPPER LIMIT	183760	7.287	720580	8.575	396696	10.328
LOWER LIMIT	45940	6.287	180145	7.575	99174	9.328
EPA SAMPLE NO.						
01 JPP-42.2-012925	70012	6.79	264489	8.08	130756	9.83
02 JPP-51.1-012925	82126	6.79	311011	8.08	153227	9.83
03 JPP-46.1-012925	83684	6.79	306437	8.08	148193	9.83
04 JPP-42.1-012925	81132	6.79	300029	8.08	147138	9.83

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.