

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/31/2025

Lab File ID: BF141336.D Time Analyzed: 17:19

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	493954	11.322	281629	13.968	286623	15.433
	UPPER LIMIT	987908	11.822	563258	14.468	573246	15.933
	LOWER LIMIT	246977	10.822	140815	13.468	143312	14.933
	EPA SAMPLE NO.						
01	357MS	473142	11.32	293043	13.97	307605	15.43
02	357MSD	457150	11.32	288493	13.97	294341	15.42

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.