

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF020725

Review By	Rahul	Review On	2/11/2025 4:20:20 PM
Supervise By	mohammad	Supervise On	2/12/2025 7:26:07 AM
SubDirectory	BF020725	HP Acquire Method	BNA_F
		HP Processing Method	bf020625

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12650,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141495.D	07 Feb 2025 09:28		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF141496.D	07 Feb 2025 09:55		RC/JU	Ok
3	PB166588BL	PB166588BL	BF141497.D	07 Feb 2025 10:21		RC/JU	Ok
4	PB166588BS	PB166588BS	BF141498.D	07 Feb 2025 10:47		RC/JU	Ok,M
5	PB166567BL	PB166567BL	BF141499.D	07 Feb 2025 11:13		RC/JU	Ok
6	PB166563BS	PB166563BS	BF141500.D	07 Feb 2025 11:40		RC/JU	Ok,M
7	PB166567BS	PB166567BS	BF141501.D	07 Feb 2025 12:06		RC/JU	Ok,M
8	PB166567BSD	PB166567BSD	BF141502.D	07 Feb 2025 12:32		RC/JU	Ok,M
9	PB166580BL	PB166580BL	BF141503.D	07 Feb 2025 12:58		RC/JU	Ok
10	PB166580BS	PB166580BS	BF141504.D	07 Feb 2025 13:25		RC/JU	Ok,M
11	DFTPP	DFTPP	BF141505.D	07 Feb 2025 14:07		RC/JU	Ok
12	SSTDCCC040	SSTDCCC040	BF141506.D	07 Feb 2025 14:33		RC/JU	Ok,M
13	PB166528TB	PB166528TB	BF141507.D	07 Feb 2025 15:00		RC/JU	Ok
14	Q1285-02	72-11977	BF141508.D	07 Feb 2025 15:31		RC/JU	Ok
15	Q1304-01	RB21140	BF141509.D	07 Feb 2025 15:58	Internal Standard Fail	RC/JU	Ok
16	Q1299-01	WC-1	BF141510.D	07 Feb 2025 16:24		RC/JU	Ok
17	Q1299-05	WC-2	BF141511.D	07 Feb 2025 16:50		RC/JU	Ok,M
18	Q1288-01	HR-01-020425	BF141512.D	07 Feb 2025 17:16	Need 2X Dilution	RC/JU	Dilution

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF020725

Review By	Rahul	Review On	2/11/2025 4:20:20 PM		
Supervise By	mohammad	Supervise On	2/12/2025 7:26:07 AM		
SubDirectory	BF020725	HP Acquire Method	BNA_F	HP Processing Method	bf020625
STD. NAME		STD REF.#			
Tune/Reschk		SP6717			
Initial Calibration Stds		SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729			
CCC		SP6725			
Internal Standard/PEM		S12650,10ul/1000ul sample			
ICV/I.BLK		SP6686			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1288-01MS	HR-01-020425MS	BF141513.D	07 Feb 2025 17:43		RC/JU	Ok,M
20	Q1288-01MSD	HR-01-020425MSD	BF141514.D	07 Feb 2025 18:09		RC/JU	Ok,M
21	Q1215-07	JPP-29.2-012825	BF141515.D	07 Feb 2025 18:35	Need 2X Dilution	RC/JU	Dilution
22	Q1216-07	JPP-21.1-012825	BF141516.D	07 Feb 2025 19:02		RC/JU	Ok,M
23	Q1216-03	JPP-18.1-012825	BF141517.D	07 Feb 2025 19:28		RC/JU	Ok,M
24	Q1241-03	JPP-3.5-013025	BF141518.D	07 Feb 2025 19:54	Need 5X Dilution	RC/JU	Dilution
25	Q1232-15	JPP-42.2-012925	BF141519.D	07 Feb 2025 20:21	Need 5X Dilution	RC/JU	Dilution
26	Q1216-15	JPP-26.1-012825	BF141520.D	07 Feb 2025 20:47	Internal Standard Fail, Need 5X Dilution	RC/JU	Dilution
27	Q1241-11	JPP-5.2-013025	BF141521.D	07 Feb 2025 21:13	Surrogate Fail	RC/JU	Ok,M
28	Q1235-07	JPP-16.1-012925	BF141522.D	07 Feb 2025 21:40		RC/JU	Ok,M
29	Q1232-07	JPP-46.1-012925	BF141523.D	07 Feb 2025 22:07		RC/JU	Ok,M
30	Q1241-19	JPP-51.4-013025	BF141524.D	07 Feb 2025 22:33		RC/JU	Ok,M
31	Q1241-15	JPP-5.4-013025	BF141525.D	07 Feb 2025 23:00	Need further 5X Dilution	RC/JU	Dilution
32	Q1242-03	JPP-6.2-013025	BF141526.D	07 Feb 2025 23:27	Surrogate Fail	RC/JU	Ok,M
33	Q1232-11	JPP-42.1-012925	BF141527.D	07 Feb 2025 23:53		RC/JU	Ok,M
34	Q1232-19	JPP-51.1-012925	BF141528.D	08 Feb 2025 00:20		RC/JU	Ok,M
35	Q1235-03	JPP-51.2-012925	BF141529.D	08 Feb 2025 00:47	Surrogate Fail	RC/JU	Ok,M

M : Manual Integration