

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG No.: Q1232

Instrument ID: BNA_F Calibration Date/Time: 02/03/2025 11:24

Lab File ID: BF141361.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.380	1.326		-3.9	
2-Fluorophenol	1.303	1.313		0.8	
Phenol-d6	1.657	1.641		-1.0	
1,4-Dichlorobenzene	1.547	1.568		1.4	20.0
2-Methylphenol	1.135	1.098		-3.3	
3+4-Methylphenols	1.393	1.446		3.8	
Nitrobenzene-d5	0.379	0.397		4.7	
Hexachloroethane	0.571	0.590		3.3	
Nitrobenzene	0.393	0.407		3.6	
Hexachlorobutadiene	0.194	0.207		6.7	20.0
2,4,6-Trichlorophenol	0.372	0.384		3.2	20.0
2-Fluorobiphenyl	1.299	1.315		1.2	
2,4,5-Trichlorophenol	0.405	0.416		2.7	
2,4-Dinitrotoluene	0.406	0.426		4.9	
2,4,6-Tribromophenol	0.183	0.200		9.3	
Hexachlorobenzene	0.238	0.249		4.6	
Pentachlorophenol	0.139	0.152		9.4	20.0
Terphenyl-d14	1.297	1.144		-11.8	

All other compounds must meet a minimum RRF of 0.010.