

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG No.: Q1232

Instrument ID: BNA_F Calibration Date/Time: 02/03/2025 17:14

Lab File ID: BF141373.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.380	1.232		-10.7	
2-Fluorophenol	1.303	1.302		-0.1	
Phenol-d6	1.657	1.598		-3.6	
1,4-Dichlorobenzene	1.547	1.570		1.5	20.0
2-Methylphenol	1.135	1.109		-2.3	
3+4-Methylphenols	1.393	1.420		1.9	
Nitrobenzene-d5	0.379	0.382		0.8	
Hexachloroethane	0.571	0.590		3.3	
Nitrobenzene	0.393	0.389		-1.0	
Hexachlorobutadiene	0.194	0.204		5.2	20.0
2,4,6-Trichlorophenol	0.372	0.380		2.2	20.0
2-Fluorobiphenyl	1.299	1.292		-0.5	
2,4,5-Trichlorophenol	0.405	0.413		2.0	
2,4-Dinitrotoluene	0.406	0.423		4.2	
2,4,6-Tribromophenol	0.183	0.194		6.0	
Hexachlorobenzene	0.238	0.244		2.5	
Pentachlorophenol	0.139	0.157		12.9	20.0
Terphenyl-d14	1.297	1.274		-1.8	

All other compounds must meet a minimum RRF of 0.010.