

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/31/2025

Lab File ID: BF141336.D Time Analyzed: 17:19

Instrument ID: BNA\_F GC Column: DB-UI ID: 0.18 (mm)

|                   | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #  | IS3 (ANT)<br>AREA # | RT #   |
|-------------------|---------------------|-------|---------------------|-------|---------------------|--------|
| 12 HOUR STD       | 140033              | 6.798 | 558386              | 8.08  | 299701              | 9.84   |
| UPPER LIMIT       | 280066              | 7.298 | 1116770             | 8.581 | 599402              | 10.339 |
| LOWER LIMIT       | 70016.5             | 6.298 | 279193              | 7.581 | 149851              | 9.339  |
| EPA SAMPLE NO.    |                     |       |                     |       |                     |        |
| 01 BELL-25-002MS  | 131554              | 6.80  | 434471              | 8.08  | 208861              | 9.84   |
| 02 BELL-25-002MSD | 142670              | 6.80  | 521314              | 8.09  | 271292              | 9.85   |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.