

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: RUTW01

Lab Code: CHEM

SAS No.: Q1232

SDG NO.: Q1232

Lab File ID: BF141335.D

DFTPP Injection Date: 01/31/2025

Instrument ID: BNA_F

DFTPP Injection Time: 16:53

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	46.2
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	38.1
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	49
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	25.8
365	Greater than 1% of mass 198	2.8
441	Present, but less than mass 443	11.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.6 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF141336.D	01/31/2025	17:19
BELL-25-002MS	Q1218-02MS	BF141339.D	01/31/2025	18:44
BELL-25-002MSD	Q1218-02MSD	BF141340.D	01/31/2025	19:11