

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF012925

Review By	Jagrut	Review On	1/31/2025 4:39:47 PM
Supervise By	mohammad	Supervise On	2/4/2025 7:21:59 AM
SubDirectory	BF012925	HP Acquire Method	BNA_F
		HP Processing Method	bf012925

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12649,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141289.D	29 Jan 2025 10:03		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF141290.D	29 Jan 2025 10:30		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF141291.D	29 Jan 2025 10:56	Compound #54 is removed from 5ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF141292.D	29 Jan 2025 11:49		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF141293.D	29 Jan 2025 15:17	The Calibration is fail for compound #77	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF141294.D	29 Jan 2025 15:45	The Calibration is Good For 8270 DOD Except com#77 and good for 625.1 Method Except com#77	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF141295.D	29 Jan 2025 16:11		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF141296.D	29 Jan 2025 16:38		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF141297.D	29 Jan 2025 17:05		RC/JU	Ok
10	SSTDICV040	ICVBF012925	BF141298.D	29 Jan 2025 17:31		RC/JU	Ok,M
11	PB166167BL	PB166167BL	BF141299.D	29 Jan 2025 18:24		RC/JU	Ok

M : Manual Integration