

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG No.: Q1232

Instrument ID: MSVOA_X Calibration Date/Time: 02/03/2025 09:52

Lab File ID: VX044802.D Init. Calib. Date(s): 01/15/2025 01/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:31 13:37

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.719	0.766		6.54	20
1,1-Dichloroethene	0.585	0.636		8.72	20
2-Butanone	0.444	0.459		3.38	20
Carbon Tetrachloride	0.497	0.454		-8.65	20
Chloroform	1.186	1.162		-2.02	20
Benzene	1.370	1.440		5.11	20
1,2-Dichloroethane	0.500	0.454		-9.2	20
Trichloroethene	0.329	0.337		2.43	20
Tetrachloroethene	0.314	0.312		-0.64	20
Chlorobenzene	1.062	1.101	0.3	3.67	20
1,2-Dichloroethane-d4	0.755	0.726		-3.84	20
Dibromofluoromethane	0.318	0.333		4.72	20
Toluene-d8	1.108	1.249		12.73	20
4-Bromofluorobenzene	0.413	0.437		5.81	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.