

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1232

Client: RU2 Engineering, LLC

Analytical Method: SW8260D

Datafile : VX044805.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0203WBS01	Vinyl chloride	20	20.2	ug/L	101			65	117	
	1,1-Dichloroethene	20	16.8	ug/L	84			74	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	17.6	ug/L	88			77	113	
	Chloroform	20	19.3	ug/L	97			79	113	
	Benzene	20	20.7	ug/L	104			82	109	
	1,2-Dichloroethane	20	18.0	ug/L	90			80	115	
	Trichloroethene	20	19.4	ug/L	97			77	113	
	Tetrachloroethene	20	18.9	ug/L	95			67	123	
	Chlorobenzene	20	19.9	ug/L	100			82	109	