

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232
 Lab File ID: VN085617.D Date Analyzed: 02/03/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:56
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	152662	13.788				
UPPER LIMIT	305324	14.288				
LOWER LIMIT	76331	13.288				
EPA SAMPLE NO.						
JPP-46.2-012925	109751	13.79				
JPP-42.1-012925	158909	13.79				
VN0203WBL01	111992	13.79				
VN0203WBS01	141185	13.79				
VN0203WBSD01	132351	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.