

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232
 Lab File ID: VX044802.D Date Analyzed: 02/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:52
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	101849	12.018				
UPPER LIMIT	203698	12.518				
LOWER LIMIT	50924.5	11.518				
EPA SAMPLE NO.						
JPP-46.1-012925	91814	12.02				
JPP-42.2-012925	90092	12.02				
JPP-51.1-012925	85623	12.02				
VX0203WBL01	100905	12.02				
VX0203WBS01	107943	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.