

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RUTW01
Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232
Lab File ID: VN085616.D BFB Injection Date: 02/03/2025
Instrument ID: MSVOA_N BFB Injection Time: 12:23
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.2
75	30.0 - 60.0% of mass 95	54.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	1.4 (1.9) 1
174	50.0 - 100.0% of mass 95	76.2
175	5.0 - 9.0% of mass 174	5.5 (7.2) 1
176	95.0 - 101.0% of mass 174	75 (98.4) 1
177	5.0 - 9.0% of mass 176	5.1 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN085617.D	02/03/2025	12:56
VN0203WBL01	VN0203WBL01	VN085618.D	02/03/2025	13:34
VN0203WBS01	VN0203WBS01	VN085619.D	02/03/2025	13:58
VN0203WBSD01	VN0203WBSD01	VN085620.D	02/03/2025	14:22
JPP-46.2-012925	Q1232-02	VN085633.D	02/03/2025	19:37
JPP-42.1-012925	Q1232-10	VN085642.D	02/03/2025	23:14