

Surrogate Summary

SDG No.: **Q1232**

Client: **RU2 Engineering, LLC**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1232-01	JPP-46.2-012925	1,2-Dichloroethane-d4	50	66.4	133	50	163
		Dibromofluoromethane	50	53.5	107	54	147
		Toluene-d8	50	50.5	101	58	134
		4-Bromofluorobenzene	50	50.7	101	29	146
Q1232-05	JPP-46.1-012925	1,2-Dichloroethane-d4	50	62.2	124	50	163
		Dibromofluoromethane	50	53.0	106	54	147
		Toluene-d8	50	51.0	102	58	134
		4-Bromofluorobenzene	50	50.0	100	29	146
Q1232-09	JPP-42.1-012925	1,2-Dichloroethane-d4	50	59.4	119	50	163
		Dibromofluoromethane	50	40.1	80	54	147
		Toluene-d8	50	50.4	101	58	134
		4-Bromofluorobenzene	50	47.8	96	29	146
Q1232-13	JPP-42.2-012925	1,2-Dichloroethane-d4	50	63.8	128	50	163
		Dibromofluoromethane	50	49.4	99	54	147
		Toluene-d8	50	50.3	100	58	134
		4-Bromofluorobenzene	50	50.1	100	29	146
Q1232-17	JPP-51.1-012925	1,2-Dichloroethane-d4	50	65.4	131	50	163
		Dibromofluoromethane	50	53.8	107	54	147
		Toluene-d8	50	51.1	102	58	134
		4-Bromofluorobenzene	50	51.0	102	29	146
VY0204SBL01	VY0204SBL01	1,2-Dichloroethane-d4	50	57.4	115	50	163
		Dibromofluoromethane	50	51.6	103	54	147
		Toluene-d8	50	49.8	100	58	134
		4-Bromofluorobenzene	50	47.5	95	29	146
VY0204SBS01	VY0204SBS01	1,2-Dichloroethane-d4	50	44.3	89	50	163
		Dibromofluoromethane	50	44.6	89	54	147
		Toluene-d8	50	45.4	91	58	134
		4-Bromofluorobenzene	50	45.0	90	29	146
VY0204SBSD01	VY0204SBSD01	1,2-Dichloroethane-d4	50	55.2	110	50	163
		Dibromofluoromethane	50	53.3	107	54	147
		Toluene-d8	50	54.0	108	58	134
		4-Bromofluorobenzene	50	56.3	113	29	146