

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RUTW01
Lab Code: CHEM Case No.: Q1232 SAS No.: Q1232 SDG NO.: Q1232
Lab File ID: VY021045.D BFB Injection Date: 02/04/2025
Instrument ID: MSVOA_Y BFB Injection Time: 09:16
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.3
75	30.0 - 60.0% of mass 95	55.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	1.3 (1.5) 1
174	50.0 - 100.0% of mass 95	88.1
175	5.0 - 9.0% of mass 174	7 (7.9) 1
176	95.0 - 101.0% of mass 174	84.8 (96.1) 1
177	5.0 - 9.0% of mass 176	5.3 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY021046.D	02/04/2025	09:47
VY0204SBL01	VY0204SBL01	VY021047.D	02/04/2025	10:15
VY0204SBS01	VY0204SBS01	VY021048.D	02/04/2025	10:52
VY0204SBSD01	VY0204SBSD01	VY021049.D	02/04/2025	11:15
JPP-46.2-012925	Q1232-01	VY021056.D	02/04/2025	14:16
JPP-46.1-012925	Q1232-05	VY021057.D	02/04/2025	14:39
JPP-42.1-012925	Q1232-09	VY021058.D	02/04/2025	15:03
JPP-42.2-012925	Q1232-13	VY021059.D	02/04/2025	15:26
JPP-51.1-012925	Q1232-17	VY021060.D	02/04/2025	15:50