



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

RU2 Engineering, LLC

Project Name: NYCDDC SANTWOBR Brooklyn Bridge BBMCR

Project # N/A

Chemtech Project # Q1235

Test Name: SVOC-TCL BNA -20

A. Number of Samples and Date of Receipt:

8 Solid samples were received on 01/30/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Diesel Range Organics, Gasoline Range Organics, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, Paint Filter, PCB, Pesticide-TCL, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL and VOCMS Group1. This data package contains results for SVOC-TCL BNA -20.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df The analysis of SVOC-TCL BNA -20 was based on method 8270E and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for JPP-51.2-012925 [2,4,6-Tribromophenol - 3% and 2-Fluorophenol - 13%], due to dirty and concentrated matrix as corrective action lab reanalyzed the sample but surrogate is failing therefore reported run is Final.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {Q1239-10MS} with File ID: BF141345.D recoveries met the requirements for all compounds except for 2-Chlorophenol[75%], Acenaphthylene[78%], Isophorone [75%] and Naphthalene[71%], due to Matrix interference therefore no corrective action is required.

The MSD {Q1239-10MSD} with File ID: BF141346.D recoveries met the acceptable requirements except for Benzaldehyde[92%], due to Matrix interference therefore no corrective action is required.

The RPD for {Q1239-10MSD} with File ID: BF141346.D met criteria except for Benzo(k)fluoranthene[22%], due to Matrix interference therefore no corrective action is required.

The Blank Spike for {PB166418BS} with File ID: BF141441.D met requirements for all samples except for Acetophenone[100%], Hexachlorocyclopentadiene[191%], The associate samples have no positive hit for these compounds therefore no corrective action was taken.

The Blank analysis did not indicate the presence of lab contamination.
The Initial Calibration met the requirements .

The Continuous Calibration File ID BF141336.D met the requirements except for 4-Nitrophenol and Pentachlorophenol, The associate samples have no positive hit for these compounds therefore no corrective action was taken.

The Tuning criteria met requirements.

Samples JPP-51.2-012925, JPP-16.1-012925 analyzed with direct 2X dilution due to viscous and concentrated matrix..

E. Additional Comments:

The Sample JPP-51.2-012925, have the concentration of target compound below method detection limits; therefore it is not reported as Hit in Form1.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____