GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1235   MATRIX: Solid

METHOD: 8081B/3541

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified.			✓
2.	Standard Summary Submitted.			✓
3.	Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD.			✓

The Initial Calibration met the requirements .
The Continuous Calibration met the requirements .

		NA	NO	YES
4.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5.	Surrogate Recoveries Meet Criteria			✓

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

6.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		✓	

If not met, list those compounds and their recoveries which fall outside the acceptable range.

The MS recoveries met the requirements for all compounds .
The MSD recoveries met the acceptable requirements .
The Blank Spike met requirements for all samples .
The RPD for {Q1239-10MSD} with File ID: PL093961.D met criteria except for Endrin aldehyde[30%] due to difference in results of MS and MSD.

		NA	NO	YES
7.	Retention Time Shift Meet Criteria (if applicable)			✓

Comments:

		NA	NO	YES
8.	Extraction Holding Time Met			✓

If not met, list number of days exceeded for each sample:

		NA	NO	YES
9.	Analysis Holding Time Met			✓

If not met, list those compounds and their recoveries which fall outside the acceptable range.

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

	NA	NO	YES

ADDITIONAL COMMENTS:

The soil samples results are based on a dry weight basis.

For samples # JPP-51.2-012925 compound (Endrin) below Method detection limits,
 therefore it is not reported as Hit in Form-1.

For samples # JPP-16.1-012925 compound (delta-BHC) below Method detection limits,
therefore it is not reported as Hit in Form-1.

_____   _____

QA REVIEW   Date