



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: RU2 Engineering, LLC **SDG No.:** Q1235
Contract: RUTW01 **Lab Code:** CHEM **Case No.:** Q1235 **SAS No.:** Q1235
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.26	0.2	130	40 - 160	CV	02/03/2025	16:08	LB134537
CRI01	Arsenic	21.7	20.0	108	40 - 160	P	02/10/2025	14:31	LB134659
	Barium	88.7	100	89	40 - 160	P	02/10/2025	14:31	LB134659
	Cadmium	5.75	6.0	96	40 - 160	P	02/10/2025	14:31	LB134659
	Chromium	10.1	10.0	101	40 - 160	P	02/10/2025	14:31	LB134659
	Lead	11.5	12.0	96	40 - 160	P	02/10/2025	14:31	LB134659
	Selenium	19.7	20.0	99	40 - 160	P	02/10/2025	14:31	LB134659
	Silver	10.6	10.0	106	40 - 160	P	02/10/2025	14:31	LB134659