

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: RU2 Engineering, LLC **SDG No.:** Q1235
Contract: RUTW01 **Lab Code:** CHEM **Case No.:** Q1235 **SAS No.:** Q1235

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166425BS							
Arsenic	ug/L	4000	3710		93	80 - 120	P
Barium	ug/L	1000	885		88	80 - 120	P
Cadmium	ug/L	1000	964		96	80 - 120	P
Chromium	ug/L	2000	2000		100	80 - 120	P
Lead	ug/L	5000	4420		88	80 - 120	P
Selenium	ug/L	10000	9040		90	80 - 120	P
Silver	ug/L	380	350		92	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1235

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1235

SAS No.: Q1235

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166487BS Mercury	ug/L	4.0	4.10		102	80 - 120	CV