

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG01312
Data File : FG015283.D
Signal(s) : FID1A.ch
Acq On : 31 Jan 2025 16:17
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 02/03/2025
Supervised By :Ankita Jodhani 02/03/2025

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG011325.M
Title :

Signal : FID1A.ch

peak #	R. T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.550	4.499	4.613	BB	661355	6845031	92.84%	8.857%
2	6.735	6.687	6.786	BB	702454	7102993	96.34%	9.191%
3	8.574	8.528	8.633	BB	667384	7092897	96.20%	9.178%
4	10.191	10.141	10.253	BB	661858	7206608	97.74%	9.325%
5	11.642	11.592	11.688	BB	642047	7372867	100.00%	9.540%
6	12.960	12.859	13.001	BV	618431	7307969	99.12%	9.456%
7	14.165	14.125	14.207	PB	581610	7094223	96.22%	9.180%
8	15.069	15.023	15.137	PV	468619	6434170	87.27%	8.326%
9	15.275	15.187	15.309	BV	545427	7024640	95.28%	9.090%
10	16.300	15.826	16.334	PV	530319	6878052	93.29%	8.900%
11	17.255	17.068	17.290	VV	507994	6921377	93.88%	8.956%
Sum of corrected areas:						77280827		

FG011325.M Sat Feb 01 02:07:44 2025