

Report of Analysis

| | | | |
|--------------------|---------------------------------------|--------------------|----------|
| Client: | RU2 Engineering, LLC | Date Collected: | 01/29/25 |
| Project: | NYCDDC SANTWOBR Brooklyn Bridge BBMCR | Date Received: | 01/30/25 |
| Client Sample ID: | JPP-16.1-012925 | SDG No.: | Q1235 |
| Lab Sample ID: | Q1235-07 | Matrix: | SOIL |
| Analytical Method: | SW8082A | % Solid: | 89.2 |
| Sample Wt/Vol: | 30.02 | Units: | g |
| Soil Aliquot Vol: | | | uL |
| Extraction Type: | | Test: | PCB |
| GPC Factor : | 1.0 | PH : | |
| Prep Method : | SW3541B | Injection Volume : | |

| | | | | |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
| PO109357.D | 1 | 01/31/25 08:15 | 02/01/25 02:41 | PB166412 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| TARGETS | | | | | | |
| 12674-11-2 | Aroclor-1016 | 19.0 | U | 3.80 | 19.0 | ug/kg |
| 11104-28-2 | Aroclor-1221 | 19.0 | U | 7.20 | 19.0 | ug/kg |
| 11141-16-5 | Aroclor-1232 | 19.0 | U | 3.80 | 19.0 | ug/kg |
| 53469-21-9 | Aroclor-1242 | 19.0 | U | 3.80 | 19.0 | ug/kg |
| 12672-29-6 | Aroclor-1248 | 19.0 | U | 8.80 | 19.0 | ug/kg |
| 11097-69-1 | Aroclor-1254 | 19.0 | U | 3.10 | 19.0 | ug/kg |
| 37324-23-5 | Aroclor-1262 | 19.0 | U | 5.10 | 19.0 | ug/kg |
| 11100-14-4 | Aroclor-1268 | 19.0 | U | 3.80 | 19.0 | ug/kg |
| 11096-82-5 | Aroclor-1260 | 19.0 | U | 3.30 | 19.0 | ug/kg |
| SURROGATES | | | | | | |
| 877-09-8 | Tetrachloro-m-xylene | 19.6 | | 32 - 144 | 98% | SPK: 20 |
| 2051-24-3 | Decachlorobiphenyl | 13.0 | | 32 - 175 | 65% | SPK: 20 |

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit