

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL012125

Review By	Abdul	Review On	1/22/2025 8:17:54 AM
Supervise By	Ankita	Supervise On	1/22/2025 8:30:39 AM
SubDirectory	PL012125	HP Acquire Method	HP Processing Method pl012125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL093724.D	21 Jan 2025 10:03		AR\AJ	Ok
2	I.BLK	I.BLK	PL093725.D	21 Jan 2025 10:16		AR\AJ	Ok
3	PEM	PEM	PL093726.D	21 Jan 2025 10:30		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL093727.D	21 Jan 2025 10:43		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PL093728.D	21 Jan 2025 10:57		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL093729.D	21 Jan 2025 11:10		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL093730.D	21 Jan 2025 11:24		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL093731.D	21 Jan 2025 11:38		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL093732.D	21 Jan 2025 11:51		AR\AJ	Ok
10	PCHLORICC1000	PCHLORICC1000	PL093733.D	21 Jan 2025 12:05		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL093734.D	21 Jan 2025 12:18		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL093735.D	21 Jan 2025 12:32		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL093736.D	21 Jan 2025 12:45		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL093737.D	21 Jan 2025 12:59		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL093738.D	21 Jan 2025 13:12		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL093739.D	21 Jan 2025 13:26		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL093740.D	21 Jan 2025 13:39		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL093741.D	21 Jan 2025 13:53		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL012125

Review By	Abdul	Review On	1/22/2025 8:17:54 AM
Supervise By	Ankita	Supervise On	1/22/2025 8:30:39 AM
SubDirectory	PL012125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL093742.D	21 Jan 2025 14:07		AR\AJ	Ok
20	PSTDICV050	ICVPL012125	PL093743.D	21 Jan 2025 14:20		AR\AJ	Ok
21	PCHLORICV500	ICVPL012125CHLOR	PL093744.D	21 Jan 2025 14:47		AR\AJ	Ok,M
22	PTOXICV500	ICVPL012125TOX	PL093745.D	21 Jan 2025 15:14		AR\AJ	Ok
23	I.BLK	I.BLK	PL093746.D	21 Jan 2025 15:41		AR\AJ	Ok
24	PEM	PEM	PL093747.D	21 Jan 2025 15:54		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PL093748.D	21 Jan 2025 16:07		AR\AJ	Ok,M
26	Q1093-01RE	RT-3407RE	PL093749.D	21 Jan 2025 16:21	F Flag coming , DCB high in 2nd column	AR\AJ	Confirms
27	I.BLK	I.BLK	PL093750.D	21 Jan 2025 16:34		AR\AJ	Ok
28	PSTDCCC050	PSTDCCC050	PL093751.D	21 Jan 2025 16:48		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL013125

Review By	Abdul	Review On	2/3/2025 9:55:31 AM
Supervise By	Ankita	Supervise On	2/3/2025 1:14:23 PM
SubDirectory	PL013125	HP Acquire Method	HP Processing Method pl012125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL093927.D	31 Jan 2025 10:38		AR\AJ	Ok
2	I.BLK	I.BLK	PL093928.D	31 Jan 2025 10:51		AR\AJ	Ok
3	PEM	PEM	PL093929.D	31 Jan 2025 11:04		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL093930.D	31 Jan 2025 11:17		AR\AJ	Ok,M
5	PCHLORCCC500	PCHLORCCC500	PL093931.D	31 Jan 2025 11:54		AR\AJ	Ok,M
6	PTOXCCC500	PTOXCCC500	PL093932.D	31 Jan 2025 12:07		AR\AJ	Ok
7	PB166365BL	PB166365BL	PL093933.D	31 Jan 2025 12:29		AR\AJ	Ok
8	PB166365BS	PB166365BS	PL093934.D	31 Jan 2025 13:08		AR\AJ	Ok
9	PB166365BSD	PB166365BSD	PL093935.D	31 Jan 2025 13:22		AR\AJ	Ok,M
10	PB166365BS	PB166365BS	PL093936.D	31 Jan 2025 13:35		AR\AJ	Ok,M
11	PB166365BS	PB166365BS	PL093937.D	31 Jan 2025 13:49		AR\AJ	Ok,M
12	Q1211-01	TAPHHA-MW01-01282	PL093938.D	31 Jan 2025 14:02	TCMX high in 1st column	AR\AJ	Ok
13	Q1211-02	TAPIAL2-MW03-01282	PL093939.D	31 Jan 2025 14:15		AR\AJ	Ok,M
14	PB166413BL	PB166413BL	PL093940.D	31 Jan 2025 14:28		AR\AJ	Ok
15	PB166413BS	PB166413BS	PL093941.D	31 Jan 2025 14:43		AR\AJ	Ok,M
16	I.BLK	I.BLK	PL093942.D	31 Jan 2025 14:57		AR\AJ	Ok
17	PSTDCCC050	PSTDCCC050	PL093943.D	31 Jan 2025 15:10		AR\AJ	Ok,M
18	PCHLORCCC500	PCHLORCCC500	PL093944.D	31 Jan 2025 15:23		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL013125

Review By	Abdul	Review On	2/3/2025 9:55:31 AM
Supervise By	Ankita	Supervise On	2/3/2025 1:14:23 PM
SubDirectory	PL013125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXC500	PTOXC500	PL093945.D	31 Jan 2025 16:34		AR\AJ	Ok,M
20	Q1232-03	JPP-46.2-012925	PL093946.D	31 Jan 2025 17:01	DCB high in 2nd column	AR\AJ	Ok,M
21	Q1232-07	JPP-46.1-012925	PL093947.D	31 Jan 2025 17:14		AR\AJ	Ok,M
22	Q1232-11	JPP-42.1-012925	PL093948.D	31 Jan 2025 17:27		AR\AJ	Ok,M
23	Q1232-15	JPP-42.2-012925	PL093949.D	31 Jan 2025 17:40		AR\AJ	Ok,M
24	Q1232-19	JPP-51.1-012925	PL093950.D	31 Jan 2025 17:53		AR\AJ	Ok,M
25	Q1235-03	JPP-51.2-012925	PL093951.D	31 Jan 2025 18:07		AR\AJ	Ok,M
26	Q1235-07	JPP-16.1-012925	PL093952.D	31 Jan 2025 18:20		AR\AJ	Ok,M
27	Q1239-01	286	PL093953.D	31 Jan 2025 18:33		AR\AJ	Ok,M
28	Q1239-04	348	PL093954.D	31 Jan 2025 18:46		AR\AJ	Ok,M
29	Q1239-07	RBR22266	PL093955.D	31 Jan 2025 18:59		AR\AJ	Ok,M
30	Q1239-10	357	PL093956.D	31 Jan 2025 19:12		AR\AJ	Ok,M
31	I.BLK	I.BLK	PL093957.D	31 Jan 2025 19:26		AR\AJ	Ok
32	PEM	PEM	PL093958.D	31 Jan 2025 19:39		AR\AJ	Ok,M
33	PSTDCC050	PSTDCC050	PL093959.D	31 Jan 2025 20:18		AR\AJ	Ok,M
34	Q1239-10MS	357MS	PL093960.D	31 Jan 2025 20:31		AR\AJ	Ok,M
35	Q1239-10MSD	357MSD	PL093961.D	31 Jan 2025 20:45	RPD Fail	AR\AJ	Ok,M
36	Q1241-03	JPP-3.5-013025	PL093962.D	31 Jan 2025 20:58		AR\AJ	Ok,M
37	Q1241-07	JPP-5.3-013025	PL093963.D	31 Jan 2025 21:11		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL013125

Review By	Abdul	Review On	2/3/2025 9:55:31 AM
Supervise By	Ankita	Supervise On	2/3/2025 1:14:23 PM
SubDirectory	PL013125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q1241-11	JPP-5.2-013025	PL093964.D	31 Jan 2025 21:24		AR\AJ	Ok,M
39	Q1241-15	JPP-5.4-013025	PL093965.D	31 Jan 2025 21:37		AR\AJ	Ok,M
40	Q1241-19	JPP-51.4-013025	PL093966.D	31 Jan 2025 21:50		AR\AJ	Ok,M
41	Q1242-03	JPP-6.2-013025	PL093967.D	31 Jan 2025 22:04		AR\AJ	Ok,M
42	Q1243-01	CL-01-01302025	PL093968.D	31 Jan 2025 22:17		AR\AJ	Ok
43	Q1244-01	EO-02-01302025	PL093969.D	31 Jan 2025 22:30		AR\AJ	Ok,M
44	I.BLK	I.BLK	PL093970.D	31 Jan 2025 22:43		AR\AJ	Ok
45	PSTDCCC050	PSTDCCC050	PL093971.D	01 Feb 2025 00:57	Comp#22 recovery low	AR\AJ	Ok,M
46	Q1215-03	JPP-29.1-012825	PL093972.D	01 Feb 2025 01:10		AR\AJ	Ok,M
47	Q1215-03MS	JPP-29.1-012825MS	PL093973.D	01 Feb 2025 01:24		AR\AJ	Ok,M
48	Q1215-03MSD	JPP-29.1-012825MSD	PL093974.D	01 Feb 2025 01:37		AR\AJ	Ok,M
49	Q1216-15	JPP-26.1-012825	PL093975.D	01 Feb 2025 01:50	DCB high in both column	AR\AJ	ReRun
50	Q1219-01	LAW-25-0015	PL093976.D	01 Feb 2025 02:03	DCB high in 1st column	AR\AJ	Ok,M
51	I.BLK	I.BLK	PL093977.D	01 Feb 2025 02:43		AR\AJ	Ok
52	PSTDCCC050	PSTDCCC050	PL093978.D	01 Feb 2025 02:56	Comp#22 recovery low	AR\AJ	Ok,M

M : Manual Integration