

### Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1235

Client: RU2 Engineering, LLC

Analytical Method: 8151A

DataFile : PS029027.D

| Lab Sample ID:    | Parameter          | Spike | Sample<br>Result | Result | Units | Rec | Rec<br>Qual | RPD | RPD<br>Qual | Low | Limits<br>High | RPD |
|-------------------|--------------------|-------|------------------|--------|-------|-----|-------------|-----|-------------|-----|----------------|-----|
| Client Sample ID: | JPP-29.1-012825MSD |       |                  |        |       |     |             |     |             |     |                |     |
| Q1215-04MSD       | 2,4-D              | 50    | 0                | 52.5   | ug/L  | 105 |             | 0   |             | 65  | 135            | 20  |
|                   | 2,4,5-TP(Silvex)   | 50    | 0                | 63.8   | ug/L  | 128 |             | 1   |             | 62  | 139            | 20  |