

### CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1235 SAS No.: Q1235 SDG NO.: Q1235

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL094022.D Time Analyzed: 22:41

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Decachlorobiphenyl   | 9.056 | 8.953     | 9.153 | 48.960             | 50.000            | -2.1 |
| Endrin               | 6.572 | 6.472     | 6.672 | 50.650             | 50.000            | 1.3  |
| gamma-BHC (Lindane)  | 4.327 | 4.227     | 4.427 | 51.980             | 50.000            | 4.0  |
| Heptachlor           | 4.915 | 4.814     | 5.014 | 51.640             | 50.000            | 3.3  |
| Heptachlor epoxide   | 5.683 | 5.582     | 5.782 | 50.990             | 50.000            | 2.0  |
| Methoxychlor         | 7.500 | 7.398     | 7.598 | 53.980             | 50.000            | 8.0  |
| Tetrachloro-m-xylene | 3.538 | 3.439     | 3.639 | 51.210             | 50.000            | 2.4  |

### CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1235 SAS No.: Q1235 SDG NO.: Q1235

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL094022.D Time Analyzed: 22:41

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Decachlorobiphenyl   | 7.910 | 7.810     | 8.010 | 53.520             | 50.000            | 7.0  |
| Endrin               | 5.636 | 5.536     | 5.736 | 56.890             | 50.000            | 13.8 |
| gamma-BHC (Lindane)  | 3.607 | 3.507     | 3.707 | 52.380             | 50.000            | 4.8  |
| Heptachlor           | 3.945 | 3.845     | 4.045 | 51.290             | 50.000            | 2.6  |
| Heptachlor epoxide   | 4.727 | 4.627     | 4.827 | 52.980             | 50.000            | 6.0  |
| Methoxychlor         | 6.610 | 6.509     | 6.709 | 56.660             | 50.000            | 13.3 |
| Tetrachloro-m-xylene | 2.774 | 2.674     | 2.874 | 51.590             | 50.000            | 3.2  |